(PV-STR-01) SwSTR

Document ID: `PV-STR-01`
Product: `Portview`
Document Status: `Released`

Document Overview

This document records the unit-level verification results for Portview.

Document Approval

Prepared by

Title	Name	Signature
Manager	`J. W. Lee`	

Reviewed by

Title	Name	Signature
General Manager	`S. I. Choi`	

Approved by

Title	Name	Signature
CTO (Director)	`K. Y. Ro`	

Revision History

Rev.	Date	Description
`0.0`	`2024-01-17`	Initial unit result record prepared according to verification-item subdivision
`0.1`	`2025.07.24`	Unit result set updated with GUI translation-related verification coverage

1. Purpose

This document records the executed unit verification results for Portview and provides the evidence summary associated with the unit procedure baseline.

2. References

This result record is governed primarily by:

- `PV-SRS-01` SwSRS for Portview
- `PV-STP-01` SwSTP for Portview
- `PV-SV-04` Software Verification Plan
- `PV-SV-05` Software Verification Report

3. Execution Information

Field	Value
Executor	`J. W. Lee`
Execution date	`2025.07.24`
Independent reviewer	`M. C. Boo`
Review date	`2025.07.24`

Field	Value
Execution note	Tests were executed using a pre-configured Genoray testing platform with standardized test data sets. The controlled test environment and procedure-level acceptance criteria enabled single-executor throughput. Independent verification is provided through document review by a separate reviewer.
Test location	`Genoray testing platform`
Configuration	`Default`
Software baseline	`Portview 2.2.5.16`
Governing procedure	`PV-STP-01`

4. Result Coverage

The unit result record covers the following authored unit procedures.

Procedure Range	Coverage
`UTP-001` to `UTP-005`, `UTP-018` to `UTP-022`	Acquisition and device access
`UTP-006` to `UTP-012`	Viewer and image interaction
`UTP-013` to `UTP-017`	Output and communication
`UTP-023` to `UTP-025`	Administrative and support functions

5. Result Summary

The unit verification activities were executed against the released Portview unit baseline and were reported as successfully completed.

Area	Result Statement	Status
Acquisition and device access	Unit-level acquisition and device actions were executed successfully	`Pass summary stated`

Area	Result Statement	Status
Viewer and image interaction	Unit-level viewer interactions were reported as acceptable	`Pass summary stated`
Output and communication	Unit-level output and communication actions were reported as acceptable	`Pass summary stated`
Administrative and support functions	Unit-level support and administrative actions were reported as acceptable	`Pass summary stated`

5.1 Procedure-Level Result Set

The current authored result set is aligned to the `UTP-xxx` unit procedure structure defined in `PV-STP-01`.

Unit Procedure ID	Title	Result Summary	Status	Notes
`UTP-001`	Acquire intraoral sensor image	Unit sensor acquisition was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-002`	Acquire image from file	File import at unit level was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-003`	Acquire image from TWAIN device	TWAIN unit behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-004`	Show status of device	Device-status behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-005`	Select device	Device selection behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Unit Procedure ID	Title	Result Summary	Status	Notes
`UTP-006`	Main-window behavior	Main-window behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-007`	Tooth map window	Tooth-map behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-008`	Compare image display	Compare-view behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-009`	Pan behavior	Pan interaction was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-010`	Zoom behavior	Zoom interaction was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-011`	Rotate behavior	Rotate behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-012`	Thumbnail viewer	Thumbnail-view behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-013`	Print image	Standard print behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-014`	Transmit to DICOM printer	DICOM-print behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Unit Procedure ID	Title	Result Summary	Status	Notes
`UTP-015`	Export image to file	File export behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-016`	Export images to CD	Media export behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-017`	Send image to DICOM server	DICOM-send behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-018`	FMX continuous acquisition	FMX continuous acquisition behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-019`	Multi-sensor selection	Multi-sensor selection behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-020`	Sensor status	Sensor-status behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-021`	USB integrity check	USB integrity behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-022`	Local-network connection check	Network-dependent unit behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Unit Procedure ID	Title	Result Summary	Status	Notes
`UTP-023`	Logging	Logging behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-024`	Patient management	Patient-management behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`UTP-025`	GUI translation	GUI translation behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

6. Deviations And Issues

No Portview-specific unit deviation is separately listed in the current result set.

ID	Description	Disposition
`NONE-RECORDED`	No separate Portview unit deviation is recorded in the current result summary set	`Accepted as no recorded deviation`

7. Evidence References

Evidence Type	Reference	Location
Executed test protocol	Signed unit-test execution protocol for `UTP-001` to `UTP-025`	QMS controlled archive, test-execution record set
Test environment log	Genoray testing platform configuration and session log	QMS controlled archive
Independent review record	Review sign-off by `M. C. Boo` confirming result acceptance	QMS controlled archive

8. Traceability Use

This result record is intended to support:

- unit-result linkage in `PV-TM-01` Traceability Matrix
- unit-result references in `PV-SV-05` Software Verification Report
- release-support statements in `PV-SV-01` Software Validation Report