

QA File:

C:\Canberra\Apex\Root\FUSRAP HISS\QA\D0000000001SRS120910 MARSOILC.QCK

Detector: DET01

Geometry: <None>

Certificate: SRS120910 MARSOIL

Measurement Date: 1/2/24 09:37:52

Live Time: 300.0 seconds Real Time: 305.0 seconds

Sample ID: HQ02JAN24

Description: DAILY QC SRS120910

Date: 10/1/21 12:00:00

Quantity: 1.00000E+00

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM Am-241 [SD: 1.20215E+00+/- 0.030]	1.18969E+00	-4.15858E-01 < : : : >
Peak FWHM Cs-137 [SD: 1.60886E+00+/- 0.061]	1.53904E+00	-1.13651E+00 < : : : >
Peak FWHM Co-60 1332 [SD: 2.01989E+00+/- 0.093]	1.98026E+00	-4.24829E-01 < : : : >
Peak centroid Am-241 [SD: 4.77030E+02+/- 0.096]	4.77153E+02	1.28999E+00 < : : : >
Peak centroid Cs-137 [SD: 5.29506E+03+/- 0.948]	5.29654E+03	1.56024E+00 < : : : >
Peak centroid Co-60 1332 [SD: 1.06622E+04+/- 1.931]	1.06652E+04	1.55187E+00 < : : : >
Peak energy Am-241 [SD: 5.95927E+01+/- 0.023]	5.95905E+01	-9.63548E-02 < : : : >
Peak energy Cs-137 [SD: 6.61796E+02+/- 0.122]	6.62006E+02	1.73344E+00 < : : : >
Peak energy Co-60 1332 [SD: 1.33274E+03+/- 0.248]	1.33316E+03	1.70027E+00 < : : : >
Activity (pCi) Am-241 [SD: 5.91513E+04+/-897.38]	5.70566E+04	-2.33426E+00 < :In : : >
Activity (pCi) Cs-137 [SD: 2.43670E+04+/-419.20]	2.48631E+04	1.18338E+00 < : : : >
Activity (pCi) Co-60 1332 [SD: 3.85146E+04+/-637.31]	3.73613E+04	-1.80951E+00 < : : : >

Note: The Sample Driven mean and standard deviation does NOT include the last measurement.

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)

Analyst and date: MA Spectrum is shifted, will adjust gain 1/2/2024