

## QA File:

C:\Canberra\Apex\Root\FUSRAP HISS\QA\G0000000003D0000000001S.QCK

Detector: DET01

Geometry: Empty Shield

Certificate: &lt;None&gt;

Measurement Date: 7/8/24 05:42:18

Live Time: 86400.0 seconds Real Time: 87135.3 seconds

Sample ID: SV08JUL24

Description: WEEKLY BKG

Date: 7/8/24 05:42:15

Quantity: 1.00000E+00

| Parameter Description<br>[Mean +/- Std. Dev.]            | Value       | Deviation/Flags<br>< LU : SD : UD : BS > |
|----------------------------------------------------------|-------------|------------------------------------------|
| WKLY-Background count rate<br>[SD: 2.05328E+00+/- 0.245] | 2.10671E+00 | 2.18314E-01<br>< : : : >                 |
| Backrate CH 1 to 4096<br>[SD: 1.31986E+00+/- 0.159]      | 1.34763E+00 | 1.74340E-01<br>< : : : >                 |
| Backrate CH 4097 to 8192<br>[SD: 4.53318E-01+/- 0.054]   | 4.66725E-01 | 2.48078E-01<br>< : : : >                 |
| Backrate CH 8193 to 12288<br>[SD: 1.92171E-01+/- 0.023]  | 2.01343E-01 | 3.99735E-01<br>< : : : >                 |
| Backrate CH 12289 to 16384<br>[SD: 8.79313E-02+/- 0.011] | 9.10185E-02 | 2.93330E-01<br>< : : : >                 |

Note: The Sample Driven mean and standard deviation does NOT include the last measurement.

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)

Analyst and date: CK 07/09/2024