

Efficiency Report

Detector Name	XLB1_D1P1
Alpha Eff	30.3 +/- 1.26 %
Alpha Spilldown	11.9 +/- 0.703 %
Beta Eff	0 +/- 0 %
Beta SpillUp	0 +/- 0 %
Alpha Voltage	---
Beta Voltage	1410
Alpha Bkg	0 +/- 0 %
Beta Bkg	0 +/- 0 %

Device: XLB1

Calibration: AB Air Filter Calibration

Detector: XLB1_D1P1

Procedure: Daily Source Check

Batch ID: Batch_4474

Count Mode: Simultaneous

Group: A

Carrier ID: 1

<u>Sample Name</u>	<u>Assay date</u>	<u>Standard</u>	<u>Live Time (m)</u>	<u>Alpha (CPM)</u>	<u>Beta (CPM)</u>	<u>Efficiency</u>	<u>Spillover</u>	<u>Decay Corrected Activity</u>
Calibration Check_14357	1/2/2024 10:02:42 AM	Th-230 120911	10	270 +/-5.19	32.1 +/-1.79	30.3 % +/-1.26 %	11.9 % +/-0.703 %	14.8 Bq

CK 1/2/2024