

Background Report

Batch Name: Batch_2741**Count Date:** 1/2/2023 1:41:58 PM**Procedure:** AB Air Filter System BKG**Preset Count Time (m):** 1,440.00**Calibration:** AB Air Filter Calibration**Count Mode:** Simultaneous**Group:** C**Operating Voltage:** 1380**Carrier:** 39

Calculated Background (cpm)

Detector Name	Alpha Bkg Rate (cpm)	Beta Bkg Rate (cpm)
XLB1_D1P1	0.00 +/- 0.00	0.39 +/- 0.28

Assay Date	Sample name	Iteration	Alpha Bkg Rate (cpm)	Beta Bkg Rate (cpm)	Guard (cpm)	Live Time (Minutes)
1/2/2023	System	1	0.00 +/- 0.00	0.39 +/- 0.28	230.45	5.00
1:41:58 PM	Background_9248					

CK 1/5/2023

5 min. system bkg conducted to allow for LCS samples to be run for secondary source check following calibration.