

Efficiency Report

Detector Name	XLB1_D1P1
Alpha Eff	29.8 +/- 0.612 %
Alpha Spilldown	14.7 +/- 0.392 %
Beta Eff	0 +/- 0 %
Beta SpillUp	0 +/- 0 %
Alpha Voltage	---
Beta Voltage	1380
Alpha Bkg	0 +/- 0 %
Beta Bkg	0 +/- 0 %

Device: XLB1

Calibration: AB Air Filter Calibration

Detector: XLB1_D1P1

Procedure: AB Air Filter Efficiency

Batch ID: Batch_2738

Count Mode: Simultaneous

Group: D

Carrier ID: 1

<u>Sample Name</u>	<u>Assay date</u>	<u>Standard</u>	<u>Live Time (m)</u>	<u>Alpha (CPM)</u>	<u>Beta (CPM)</u>	<u>Efficiency</u>	<u>Spillover</u>	<u>Decay Corrected Activity</u>
Efficiency Calibration_9244	1/2/2023 11:59:26 AM	Th-230 120912	41.58333333333333	265 +/-2.52	38.9 +/-0.967	29.8 % +/-0.612 %	14.7 % +/-0.392 %	14.8 Bq

CK 1/5/2023

BQW 1/5/2023