

QA File:

C:\Canberra\Apex\Root\FUSRAP HISS\QA\D0000000001B.QCK

Detector: DET01

Geometry: <None>

Certificate: <None>

Measurement Date: 8/1/24 06:57:00

Live Time: 300.0 seconds Real Time: 302.7 seconds

Sample ID: AW01AUG24

Description: DAILY BKG

Date: 8/1/24 06:56:58 Quantity: 1.00000E+00

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Background count rate [SD: 2.10180E+00 +/- 0.094]	2.13333E+00	3.36289E-01 < : : : >

Note: The Sample Driven mean and standard deviation does NOT include the last measurement.

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)

Analyst and date: BQW 08/01/2024