



QA SUMMARY REPORT
Review Of QA Results - Pulser Check

Date : 8/1/2024

Time : 06:15:54

CHAMBER	DEVICE	PARAMETER	FLAG	DATE
DET 1	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:05
DET 2	Alpha Analyst10464	Peak FWHM	Action	8/1/2024 06:04:08
DET 3	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:09
DET 4	Alpha Analyst10464	Peak FWHM	Investigate	8/1/2024 06:04:11
DET 5	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:13
DET 6	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:16
DET 7	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:18
DET 8	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:20
DET 9	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:22
DET 10	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:25
DET 11	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:27
DET 12	Alpha Analyst10464	ALL	Passed	8/1/2024 06:04:30
DET 13	Alpha Analyst10465	ALL	Passed	8/1/2024 06:04:31
DET 14	Alpha Analyst10465	ALL	Passed	8/1/2024 06:04:33
DET 15	Alpha Analyst10465	ALL	Passed	8/1/2024 06:04:34
DET 16	Alpha Analyst10465	ALL	Passed	8/1/2024 06:04:36

APPROVED BY: ALAPPROVAL DATE: 08/01/2024