

QA File:

C:\Canberra\Apex\Root\FUSRAP HISS\QA\G0000000003D0000000001S.QCK

Detector: DET01

Geometry: Empty Shield

Certificate: <None>

Measurement Date: 5/10/24 13:54:41

Live Time: 86400.0 seconds Real Time: 87142.0 seconds

Sample ID: FL10MAY24

Description: SYSTEM BKG

Date: 5/10/24 13:54:38

Quantity: 1.00000E+00

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
WKLY-Background count rate [SD: 2.05393E+00+/- 0.261]	2.03882E+00	-5.79302E-02 < : : : >
Backrate CH 1 to 4096 [SD: 1.32032E+00+/- 0.170]	1.31155E+00	-5.16265E-02 < : : : >
Backrate CH 4097 to 8192 [SD: 4.53552E-01+/- 0.058]	4.48576E-01	-8.64267E-02 < : : : >
Backrate CH 8193 to 12288 [SD: 1.92133E-01+/- 0.024]	1.91215E-01	-3.75557E-02 < : : : >
Backrate CH 12289 to 16384 [SD: 8.79309E-02+/- 0.011]	8.74768E-02	-4.07000E-02 < : : : >

Note: The Sample Driven mean and standard deviation does NOT include the last measurement.

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)

Analyst and date: MLA 05/13/2024