



QA SUMMARY REPORT

Review Of QA Results - System Background

Date : 6/25/2024

Time : 06:01:30

Geometry : Shelf 2

CHAMBER	DEVICE	PARAMETER	FLAG	DATE
DET 1	Alpha Analyst10464	ALL	Passed	6/24/2024 05:51:55
DET 2	Alpha Analyst10464	ALL	Passed	6/24/2024 05:51:57
DET 3	Alpha Analyst10464	ALL	Passed	6/24/2024 05:51:58
DET 4	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:00
DET 5	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:02
DET 6	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:04
DET 7	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:05
DET 8	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:07
DET 9	Alpha Analyst10464	Bkg cpm	Investigate	6/24/2024 05:52:09
DET 10	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:10
DET 11	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:12
DET 12	Alpha Analyst10464	ALL	Passed	6/24/2024 05:52:13
DET 13	Alpha Analyst10465	ALL	Passed	6/24/2024 05:52:15
DET 14	Alpha Analyst10465	ALL	Passed	6/24/2024 05:52:16
DET 15	Alpha Analyst10465	Bkg cpm	Action	6/24/2024 05:52:17
DET 16	Alpha Analyst10465	ALL	Passed	6/24/2024 05:52:18

APPROVED BY: MAAPPROVAL DATE: 06/25/2024

Detector 15 is outside the action limit. The system background will be recounted on Detector 15.