



QA SUMMARY REPORT
Review Of QA Results - Pulser Check

Date : 7/10/2024

Time : 06:45:05

CHAMBER	DEVICE	PARAMETER	FLAG	DATE
DET 1	Alpha Analyst10464	ALL	Passed	7/10/2024 06:34:49
DET 2	Alpha Analyst10464	ALL	Passed	7/10/2024 06:34:50
DET 3	Alpha Analyst10464	ALL	Passed	7/10/2024 06:34:52
DET 4	Alpha Analyst10464	Peak FWHM	Investigate	7/10/2024 06:34:54
DET 5	Alpha Analyst10464	ALL	Passed	7/10/2024 06:34:56
DET 6	Alpha Analyst10464	ALL	Passed	7/10/2024 06:34:57
DET 7	Alpha Analyst10464	Peak FWHM	Investigate	7/10/2024 06:34:59
DET 8	Alpha Analyst10464	ALL	Passed	7/10/2024 06:35:00
DET 9	Alpha Analyst10464	ALL	Passed	7/10/2024 06:35:02
DET 10	Alpha Analyst10464	ALL	Passed	7/10/2024 06:35:04
DET 11	Alpha Analyst10464	ALL	Passed	7/10/2024 06:35:05
DET 12	Alpha Analyst10464	ALL	Passed	7/10/2024 06:35:07
DET 13	Alpha Analyst10465	ALL	Passed	7/10/2024 06:35:08
DET 14	Alpha Analyst10465	ALL	Passed	7/10/2024 06:35:09
DET 15	Alpha Analyst10465	ALL	Passed	7/10/2024 06:35:12
DET 16	Alpha Analyst10465	ALL	Passed	7/10/2024 06:35:13

APPROVED BY: ALAPPROVAL DATE: 07/10/2024