NI PXI/PXIe-2569 Specifications

100-Channel SPST Relay Module

このドキュメントには、日本語ページも含まれています。

This document lists specifications for the NI PXI/PXIe-2569 (NI 2569) general-purpose relay module. All specifications are subject to change without notice. Visit ni.com/manuals for the most current specifications.

Refer to the NI Switches Help for detailed topology information.



Caution Refer to the *Read Me First: Safety and Electromagnetic Compatibility* document for important safety and compliance information.



Caution When operating this product, use shielded cables and accessories.



Caution The protection provided by the NI 2569 can be impaired if it is used in a manner not described in this document.

About These Specifications

Specifications characterize the warranted performance of the instrument under the stated operating conditions.

Typical Specifications are specifications met by the majority of the instrument under the stated operating conditions and are tested at 23 °C. Typical specifications are not warranted.

All voltages are specified in DC, AC_{pk}, or a combination unless otherwise specified.

Clean devices and terminal blocks by brushing off light dust with a soft, nonmetallic brush. Remove other contaminants with a soft, lint-free, dampened cloth. Do not use detergent or chemical solvents. The unit must be completely dry and free from contaminants before returning to service.

Input Characteristics

All input characteristics are DC, AC_{rms}, or a combination unless otherwise specified.

Maximum switching voltage

Channel-to-channel 100 V Channel-to-ground 100 V, CAT I





Caution This module is rated for Measurement Category I and intended to carry signal voltages no greater than 100 V. This module can withstand up to 500 V impulse voltage. Do not use this module for connection to signals or for measurements within Categories II, III, or IV. Do not connect to MAINs supply circuits (for example, wall outlets) of 115 or 230 VAC. Refer to the Read Me First: Safety and Electromagnetic Compatibility document for more information on measurement categories.

When hazardous voltages (>42.4 $V_{nk}/60$ VDC) are present on any relay terminal, safety low-voltage (<42.4 V_{pk}/60 VDC) cannot be connected to any other relay terminal.



Caution The switching power is limited by the maximum switching current, the maximum voltage, and must not exceed 60 W, 62.5 VA.

Maximum switching power (per channel)	60 W, 62.5 VA (DC to 60 Hz)
Maximum current (switching or carry, per channel)	1 A
Simultaneous channels at maximum current (≤35 °C)	.50



Note Switching inductive loads (for example, motors and solenoids) can produce high-voltage transients in excess of the module's rated voltage. Without additional protection, these transients can interfere with module operation and impact relay life. For more information about transient suppression, visit ni.com/info and enter the Info Code induct.

Module Load Derating at >35 °C

Load derating is dependent on the ambient temperature and the sum of the current squared of each channel simultaneously carrying a signal. The result must fall within the shaded region of Figure 1. The following examples represent this calculation.

Example 1

50 channels carry 0.75 A while

10 channels carry 0.5 A

$$(50 \times 0.75^2) + (10 \times 0.5^2) = 30.6 \text{ A}^2 \times \text{channels}$$

The module in Example 1 can be used at ambient temperatures between 0 °C and 55 °C.

Example 2


60 channels carry 0.75 A while

35 channels carry 0.5 A

$$(60 \times 0.75^2) + (35 \times 0.5^2) = 42.5 \text{ A}^2 \times \text{channels}$$

The module in Example 2 can be used at ambient temperatures between 0 °C and 45 °C.

Figure 1. Module Load Derating



DC path resistance

DC path resistance typically remains low for the life of the relay. At the end of relay life, the path resistance rapidly rises above 1 Ω . Load ratings apply to relays used within the specification before the end of relay life.

Thermal EMF (typical at 23 °C)....≤12 µV

Bandwidth (-3 dB, typical at 23 °C)

50 Ω termination..... \geq 20 MHz

Crosstalk (typical at 23 °C, 50 Ω termination)

Channel-to-channel 10111

10 KHz	≤-85 aB
100 kHz	≤-65 dB
1 MHz	≤-45 dB
10 MHz	<-25 dB

Isolation (typical at 23 °C, 50 Ω termination)

Open channel

10 kHz	≥85 dB
100 kHz	≥65 dB
1 MHz	≥45 dB
10 MHz	≥25 dB

Dynamic Characteristics

Relay operate time

Typical	1	ms	
Maximum	3	.4 n	n



Note Certain applications may require additional time for proper settling. Refer to the *NI Switches Help* for information about including additional settling time.

Expected relay life



Note The relays used in the NI 2569 are field replaceable. Refer to the NI Switches Help for information about replacing a failed relay.

Trigger Characteristics

Input trigger



Note The NI 2569 can recognize trigger pulse widths less than 150 ns if you disable digital filtering. For information about disabling digital filtering, refer to the *NI Switches Help*.

Output trigger

Physical Characteristics

Relay type	Electromechanical, latching
Relay contact material	Palladium-ruthenium, gold covered
I/O connector	200 POS LFH Matrix 50, receptacle

Power	requirement
TOWEI	requirement

PXI	6 W at 5 V
	2.5 W at 3.3 V
PXI Express	7.5 W at 12 V
	2.5 W at 3.3 V
Dimensions (L × W × H)	3U, one slot,
	PXI/cPCI module, PXIe compatible
	$21.6 \times 2.0 \times 13.0 \text{ cm}$
	$(8.5 \times 0.8 \times 5.1 \text{ in.})$
Weight	289 g (10.2 oz)

Environment

Operating temperature	.0 °C to 55 °C
Storage temperature	20 °C to 70 °C
Relative humidity	.5% to 85%, noncondensing
Pollution Degree	.2
Maximum altitude	. 2,000 m
Indoor use only.	

Shock and Vibration

Operational Shock	30 g peak, half-sine, 11 ms pulse (Tested in accordance with IEC 60068-2-27. Test profile developed in accordance with MIL-PRF-28800F.)
Random Vibration	
Operating	5 to 500 Hz, 0.3 g _{rms}
Nonoperating	5 to 500 Hz, 2.4 g _{rms}
	(Tested in accordance with IEC 60068-2-64.
	Nonoperating test profile exceeds the requirements of MIL-PRF-28800F, Class 3.)

Diagrams

Figure 2 shows the NI 2569 power-on state.

Figure 2. NI 2569 Power-On State




Figure 3 shows the NI 2569 pinout.

Figure 3. NI 2569 Pinout

-			
)	
CH0 COM0 150 COM2 149	<u> </u>	51	
CH0 COM2 151 COM2 152	- 66 -	50	COM1
CH2 • 152	- 66 -	53 . 0115	COM3
CH4 • 001014 0 153	• 11 •	48 0115	COM5
CHE a COIVIO 154		54 47 ○ CH7	COM7
COM8 • 140		55 CH9	COM9
COM10 • 145		56 45 → CH11	
01110 0 001440 144		45 CH11 57 CH13	COM11
CH12 • COM12 • 157 COM14 • 158	- 96 -	58 43 • CH15	COM13
CH14 COM14 158 COM16 159	- 66 -	59 01147	COM15
UD 10 0	- 11 -	92	COM17
CH18 • CONTO 160	• 11 •	41 0 01113	COM19
CH20 COM20 140 COM20 161		61 40 CH21 62 CH23	COM21
CH22 COM22 139 COM24 138		62 39	COM23
CH24 COM24 138 CH24 COM24 138		39 ° CH25 63 ° CH25	COM25
COM26 ∘ 137	-	164 CH27	
		37 ° CH27 65 ○ CH29	COM27
CH28 COM28 135 CH30 COM30 135 CH30 COM30 136	- 66 -	66 - 01104	COM29
CH30 COM32 134 CH32 COM32 134	- 66 -	35 ° CH31 ° CH33	COM31
	- 11 -	00	COM33
CH34 COM34 133 CH34 132		33 0 01100	COM35
CH36 COM36 132 CH36 COM36 169		69 32 → CH37 70	COM37
CH38 COM38 131 CH38 COM38 131		32 ° CH37 70 ° CH39	СОМ39
CH40 COM40 130 CH40 171		71 30 ○ CH41	COM41
COM42 0 123	-	1/2	
COM44 a 128		73 ○ CH45	COM43
CH44 COM44 173 COM46 127 CH46 COM46 173	- 66 -	28 CH45 74 CH47	COM45
COM48 • 126	- % -	27 CH47 75 CH49	COM47
CH48 COM50 175	- 66 -	76	COM49
CH50 COM52 176	- %	77 CHE2	COM51
CH52 • 177	- 11 -	78 CH55	COM53
CH54 COM54 123	- 11 -	23 000	COM55
CH54 COM56 • 122 CH56 • COM58 • 121 R80 COM60 • 120 CO	0 11 0	79 CH57	COM57
CH58 COM58 121		21 001100	COM59
CLICO a 181		20 0 0 0 0 0 0	COM61
COM62 • 182		82 19	COM63
CH62 COM64 118 CH64 COM66 117		83 18 ○ CH65	COM65
CLICC - CONIOO 184		18 ° CH65 84 ○ CH67	
COM68 a 116		85 a CHEO	COM67
COM70 ∘ 115		16 CH69 86 ○ CH71	COM69
COM72 0 114	- % -	15 CH71 87 CH73	COM71
CH72 • COM74 • 113	- % -	88 0 CH75	COM73
CH74 COM76 188	- 99 -	89 . 01177	COM75
CH/6 0 111	- 11 -	12 01177	COM77
CUIVI/8 190	- 11 -	 :: 	COM79
CH/8 COM80 110 CH80 COM80 191	• 11 •	10 00101	COM81
CH80 COM82 109 CH82 COM82 192		92 9 CH83 93 CH85	COM83
CH84 COM84 - 108 COM86 - 107		8 - 01103	COM85
CH86 COM86 107 CH86 COM88 106		94 7	COM87
		95 6 CH89	
COM00 a 105		6 ° CH89 ° ° CH91	COM89
CH90 COM90 196 COM92 104 CH92 COM92 197	- 99 -	97 4 ○ CH93	COM91
OT 102 - COMO4 a 103	- 66 -	98 0 CH95	COM93
CH94 • COMOS a 102	- 66 -	99	COM95
CH96 COM96 199	- 10 -	2 01137	COM97
CH98 ← COM98 ← 101 CH98 ← COM98 ← 101 200	6 90 0	100 CH99 ○	COM99
	<u></u>	J	



Note For topology-specific connection information, refer to your device in the NI Switches Help and the installation instructions for any associated cables or terminal blocks.

Accessories

Visit ni.com for more information about the following accessories.

Table 1. NI Accessories for the NI 2569

Accessory	Part Number
LFH200 to 50-pin DSUB switch cable (CH-Com twisted), 1 m	779038-02
LFH200 connector to bare wire switch cable, 2 m	779038-01
NI TBX-50B, 50-pin DSUB screw terminal block	782866-01

Table 2. Third-Party Accessories for the NI 2569

Accessory	Manufacturer	Part Number
Terminal sticks (four required per module)	Molex	71715-4002
Plug connector subassembly	Molex	71719-3000
Backshell only	Jevons	JDC200B-832
Mass interconnect cable assembly, 20 in.	Virginia Panel	540105010105
Mass interconnect cable assembly, 36 in.	Virginia Panel	540105010205
Mating ITA module* (one required per module)	Virginia Panel	510108131
Mating ITA PC* (198 required per module)	Virginia Panel	720101101
DAK assembly NI PCB, 200 Pin LFH, male	MAC Panel	561036
* PCB mount, additional cover, or enclosure required. See previous safety caution.		



Caution You *must* install mating connectors according to local safety codes and standards and according to the specifications provided by the connector manufacturer. You are responsible for verifying safety compliance of third-party connectors and their usage according to the relevant standard(s), including UL and CSA in North America and IEC and VDE in Europe.

Table 3. Third-Party Accessories for the LFH200 to 50-pin D-SUB Switch Cable

Accessory	Manufacturer	Part Number
VARIOFACE module, with screw connection and 50 position D-SUB pin strip	Phoenix Contact	FLK-D50 SUB/S
VARIOFACE module, with screw connection and 50 position D-SUB pin strip	Phoenix Contact	FLKM-D50 SUB/S
VARIOFACE module, with screw connection and 50 position D-SUB pin strip	Phoenix Contact	FLKMS-D50 SUB/S
VARIOFACE module, with screw connection and 50 position D-SUB pin strip, with LED indicators	Phoenix Contact	FLKM-D50 SUB/S/LA

Compliance and Certifications

Safety

This product meets the requirements of the following standards of safety for electrical equipment for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA 61010-1



Note For UL and other safety certifications, refer to the product label or the *Online* Product Certification section.

Electromagnetic Compatibility

This product meets the requirements of the following EMC standards for electrical equipment for measurement, control, and laboratory use:

- EN 61326-1 (IEC 61326-1): Class A emissions; Basic immunity
- EN 55011 (CISPR 11): Group 1, Class A emissions
- AS/NZS CISPR 11: Group 1, Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



Note For EMC declarations and certifications, refer to the *Online Product* Certification section.

CE Compliance (€

This product meets the essential requirements of applicable European Directives as follows:

- 2006/95/EC; Low-Voltage Directive (safety)
- 2004/108/EC; Electromagnetic Compatibility Directive (EMC)

Online Product Certification

To obtain product certifications and the Declaration of Conformity (DoC) for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Environmental Management

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *Minimize Our Environmental Impact* web page at ni.com/environment. This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.

Waste Electrical and Electronic Equipment (WEEE)



EU Customers At the end of the product life cycle, all products *must* be sent to a WEEE recycling center. For more information about WEEE recycling centers, National Instruments WEEE initiatives, and compliance with WEEE Directive 2002/96/EC on Waste Electrical and Electronic Equipment, visit ni.com/environment/weee.

电子信息产品污染控制管理办法 (中国 RoHS)



中国客户 National Instruments 符合中国电子信息产品中限制使用某些有害物质指令(ROHS)。关于 National Instruments 中国 ROHS 合规性信息,请登录 ni.com/environment/rohs_china。(For information about China ROHS compliance, go to ni.com/environment/rohs_china.)

Refer to the NI Trademarks and Logo Guidelines at ni.com/trademarks for more information on National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products/technology, refer to the appropriate location: Help» Patents in your software, the patents.txt file on your media, or the National Instruments Patents Notice at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the readme file for your NI product. Refer to the Export Compliance Information at ni.com/legal/export-compliance for the National Instruments global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data.

© 2004-2013 National Instruments, All rights reserved.

373776J Mar13

NI PXI/PXIe-2569 仕様

100 チャンネル SPST リレーモジュール

このドキュメントには、NI PXI/PXIe-2569 (NI 2569) 汎用リレーモジュールの仕様が記載されています。すべての仕様は事前の通知なしに変更されることがあります。最新の仕様については、ni.com/manuals を参照してください。

トポロジ......100-SPST (ラッチ型)、50-DPST

トポロジ情報については、『NIスイッチヘルプ』を参照してください。



注意 安全規格の詳細については、『はじめにお読みください:安全対策と電磁両立性について』を参照してください。



注意 この製品を使用する場合、シールドされたケーブルおよびアクセサリを使用してください。



注意 このドキュメントに記載されている以外の方法で NI 2569 を使用した場合、その保護機能が損なわれる可能性があります。

仕様値について

「仕様」は、記載された動作条件下で保証される計測器の性能を示します。

「標準仕様」は、記載された動作条件下で大多数の計測器が満たす仕様を示し、 23 ℃で検証されています。標準仕様は保証されている値ではありません。

すべての電圧は特に注釈のない限り、DC、ACpk、もしくはその組み合わせとします。

柔らかな非金属のブラシを使用してデバイスと端子台のほこりを取り除きます。やわらかい糸くずのない湿った布でその他の汚れを取り除きます。洗剤や化学溶剤は使用しないでください。動作させる前に、端子台を完全に乾燥させ汚染物質が付着していないか確認します。

入力特性

すべての入力特性は特に注釈のない限り、DC、AC $_{
m ms}$ 、もしくはその組み合わせとします。

最大スイッチ電圧

チャンネル間	100 V	
チャンネル / グランド間	100 V、	CATI





注意 このモジュールは、Measurement Category I に準拠し、100 V 未満 の信号電圧を許容するように設計されています。また、このモジュールは、 最大 500 V のインパルス電圧に対して耐性があります。Category II、III、ま たは IV の信号を、このモジュールに接続または測定しないでください。ま た、115 または 230 VAC の MAINS 電源回路(例:壁コンセント)に接続し ないでください。測定カテゴリの詳細については、『はじめにお読みくださ い:安全対策と電磁両立性について』を参照してください。危険電圧 (>42.4 V_{pv}/60 VDC) がリレー端子に接続されている場合、安全低電圧 (<42.4 V_{pv}/60 VDC)をその他のリレー端子に接続することはできません。



注意 スイッチ電力は、最大スイッチ電流と最大電圧によって制限されて います。60 W、62.5 VA を超えないように注意してください。

最大スイッチ電力 (チャンネルあたり)......60 W、62.5 VA (DC ~ 60 Hz)

最大電流(スイッチまたは許容、 チャンネルあたり)......1 A

同時に使用可能なチャンネル数 最大電流 (≤35 ℃)......50



メモ 誘導性負荷(モーターおよびソレノイド)を切り替えることで、モ ジュールの定格電圧を超える高電圧過渡信号が生じます。追加の保護を施 さない場合、これらの過渡によってモジュール動作が干渉され、リレー寿 命に影響を与えます。過渡減衰についての詳細は、ni.com/jp/infoで Info Code のフィールドに「ipf8kh」と入力します。

>35 ℃ 時のモジュール負荷の低下

負荷の低下は、周囲温度、および印加されている各チャンネルに対する電流値の 2 乗の合計に依存します。結果は、図 1 の灰色の領域内に位置している必要がありま す。以下の例はこの計算式を示します。

例 1

50 チャンネル使用時 0.75 A 10 チャンネル使用時 0.5 A

(50 × 0.75²) + (10 × 0.5²) = 30.6 A² × チャンネル数


例 1 のモジュールは、0 ~ 55 ℃の周囲温度で適用できます。

例 2

60 チャンネル使用時 0.75 A 35 チャンネル使用時 0.5 A

(60 × 0.75²) + (35 × 0.5²) = 42.5 A² × チャンネル数

図1. モジュール負荷の低下



最小スイッチ負荷......20 mV/10 mA

DC パス抵抗

初期.....<0.55Ω 寿命末期.....≥1Ω

通常、DC パス抵抗は、リレーの寿命が続く間小さい値を保持します。 リレーの寿命 末期時には、パスの抵抗は急速に1Ω以上になります。負荷定格は寿命末期以前の仕 様範囲内で使用されるリレーに適用されます。

接触電位 (23 ℃時標準)......≤12 uV

帯域幅(-3 dB、23 ℃時標準)

50 Ω 終端.....≥20 MHz

クロストーク(23 ℃時標準、50 Ω 終端)

チャンネル間

10 kHz	≤-85 dB
100 kHz	≤-65 dB
1 MHz	≤-45 dB
10 MHz	≤-25 dB

絶縁 (23℃時標準、50 ○終端)

オープンチャンネル

10 kHz	≥85 dB
100 kHz	≥65 dB
1 MHz	≥45 dB
10 MHz	>25 dB

動特性

リレー動作(セット)時間



メモ アプリケーションによっては、より長い整定時間が必要な場合があります。整定時間の追加についての情報は、『NI スイッチヘルプ』を参照してください。

リレー寿命

機械的......1 × 108 サイクル

電気的

10 VDC、100 mADC 抵抗2.5 × 10⁶ サイクル

10 VDC、1 ADC 抵抗......1 × 10⁶ サイクル

30 VDC、1 ADC 抵抗5 × 10⁵ サイクル

60 VDC、1 ADC 抵抗1 × 105 サイクル



メモ NI 2569 で使用されているリレーは、現場交換が可能です。破損したリレーの交換についての情報は、『NI スイッチヘルプ』を参照してください。

トリガ特性

入力トリガ



メモ NI 2569 は、デジタルフィルタを無効にすることによって、150 ns 未満のトリガパルス幅を認識することができます。デジタルフィルタを無効にする方法については、『NI スイッチヘルプ』を参照してください。

出力トリガ

物理特性

所要電力

2.5 W (3.3 V 時) 2.5 W (3.3 V 時) PXIe 互換 21.6 × 2.0 × 13.0 cm $(8.5 \times 0.8 \times 5.1 \text{ in.})$

重量......289 g(10.2 oz)

環境

動作温度......0~55℃ 保管温度......-20 ~ 70 ℃ 最大使用高度 2,000 m 室内使用のみ。

耐衝撃 / 振動

(IEC 60068-2-27 に準拠して試験済み。 MIL-PRF-28800F に準拠してテストプロファ イルを確立。) ランダム振動 動作時......5 ~ 500 Hz、0.3 g_{rms} 非動作時......5 ~ 500 Hz、2.4 g_{rms} (IEC 60068-2-64 に準拠して試験済み。 非動作時のテストプロファイルは MII-PRF-28800F、Class 3 の要件を上回る。)

図 2 は、電源投入状態の NI 2569 を示しています。

図 2. NI 2569 電源投入時の状態




図 3 は、NI 2569 のピン配列を示しています。

		_		١		
COM0 ∘ 15	0	J.		51 50	CH1	00144
CH0 COM2 15				50 52 49	CH3 °	COM1
COM4 • 14	8 0			53	CH5	COM3
CH4 • COM6 • 15	7 0			48 54 47	CH7 °	COM5
CH6 COM8 14	_			55 -	CH9 °	COM7
CH10 COM10 15	5 °			46	CH11	COM9
CH10 • COM10 - 14	4 I ~			57	CH13	COM11
CH12 COM12 6 15	3 0			58	CH15	COM13
CH14 • 14				59	CH17	COM15
CH16 • COM19 = 14	1 °			42	CH19	COM17
CH18 • COMOD = 14	0 1			61	CH21	COM19
CH20 ◆ 10	9 0		0	40		COM21
CH22 COM22 13	2 0	J.	0-	39	CH23	COM23
CH22 • COM24 • 13 CH24 • COM24 • 13	3 0	J.	0-	38	CH25	COM25
CH26 COM26 - 13 CH26 COM28 - 13	4 0	, , ,	0-	37	CH27	COM27
CH28 • COIVIZE 16	5 0	J.	<u> </u>	36	CH29	COM29
CUIVISU TIE	6	J.	-	35	CH31	COM31
CH32 COM32 ∞ 13	7			34	CH33	COM33
CU24 COM34 ∘ 13	8 -			68 33	CH35	COM35
CH26 COM36 0 15	9			33 69 32	CH37	COM37
COM38 • 13	1 0	— r		32 70 31	CH39	
COM40 • 13	0	— r		31 71 30	CH41	COM39
CH40 COM42 17	9 0			30 72 20	°	COM41
COM44 0 12	8 0			73	CH45	COM43
CH44 • COM46 • 17	7 0			74	CH47	COM45
CH44 COM46 • 12 CH46 COM48 • 12	6			75	CH49	COM47
CH40 0 1/	5 1 ~			76	CH51	COM49
CH50 ○ 1/2	6 4			25 77	CH53	COM51
CH52 COM52 12 COM54 12	7 3			24		COM53
CH54 • COM51 17	8 0		0-	20	CH55	COM55
CH56 • 1/	9 0	7,	0-	22	CH57	COM57
CH58 COM58 12	0 0	J.	0-	21	CH59	COM59
CH60 COM60 18		J.	0-	20 -	CH61	COM61
CH62 COIVI62 18	2	J.	<u> </u>	19	CH63	COM63
CH64 COM64 - 18		J.	-	10	CH65	COM65
CUICO - COIVI66 0 18	4 _			17	CH67	COM67
COM68 ∘ 18	5 0	— i		85 16 86	CH69	COM69
CH68 • COM70 • 11 CH70 • COM70 • 11	5	— i		86 15 87	CH71	COM71
CH72 COM72 0 118	7 0	— ·		14	CH73	COM73
COM74 • 11	3 0	— i		13	CH75	
COM76 • 11	2 0	— i		89 12	CH77	COM75
COM78 0 11	1 0			90 0	° CH79	COM77
COM80 -11	o o			91 0	°	COM79
COMP2 ~ 10	9			92	CH83	COM81
CH82 COM84 0 10	8 0			93 0	CH85	COM83
CH84 COM86 19	7 0			0.4	CH87	COM85
COMPS o 10	6			7	CH89	COM87
CH88 COM90 19	5 5	, II	0-	96	CH91	COM89
CH90 0 COM00 a 10	4			5		COM91
CLICO a CONICE 19	7 0	J	0-	4	CH93	СОМ93
CH92 COM94 - 10			<u>~</u>	3	CH95	COM95
CH96 • COIVI96 • 19	9 0	J.	0-	2	CH97	COM97
CH98 ← COM98 ← 10 CH98 ← 10		J.	o-	100	CH99	COM99
-		_		J		



メモ 各トポロジ特有の接続方法については、『NI スイッチヘルプ』でご 使用のデバイストピックから関連ケーブルおよび端子台の取り付け手順を 参照してください。

アクセサリ

以下のアクセサリの詳細については、ni.comを参照してください。

表 1. NI 2569 対応の NI アクセサリ

アクセサリ	製品番号
LFH200 コネクタ⇔ 50 ピン D-SUB スイッチケーブル (チャンネル / コモン間ツイスト)、1 m	779038-02
LFH200 コネクタ⇔裸線スイッチケーブル、2 m	779038-01
NI TBX-50B、50 ピン D-SUB ネジ留め式端子台	782866-01

表 2. NI 2569 対応の他社製アクセサリ

アクセサリ	製造元	製品番号		
端子コネクタ(モジュールあたり4つ必要)	Molex	71715-4002		
プラグコネクタサブアセンブリ	Molex	71719-3000		
バックシェルのみ	Jevons	JDC200B-832		
マスインターコネクトケーブルアセンブリ、20 in.	Virginia Panel	540105010105		
マスインターコネクトケーブルアセンブリ、36 in.	Virginia Panel	540105010205		
メイト ITA モジュール * (モジュールあたり 1 つ必要)	Virginia Panel	510108131		
メイト ITA パッチコード・ (モジュールあたり 198 本必要)	Virginia Panel	720101101		
DAK アセンブリ NI PCB、200 ピン LFH、オス	MAC Panel	561036		
PCB マウント、別のカバーまたはケースが必要です。前述の安全に関する注意事項を参照				

してください。



注意 地域の安全コードと基準、およびコネクタの製造元によって提供され た規格に従って必ずメイトコネクタを取り付けてください。他社製コネクタ の安全適合指令、また該当する基準(北米では UL および CSA、ヨーロッパ では IEC および VDE を含む)に従った使用方法を確認してください。

表 3. LFH200 コネクタ⇔ 50 ピン D-SUB スイッチケーブル対応の他社製アクセサリ

アクセサリ	製造元	製品番号
VARIOFACE モジュール、ネジ接続および 50 D-SUB ピンストリップ付	Phoenix Contact	FLK-D50 SUB/S
VARIOFACE モジュール、ネジ接続および 50 D-SUB ピンストリップ付	Phoenix Contact	FLKM-D50 SUB/S
VARIOFACE モジュール、ネジ接続および 50 D-SUB ピンストリップ付	Phoenix Contact	FLKMS-D50 SUB/S
VARIOFACE モジュール、ネジ接続および 50 D-SUB ピンストリップ付、LED 表示器付	Phoenix Contact	FLKM-D50 SUB/S/LA

認可および準拠

安全性

この製品は、計測、制御、実験に使用される電気装置に関する以下の規格および安全 性の必要条件を満たします。

- IEC 61010-1, EN 61010-1
- UL 61010-1、CSA 61010-1



メモ UL およびその他の安全保証については、製品ラベルまたは「オンラ イン製品認証しセクションを参照してください。

雷磁面立性

この製品は、計測、制御、実験に使用される電気装置に関する以下の EMC 規格の必 要条件を満たします。

- EN 61326-1 (IEC 61326-1): Class A エミッション、基本イミュニティ
- EN 55011 (CISPR 11): Group 1、Class A エミッション
- AS/NZS CISPR 11: Group 1、Class A エミッション
- FCC 47 CFR Part 15B: Class A エミッション
- ICES-001: Class A エミッション



メモ EMC 宣言および認証については、「オンライン製品認証」セクショ ンを参照してください。

CE マーク準拠(🤇

この製品は、該当する EC 理事会指令による基本的要件に適合しています。

- 2006/95/EC、低電圧指令(安全性)
- 2004/108/EC、電磁両立性指令(EMC)

オンライン製品認証

この製品の製品認証および適合宣言(DOC)を入手するには、ni.com/certification(英語)にアクセスして型番または製品ラインで検索し、保証の欄の該当するリンクをクリックしてください。

環境管理

ナショナルインスツルメンツは、環境に優しい製品の設計および製造に努めています。NI は、製品から特定の有害物質を除外することが、環境および NI のお客様にとって有益であると考えています。

環境に関する詳細は、ni.com/environment からアクセス可能な「Minimize Our Environmental Impact」ページ(英語)を参照してください。このページには、ナショナルインスツルメンツが準拠する環境規制および指令、およびこのドキュメントに含まれていないその他の環境に関する情報が記載されています。

廃電気電子機器(WEEE)



欧州のお客様へ 製品寿命を過ぎたすべての製品は、必ず WEEE リサイクルセンターへ送付してください。WEEE リサイクルセンターおよびナショナルインスツルメンツの WEEE への取り組み、および廃電気電子機器のWEEE 指令 2002/96/EC 準拠については、ni.com/environment/weee (英語)を参照してください。

电子信息产品污染控制管理办法 (中国 RoHS)



中国客户 National Instruments 符合中国电子信息产品中限制使用某些有害物质指令(RoHS)。关于 National Instruments 中国 RoHS 合规性信息,请登录 ni.com/environment/rohs_china。(For information about China RoHS compliance, go to ni.com/environment/rohs_china.)

National Instruments の商標の詳細については、ni.com/trademarks に掲載されている「NI Trademarks and Logo Guidelines」をご覧下さい。本文書中に記載されたその他の製品名および企業名は、それぞれの企業の商標または商号です。National Instruments の製品 / 技術を保護する特許については、ソフトウェアで参照できる特許情報 (ヘルプー特許情報)、メディアに含まれている patents.txt ファイル、または「National Instruments Potent Notice」 (ni.com/patents) のうち、該当するリソースから参照してください。エンドユーザ使用許諾契約 (EULA) に関する情報および他社製品の法的注意事項はで使用の NI 製品の Readme ファイルにあります。ナショナルインスツルメンツの輸出関連法規遵守に対する方針について、また必要な HTS コード、ECCN、その他のインボート / エクスボートデータを取得する方法については、「輸出関連法規の遵守に関する情報」 (ni.com/legal/export-compliance)を参照してください。

© 2004-2013 National Instruments. All rights reserved.