

NI PXI-2722 Specifications

16-Bit Resistor Module

This document lists specifications for the National Instruments PXI-2722 16-bit resistor module. All specifications are subject to change without notice. Visit ni.com/manuals for the most current specifications.

Topology Independent

Refer to the *NI Switches Help* for detailed topology information.



Caution To ensure the specified EMC performance, operate this product only with shielded cables and accessories.



Caution Refer to the *Read Me First: Safety and Electromagnetic Compatibility* document at ni.com/manuals for important safety and compliance information.



Caution Using the NI PXI-2722 in a manner not described in this document may impair the protection the NI PXI-2722 provides.

About These Specifications

Specifications characterize the warranted performance of the instrument under the stated operating conditions.

Typical Specifications are specifications met by the majority of the instrument under the stated operating conditions and are tested at 23 °C. Typical specifications are not warranted.

All voltages are specified in DC, AC_{pk} , or a combination unless otherwise specified.

Input Characteristics

Number of channels 5

Maximum voltage

Channel-to-channel 60 VDC, 30 VAC_{rms} ,
CAT I

Channel-to-ground..... 60 VDC, 30 VAC_{rms} ,
CAT I

Maximum current 0.3 A (per channel)

Maximum power0.25 W
(per channel)

Resistor accuracy.....0.5%

TCR 100 ppm

Resolution..... 0.25 Ω steps

Range..... 0.25 Ω to 16,320 Ω (per
channel)

Typical channel accuracy

Channel set between

20 and 40 Ω <10.0%

Channel set between

40 and 120 Ω <5.0%

Channel set between

120 and 400 Ω <2.0%

Channel set between

400 and 16,000..... <1.0%

Channel accuracy is a function of the overall resistor accuracy, plus relay and trace path resistance, plus any affects of temperature.

Typical minimum resistance2 Ω

Typical DC offset resistance
(per channel)

Initial<2 Ω

End-of-life..... $\geq 3 \Omega$

Channel resistance is a combination of relay contact resistance, trace resistance, and resistor value. Contact resistance typically remains low for the life of a relay. At the end of relay life, the offset resistance may rapidly rise above 3 Ω .

The following figures show the expected resistance error and the expected resistance.




Figure 1. Expected resistance error between 0 and 16000 Ω




Figure 2. Expected resistance error between 0 and 100 Ω




Figure 3. Expected resistance between 0 and 16000 Ω




Figure 4. Expected resistance between 0 and 50 Ω

Dynamic Characteristics

Simultaneous drive limit..... 89 relays

Relay operate/release time 250 μs



Note Certain applications may require additional time for proper settling. Refer to the *NI Switches Help* for information about including additional settling time.

Typical relay life

Low power load (<50 mW) 1×10^8 cycles

Full power load (<250 mW) 1×10^6 cycles



Note Reed relays are highly susceptible to damage caused by switching capacitive and inductive loads. Capacitive loads can cause high inrush currents while inductive loads can cause high flyback voltages. The addition of appropriate resistive protection can greatly improve contact lifetime. For more information about adding protection circuitry to a capacitive load, visit ni.com/info and enter the Info Code `relaylifetime`. For information about inductive loads, enter the Info Code `relayflyback`.



Note The relays used in the NI PXI-2722 are field replaceable. Refer to the *NI Switches Help* for information about replacing a failed relay.

Trigger Characteristics

Input trigger

Sources PXI trigger lines 0–7,

Minimum pulse width 150 ns



Note The NI PXI-2722 can recognize trigger pulse widths that are less than 150 ns by disabling digital filtering. For information about disabling digital filtering, refer to the *NI Switches Help*.

Physical Characteristics

Relay types..... Reed



Note NI advises against installing reed relay modules directly adjacent to an embedded controller with a magnetic hard drive because of the sensitivity of reed relays and the possibility of interference.

Front panel connector 37-pin D-SUB connector

DMM port 2 × 2 Micro-Fit connector

PXI power requirement..... 1.2 W at 3.3 V
4.2 W typical,
11 W maximum
at 5 V

Dimensions (L × W × H) 3 U, one slot,
PXI/cPCI module,
21.6 × 2.0 × 13.0 cm
(8.5 × 0.8 × 5.1 in.)

Weight..... 266 g (9.3 oz)

Environment

Operating temperature 0 °C to 55 °C

Storage temperature –20 °C to 70 °C

Relative humidity..... 5% to 85%
noncondensing

Pollution Degree 2

Maximum altitude..... 2,000 m

Indoor use only.

Shock and Vibration

Operational shock.....30 g peak, half-sine,
11 ms pulse
(Tested in accordance
with IEC 60068-2-27.
Test profile developed
in accordance with
MIL-PRF-28800F.)

Random vibration

Operating5 to 500 Hz, 0.3 g_{rms}

Nonoperating5 to 500 Hz, 2.4 g_{rms}

(Tested in accordance
with IEC 60068-2-64.
Nonoperating test profile
exceeds the requirements
of MIL-PRF-28800F,
Class 3.)

Diagrams

Refer to the following figure for the power-on-state diagram of the NI PXI-2722.




Figure 5. NI PXI-2722 Power-On-State

The following figure shows the NI PXI-2722 connector pinout.




Figure 6. NI PXI-2722 Connector Pinout



Note For topology-specific connection information, refer to your device in the *NI Switches Help* and the installation instructions for any associated cables or terminal blocks.

Accessories

Visit ni.com for more information about the following accessories.

Table 1. Accessories Available for the NI PXI-2722

| Accessory | Part Number |
|---|-------------|
| 37-pin female to male shielded I/O cable, 1 m | 778621-01 |
| 37-pin female to male shielded I/O cable, 2 m | 778621-02 |
| 37-pin solder cup terminals, D-SUB female shell w/strain relief | 779184-01 |
| High-voltage DIN-rail screw terminal block | 779491-01 |
| Crimp-and-Poke 37-pin D-SUB custom connectivity accessory | 779185-01 |
| Solder cup 37-pin D-SUB custom connectivity accessory | 779184-01 |
| 2 x 2 microfit DMM connector | 782465-01 |



Caution You *must* install mating connectors according to local safety codes and standards and according to the specifications provided by the connector manufacturer. You are responsible for verifying safety compliance of third-party connectors and their usage according to the relevant standard(s), including UL and CSA in North America and IEC and VDE in Europe.

Compliance and Certifications

Safety

This product meets the requirements of the following standards of safety for electrical equipment for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA 61010-1



Note For UL and other safety certifications, refer to the product label or the [Online Product Certification](#) section.

Electromagnetic Compatibility

This product meets the requirements of the following EMC standards for electrical equipment for measurement, control, and laboratory use:

- EN 61326-1 (IEC 61326-1): Class A emissions; Basic immunity
- EN 55011 (CISPR 11): Group 1, Class A emissions
- AS/NZS CISPR 11: Group 1, Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



Note In the United States (per FCC 47 CFR), Class A equipment is intended for use in commercial, light-industrial, and heavy-industrial locations. In Europe, Canada, Australia and New Zealand (per CISPR 11) Class A equipment is intended for use only in heavy-industrial locations.



Note Group 1 equipment (per CISPR 11) is any industrial, scientific, or medical equipment that does not intentionally generates radio frequency energy for the treatment of material or inspection/analysis purposes.



Note For EMC declarations and certifications, refer to the [Online Product Certification](#) section.

CE Compliance

This product meets the essential requirements of applicable European Directives as follows:

- 2006/95/EC; Low-Voltage Directive (safety)
- 2004/108/EC; Electromagnetic Compatibility Directive (EMC)

Online Product Certification

To obtain product certifications and the Declaration of Conformity (DoC) for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Environmental Management

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *NI and the Environment* Web page at ni.com/environment. This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.

Waste Electrical and Electronic Equipment (WEEE)



EU Customers At the end of the product life cycle, all products *must* be sent to a WEEE recycling center. For more information about WEEE recycling centers, National Instruments WEEE initiatives, and compliance with WEEE Directive 2002/96/EC on Waste Electrical and Electronic Equipment, visit ni.com/environment/weee.

电子信息产品污染控制管理办法（中国 RoHS）



中国客户 National Instruments 符合中国电子信息产品中限制使用某些有害物质指令 (RoHS)。关于 National Instruments 中国 RoHS 合规性信息, 请登录 ni.com/environment/rohs_china。(For information about China RoHS compliance, go to ni.com/environment/rohs_china.)

LabVIEW, National Instruments, NI, ni.com, the National Instruments corporate logo, and the Eagle logo are trademarks of National Instruments Corporation. Refer to the *Trademark Information* at ni.com/trademarks for other National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products/technology, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your media, or the *National Instruments Patent Notice* at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the NI-272x Reference VIs Readme. Refer to the *Export Compliance Information* at ni.com/legal/export-compliance for the National Instruments global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data..