



8930 S. Beck Avenue, Suite 107 Tempe, AZ 85284

Phone: (480) 240-3400

Fax: (480) 240-3401

E-mail: ftimarket@ftimeters.com

Website: www.ftimeters.com

Certificate #68005930

Customer: FUERZA AERA DE CHILE
Job No: 135025.1-1
P.O. Number: VERONICA CC
Meter Model No: FT-12AEYS-LEA-5
Meter Serial No: 180612M67753
Linearity% +/- 0.2960 (Over 10:1 Range)
Calib Inv No: 51868
Report No: 251223B
Calibrator Recal Date: 6/23/2026

Cust Part No: N/A
Meter ID (in): 0.750
End Fitting: AN
Bearing Type: BALL
Pickoff Type: RF
Pickoff P/N: 27-31949-101
Signal Output: N/A
Procedure: CP-08

Ambient Temp: 75 °F
Ambient Pres: 14.1 PSIA
Humidity: 26%
Calib Fluid: MIL-PRF-7024E
Uncertainty: +/- .05%
Fluid Visc: 1.269 cSt
Density: 6.323 Lb/Gal
Average K Factors: 5877.232 P/Gal

Notes:

* Mass flow rate based on density of 6.32319 Lb/Gal

Meter Freq Hz	Vol Flow Rate GPM	Mass Flow Rate PPH	Meter K Factor P/Gal	Meter K Factor P/Lb	Fluid Density Lb/Gal	Roshko Corrected F/V (70°F)	Strout al Corrected KFactor (70°F)	Reynolds Number	Pressure Drop PSIA	Freq / Viscosity Hz/cSt	Meter Temp °F
2505.226	25.49846	9673.893	5895.006	932.284	6.323	1974.171	5896.180	84717.09	-	1973.909	77.5
1922.273	19.58114	7428.911	5890.178	931.520	6.323	1515.442	5891.374	65084.88	-	1515.236	77.6
1528.454	15.57920	5910.611	5886.517	930.942	6.323	1205.279	5887.727	51796.16	-	1205.113	77.7
1201.261	12.25967	4651.212	5879.083	929.766	6.323	947.468	5880.302	40768.34	-	947.337	77.8
940.043	9.60583	3644.368	5871.699	928.598	6.323	741.498	5872.921	31945.82	-	741.395	77.8
729.100	7.45698	2829.112	5866.450	927.768	6.323	575.096	5867.670	24798.94	-	575.017	77.8
571.362	5.84520	2217.619	5864.935	927.528	6.323	450.634	5866.149	19436.97	-	450.572	77.8
443.899	4.54089	1722.775	5865.359	927.595	6.323	350.048	5866.565	15097.39	-	350.000	77.7
358.258	3.66803	1391.620	5860.211	926.781	6.323	282.472	5861.408	12193.57	-	282.433	77.7
277.635	2.83858	1076.932	5868.457	928.085	6.323	218.878	5869.649	9435.13	-	218.848	77.6
218.671	2.23343	847.344	5874.483	929.038	6.323	172.375	5875.671	7422.91	-	172.352	77.6
169.799	1.73145	656.898	5884.055	930.552	6.323	133.847	5885.243	5754.43	-	133.829	77.6
135.610	1.37781	522.728	5905.471	933.939	6.323	106.897	5906.665	4579.09	-	106.882	77.6
106.171	1.07638	408.369	5918.234	935.957	6.323	83.694	5919.431	3577.43	-	83.683	77.6
81.948	0.82832	314.257	5935.974	938.763	6.323	64.600	5937.176	2753.02	-	64.591	77.6
65.600	0.66499	252.291	5918.853	936.055	6.323	51.708	5920.047	2209.99	-	51.701	77.6
50.339	0.51135	194.000	5906.696	934.133	6.323	39.673	5907.879	1699.12	-	39.668	77.5
38.666	0.39495	149.842	5873.938	928.952	6.323	30.469	5875.108	1312.21	-	30.465	77.5
29.864	0.30776	116.760	5822.189	920.768	6.324	23.529	5823.339	1022.32	-	23.526	77.4
23.775	0.24779	94.011	5756.843	910.434	6.324	18.729	5757.972	823.01	-	18.727	77.3

Calibration Technician: M. SKINNER

Date: February 23, 2026

Certified By: M. Skinner

Calibration Technician

On Behalf of Flow Technology, LLC

Statement of Calibration and Conformance

We hereby certify that the above unit conforms to the applicable specifications as referenced in, or furnished with this purchase order and has been calibrated using standards traceable to the National Institute of Standards and Technology and through the Certificate Number listed above. Evidence of traceability is archived within the manufacturing plant and is subject to examination upon request. FTI Flow Technology, LLC's calibration systems is in compliance with ANSI/NCCL Z540-1-1994. Results apply only to the unit calibrated and shall not be reproduced, except in full, without the written approval of FTI Flow Technology, LLC.