



WS Technologies Inc.

Certificate of Calibration

FUERZA AEREA DE CHILE

Certificate #: 10385 - 27012026

Calibration Date: January 26, 2026

Calibration Due: January 26, 2028

Equipment: Beacon Tester

Model Series: BT200-2100Y

Serial Number: 10385

WS Technologies Inc. certifies that this instrument has been calibrated to manufacturer specifications and accuracies using instruments and standards which are traceable to the National Institute of Standards and Technology. The instruments or standards used for this calibration are calibrated on a schedule which is established to maintain traceability at the required accuracy level.

Claudia Valverde

This certificate shall not be reproduced, except in full, without the written approval from WS Technologies Inc.



WS Technologies Inc.

Calibration Standards and Testing Instruments Used:

Test Instrument	Serial #	Cal Date	Due Date
Trimble Thunderbolt	31390046	Dec 23, 2025	Dec 23, 2027
Keysight N1913A Power Meter	MY58520009	Dec 11, 2025	Dec 11, 2026
Keysight E9301B Sensor Assembly	MY58170004	Dec 11, 2025	Dec 11, 2027
Extech RH300 Psychrometer	T 2532162	Jan 09, 2026	Jan 09, 2027
WST Beacon Simulator	TJ153-1	Apr 21, 2025	Apr 21, 2026

Calibration Measurement Data Taken On January 26, 2026

Calibration Procedure:	PP-910-BT200				
Beacon Tester Model:	BT200-2100Y	Serial Number:	10385		
Date of Calibration:	January 26, 2026	Calibration Due Date:	January 26, 2028		
Laboratory Temperature:	21°C ± 3°C	Laboratory Humidity:	15% ± 10%		
As Found Condition:	In tolerance	As Left Condition:	In tolerance		
Parameter	Reference	Lower Limit*	Upper Limit*	Measured Data	Pass
15 Hex ID	2788334E1EFFBFF	Must be 2788334E1EFFBFF		2788334E1EFFBFF	✓
406 Frequency	406.050000 MHz	406.049950 MHz	406.050050 MHz	406.050001 MHz	✓
406 Power	38.0 dBm	37.5 dBm	38.5 dBm	37.6 dBm	✓
406 Positive Phase	1.10 rad	1.05 rad	1.15 rad	1.11 rad	✓
406 Negative Phase	-1.10 rad	-1.15 rad	-1.05 rad	-1.10 rad	✓
406 Modulation Rise Time	200.0 µs	178.0 µs	222.0 µs	198.0 µs	✓
406 Modulation Fall Time	200.0 µs	178.0 µs	222.0 µs	207.9 µs	✓
406 Modulation Symmetry	0.0 %	0.0 %	5.0 %	0.0 %	✓
406 Modulation Bitrate	400.0 bps	399.6 bps	400.4 bps	400.0 bps	✓
406 CW Preamble	160.0 ms	158.0 ms	162.0 ms	159.7 ms	✓
406 Spectral Mask	Pass	Must be Pass		Pass	✓
121 Frequency	121.500000 MHz	121.499940 MHz	121.500060 MHz	121.499982 MHz	✓
121 Peak Power	21.0 dBm	19.7 dBm	22.3 dBm	21.2 dBm	✓
121 Sweep Direction	Upwards	Must be Upwards		Upwards	✓
121 Audio Frequency - Lower	500 Hz	465 Hz	535 Hz	500 Hz	✓
121 Audio Frequency - Upper	1500 Hz	1455 Hz	1545 Hz	1500 Hz	✓
121 Duty Cycle	40 %	37 %	43 %	42 %	✓
121 Modulation Factor	85 %	78 %	92 %	84 %	✓
121 Sweep Repetition Rate	3.0 Hz	2.9 Hz	3.1 Hz	3.0 Hz	✓
243 Frequency	243.000000 MHz	242.999940 MHz	243.000060 MHz	242.999987 MHz	✓
243 Peak Power	21.0 dBm	19.7 dBm	22.3 dBm	21.1 dBm	✓
243 Sweep Direction	Upwards	Must be Upwards		Upwards	✓
243 Audio Frequency - Lower	500 Hz	465 Hz	535 Hz	500 Hz	✓
243 Audio Frequency - Upper	1500 Hz	1455 Hz	1545 Hz	1500 Hz	✓
243 Duty Cycle	40 %	37 %	43 %	41 %	✓
243 Modulation Factor	85 %	78 %	92 %	86 %	✓
243 Sweep Repetition Rate	3.0 Hz	2.9 Hz	3.1 Hz	3.0 Hz	✓
AIS1 Frequency	161.975000 MHz	161.974940 MHz	161.975060 MHz	161.975019 MHz	✓
AIS1 Power	35.0 dBm	34.0 dBm	36.0 dBm	35.1 dBm	✓
AIS2 Frequency	162.025000 MHz	162.024940 MHz	162.025060 MHz	162.024985 MHz	✓
AIS2 Power	35.0 dBm	34.0 dBm	36.0 dBm	34.9 dBm	✓
AIS Message CRC	Pass	Must be Pass		Pass	✓

* Limits derived from both test equipment uncertainties and Beacon Tester stated uncertainties

Technician: Brittany Meek

Signature:

Date: January 26, 2026