

BARFIELD INC.
4101 N.W. 29TH ST.
Miami FL - 33142
(800) 321-1039



**CERTIFICATE OF CALIBRATION/
FUNCTIONAL TEST**

W.O. No.: WO1253732

Customer Name FUERZA AEREA DE CHILE

Date of Certification 4/21/2026 Certification Due Date: 4/21/2027
Part Number: 101-01175 Received in Tolerance: NO
Serial Number: 00792 Returned in Tolerance: YES
Description: DIGITAL PITOT STATIC, T/S, DPS1000 Temperature: 74°
Manufacturer: BARFIELD INC. Humidity: 43%

Calibration Procedure Used: 20-101-01175

Deviation: N/A

STANDARDS USED TO CERTIFY EQUIPMENT

Equipment	Serial Number.	Last Date Calibrated	Next Calibration Date
3885E	50614	3/26	9/26
3886E	54891	3/26	9/26
1254EA	1066400001	3/26	3/27
4515E-1	MY40019256	9/25	9/26
4515E-2	MY45026735	4/26	4/27

Comments UNIT FAILED TEST DUE TO DEFECTIVE SD CARD.

N.B: RECEIVED WITH ACCESSORY BAG.

Results: REPLACED LISTED PART, CALIBRATED AND TESTED W/I MFG SPECS.

This is to certify that the above listed part number meets or exceeds all specifications as stated in the referenced procedure at the points tested (unless otherwise noted). Calibration Standards used in the calibration of this product are traceable to the National Institute of Standards and Technology (NIST). Unless otherwise stated the uncertainty of the measurement standards shall not exceed 25% of the acceptable tolerance. This calibration is in accordance with Barfield Inc. Quality Management System (QPM-7.6) which is ISO 9001:2015/AS9100 D certified and compliant with MIL-STD-45662A, ANSI/NCSS Z540.1-1994 (R2002) and ISO 10012.

This certificate applies only to item identified and shall not be reproduced, without the permission from Barfield Inc.

Certified by: Freddy Duque

Title/Stamp Q.C. INSPECTOR

F.D. 01045

Dated: 4/21/2026

DPS1000 (MOD "B" - Units)

Outgoing - As Returned to Customer

Calibration Confirmation

UUT S/N:	00792	ACS Version:	0	Date Last Calibrated:	3/26/2020
Customer:	TA SPA-DESAROLLO DE TECNOLC	Limit Parameters:	1	Months in Service:	73
Date:	4/21/2026	Primary Standard:	Ruska 2465	Calibrated by:	C.CASTRO
W/O:	WO1253732	Working Standards:	0	MOD:	A, B, C
HUM Version:		5.0.0	PCM version:		6.0

Ps Performance Data

Ps Channel Leak Test:		-9.4	ft/min @ 5,000 ft	Tol (ft/min)		25	P	
Ps Channel Leak Test:		-12.3	ft/min @ 20,000 ft			50	P	
Test Point	Primary Standard Ps Channel inHg	Equivalent Pressure Altitude ft	UUT Reading inHg	UUT Deviation Note(1) inHg	UUT Deviation ft	UUT Deviation Tolerance (±) inHg	UUT Deviation Tolerance (±) ft	Status P: Pass OT: Note(2) Fail: Note(3)
1	2.9648	53000	2.9645	-0.0003	2	0.0030	21	P
2	3.0369	52500	3.0366	-0.0003	2	0.0020	14	P
3	3.4247	50000	3.4243	-0.0004	2	0.0020	12	P
4	7.0406	35000	7.0408	0.0002	-1	0.0020	6	P
5	11.1035	25000	11.1038	0.0003	-1	0.0020	4	P
6	14.9421	18000	14.9424	0.0003	-1	0.0020	3	P
7	19.0293	12000	19.0298	0.0005	-1	0.0020	3	P
8	23.9782	6000	23.9791	0.0009	-1	0.0020	2	P
9	26.8167	3000	26.8177	0.0010	-1	0.0020	2	P
10	29.9213	0	29.9224	0.0011	-1	0.0020	2	P
11	32.1480	-2000	32.1495	0.0015	-1	0.0020	2	P
12	38.0000	-6767	38.0014	0.0014	-1	0.0038	3	P
Ref Clock 20 MHz						20.00	MHz	
Ps1 Tcomp ADC Ref						1.2496	Vdc	

Pt Performance Data

Tol (kts/min)										
Pt Channel Leak Test:			-0.6	kts/min @ 200 kts		1.0		P		
Test Point	Primary Standard Ps Channel inHg	Primary Standard Pt Channel inHg	Primary Qc Pressure Qc = Pt-Ps inHg <diff>	Equivalent CAS kts	UUT Reading Qc inHg	UUT Deviation Note(1) inHg	UUT Deviation kts	UUT Deviation Tolerance (±) inHg	UUT Deviation Tolerance (±) kts	Status P: Pass OT: Note(2) FAIL: Note(3)
1	29.9213	29.9213	0.0000	0	0.0000	0.0000	0.0	0.0010	4.6	P
2	29.9213	29.9405	0.0192	20	0.0207	0.0015	0.8	0.0019	1.0	P
3	29.9213	30.0411	0.1198	50	0.1212	0.0014	0.3	0.0026	0.5	P
4	29.9213	30.4027	0.4814	100	0.4826	0.0012	0.1	0.0032	0.3	P
5	29.9213	33.0215	3.1002	250	3.1023	0.0021	0.1	0.0056	0.2	P
6	29.9213	38.3063	8.3850	400	8.3908	0.0058	0.1	0.0084	0.2	P
7	29.9213	43.6969	13.7756	500	13.7817	0.0061	0.1	0.0148	0.2	P
8	29.9213	50.9962	21.0749	600	21.0810	0.0061	0.1	0.0195	0.2	P
9	29.9213	55.5106	25.5893	650	25.5952	0.0059	0.1	0.0236	0.2	P
10	29.9213	69.9213	40.0000	N/A	40.0006	0.0006	N/A	0.0395	N/A	P

Calibration Coefficients - As Returned to Customer

Ps Calibration Coefficients		Pt Calibration Coefficients	
Ps Transducer P/N:	615-00010	Qc Transducer P/N:	615-00030
Ps Transducer S/N:	229844	Qc Transducer S/N:	226093
ps_a1 Offset (inHg):	-0.001241	qc_a1 Offset (inHg):	0.000059
ps_a2 Span:	0.999957	qc_a2 Span:	0.999005
ps_cc Correlation Coef:	1.000000	qc_cc Correlation Coef:	1.000000
Ref Clock 20 MHz:	19.999990		
Ps1 Tcomp ADC Ref:	1.2496		

Operational Hours:	9:08
Power ON cycles	11
Pressure Pump:	1:40
Vacuum Pump:	1:40

Note (1)	UUT Status of Pass, OT, FAIL is determined by UUT Deviation inHg (not Deviation in ft or kts).
Note (2)	"OT": Out of Tolerance. If Outgoing Validation results in 1 or more Test Points Status: "OT", the calibration must be performed again. All Test Points must result in "P" for Pass within UUT Deviation Tolerance inHg specification.
Note (3)	If 1 or more Test Points results in "FAIL", then transducer is significantly (>= 2x) out of Tolerance. When Status: "FAIL", troubleshoot to correct problem. Transducer may require replacing.
Note (4)	Status: "P" must be obtained for ALL test points. If 1 or more result in Status: "OT" Out of Tolerance, the calibration must be performed again. If 1 or more result in Status: "FAIL", troubleshooting and possible transducer replacement required.
Note	MOD "B" units configured with Esterline Ps transducer, accuracy rated up to 53,000 ft.

DPS1000 (MOD "B" - Units)

Incoming - As Received

Calibration Check

UUT S/N: 00792	ACS Version: 0	Date Last Calibrated: 3/26/2020
Customer: PA-DESAROLLO DE TECN	Limit Parameters: 1	Months in Service: 73
Date: 4/21/2026	Primary Standard: Ruska 2465	Calibrated by: C.CASTRO
W/O: WO1253732	Working Standards: 0	MOD: A, C

HUIM Version: 5.0.0

PCM version: 6.0

Ps Performance Data

Ps Channel Leak Test: -4.40 ft/min @ 5,000 ft	Tol (ft/min) 25	P
Ps Channel Leak Test: -12.30 ft/min @ 20,000 ft	50	P

Test Point	Primary Standard Ps Channel inHg	Equivalent Pressure Altitude ft	UUT Reading inHg	UUT Deviation Note(1) inHg	UUT Deviation ft	UUT Deviation Tolerance (±) inHg	UUT Deviation Tolerance (±) ft	Status P: Pass OT: Note(2) FAIL: Note(3)
1	2.9648	53000	2.9660	0.0012	-8	0.0068	48	P
2	3.0369	52500	3.0383	0.0014	-10	0.0053	36	P
3	3.4247	50000	3.4261	0.0014	-9	0.0053	32	P
4	7.0406	35000	7.0425	0.0019	-6	0.0053	16	P
5	11.1035	25000	11.1055	0.0020	-4	0.0053	11	P
6	14.9421	18000	14.9442	0.0021	-3	0.0053	9	P
7	19.0293	12000	19.0315	0.0022	-3	0.0053	7	P
8	23.9782	6000	23.9807	0.0025	-3	0.0053	6	P
9	26.8167	3000	26.8191	0.0024	-2	0.0053	5	P
10	29.9213	0	29.9241	0.0028	-3	0.0053	5	P
11	32.1480	-2000	32.1508	0.0028	-2	0.0053	5	P
12	38.0000	-6767	38.0034	0.0034	-2	0.0068	5	P

Pt Performance Data

Pt Channel Leak Test: -0.10 kts/min @ 200 kts							Tol (kts/min) 1.0			P
Test Point	Primary Standard Ps Channel inHg	Primary Standard Pt Channel inHg	Primary Qc Pressure Qc = Pt-Ps inHg <diff>	Equivalent CAS kts	UUT Reading Qc inHg	UUT Deviation Note(1) inHg	UUT Deviation kts	UUT Deviation Tolerance (±) inHg	UUT Deviation Tolerance (±) kts	Status P: Pass OT: Note(2) FAIL: Note(3)
1	29.9213	29.9213	0.0000	0	0.0001	0.0001	1.4	0.0015	5.6	P
2	29.9213	29.9405	0.0192	20	0.0208	0.0016	0.8	0.0029	1.6	P
3	29.9213	30.0411	0.1198	50	0.1215	0.0017	0.4	0.0036	0.8	P
4	29.9213	30.4027	0.4814	100	0.4831	0.0017	0.2	0.0042	0.4	P
5	29.9213	33.0215	3.1002	250	3.1057	0.0055	0.2	0.0066	0.3	P
6	29.9213	38.3063	8.3850	400	8.3990	0.0140	0.3	0.0116	0.3	OT
7	29.9213	43.6969	13.7756	500	13.7955	0.0199	0.3	0.0178	0.3	OT
8	29.9213	50.9962	21.0749	600	21.1016	0.0267	0.3	0.0225	0.3	OT
9	29.9213	55.5106	25.5893	650	25.6200	0.0307	0.3	0.0266	0.3	OT
10	29.9213	69.9213	40.0000	N/A	40.0398	0.0398	N/A	0.0425	N/A	P

Calibration Coefficients - As Received

Ps Coefficients	
Ps Transducer P/N:	615-00010
Ps Transducer S/N:	229844
ps_a1 Offset (inHg):	0.000000
ps_a2 Span:	1.000000
ps_cc Correlation Coef:	1.000000
Ref Clock 20 MHz:	20.000000
Ps1 Tcomp ADC Ref:	1.2500

Pt Coefficients	
Qc Transducer P/N:	615-00030
Qc Transducer S/N:	226093
qc_a1 Offset (inHg):	0.000000
qc_a2 Span:	1.000000
qc_cc Correlation Coef:	1.000000

Note (1)	UUT Status of Pass, OT, FAIL is determined by UUT Deviation inHg (not Deviation in ft or kts).
Note (2)	"OT": Out of Tolerance.
Note (3)	If 1 or more Test Points results in "FAIL", then transducer is significantly ($\geq 2x$) out of Tolerance. When Status: "FAIL", troubleshoot to correct problem. Transducer may require replacing. Note: The customer should be notified if "FAIL" status occurs. The customer should be provided a copy of this Incoming (As-Received) results so they can evaluate the impact it may created on recent aircraft evaluated with this unit.
Note	MOD "B" units are configured with Esterline Ps transducer, accuracy rated up to 53,000 ft.



varCompany Name

4101 NW 29th Street

Miami, FL, 33142

Repair Station no:

Teardown

WO: WO1253732

Cust. RO: 61793

To:

DTS SPA-DESAROLLO DE TECNOLOGIAS Y
SISTEMAS
RODRIGO DE ARAYA #1263
SANTIAGO 7810161
CHILE

Ship To:

DESAROLLO DE TECNOLOGIAS Y SISTEMAS SpA
RODRIGO DE ARAYA #1263
SANTIAGO 7810161
CHILE

Incoming PN: 101-01175

Descr: DIG. AIR DATA T/S, DPS1000

S/N: 00792

Incoming Work Req: CALIBRATION

OutgoingPN: 101-01175

Mfg: 21844

Warranty: No

Outgoing Work Req: REPAIRED

Removal Date:

Tail Number:

Station:

Position:

Manuals: TYPE: CP PUB: CP 20-101-01175 REV: A DATE: 4/1/2022

SYMPTOMS:

REASON FOR REMOVAL

FOR REPAIR AND CALIBRATION

FAULTS:

FAULT PRELIMINARY

THIS IS A PRELIMINARY QUOTE, ADDITIONAL PARTS COULD BE ADDED.

UNIT FAILS TEST(S) DUE TO THE FOLLOWING:

unit came without carrying bag for hoses. recommend carrying bag for hoses

warranty seal lifted. unit not covered by warranty.

UNIT RECEIVED BLACK SCREEN DUE TO CORRUPT SD CARD

REQUIRES REPLACEMENT.

UNIT FAILED TEST DUE TO OUT OF CALIBRATION

CUSTOMER REPORTED: FOR REPAIR AND CALIBRATION.

CORRECTIVE ACTIONS:

REPAIRED

REPLACED NEW SD CARD.

CALIBRATION FIXED FAULT. UNIT CALIBRATE AND
TESTED W/ MFG. SPECS.

Technician: LENET PETIONET

Date: 5/18/2026

Inspector: FREDDY DUGUE

Signature: 