8930 S. Beck Avenue, Suite 107 Tempe, AZ 85284

Phone: (480) 240-3400 **Fax:** (480) 240-3401 **E-mail:** ftimarket@ftimeters.com **Website:** www.ftimeters.com

Certificate #68006417

Customer:	FUERZA AERA DE CHILE				
Job No:	139016.1-1	Cust Part No:	N/A	Ambient Temp:	75 °F
P.O. Number:	62471	Meter ID (in)	0.750	Ambient Pres:	14.1 PSIA
Meter Model No:	FT-12AEYS-LEA-5	End Fitting:	AN	Humidity:	40%
Meter Serial No:	12013005	Bearing Type:	BALL	Calib Fluid:	MIL-PRF-7024E
Linearity% +/-	0.2165 (Over 10:1 Range)	Pickoff Type:	RF	Uncertainty:	+/- .05%
Calib Inv No:	51868	Pickoff P/N:	27-31949-101	Fluid Visc:	1.274 cSt
Report No:	260623A	Signal Output:	N/A	Density:	6.328 Lb/Gal
Calibrator Recal Date:	12/23/2026	Procedure:	CP-08	Average K Factors:	6175.007 P/Gal

Notes:

* Mass flow rate based on density of 6.32833 Lb/Gal

Meter Freq Hz	Vol Flow Rate GPM	Mass Flow Rate PPH	Meter K Factor P/Gal	Meter K Factor P/Lb	Fluid Density Lb/Gal	Roshko Corrected F/V (70°F)	Strouhal Corrected KFactor (70°F)	Reynolds Number	Pressure Drop PSIA	Freq / Viscosity Hz/cSt	Meter Temp °F
2617.112	25.49170	9679.193	6159.915	973.387	6.328	2053.843	6160.770	84349.11	-	2053.653	75.2
2006.840	19.53087	7415.868	6165.130	974.211	6.328	1575.124	6165.991	64633.86	-	1574.977	75.2
1584.976	15.41397	5852.679	6169.635	974.923	6.328	1244.064	6170.498	51011.84	-	1243.948	75.2
1237.099	12.01409	4561.746	6178.244	976.284	6.328	971.100	6179.112	39763.67	-	971.009	75.3
971.539	9.42548	3578.853	6184.546	977.279	6.328	762.623	6185.415	31195.30	-	762.551	75.3
761.778	7.38797	2805.209	6186.644	977.611	6.328	597.971	6187.513	24451.90	-	597.915	75.3
607.085	5.89359	2237.795	6180.455	976.633	6.328	476.516	6181.321	19504.95	-	476.472	75.2
476.917	4.63497	1759.898	6173.725	975.570	6.328	374.330	6174.588	15338.92	-	374.295	75.2
373.453	3.63349	1379.634	6166.855	974.484	6.328	293.110	6167.715	12024.14	-	293.082	75.2
291.914	2.84096	1078.710	6165.113	974.209	6.328	229.107	6165.972	9401.25	-	229.086	75.2
224.465	2.18231	828.622	6171.402	975.202	6.328	176.167	6172.261	7221.50	-	176.150	75.2
177.899	1.72597	655.351	6184.325	977.245	6.328	139.620	6185.186	5711.42	-	139.607	75.2
143.833	1.38656	526.478	6223.987	983.512	6.328	112.887	6224.854	4588.41	-	112.876	75.2
111.777	1.07823	409.404	6220.020	982.885	6.328	87.728	6220.887	3568.09	-	87.720	75.2
88.660	0.85357	324.102	6232.178	984.806	6.328	69.586	6233.047	2824.70	-	69.580	75.2
69.049	0.66332	251.861	6245.762	986.953	6.328	54.193	6246.632	2195.04	-	54.188	75.2
54.366	0.52559	199.565	6206.314	980.719	6.328	42.672	6207.182	1739.38	-	42.668	75.2
41.731	0.40552	153.976	6174.498	975.692	6.328	32.763	6175.372	1342.36	-	32.760	75.3
30.373	0.29914	113.583	6092.137	962.677	6.328	23.852	6093.011	990.47	-	23.850	75.4
24.914	0.24834	94.295	6019.248	951.159	6.328	19.574	6020.133	822.68	-	19.572	75.5

Calibration Technician: J. ESPERAS

Date: August 13, 2026

Certified By: C. Bagay

Calibration Technician

On Behalf of Flow Technology, LLC

Statement of Calibration and Conformance

We hereby certify that the above unit conforms to the applicable specifications as referenced in, or furnished with this purchase order and has been calibrated using standards traceable to the National Institute of Standards and Technology and through the Certificate Number listed above. Evidence of traceability is archived within the manufacturing plant and is subject to examination upon request. FTI Flow Technology, LLC's calibration systems is in compliance with ANSI/NCSL Z540-1-1994. Results apply only to the unit calibrated and shall not be reproduced, except in full, without the written approval of FTI Flow Technology, LLC.