

Phone: (480) 240-3400

Fax: (480) 240-3401

E-mail: ftimarket@ftimeters.com

Website: www.ftimeters.com

Certificate #68005535

Customer: FUERZA AERA DE CHILE
Job No: 132175.1-1
P.O. Number: 60078
Meter Model No: FT-12AEYS-LEA-5
Meter Serial No: 12013005
Linearity% +/- 0.2574 (Over 10:1 Range)
Calib Inv No: 51868
Report No: 250611A
Calibrator Recal Date: 12/11/2025

Cust Part No: N/A
Meter ID (in): 0.750
End Fitting: AN
Bearing Type: BALL
Pickoff Type: RF
Pickoff P/N: 27-31949-101
Signal Output: N/A
Procedure: CP-08

Ambient Temp: 75 °F
Ambient Pres: 14.1 PSIA
Humidity: 32%
Calib Fluid: MIL-PRF-7024E
Uncertainty: +/- .05
Fluid Visc: 1.268 cSt
Density: 6.323 Lb/Gal
Average K Factors: 6176.279 P/Gal

Notes:

* Mass flow rate based on density of 6.32305 Lb/Gal

Meter Freq Hz	Vol Flow Rate GPM	Mass Flow Rate PPH	Meter K Factor P/Gal	Meter K Factor P/Lb	Fluid Density Lb/Gal	Roshko Corrected F/V (70°F)	Strouhal Corrected KFactor (70°F)	Reynolds Number	Pressure Drop PSIA	Freq / Viscosity Hz/cSt	Meter Temp °F
2596.629	25.31574	9604.361	6154.185	973.294	6.324	2045.845	6155.402	84095.56	-	2045.576	77.4
2025.459	19.73208	7486.017	6158.881	974.036	6.323	1597.291	6160.150	65607.09	-	1597.072	77.7
1584.263	15.41573	5848.468	6166.152	975.186	6.323	1249.818	6167.443	51274.33	-	1249.644	77.8
1242.200	12.07058	4579.373	6174.685	976.536	6.323	980.079	6175.984	40152.58	-	979.941	77.9
980.277	9.51818	3611.035	6179.399	977.281	6.323	773.423	6180.699	31661.98	-	773.315	77.9
772.974	7.50000	2845.373	6183.790	977.976	6.323	609.784	6185.083	24945.32	-	609.699	77.8
599.432	5.82312	2209.191	6176.402	976.807	6.323	472.810	6177.685	19365.08	-	472.745	77.8
473.685	4.60493	1747.031	6171.890	976.094	6.323	373.562	6173.162	15311.32	-	373.510	77.7
369.607	3.59470	1363.768	6169.195	975.667	6.323	291.433	6170.457	11950.29	-	291.393	77.7
290.229	2.82232	1070.739	6170.010	975.796	6.323	228.821	6171.268	9381.65	-	228.790	77.6
226.428	2.19978	834.559	6175.929	976.733	6.323	178.501	6177.182	7311.54	-	178.477	77.6
178.510	1.73144	656.880	6185.946	978.317	6.323	140.717	6187.197	5754.53	-	140.698	77.6
140.576	1.35720	514.897	6214.715	982.867	6.323	110.815	6215.972	4510.72	-	110.800	77.6
111.571	1.07511	407.879	6226.578	984.743	6.323	87.956	6227.842	3573.44	-	87.944	77.6
88.024	0.84724	321.428	6233.711	985.871	6.323	69.394	6234.977	2816.09	-	69.385	77.6
69.274	0.66627	252.770	6238.406	986.613	6.323	54.612	6239.673	2214.56	-	54.605	77.6
52.925	0.51111	193.905	6213.032	982.600	6.323	41.722	6214.291	1698.76	-	41.716	77.6
41.929	0.40690	154.372	6182.665	977.798	6.323	33.052	6183.916	1352.37	-	33.048	77.6
32.797	0.32147	121.962	6121.201	968.077	6.323	25.853	6122.439	1068.42	-	25.849	77.6
25.105	0.24985	94.789	6028.800	953.464	6.323	19.789	6030.016	830.35	-	19.786	77.6

Calibration Technician: J. ESPERAS

Date: August 6, 2025

Certified By: C. Bagay

Calibration Technician

On Behalf of Flow Technology, LLC

Statement of Calibration and Conformance

We hereby certify that the above unit conforms to the applicable specifications as referenced in, or furnished with this purchase order and has been calibrated using standards traceable to the National Institute of Standards and Technology and through the Certificate Number listed above. Evidence of traceability is archived within the manufacturing plant and is subject to examination upon request. FTI Flow Technology, LLC's calibration systems is in compliance with ANSI/NCCL Z540-1-1994. Results apply only to the unit calibrated and shall not be reproduced, except in full, without the written approval of FTI Flow Technology, LLC.