

Phone: (480) 240-3400

Fax: (480) 240-3401

E-mail: ftimarket@ftimeters.com

Website: www.ftimeters.com

Certificate #68005931

Customer: FUERZA AERA DE CHILE
Job No: 135025.1-1
P.O. Number: VERONICA CC
Meter Model No: FT-12AEYS-LEA-5
Meter Serial No: 180221M65356
Linearity% +/- 0.4330 (Over 10:1 Range)
Calib Inv No: 51868
Report No: 251223B
Calibrator Recal Date: 6/23/2026

Cust Part No: N/A
Meter ID (in): 0.750
End Fitting: AN
Bearing Type: BALL
Pickoff Type: RF
Pickoff P/N: 27-31949-101
Signal Output: N/A
Procedure: CP-08

Ambient Temp: 75 °F
Ambient Pres: 14.1 PSIA
Humidity: 26%
Calib Fluid: MIL-PRF-7024E
Uncertainty: +/- .05%
Fluid Visc: 1.269 cSt
Density: 6.324 Lb/Gal
Average K Factors: 5696.683 P/Gal

Notes:

* Mass flow rate based on density of 6.32356 Lb/Gal

Meter Freq Hz	Vol Flow Rate GPM	Mass Flow Rate PPH	Meter K Factor P/Gal	Meter K Factor P/Lb	Fluid Density Lb/Gal	Roshko Corrected F/V (70°F)	Strouhal Corrected KFactor (70°F)	Reynolds Number	Pressure Drop PSIA	Freq / Viscosity Hz/cSt	Meter Temp °F
2437.071	25.68083	9743.649	5693.906	900.428	6.324	1918.320	5694.983	85228.36	-	1918.079	77.1
1894.179	19.96290	7574.192	5693.095	900.300	6.324	1491.817	5694.200	66288.64	-	1491.624	77.3
1468.025	15.48503	5875.227	5688.170	899.521	6.324	1156.626	5689.293	51438.88	-	1156.474	77.4
1158.224	12.23083	4640.539	5681.825	898.517	6.323	912.743	5682.959	40637.89	-	912.622	77.5
921.268	9.74146	3696.039	5674.310	897.329	6.323	726.136	5675.451	32372.37	-	726.038	77.5
708.432	7.49693	2844.433	5669.782	896.613	6.323	558.337	5670.919	24911.49	-	558.262	77.5
555.456	5.88311	2232.129	5664.930	895.846	6.324	437.687	5666.056	19545.18	-	437.629	77.4
438.432	4.64322	1761.701	5665.447	895.928	6.324	345.458	5666.571	15425.25	-	345.413	77.4
344.780	3.64661	1383.572	5672.878	897.103	6.324	271.663	5674.003	12114.30	-	271.627	77.4
266.502	2.81241	1067.067	5685.549	899.106	6.324	209.983	5686.676	9342.92	-	209.955	77.4
212.233	2.23456	847.823	5698.653	901.179	6.324	167.225	5699.782	7423.35	-	167.203	77.4
168.551	1.76982	671.491	5714.203	903.638	6.324	132.809	5715.336	5879.52	-	132.791	77.4
132.923	1.38914	527.059	5741.228	907.911	6.324	104.743	5742.370	4615.18	-	104.729	77.4
101.940	1.06119	402.631	5763.675	911.461	6.323	80.336	5764.827	3526.00	-	80.325	77.5
82.288	0.85506	324.422	5774.216	913.128	6.323	64.855	5775.374	2841.34	-	64.847	77.5
62.595	0.65197	247.366	5760.520	910.962	6.323	49.334	5761.676	2166.49	-	49.328	77.5
49.195	0.51452	195.216	5736.762	907.205	6.323	38.771	5737.911	1709.67	-	38.766	77.5
37.634	0.39569	150.128	5706.671	902.447	6.323	29.658	5707.811	1314.71	-	29.654	77.5
29.658	0.31428	119.241	5662.224	895.418	6.324	23.370	5663.348	1044.10	-	23.367	77.4
22.602	0.24279	92.117	5585.624	883.304	6.324	17.807	5586.725	806.46	-	17.804	77.4

Calibration Technician: M. SKINNER

Date: February 23, 2026

Certified By: M. Skinner

Calibration Technician

On Behalf of Flow Technology, LLC

Statement of Calibration and Conformance

We hereby certify that the above unit conforms to the applicable specifications as referenced in, or furnished with this purchase order and has been calibrated using standards traceable to the National Institute of Standards and Technology and through the Certificate Number listed above. Evidence of traceability is archived within the manufacturing plant and is subject to examination upon request. FTI Flow Technology, LLC's calibration systems is in compliance with ANSI/NCCL Z540-1-1994. Results apply only to the unit calibrated and shall not be reproduced, except in full, without the written approval of FTI Flow Technology, LLC.