

Thickness of the electrode array $0,75 \pm 0,5$

Contact Nr.1
 $\varnothing 2,7 \pm 0,1$

Contact Nr.5
 $\varnothing 1,0 \pm 0,1$

Contact Nr.4

20

The last 4 Contact sleeves are not functional

28

10

Contact Nr.26

Contact Nr.29

Contact Nr.50

Back side label

Contact Nr.53

The last 7 Contact sleeves are not functional

53

Contact Nr.29

$47 \pm 0,25$

417 ± 15

$80 \pm 0,25$

10 Electrode cable, silicone

20 Electrode array, 53(32+21) contacts, Ptlr

30 Electrode contact sleeve

40 Inner color code

50 Outer color code 1

60 Outer color code 2



hamed.karkan@cortec-neuro.com

Signature

Name

Date

Title:

Drawn H.Karkan 11.07.2018

Checked & Approved C.Henke 11.07.2018

Changes:

First Angle Projection Drawing number:

Rev.:

2A-61-094X-X

1.00

This document is confidential. It contains designs and/or information that belongs to CorTec GmbH.

This document may not, in whole or part, be duplicated or shared or used for the manufacture of any part(s) herein.

or for any other purpose, without the prior permission of CorTec GmbH

ACAD_Template.dwg | version: 1.00

Template approved: C.V.B.

Date of approval: 28.06.2018

Unit: mm

Material: See table of materials

Tolerance: ISO 2768-mK

if not stated otherwise

Replaces:

3

4

5

6

2

1