#### Renishaw plc

Old Town, Wotton-under-Edge, Gloucestershire GL12 8JR United Kingdom Tel +44 (0) 1453 844302 Fax +44 (0) 1453 844236 Email uk@renishaw.com

www.renishaw.com



# Renishaw Raman instrument service acceptance form

Customer name: UCL (Dept of Physics)

System serial no.: H42218

## Specifics of service

This is to certify that the following service specifics have been performed:

- → Video camera re-centred and aligned to eye pieces.
- → Core white light alignment found to be off, therefore realigned.
- → 1200 l/mm and 2400 l/mm gratings checked and aligned for optimum dispersion, focussing and calibration.
- → Failed Modu-Laser (no emission) and fan assembly removed from system and replaced by new spare (s/n ML0892RM011ACE). **NOTE** replacement laser has a different interlock connector. The existing interlock cable (9 pin D-type) needs to be replaced with a 3 pin Din connector (connecting two existing wires to pins 1 and 3 polarity unimportant). The laser will not operate without a functioning interlock cable.
- → 514nm and 488nm lines aligned into and through system, including pin hole.
- → 785nm laser found to be low on power (60mW). Laser tuned to maximum output of 180mW (specification is 300mW for a new laser).
- → 488nm Fibre optic probe aligned and works well, with good laser blocking. Note internal Fibre optic probe USB camera in not compatible with Windows 7).
- → All test data, laser powers (indication values only) and registry files saved to computer.

#### For and behalf of Renishaw:

Name of service engineer	Signature	Date
Richard Wrighton	And I	09.06.2016

### **Acceptance of service**

I hereby accept that the service specifics described above have been successfully completed.

For and behalf of customer:

Name of customer	Signature	Date
Name of distributor	Signature	Date

For support issues please contact Raman.support@Renishaw.com

GE/8012/02 Issue 1 March 2002 The copyright in this document is owned by Renishaw plc Spectroscopy Products Division. This document shall not be copied, either wholly or in part, by any means, without the prior written consent of Renishaw plc Spectroscopy Products Division, whether given by contract or otherwise.







