

Bit depth

Detectors in Fluorescence Microscopy

The detectors used in fluorescence microscopy are **monochromatic**.

Microscope cameras are **not able to distinguish between different wavelengths** (they just collect photons), you need **fluorescence filters** to separate your fluorophores.

The detector converts photons in digital numbers (linear relation).

