

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/341236535>

Automating detection and localization of myocardial infarction using shallow and end-to-end deep neural networks

Article in *Applied Soft Computing* · May 2020

DOI: 10.1016/j.asoc.2020.106383

CITATIONS

58

READS

2,012

4 authors, including:



Vahab Vahdat

69 PUBLICATIONS 333 CITATIONS

SEE PROFILE



Seyedmohammad Salehi

University of Delaware

20 PUBLICATIONS 141 CITATIONS

SEE PROFILE

Automating Detection and Localization of Myocardial Infarction using Shallow and End-to-End Deep Neural Networks

Kamal Jafarian^{1,2}, Vahab Vahdat^{*3,4}, Seyedmohammad Salehi⁵, Mohammadsadegh Mobin⁶

¹Dept. of Electronic Systems, Aalborg University, Denmark

²Petricore Norway AS, Trondheim, Norway

³Harvard Medical School, Boston, MA, USA

⁴Massachusetts General Hospital, Institute for Technology Assessment, Boston, MA, USA

⁵Dept. of Computer and Information Sciences, University of Delaware, Newark, DE, USA

⁶Dept. of Industrial and Systems Engineering, Oakland University, Rochester, MI, USA

Abstract

Myocardial infarction (MI), also known as a heart attack, is one of the common cardiac disorders caused by prolonged myocardial ischemia. For MI patients, specifying the exact location of a heart muscle suffering from blood shortage or stoppage is of crucial importance. Automatic localization systems can support physicians for better decisions in emergency situations. Using 12-lead electrocardiogram, in this paper, two MI detection and localization methods are proposed with general and end-to-end deep machine learning techniques. For the feature extraction phase, the general approach performs a Discrete Wavelet Transform (DWT) and Principal Component Analysis (PCA) on the pre-processed signals followed by a shallow neural network (NN) for the classification phase. However, in the end-to-end residual deep learning technique, a Convolutional Neural Network (CNN) is directly employed on the pre-processed input signals. For specifying the infarcted region of myocardium, 6 classes of subdiagnosis are considered. Proposed models are verified with the Physikalisch-Technische Bundesanstalt (PTB) dataset, where the data of each patient is first grouped and then carefully partitioned to training, testing, and validation datasets. The results of 5-fold cross validation indicate that the general model achieves over 98% accuracy for MI detection and MI localization with fewer number feature sets compared to previous studies. Moreover, the end-to-end CNN model shows superior performance by achieving perfect results. Thus, with the larger size of CNN models, one may choose a perfect system that requires larger memory compared to another system that requires smaller memory and accepts $\sim 2\%$ false positives and false negatives.

Keywords: Myocardial infarction, EkG Signal processing, Deep residual learning, Artificial neural networks, Discrete wavelet transform, Principal Component Analysis

ACRONYMS

ALMI	Antero-Lateral myocardial infarction
AMA	Acute Marginal Artery
AMI	Anterior myocardial infarction
ANN	Artificial Neural Networks
ASMI	Antero-Septal myocardial infarction
BW	Baseline Wander

*Corresponding author

Email address: kamal.jafarian@petricore.com, vvahdatzad@mg.harvard.edu, mobin@oakland.edu, salehi@udel.edu (Kamal Jafarian^{1,2}, Vahab Vahdat^{*3,4}, Seyedmohammad Salehi⁵, Mohammadsadegh Mobin⁶)

CNN	Convolutional Neural Network
CSI	Cubic Spline Interpolation
DFT	Discrete Fourier Transform
DWT	Discrete Wavelet Transform
DNN	Deen Neural Network
EWT	Emperical Wavelet Transform
ECG	Electrocardiogram
FAWT	Fexible Analytic Wavelet Transform
ILMI	Infero-Lateral myocardial infarction
IMI	Inferior myocardial infarction
IPLMI	Infero-Postero-Lateral myocardial infarction
kNN	k-Nearest Neighbors
LA	Left Atrium
LAD	Left Anterior Descending
LCA	Left Coronary Artery
LCX	Left Circumflex
LR	Logistic Regression
LSTM	Long Short-Term Memory
LV	Left Ventricle
MI	Myocardial Infarction
ML	Machine Learning
N_{pc}	Number of Principal components
PCA	Principal Component Analysis
PCs	Principal Components
PDA	Posterior Descending Artery
PTB	Physikalisch-Technische Bundesanstalt
RA	Right Atrium
RBF	Radial asis function
RCA	Right Coronary Artery
RNN	Recurrent Neural Network
RV	Right Ventricle
SVM	Support Vector Machine
SWT	Stationary Wavelet Transform
VCG	Vector-cardiogram
WHO	World Health Organization
WT	Wavelet Transform

NOTATIONS

Ψ	Basis function in WT
τ	Position or time shift factor in WT
a	Scaling factor in WT
A_L	approximation coefficient
C_{kg}	Number of samples belonging to class k but labeled as class g
D_L	Sub-band coefficients
Acc	Model accuracy
F_P	False Positive
F_N	False Negative
n_k	Class k samples proportion in the data
T_P	True Positive

T_N	True Negative
S_n	Sensitivity
S_p	Specificity

1. Introduction

According to a 2016 World Health Organization (WHO) report [1], coronary heart disease, also known as ischemic heart disease, is the leading cause of death in the world. The WHO study shows over 17.7 million people die annually from cardiovascular diseases of which 80% are caused by heart attacks. Heart attacks, also called Myocardial Infarction (MI), is the result of partial or complete occlusion toward coronary arteries, leading to insufficient blood flow to the heart [2]. Interested readers are referred to Appendix A for an introduction to the anatomy of coronary and MI. According to emergency guidelines [3, 4], MI detection and localization must be conducted instantly on suspicious patients. The immediate assessment and therapeutic priority determination of MI not only can decrease the risk of subject's death, but also can lead to re-perfusion, which prevents undergoing necrosis [3].

Patients with MI can be diagnosed by several methods including electrocardiogram (ECG) [5, 6, 7, 8, 9], echocardiography [10], magnetic resonance imaging (MRI) [11], and variation in cardiac biomarkers such as creatine Kinase MB (CK-MB) [12], Troponins [13], and Myoglobin [14]. In practice, the first MI diagnostic method for patients with urgent need is ECG. ECG machines are frequently available inside the ambulance and emergency rooms, providing a prompt and highly accurate diagnosis of heart failures, if being interpreted correctly [3].

Several studies have been carried out to automate MI detection by analyzing ECG signals using different machine learning and artificial intelligence methods [15, 16, 17]. The literature presents several flavors of methods for ECG signal pre-processing, feature extraction, and signal classification. However, there are still challenges in the accuracy and computational complexity of the methods that have been carried out. Sun *et al.* [18] employed the latent topic multiple instance learning (LTML) classifier to detect the presence of MI using polynomial coefficients fitted on one of the channels of 12 lead ECG segments called ST. Chang *et al.* [19] examined the MI detection by Hidden Markov, Gaussian Mixture, and Support Vector Machine (SVM) classifiers on 12-lead ECG data. Waduud *et al.* [20] conducted an experiment on ST amplitude of 12-lead ECG and determined the culprit arteries are the left anterior descending and the right coronary artery. Dohare *et al.* [21] used SVM for detecting MI by determining over two hundred parameters from average beat with 10 second length in 12-lead ECG signals.

While detecting MI is crucial, the focus of new studies is on extracting additional information from the ECG signals. For example, Adam *et al.* [22] used k-Nearest Neighbor (kNN) to classify ECG signals into four categories which are MI, dilated cardiomyopathy, hypertrophic cardiomyopathy, and normal. Le *et al.* [23] determined complex spatio-temporal patterns to detect and localize MI using vector-cardiogram (VCG) and the random-walk network along with three orthogonal planes of the body. Arif *et al.* [24] implemented kNN classifier and extracted time domain features of ECG beats including T wave amplitude, Q wave and ST level deviation. In order to fully detect and localize the culprit artery, a two-class and a multi-class classifier to ten regions are constructed and evaluated respectively. Acharya *et al.* [25] studied MI anatomical location by transforming 12-lead ECG signals to a 4-level discrete wavelet transform. From each lead, 12 nonlinear features including entropy and largest Lyapunov exponent are reextracted and the optimal features are selected using a t-test feature ranking methodology. Finally, using a kNN classifier, authors evaluated their model performance under different scenarios. Sharma *et al.* [26] constructed different classifiers including kNN, the linear, and kernel-based SVM to first detect infarcted subjects and then localize MI pathologies into six different regions.

With the rise of deep learning methods in recent years, the majority of the research in MI detection has shifted to implementing these methods. One of the advantages of deep learning models is achieving high accuracy performance without extensive signal processing. In turn, the signals can be classified efficiently, often with fewer number of signals. For instance, Nezamabadi *et al.* [27] utilized a multi-scale deep convolutional neural network to extract features from chest HRCT signal. To optimize the classification performance, Wu *et al.* [28] developed a new deep feature learning method. The optimized features are fed

into the softmax regression to build a multi-class classifier. Table 1 illustrates a survey of the most recent research studies on MI detection and localization categorized by study characteristics, signal transformations, features extraction, and classifications methods.

Table 1: Review of the most recent literature for detection and localization of Myocardial Infarction (**P.M.**=Proposed Method)

Research Studies			[25]	[29]	[30]	[31]	[21]	[32]	[33]	[34]	[35]	[36]	[37]	[38]	[39]	[40]	[41]	P.M. 1	P.M. 2		
Features																					
Study Characteristics	Study type	Detection	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		Localization	✓					✓							✓	✓		✓	✓		
	Data types & preparation	Number of leads	12	1	1	3	12	12	1	1	3	3	1	8	12	12	1	12	12		
		Beat analysis	✓			✓			✓	✓			✓	✓		✓		✓			
		Record analysis					✓	✓			✓	✓			✓		✓		✓	✓	
Intra-information segmentation		–	–	–	✓	✓	–	–	No	–	✓	No	No	No	✓	–	✓	✓	✓		
Morphological analysis	ST, QRS, P wave segmentation																				
Transformation & decomposition methods			DWT	✓																✓	
			FAWT	✓																	
			DFT																	✓	
			EWT																		✓
Feature Extraction			Statistical feature investigation	✓																✓	
			PCA																		✓
Classification methods	Traditional ML	ANN	✓																✓		
		Random Forrest	✓																		
		SVM	✓																		
		Logistic Regression																	✓		
		kNN	✓																	✓	
	Deep learning	Deep Neural Network																	✓		
		Decision Tree																		✓	
		CNN	✓		✓		✓	✓	✓	✓		✓	✓						✓		
	RNN																	✓			
	LSTM																	✓			

(-) indicates the studies that the data segmentation strategy is not well-defined or could not be justified

As discussed, many studies used traditional machine learning and deep learning techniques for detecting and localizing heart disorders such as MI, but there are major drawbacks associated with previous studies that this research seeks to address. The first shortcoming of the previous studies stems from ignoring the importance of data partitioning in the classification methods. If the data in the test dataset has been partially observed in the training dataset, model performance can be overestimated. As a result, even if the accuracy of the classifier is very high, it may fail to categorize future unseen observations reliably. Through the investigation of the current literature, only a few studies reported strategies to group the records of each patients prior to splitting data into test and validation sets. Most of the studies did not segregate the data properly or did not discuss the segmentation process explicitly (shown by “-” in Table 1). The second drawback with previous studies is the lack of comprehensive modelling. Although previous studies have shown promising results, the majority of the researchers have developed models to detect the MI rather

than detect and localize.

Following the drawbacks associated with previous studies, the major contributions of this research that fill the gap of the current literature are as follows:

- Previous studies overlooked the importance of proper signals partitioning for training and testing datasets. In this study, the signals of each patient are grouped to ensure that they can only be assigned to one of the partitions (i.e., train, validation, or test). Hence, the patient data will not partially be examined by the classifiers in the training phase, providing more robust performance.
- In order to build a comprehensive framework to detect and localize MI simultaneously, two classification methods are implemented and verified using the Physikalisch-Technische Bundesanstalt (PTB) database of PhysioNet bank [42]. In the first method, a classic ECG processing method with a novel configuration including a shallow neural network (NN) is utilized in both detecting and locating MI into 6 specific infarcted regions. In the second method, an end-to-end convolutional neural network is deployed. Building both traditional and deep neural networks enable the opportunity to compare the complexity, efficiency, and accuracy of both models within one framework.
- For the traditional neural network model, several innovative techniques are used. While there is a known trade-off between the size of feature vector and the obtained accuracy from the classifier, this research builds a highly accurate model with fewer number of features extracted from ECG signals. For this purpose, a different combination of Principal Component Analysis (PCA) and Discrete Wavelet Transform (DWT) is employed to extract distinctive features from ECG signals. Using PCA, the number of signals is reduced to 7 signals from 12 (original ECG signals) in MI detection and 8 signals in MI localization phase, respectively. The results indicate that the accuracy of the classifiers reaches over 99% in detecting and 98% in locating the MI.
- By utilizing end-to-end deep neural networks, we are able to build efficient models that require only few steps to detect and localize MI from the raw signals. This advantage can harness the use of these classification method on big datasets and ease the required hardware design. Our proposed end-to-end MI detector and localizer classifier achieves perfect results for MI patients.

The rest of the paper is organized as follows. Section 2 is dedicated to our method, providing a comprehensive explanation of the proposed framework for detecting and localizing MI. Section 3 presents an overview of the dataset that is used in this research. The results of the proposed models are presented in Section 4, followed by evaluation and performance discussion in Section 5. Finally, Section 6 concludes the paper with future research directions.

2. Methodology

For detection and localization of MI, two automatic methods (i.e., classic multi-lead ECG processing system and end-to-end deep neural network) are implemented and compared. The first step in utilizing each of these methods is signal pre-processing, as shown in Figure 1. The following steps differ based on the type of technique. In classic approach, several transformations and feature extraction layers should be performed. In contrast, in end-to-end deep learning approach, the pre-processed signals are directly fed into the deep network without additional processing. In the following sections, details pertaining to each of these methods are further explored.

2.1. Pre-processing Phase

ECG signals can be corrupted by two major artifacts; first, high-frequency noise such as electromyogram (EMG) induced noise, mechanical forces on the electrodes, and power line interference, and second, baseline wander (BW) that may be caused by patient movement and respiration or destabilized instruments during the ECG recording [43, 44]. Prior to analyzing the ECG signals, it is required to remove the noise from the signals while preserving signals' main informative values. In this study, as shown in Figure 2, the residual




Figure 1: A schematic diagram of steps for different approaches investigated in this study.

power-line interference and its harmonics are suppressed by applying a 60-Hz notch filter with a 3-dB bandwidth of 5 Hz, also known as band-rejection filter. Moreover, a third order cubic spline interpolation (CSI) is employed to remove BW [45, 46] which is capable of recovering the isoelectric level in the detrend signal. In this technique, a cubic polynomial is fitted to a set of representative points of the original ECG signal. As a result, the signal's low frequency trend can be obtained considering the reference points. The fitted baseline is then subtracted from the ECG signal to remove this artifact.




Figure 2: Signal pre-processing including baseline wandering and power line artifact rejection.

2.2. Classic Approach

After the pre-processing phase, a set of DWT and PCA transformations are performed to extract distinctive features from the signals. Utilizing the extracted features, the shallow neural networks are trained for detection and localization of MI, as briefly discussed in the following Subsections.

2.2.1. Feature Extraction

After de-noising ECG signals, signals are decomposed into the wavelet coefficients. Wavelet transform (WT) is a mathematical convolution of a signal with a specific wavelet function to represent the time-frequency characteristics of a signal. The WT of function $f(t)$ with the wavelet basis function $\Psi(\cdot)$ at scale a and position or time τ is given by:

$$W_{f(a,\tau)} = \frac{1}{\sqrt{a}} \int_{-\infty}^{+\infty} f(t) \Psi\left(\frac{t-\tau}{a}\right) dt \quad (1)$$

In order to implement the wavelets on digital signals, discrete wavelet transform (DWT) is used by a set of predefined discrete scales. To decompose signals, dyadic WT is employed that characterizes basic wavelet shape and covers the entire region of interest. Dyadic WT based on the squared dilation of the mother wavelet at the scale 2^j (or level j), is shown in Equation 2.

$$\Psi_{2^j}(t) = \frac{1}{\sqrt{2^j}} \Psi\left(-\frac{t}{2^j}\right) \quad (2)$$

Multi-resolution decomposition of an ECG signal using dyadic WT at level L can be shown by $L+1$ sub-band coefficients including *one* approximate coefficient (A_L), and L details sub-band coefficients, namely, D_L, D_{L-1}, \dots, D_1 [47]. In this paper, a 6-level wavelet decomposition is performed for each lead of ECG signal with the *Daubechies 4* mother wavelet function. By using $L = 6$, an approximation sub-band matrix $A_6 = \{A_6^1, A_6^2, \dots, A_6^m\}$ and six details sub-band matrices $D_i = \{D_i^1, D_i^2, \dots, D_i^m\}$ are constructed, where $m = 12$ is the leads number and $i = 1, 2, \dots, 6$ is the level of coefficients [26].

A distinctive set of features can be extracted out of the constructed matrices $\{A_6, D_i; i = 1, 2, \dots, 6\}$. Different statistical and spectral feature sets are employed for this study (e.g., Standard deviation, signal power, and dominant frequency). Our analysis shows that the only feature set that has significant impact on the results is standard deviation of transformed signals. Other feature sets are eventually removed from the analysis while the entire set of final features is still statistically significant with p-value of less than 0.01.

In our study, most of the region of interest lies in the lower frequency sub-band coefficients, as also inspected by previous studies [48]. As a result, we only take $\{A_6, D_6, D_5, D_4\}$ for further feature extraction. The first three level details coefficients $\{D_3, D_2, D_1\}$ hold relatively high-frequency contents (larger than 62.5Hz) in comparison with the natural heart activities. This also demonstrates that implementation of a high-frequency noise rejection filter was unnecessary in the pre-processing step.

2.2.2. Principal Component Analysis

After decomposing the signals, PCA is used to extract the temporal deviations of ECG signals. PCA serves as a dimension reduction method to represent underlying variations of the data in a fewer linearly independent features [49]. Using PCA, the constructed matrices including A_6, D_6, D_5 and D_4 are projected into a new mutually uncorrelated space [50]. As the number of selected principal components (N_{pc}) increases, the variation of signal decreases, albeit cumulative variation covered by the model increases. Several frequency and time-domain features can be extracted from the developed signals. For this purpose, standard deviation of each derived signal is computed to quantify the amplitude dispersion from the mean value. Different values for N_{pc} are tested for feature extraction, and the classifier's performance is assessed. Thus, each matrix in $\{A_6, D_6, D_5, D_4\}$ produces k standard deviations, and entirely $4 \times k$ features are extracted for classification.

While combining PCA and DWT has been used by other researchers, we revised the sequence of combining these methods to achieve better results. Previously, PCA has been widely used in pre-processing phase for dimension reduction purpose and then followed by DWT on the reduced number of signals. However, using PCA in pre-processing phase, as shown in Figure 3, could impair the extraction of main information in the signal. This is as a result of the fact that the 12-lead ECG signals have disparate frequency resolutions. Thus, the lower frequency lead ECGs are always suppressed by PCA, even if carrying useful information. In this research, PCA is performed on the distinctive set of features that is extracted from DWT, or specifically $\{A_6, D_6, D_5, D_4\}$. Each set of these features have similar frequency resolutions and applying PCA will find




Figure 3: Applying PCA on original ECG signal. While PCA can reduce the dimensions of study, it can inversely affect full extraction of information from the signal.

155 maximum variation correspond to that frequency resolution, preserving information from different range of signal frequencies. As shown in Figure 4, applying PCA on the extracted feature will not disregard the important information on each signal, and therefore, will improve the accuracy of the classification model.




Figure 4: Implementation of PCA method in this research. Instead of using PCA on the original ECG signals, it is used on distinctive set of features that are extracted from DWT.

2.2.3. Classification using Shallow Neural Network

160 In the first approach, shallow neural networks (NN) classifier is used to detect and localize MI. NN is a network of neurons that receive series of inputs and use a propagation function to produce outputs based on their internal states. Outputs of neurons will be served as inputs of next-layer neurons forming a directed, weighted graph. The properties of the trained NN is as follows. A three-layer neural network is constructed by $4 \times N_{pc}$ input nodes from the extracted features. The number of hidden layers are the summation of output layer and $\frac{2}{3}$ of the input layer size, as recommended by [51]. The output layer consists of two nodes for detection of infarcted patients from healthy individuals, and six nodes for localization of MI

165

cases. Layers in the network are fully connected. In addition, the network is trained such that each input vector results in an output vector with a non-zero element corresponding to the predicted class label. The NN classifiers are trained by Levenberg-Marquardt back-propagation algorithm [52, 53]. Unlike other back-propagation algorithms, Levenberg-Marquardt algorithm calculates the approximation of loss, the gradient, and the Hessian matrices rather than computing the exact values. Then, in order to minimize the loss at each iteration, the damping parameter is adjusted. The learning procedure is continued either until 30 validation failures are obtained or when the maximum number of iterations (i.e., 100 epochs) is reached.

2.3. End-to-End Deep Residual Learning with Dilated Convolutions

Deep residual learning, originally conceived for image recognition applications, was proposed by He *et al.* [54] in order to overcome the performance degradation that occurs in convolutional neural networks (CNNs) when the number of hidden layers is large. In short, the idea behind deep residual learning is as follows. Let l designate a particular layer of a neural network such that \mathbf{x}_{l-1} represents its input. The authors of [54] stated that when networks are too deep, instead of optimizing the original mapping $\mathcal{F}_l^{l+k}(\mathbf{x}_{l-1})$ between layers l and $l+k$ ($k \in \mathbb{N}$), it might be easier to optimize the residual mapping $\mathcal{H}_l^{l+k}(\mathbf{x}_{l-1}) = \mathcal{F}_l^{l+k}(\mathbf{x}_{l-1}) + \mathbf{x}_{l-1}$. Such a residual mapping can be achieved by means of identity mapping, namely, by using identity shortcut connections skipping $k+1$ layers.

A diagram of our end-to-end deep residual neural network with dilated convolutions can be seen in Figure 5. The proposed deep residual learning architecture for MI detection and localization is originally based on research studies for keyword spotting presented in [55] and [56]. However, for this research, the architecture is significantly revised.




Figure 5: Diagram of the proposed end-to-end deep residual neural network for MI detection and localization (1-C stands for one-dimensional convolutional layer)

In this research, we consider a front-end based on one-dimensional convolutional layers (1-C), which to the best of our knowledge, is novel in the context of MI detection and localization. Our analysis shows that such front-end provides better performance compared to directly feeding the multi-lead ECG signal matrix into the first two-dimensional (2D) convolutional layer of the deep residual neural network as in [38].

We have extensively analyzed the proposed model with different number of filters, kernel sizes, and stride numbers to find the high performance 1D convolutional layers. The parameters of the proposed CNN model are 20 filters, zero bias vectors, a kernel size of 100 samples, and a stride of 50 samples. It is noteworthy that this set-up yields a *pseudo*-time-frequency representation of the ECG signal, where the number of filters of the layers would correspond to the number of frequency bins. Moreover, the kernel size and stride match the length of the analysis window and its hop size, respectively. Therefore, with sampling rate of 100 Hz, it is equivalent to using 20 *pseudo*-frequency bins and a 1-second long analysis window with 50% overlap.

After applying a rectified linear unit (ReLU) activation function, each lead is represented by a 9×20 *pseudo*-time-frequency matrix. Then, the twelve *pseudo*-time-frequency matrices (one from each of the 12 leads) are stacked across the depth dimension and the resulting $9 \times 20 \times 12$ feature volume is fed into the first 2D convolutional layer of our architecture. Following the shallowest 2D convolutional layer, there is a total of three residual blocks with identity mapping. In each residual block, there are two convolutional layers, each followed by a ReLU activation function, and one batch normalization layer for regularization purposes [57]. The six convolutional layers of the residual blocks apply dilated convolutions to increase the receptive

field of the network [58]. For each layer, the dilation rate depends on the layer position within the network. Thus, the dilation rates of the convolutional layers in the first and second residual blocks are (1, 1) and (2, 2), respectively. Additionally, the dilation rates of the first and second convolutional layers of the third residual block are, respectively, (4, 4) and (8, 8). A non-residual convolutional layer with (16, 16) convolution dilation, another batch normalization layer, and an average pooling layer are appended to the third residual block. Then, for MI detection and classification, a fully-connected layer with softmax activation is employed. Finally, it is worth mentioning that each of the 2D convolutional layers have 7 filters (equal to the number of neurons in the fully-connected layer), zero bias vectors, and a kernel size of 3×3 .

3. Dataset and Materials

The national Metrology Institute of Germany, also known as Physikalisch-Technische Bundesanstalt (PTB), provides a free digitized ECG database for the purpose of research and education through PhysioNet bank [59, 42]. The PTB database has been widely used in MI detection and localization studies [24, 19, 23, 60, 26]. There are 549 records of conventional 12-leads ECG signals known as I, II, III, aVR, aVL, aVF, V1-V6. Each signal is recorded at 1000 Hz sampling rate and 16 bits amplitude resolution over a range of $\pm 16.384\text{mV}$.

The records belong to 290 subjects aged from 17 to 85 years old. 72% of subjects are male with mean age of 57.2 and 28% of subjects are female with mean age of 61.6 years old. The clinical summary for 22 subjects were not available in the dataset and excluded from this study. 148 subjects or correspondingly 368 out of 549 records are labelled as MI cases and the number of subjects that are labeled as healthy cases are 52 (80 records). Other subjects are labeled with heart diseases unrelated to MI such as branch block and consequently are excluded from this study.

For this study, the data is segmented into 5-second long samples from 12-lead signals, constructing 5968 segments labelled with six different classes. In order to balance the data distribution and keep the prior probabilities equal, an adjustable overlap percentage is used for signal windowing as shown in Table 2. Note that this data is completely de-identified and publicly available; therefore no ethics statement was acquired for this investigation.

Table 2: The distribution of data between the classes.

MI Location	No. of Segments	Overlap (%)
Normal	1886	10.0
Anterior (AMI)	1861	50.0
Antero-Lateral (ALMI)	1775	50.0
Antero-Septal (ASMI)	2005	20.0
Inferior (IMI)	2000	10.0
Infero-Lateral (ILMI)	1853	40.0
Infero-Postero-Lateral (IPLMI)	1861	80.0

In order to properly train and validate the classification methods, the data is divided into three sets. 60% of the data is allocated as the training dataset, 10% is separated for validation and the rest is considered as the test dataset. As mentioned earlier, unlike many of the previous studies, we employed a patient-independent approach in which samples from a given subject are ensured to be only included in one of the three datasets. Otherwise, although the nominal accuracy of the constructed classification methods might be very high, the model is highly over-fitted and there is no guarantee that the model reaches similar accuracy for an unseen cases. Additionally, in order to suppress the effect of initialization in the constructed networks, a 5-fold cross validation method is implemented. Feature scaling and normalization are also conducted to standardize the features with a zero mean and unit standard deviation.

4. Results

In this section, results of the classic model and the end-to-end deep residual network on the PTB data-set are analyzed and compared. The classifiers' performance are measured in terms of accuracy (Acc), sensitivity (S_n), and specificity (S_p). The Acc of a classifier is the ratio of the correct predictions to the total number of predictions. The S_n , also called recall or true positive rate (TPR), measures the proportion of positive predictions that are correctly identified over all positive cases. On the other hand, S_p , also called true negative rate (TNR), is an indicator of the proportion of negative predictions that are correctly identified as such. Acc , S_n , and S_p are defined by Equations 3, 4 and 5, respectively. In the binary classification settings, T_P , T_N , F_N and F_P indicate the number of true positives, true negatives, false positives and false negatives, respectively.

$$Acc = \frac{T_P + T_N}{T_P + F_N + T_N + F_P} \quad (3)$$

$$S_n = \frac{T_P}{T_P + F_N} \quad (4)$$

$$S_p = \frac{T_N}{T_N + F_P} \quad (5)$$

These performance metrics can be expanded for classification problems with more than two classes. Let n represent the number of samples in a given data-set and k be the number of classes. Therefore, n_k denotes the sample proportion of k -th class, while n'_k indicates the number of samples labeled as the k -th class. Using these information and constructing the confusion matrix for a multi-class problem, c_{kg} represents the number of samples belonging to class k but labeled as class g . Consequently, the diagonal elements c_{kk} represent the number of correctly classified samples in class k , and non-diagonal elements denote the numbers of miss-classification samples. According to the above definition, sensitivity and specificity of each class and the overall accuracy of the classifier can be calculated as follows [61]. Note that the overall sensitivity and specificity of the classifier is the average performance of each class.

$$Acc = \frac{\sum_{k=1}^K C_{kk}}{n} = \sum_{k=1}^K S_{n_k} \frac{n_k}{n} \quad (6)$$

$$S_{n_k} = \frac{C_{kk}}{n_k} \quad (7)$$

$$S_{p_k} = \frac{\sum_{\substack{g=1 \\ g \neq k}}^K (n_g - C_{gk})}{n - n_k} \quad (8)$$

Table 3: Shallow NN classifier performance (average, standard deviation, and maximum) with respect to different number of PCs (N_{pc}) for MI detection. $N_{pc} = 7$ is selected for feature extraction.

N_{pc}	Accuracy			Sensitivity			Specificity		
	Average	Std.	Max.	Average	Std.	Max.	Average	Std.	Max.
2	90.63	1.31	91.25	90.35	1.73	92.22	90.92	1.94	93.33
3	96.55	0.79	98.25	96.60	1.88	97.56	96.50	1.58	97.89
4	97.30	0.92	99.17	96.85	1.94	97.67	97.75	1.21	98.33
5	98.09	1.15	99.58	97.60	2.45	98.63	98.58	1.27	99.25
6	98.88	0.82	99.58	98.60	2.38	99.17	99.17	0.66	100.00
7	99.11	0.88	99.73	98.83	1.61	99.46	99.39	1.02	100.00
8	97.88	1.03	99.58	97.77	2.85	98.33	98.00	1.63	100.00

4.1. Results Pertaining to Classic Approach

4.1.1. MI Detection

In order to build high performance detection system with lowest number of extracted features, the model is trained with different N_{pc} and evaluated on the validation data-set. Using 5-fold cross validation, the accuracy, sensitivity, and specificity of classification performance for each N_{pc} is presented in Table 3. As N_{pc} increases, the overall performance of the classifier improves. However, the increase in N_{pc} may introduce unnecessary variation in the model, leading to overfitting. As a result, although the model performance may be higher in training phase, it may decrease once the model is assessed on the validation data-set. More specifically, in the cases of N_{pc} equals to 5, 6, 7, training procedures are executed properly with the average accuracy higher than 98.09% and increasing N_{pc} to 8, have not improved the accuracy of the model. Thus, N_{pc} is selected for the analysis acquiring average accuracy of 99.11%, sensitivity of 98.83%, and specificity of 99.39%. The standard deviation of each constructed signal resulting from PCA is also calculated and used for the feature set. Therefore, based on the given results, 12-lead ECG selected sub-band coefficients are transformed by PCA, and only the first seven PCs have be taken in the feature extraction step.




Figure 6: Confusion matrices of MI detection from shallow NN. From left to right and top to bottom: confusion matrices from the 1st, 2nd, 3rd, 4th, and 5th cross-validation folds.

Table 4: The performance of NN classifier for MI detection over 5-fold cross-validation ($N_{pc} = 7$)

Performance metric	Fold 1	Fold 2	Fold 3	Fold 4	Fold 5	Average	Std.
Accuracy	98.43	98.10	98.08	98.95	97.52	98.21	0.52
Sensitivity	98.50	98.50	98.49	98.51	97.66	97.50	0.37
Specificity	98.37	97.50	97.51	99.32	97.37	98.01	0.83
Positive predictive value	98.33	98.33	98.17	99.17	97.50	98.30	0.59
Negative predictive value	98.53	97.74	97.95	98.78	97.54	98.11	0.53

5-fold patient independent cross-validation test is conducted on features extracted from seven PCs and the results are shown in Table 4 and Figure 6. Overall accuracy of 98.21%, sensitivity of 97.50% and specificity of 98.01% are obtained for diagnosing the infarcted subjects.

4.1.2. MI Localization

The same procedure to compute distinctive features as carried out for MI detection is performed. Using different N_{pc} in feature extraction phase, the performance of the shallow NN classifier is computed on the

Table 5: The performance of shallow NN for locating MI with respect to the different N_{pc} .

Performance metric	MI Location	Number of principal components						
		2	3	4	5	6	7	8
Sensitivity	ILMI	0.813	0.891	0.935	0.959	0.969	0.974	0.979
	AMI	0.802	0.880	0.932	0.956	0.967	0.976	0.984
	IPLMI	0.822	0.898	0.944	0.965	0.974	0.978	0.986
	ASMI	0.799	0.881	0.933	0.956	0.967	0.972	0.978
	IMI	0.797	0.881	0.932	0.955	0.969	0.978	0.986
	ALMI	0.804	0.881	0.935	0.960	0.969	0.978	0.984
Specificity	ILMI	0.970	0.984	0.990	0.994	0.996	0.996	0.996
	AMI	0.960	0.972	0.984	0.992	0.992	0.992	0.994
	IPLMI	0.965	0.984	0.993	0.997	0.997	0.998	0.998
	ASMI	0.966	0.981	0.990	0.992	0.996	0.996	0.996
	IMI	0.961	0.982	0.989	0.991	0.996	0.996	0.996
	ALMI	0.966	0.981	0.991	0.995	0.994	0.995	0.995
Accuracy	ILMI	0.872	0.933	0.947	0.968	0.975	0.984	0.989
	AMI	0.846	0.921	0.955	0.965	0.977	0.985	0.989
	IPLMI	0.952	0.974	0.991	0.989	0.997	0.995	0.997
	ASMI	0.807	0.888	0.937	0.963	0.961	0.975	0.979
	IMI	0.818	0.882	0.938	0.960	0.972	0.979	0.986
	ALMI	0.833	0.886	0.942	0.968	0.979	0.985	0.989

validation data-set. As shown in Table 5, the MI location accuracy of each region is individually assessed.

Accordingly, $N_{pc} = 8$ is selected for MI localization classifier.




Figure 7: Confusion matrices of MI localization from shallow NN. From left to right and top to bottom: confusion matrices from the 1st, 2nd, 3rd, 4th, and 5th cross-validation folds.

Table 6: Performance of multiclass shallow NN for MI localization over 5-fold cross-validation ($N_{pc} = 8$).

Performance metric	Fold 1	Fold 2	Fold 3	Fold 4	Fold 5	Average	Std.
Accuracy(%)	96.81	99.27	98.06	97.58	99.42	98.22	1.11
Sensitivity(%)	96.85	98.78	98.05	97.62	99.43	98.14	1.00
Specificity(%)	99.04	99.81	99.46	98.93	99.80	99.40	0.41

Similarly, patient-independent 5-fold cross-validation datasets are constructed and the NN is evaluated. The confusion matrices corresponding to each of the cross-validation datasets are given in Figure 7. The

error in classification of patients in different datasets is shown in Table 6. The accuracy of 98.22%, sensitivity of 98.14% ,and specificity of 99.40% are obtained.

4.2. End-to-End Deep Neural Network Results

Keras [62] was used to implement the end-to-end deep residual learning architecture depicted in Figure 5. The proportion of model parameters from the twelve one-dimensional convolutional layers is 5,997. This number of parameters is relatively low, compared to the other deep residual learning architectures such as [55] and [56] with more than 200,000 parameters. After random initialization of parameters, deep residual learning models were trained for a total of 20 epochs by considering a multi-class cross-entropy loss function. We used the Adam optimizer [63] with the learning rate and the learning rate decay of 0.001 and 0.00, respectively, $\beta_1 = 0.9$ and $\beta_2 = 0.999$. Moreover, the size of the minibatch, number of samples per gradient update, was set to 32 training examples.

Table 7: Performance of deep end-to-end NN for MI detection and localization over 5-fold cross-validation.

Performance metric	Fold 1	Fold 2	Fold 3	Fold 4	Fold 5	Average	Std.
Accuracy (%)	100	99.98	99.98	100	100	99.99	0.01
Sensitivity (%)	100	100	100	100	100	100	0.00
Specificity (%)	100	100	100	100	100	100	0.00




Figure 8: Confusion matrices for MI detection and localization using deep end-to-end NN. From left to right and top to bottom: confusion matrices from the 1st, 2nd, 3rd, 4th, and 5th folds.

Performance metrics, in percentages, and confusion matrices are shown in Table 7 and Figure 8, where we have considered 5-fold cross-validation. Our results indicate that the deep end-to-end residual NN are able to achieve *semi*-flawless MI detection and localization over different folds in cross-validation. With miss-classifying only 1 detection case in 2nd and 3rd folds, the average (i.e., across folds) MI detection and localization accuracy, sensitivity, and specificity is 99.99%, 100%, and 100% respectively.

Compared to the traditional machine learning models, deep learning models are more dependent on the size of training dataset [64, 65]. Therefore, we investigate the impact of training size on the accuracy of the model. To that end, 10% of the dataset is reserved for validation and the remaining dataset is divided between training and test datasets. For instance, with 30% training set size, the amount of test dataset is 60%. Fig. 9 illustrates this investigation. The threshold at which the model reaches the minimum acceptable

accuracy ($Acc > 99\%$) is 55% of this dataset size. This result indicates that our deep learning model is efficient and does not suffer from overfitting.




Figure 9: Training-set size versus classification accuracy for end-to-end deep learning model

5. Evaluation and Discussion

In this section, we first investigate the importance of applying PCA transformation on the overall performance of the classic approach. Then, we compare the computational time of the shallow and deep end-to-end NN models. Finally, we summarize the state-of-the-art proposed methods along with their reported performance metrics.

In the classification and pattern recognition literature, there is often a positive correlation between the feature vector size and the obtained accuracy. To verify this conjecture, we trained the model with and without PCA transformation and compared the performance of the shallow NN classifiers. In the model without PCA reduction, the entire 48 features are fed into the model which resulted in Accuracy of 90.37%, Sensitivity of 88.71%, and Specificity of 91.88%. However, with PCA reduction, the model achieves higher performance for MI detection ($Acc = 98.21\%$, $S_n = 97.50\%$, and $S_p = 98.01\%$). With the similar results obtained from MI localization comparison, we conclude that while PCA technique reduces the number of ECG signals from 48 to 28 (for detection) and 32 (for localization), the performance of the classification has not degraded but improved.

In the second method (deep end-to-end NN), we use a single 1D convolutional layer per ECG lead to extract a *pseudo*-time-frequency representation. This information is utilized to create a compact and discriminative input feature volume to attain the strengths of a deep residual learning architecture. The superiority in performance of the proposed end-to-end method compared to the classic one is significant. However, due to larger number of hidden layers and parameters in the deep NN, the computational time to train deep NN is remarkably longer. The computational time of training phase for deep NN was 1475.5 ± 110.37 seconds, compared to 480.5 ± 53.2 seconds for shallow NN. The computational time for pre-processing, frame segmentation and feature extraction for each sample took 0.46 and 0.09 seconds for classic and end-to-end approaches, respectively. However, once the classifiers are trained, both methods would detect and localize MI patient from the test data set within milliseconds (i.e., 0.01 to 0.02 seconds). Note

that the models are run on a personal computer with Intel Core i5-7500 Quad-core 3.40 GHz and Nvidia GeForce GTX 1070.

In order to signify the performance of our proposed model, a summary of recent MI detection and localization literature is provided in Table 8. This summary collects the reported results of studies where PTB database is used. The table compares these studies based on feature extraction and classification methods, number of leads, the reported features, and performance metrics of MI detection and localization methods. Note that, end-to-end techniques require no feature extraction methods, since the classifier can automatically generate the features. In this table, training set ratio is also reported as a noteworthy information to assess the complexity of classification methods used in each study. Since our two proposed methods are significantly different after pre-processing phase, each method is reported individually.

Table 8: Summary of recent studies for automated MI detection and localization where 12-lead ECG signals are obtained from PTB database; **Ftrs**: number of hand-crafted features, **S_n** : Sensitivity, **S_p** : Specificity, **Acc**: Accuracy, **n_k** : Number of classes, **P.M.**: Proposed Method

Ref.	Feature extraction method	Classification method	Training set ratio	number of leads	MI Detection				MI Localization				
					Ftrs	S_n (%)	S_p (%)	Acc(%)	Ftrs	S_n (%)	S_p (%)	Acc(%)	n_k
[18]	ST segmentation	LTMIL+kNN [†] , LTMIL+SVM, LTMIL+NN, LTMIL+RF	90%	12	74	92.30	88.10	–					
[24]	Time-domain	Pruning algorithm, kNN [†]	50%	12	36	99.97	99.90	–	36	96.72	97.11	–	10
[23]	VCG	Classification and regression tree (CART)	–	3	10	–	–	91.00		88	92	89.1	3
[66]	VCG, PCA	RBF-NN	70%	3	9	94.23	99.09	–	9	–	–	95.35	10
[26]	Multi-scale energy DWT	SVM [†] , kNN	25%	12	60	93.00	99.00	96.00	72	–	–	99.58	6
[60]	Polynomial function fitting	Decision tree [†] , Random tree, Naive Bayes, SVM	50%	2	88	–	–	94.40					
[25]	DWT	kNN	90%	12	47	99.45	96.27	98.80	25	99.55	99.16	98.74	10
[31]	SWT	CNN [†] , SVM	99%	3	84	85.33	84.09	84.54					
[21]	PCA	SVM	50%	12	14	96.60	96.60	96.60					
[36]	DFT	Logistic Regression [†] , Threshold based model	80%	3	6	96.50	92.70	95.60					
[34]	‡	13-layer deep neural network	63%	1	0	99.79	97.44	99.34					
[32]	‡	Multi-branch CNN	80%	12	0	98.73	99.35	98.79	0	–	–	94.82	5
[40]	Fourier-Bessel series expansion, EWT	Deep layer least square SVM	70%	12	108	99.87	99.60	99.74					
[38]	‡	Fully connected CNN	90%	8					0	93.3	89.7	–	3
P.M.1	DWT, PCA	NN	60%	12	28	97.5	98.01	98.21	32	98.14	99.40	98.22	6
P.M.2	‡	Deep end-to-end NN	60%	12					0	100	100	100	7

(†) indicates the classifier with the highest performance that is being reported in this table.

(‡) indicates the studies that did not use any feature extraction methods.

(–) indicates the information that was not reported in the original study or could not be computed from the study.

One of the early studies on automating MI classification was proposed by Arif *et al.* [24]. Authors employed 36 time-domain features, a pruning algorithm, and kNN for 10 classes of MI and reported 96.72% and 99.97% specificity from their classifiers. Acharya *et al.* [25] used 125,652 healthy and 485,753 MI beats and built a kNN classifier with 47 features to detect MI and 25 features to localize 10 MI classes, respectively. Le *et al.* [23] proposed a method based on capturing the complex cardiac excitation and propagation dynamics as a random-walk network reconstructed from VCG signals. For this purpose, they extracted 161 features from VCG signals and their random-walk representation for MI localization. Using

statistical tests (i.e., Anderson–Darling and Kolmogorov Smirnov), the number of features were reduced to 10 and fed into a hierarchical classification and regression tree (CART), resulting a sensitivity of $\sim 88\%$ and a specificity of $\sim 92\%$ (with standard deviation $< 5\%$). Dohare *et al.* [21] used PCA to reduce the number of features from 220 to 14 and classified MI patients using RBF-SVM classifier. The average computational time for signal processing, training and testing the classifier was 24.5, 12.50, and 0.011 seconds, respectively.

Sharma *et al.* [26] employed multi-scale energy and the Eigen-space approach for feature extraction. Authors used kNN, linear, and radial basis function (RBF) SVM classifiers among which RBF-SVM achieved the highest accuracy. For MI detection, 60 extracted features resulted in 96% accuracy and for MI localization, 72 extracted features resulted 99.58% accuracy for SVM classifier with 6 classes. Computational time for training the MI detection and localization classifiers were 56.67 and 58.5 seconds, respectively. The method carried out by Noorian *et al.* [66] converted the signals to VCG signals and only used the magnitude vector which forfeited the valuable spatial information of the leads, particularly in MI localization. Authors reported 95.35% accuracy from one single run, which was experimented on a subset of 50 out of 368 MI records.

Exploring the MI detection literature that used CNN, Liu *et al.* [32] used 1D convolutional layers for the definition of feature extraction branches on a lead basis. Authors only distinguished among five different locations of the MI with the accuracy of 94.82%. In our second proposed method, we used up to three 1D convolutional and additional three mean-pooling layers per ECG lead providing a vector to a much simpler classifier consisting of a fully-connected layer with softmax activation. The relatively high computational complexity of the feature extraction branches in [32] causes large number of parameters. In summary, our second proposed method, deep residual neural network model, is able to achieve near perfect MI detection and localization performance with less computational complexity as well as detecting an additional MI location (i.e., infero-postero-lateral).

Deep residual network also has been previously employed for MI detection and localization by Strodthoff *et al.* [38]. Apart from healthy subjects, authors only distinguish between two rough classes colloquially designated as anterior myocardial infarction (aMI) and inferior myocardial infarction (iMI). In their approach, the multi-lead ECG signal is arranged into a matrix that serves as input to the first two-dimensional convolutional layer of the deep residual learning architecture. However, in our work, inspired by speech waveform processing with neural networks [67], we found a more robust front-end consisting of the use of 1D convolutional layers to obtain an intermediate two-dimensional representation for each lead of the ECG signal. These intermediate representations are then stacked across the depth dimension and the resulting feature volume is fed into the first two-dimensional convolutional layer of our deep residual neural network, which, also unlike [38], uses dilated convolutions and has a different structure.

6. Conclusion

The timely diagnosis of myocardial Infarction (MI) is very critical to save lives. To automate diagnose and localize MI, two classification methods are developed and evaluated, using 12-lead electrocardiograms. Machine learning and deep learning methods have shown several similarities and differences in their implementation, applications, and associated performance. This research aimed to highlight the differences in the prerequisite steps of classic and deep machine learning methods for processing ECG signals and evaluate their outcomes. In the classic approach, after the ECG signals are preprocessed, the discrete wavelet transform and principal component analysis are employed as the transformations techniques to extract distinctive feature sets from the signals. A two-class Neural Network classifier is developed to diagnose infarcted subjects by employing 28 features. Then, a multi-class NN is constructed to localize the infarcted part of myocardium by employing 32 features. The average accuracy of 98.21% and 98.22% is achieved in detection and localization, respectively. Our proposed MI detector outperforms other reported classifiers that have used the same database.

With the rise of deep learning, several studies report improved results compared to classical machine learning methods for the aforementioned problem. Therefore, we also developed a novel end-to-end deep residual learning model that achieves *semi*-perfect classification for MI detection and perfect MI localization. In this method, only preprocessing step was required prior to applying end-to-end deep learning architecture

with dilated convolutions. Consequently, the proposed framework can be confidently examined and utilized by clinical datasets that are largely unstructured and the distribution of classes are highly imbalanced.

While the proposed model shows the benefits of using deep learning technique to detect and localize the MI from ECG signals that are previously recorded, a future challenge is to develop a hardware that proactively detects MI disorders. One recommendation is attaching an extra electronic circuit to the standard ECG monitoring machines in the emergency room so that the ECG machine can simultaneously detect and localize any MI disorders while being used for monitoring and recording patient’s ECG signals.

Acknowledgements

Authors want to thank the reviewers and editor-in-chief of Applied Soft Computing, for their insightful comments. We are also grateful to Dr. Ivan Lopez-Espejo, Department of Electronic Systems, Aalborg University, Denmark and Neda Derakhshani, Data Scientist at CVS-Health, for their help with developing end-to-end approach.

Appendix A. Coronary Anatomy and Myocardial Infarction

A human heart consists of four major chambers (i.e., right atrium (RA), right ventricle (RV), left atrium (LA), and left ventricle (LV)). The right and left side of the heart has different functions. The right side of the heart receives de-oxygenated blood from the veins and pumps that through the pulmonary artery into the lungs for purification. On the other hand, the left side of the heart receives the oxygenated blood from the lungs and pumps that toward the body organs through the aorta [2]. The heart functionality is highly dependent on the coronary circulation system that supplies oxygenated blood directly toward cardiac muscles to keep the heart nourished and oxygenated. The coronary circulation system is comprised of several small vessels and coronary heart disease can cause myocardial infarction (MI) leading to the insufficient supply of nutrients and the oxygen-rich blood to a section of heart muscle [2].




Figure A.10: Coronary arteries structure. LCA and RCA are the main arteries which supply several branches [68].

According to Figure A.10, two principal coronary arteries that are derived from the aorta are left coronary artery (LCA) and right coronary artery (RCA). The LCA basically supplies the left side of the myocardium consisting of the left ventricle and the left atrium. Moreover, LCA branches off into the left anterior descending (LAD) which supplies the anterior wall of the left ventricle, and left circumflex (LCX) that equips the lateral wall of the left ventricle [2, 4]. In addition, the RCA divides into smaller branches, the posterior descending artery (PDA) and the acute marginal artery (AMA) which supply blood mainly to

the right atrium and ventricle walls. The RCA occlusion can cause conduction abnormalities and increase the risk of bradycardia and the atrioventricular block [2, 4]. The occlusion of LAD, which is the most commonly occluded artery, can cause right or left bundle branch block, ventricular tachycardia and cardiac dysrhythmias [4].

The electrical impulses generated by the myocardium regulate the beating of the heart or its rhythm. This rhythmical system is vulnerable to cardiovascular diseases, especially in cases that a patient suffers from ischemia of heart muscles [2]. Therefore, with the progressing of related research as well as developing new technologies, a better understanding of the pathophysiology of the MI has been obtained using 12-lead ECG as a principal method to detect and localize MI. By employing this technique, more efficient and effective treatment plan can be built for the patients suffering from the MI or other acute coronary syndromes [69, 2, 4].




Figure A.11: ECG 12-lead viewpoints. There are six limb leads permit vertical viewpoints and six chest leads permit horizontal viewpoints to the electrical activity of the heart [70].

Table A.9: ECG leads and corresponding viewpoints to the electrical activity of the heart.

Views of the Heart	Leads
Interior	II, III, aVF
Lateral	I, aVL, V5, V6
Anterior	V3, V4
Septal	V1, V2

Electrocardiogram monitors the electrical activity and rhythm of the heart. There are six limb leads called lead I, II, III, aVL, aVR and aVF, which enable vertical viewpoints to the electrical activity of the heart, and six chest or pre-cordial leads, V1 to V6, which allow horizontal viewpoints to the electrical activity of the heart [69]. The arrangement of ECG leads can yield a distinctive anatomical connection which is provided in Table A.9 and Figure A.11.

References

References

- [1] WHO, Cardiovascular diseases (2017).
URL http://www.who.int/cardiovascular_diseases/en/
- [2] J. Hall, Guyton and Hall Textbook of Medical Physiology, Elsevier, 2015 (2015).
- [3] A. T. F. Members, F. V. de Werf, J. Bax, A. Betriu, C. Blomstrom-Lundqvist, F. Crea, V. Falk, G. Filippatos, K. Fox, K. Huber, Management of acute myocardial infarction in patients presenting with persistent ST-segment elevation: The task force on the management of ST-segment elevation acute myocardial infarction of the European Society of Cardiology, *European Heart Journal* 29 (2008) 2909–2945 (2008).
- [4] A. Kucia, T. Quinn, Acute cardiac care: A practical guide for nurses, John Wiley & Sons, 2013 (2013).
- [5] Y.-C. Yeh, W.-J. Wang, C. W. Chiou, Cardiac arrhythmia diagnosis method using linear discriminant analysis on ECG signals, *Measurement* 42 (2009) 778–789 (2009).
- [6] Y.-C. Yeh, W.-J. Wang, C. W. Chiou, A novel fuzzy c-means method for classifying heartbeat cases from ECG signals, *Measurement* 43 (2010) 1542–1555 (2010).
- [7] L. Lu, J. Yan, C. W. de Silva, Feature selection for ECG signal processing using improved genetic algorithm and empirical mode decomposition, *Measurement* 94 (2016) 372–381 (2016).
- [8] X. Wan, Y. Li, C. Xia, M. Wu, J. Liang, N. Wang, A T-wave alternans assessment method based on least squares curve fitting technique, *Measurement* 86 (2016) 93–100 (2016).
- [9] S. Sahoo, B. Kanungo, S. Behera, S. Sabut, Multiresolution wavelet transform based feature extraction and ECG classification to detect cardiac abnormalities, *Measurement* 108 (2017) 55–66 (2017).
- [10] C. H. Peels, C. A. Visser, A. J. F. Kupper, F. C. Visser, J. P. Roos, Usefulness of two-dimensional echocardiography for immediate detection of myocardial ischemia in the emergency room, *The American Journal of Cardiology* 65 (1990) 687–691 (1990).
- [11] D.-Y. Tsai, K. Kojima, Measurements of texture features of medical images and its application to computer-aided diagnosis in cardiomyopathy, *Measurement* 37 (2005) 284–292 (2005).
- [12] R. J. D. Winter, R. W. Koster, A. Sturk, G. T. Sanders, Value of myoglobin, troponin T, and CK-MB mass in ruling out an acute myocardial infarction in the emergency room, *Circulation* 92 (1995) 3401–3407 (1995).
- [13] J. McCord, R. M. Nowak, P. A. McCullough, C. Foreback, S. Borzak, G. Tokarski, M. C. Tomlanovich, G. Jacobsen, W. D. Weaver, Ninety-minute exclusion of acute myocardial infarction by use of quantitative point-of-care testing of myoglobin and troponin I, *Circulation* 104 (2001) 1483–1488 (2001).
- [14] M. Plebani, M. Zaninotto, Diagnostic strategies using myoglobin measurement in myocardial infarction, *Clinica Chimica Acta* 272 (1998) 69–77 (1998).
- [15] P. Kora, ECG based myocardial infarction detection using hybrid firefly algorithm, *Computer Methods and Programs in Biomedicine* 152 (2017) 141–148 (2017).
- [16] M. Hammad, A. Maher, K. Wang, F. Jiang, M. Amrani, Detection of abnormal heart conditions based on characteristics of ECG signals, *Measurement* 125 (2018) 634–644 (2018).
- [17] S. L. Kasar, M. S. Joshi, A. Mishra, S. Mahajan, P. Sanjeevikumar, Performance enhancement for detection of myocardial infarction from multilead ECG, *Artificial Intelligence and Evolutionary Computations in Engineering Systems* (2018) 697–705 (2018).
- [18] L. Sun, Y. Lu, K. Yang, S. Li, ECG analysis using multiple instance learning for myocardial infarction detection, *IEEE Transactions on Biomedical Engineering* 59 (2012) 3348–3356 (2012).
- [19] P.-C. Chang, J.-J. Lin, J.-C. Hsieh, J. Weng, Myocardial infarction classification with multi-lead ECG using hidden Markov models and Gaussian mixture models, *Applied Soft Computing* 12 (2012) 3165–3175 (2012).
- [20] M. A. Waduud, E. N. Clark, A. Payne, C. Berry, M. Sejersten, P. Clemmensen, P. W. Macfarlane, Location of the culprit artery in acute myocardial infarction using the ECG, in: *2011 Computing in Cardiology, IEEE*, 2011, pp. 417–420 (2011).
- [21] A. K. Dohare, V. Kumar, R. Kumar, Detection of myocardial infarction in 12 lead ECG using support vector machine, *Applied Soft Computing* 64 (2018) 138–147 (2018).
- [22] M. Adam, S. L. Oh, V. K. Sudarshan, J. E. Koh, Y. Hagiwara, J. H. Tan, R. S. Tan, U. R. Acharya, Automated characterization of cardiovascular diseases using relative wavelet nonlinear features extracted from ECG signals, *Computer Methods and Programs in Biomedicine* 161 (2018) 133–143 (2018).
- [23] T. Q. Le, S. T. Bukkapatnam, B. A. Benjamin, B. A. Wilkins, R. Komanduri, Topology and random-walk network representation of cardiac dynamics for localization of myocardial infarction, *IEEE Transactions on Biomedical Engineering* 60 (2013) 2325–2331 (2013).
- [24] M. Arif, I. A. Malagore, F. A. Afsar, Detection and localization of myocardial infarction using k-nearest neighbor classifier, *Journal of Medical Systems* 36 (2012) 279–289 (2012).
- [25] U. R. Acharya, H. Fujita, V. K. Sudarshan, S. L. Oh, M. Adam, J. E. Koh, J. H. Tan, D. N. Ghista, R. J. Martis, C. K. Chua, Automated detection and localization of myocardial infarction using electrocardiogram: A comparative study of different leads, *Knowledge-Based Systems* 99 (2016) 146–156 (2016).
- [26] L. Sharma, R. Tripathy, S. Dandapat, Multiscale energy and eigenspace approach to detection and localization of myocardial infarction, *IEEE Transactions on Biomedical Engineering* 62 (2015) 1827–1837 (2015).
- [27] K. Nezamabadi, Z. Naseri, H. A. Moghaddam, M. Modarresi, N. Pak, M. Mahdizade, Lung hrct pattern classification for cystic fibrosis using convolutional neural network, *Signal, Image and Video Processing* 13 (6) (2019) 1225–1232 (Sep

2019). doi:10.1007/s11760-019-01447-y.

URL <https://doi.org/10.1007/s11760-019-01447-y>

- [28] J. Wu, Y. Bao, S.-C. Chan, H. Wu, L. Zhang, X.-G. Wei, Myocardial infarction detection and classification — a new multi-scale deep feature learning approach, in: 2016 IEEE International Conference on Digital Signal Processing (DSP), IEEE, 2016, pp. 309–313 (2016).
- [29] U. R. Acharya, H. Fujita, S. L. Oh, Y. Hagiwara, J. H. Tan, M. Adam, Application of deep convolutional neural network for automated detection of myocardial infarction using ecg signals, *Information Sciences* 415 (2017) 190–198 (2017).
- [30] M. Kumar, R. Pachori, U. Acharya, Automated diagnosis of myocardial infarction ecg signals using sample entropy in flexible analytic wavelet transform framework, *Entropy* 19 (9) (2017) 488 (2017).
- [31] T. Reasat, C. Shahnaz, Detection of inferior myocardial infarction using shallow convolutional neural networks, in: 2017 IEEE Region 10 Humanitarian Technology Conference (R10-HTC), IEEE, 2017, pp. 718–721 (2017).
- [32] W. Liu, Q. Huang, S. Chang, H. Wang, J. He, Multiple-feature-branch convolutional neural network for myocardial infarction diagnosis using electrocardiogram, *Biomedical Signal Processing and Control* 45 (2018) 22–32 (2018).
- [33] H. W. Lui, K. L. Chow, Multiclass classification of myocardial infarction with convolutional and recurrent neural networks for portable ecg devices, *Informatics in Medicine Unlocked* 13 (2018) 26–33 (2018).
- [34] N. Liu, L. Wang, Q. Chang, Y. Xing, X. Zhou, A simple and effective method for detecting myocardial infarction based on deep convolutional neural network, *Journal of Medical Imaging and Health Informatics* 8 (7) (2018) 1508–1512 (2018).
- [35] D. Rajan, D. Beymer, G. Narayan, Generalization studies of neural network models for cardiac disease detection using limited channel ecg, in: 2018 Computing in Cardiology Conference (CinC), Vol. 45, IEEE, 2018, pp. 1–4 (2018).
- [36] D. Sadhukhan, S. Pal, M. Mitra, Automated identification of myocardial infarction using harmonic phase distribution pattern of ecg data, *IEEE Transactions on Instrumentation and Measurement* 67 (10) (2018) 2303–2313 (2018).
- [37] L. D. Sharma, R. K. Sunkaria, Inferior myocardial infarction detection using stationary wavelet transform and machine learning approach, *Signal, Image and Video Processing* 12 (2) (2018) 199–206 (2018).
- [38] N. Strodthoff, C. Strodthoff, Detecting and interpreting myocardial infarction using fully convolutional neural networks, *Physiological Measurement* 40 (1) (2019) 015001 (jan 2019). doi:10.1088/1361-6579/aaf34d.
- [39] U. B. Baloglu, M. Talo, O. Yildirim, R. San Tan, U. R. Acharya, Classification of myocardial infarction with multi-lead ecg signals and deep cnn, *Pattern Recognition Letters* 122 (2019) 23–30 (2019).
- [40] R. K. Tripathy, A. Bhattacharyya, R. B. Pachori, A novel approach for detection of myocardial infarction from ecg signals of multiple electrodes, *IEEE Sensors Journal* 19 (12) (2019) 4509–4517 (2019).
- [41] J. Zhang, F. Lin, P. Xiong, H. Du, H. Zhang, M. Liu, Z. Hou, X. Liub, Automated detection and localization of myocardial infarction with staked sparse autoencoder and treebagger, *IEEE Access* 7 (2019) 70634–70642 (2019).
- [42] A. L. Goldberger, L. A. Amaral, L. Glass, J. M. Hausdorff, P. C. Ivanov, R. G. Mark, J. E. Mietus, G. B. Moody, C.-K. Peng, H. E. Stanley, Physiobank, physiotoolkit, and physionet, *Circulation* 101 (2000) 215–220 (2000).
- [43] M. Blanco-Velasco, B. Weng, K. E. Barner, ECG signal denoising and baseline wander correction based on the empirical mode decomposition, *Computers in Biology and Medicine* 38 (2008) 1–13 (2008).
- [44] X. Hu, Z. Xiao, N. Zhang, Removal of baseline wander from ECG signal based on a statistical weighted moving average filter, *Journal of Zhejiang University-SCIENCE C* 12 (2011) 397–403 (2011).
- [45] K. Daqrouq, ECG baseline wandering reduction using discrete wavelet transform, *Asian Journal of Information Technology* 4 (2005) 989–995 (2005).
- [46] A. Fasano, V. Villani, ECG baseline wander removal with recovery of the isoelectric level, in: 2015 Computing in Cardiology Conference (CinC), IEEE, 2015, pp. 577–580 (2015).
- [47] S. G. Mallat, A theory for multiresolution signal decomposition: The wavelet representation, *IEEE Transactions on Pattern Analysis and Machine Intelligence* 11 (1989) 674–693 (1989).
- [48] L. Sharma, S. Dandapat, A. Mahanta, ECG signal denoising using higher order statistics in wavelet subbands, *Biomedical Signal Processing and Control* 5 (2010) 214–222 (2010).
- [49] J. Shlens, A tutorial on principal component analysis (2014). [arXiv:arXiv:1404.1100](https://arxiv.org/abs/1404.1100).
- [50] F. Castells, P. Laguna, L. Sörnmo, A. Bollmann, J. M. Roig, Principal component analysis in ECG signal processing, *EURASIP Journal on Advances in Signal Processing* 2007 (1) (2007) 074580 (2007).
- [51] S. Xu, L. Chen, A novel approach for determining the optimal number of hidden layer neurons for FNN’s and its application in data mining, in: International Conference on Information Technology and Applications: iCITA, 2008, pp. 683–686 (2008).
- [52] D. W. Marquardt, An algorithm for least-squares estimation of nonlinear parameters, *Journal of the Society for Industrial and Applied Mathematics* 11 (1963) 431–441 (1963).
- [53] M. T. Hagan, M. B. Menhaj, Training feedforward networks with the Marquardt algorithm, *IEEE transactions on Neural Networks* 5 (1994) 989–993 (1994).
- [54] K. He, X. Zhang, S. Ren, J. Sun, Deep residual learning for image recognition, in: Proceedings of CVPR 2016 – Conference on Computer Vision and Pattern Recognition, June 26–July 1, Las Vegas, USA, 2016, pp. 770–778 (2016).
- [55] R. Tang, J. Lin, Deep residual learning for small-footprint keyword spotting, in: Proceedings of ICASSP 2018 – 43rd IEEE International Conference on Acoustics, Speech and Signal Processing, April 15–20, Calgary, Canada, 2018, pp. 5484–5488 (2018).
- [56] I. López-Espejo, Z.-H. Tan, J. Jensen, Keyword spotting for hearing assistive devices robust to external speakers, in: Proceedings of INTERSPEECH 2019 – 20th Annual Conference of the International Speech Communication Association, September 15–19, Graz, Austria, ISCA, 2019, pp. 3223–3227 (2019).
- [57] P. Luo, X. Wang, W. Shao, Z. Peng, Towards understanding regularization in batch normalization, in: International Conference on Learning Representations, 2019 (2019).

- [58] F. Yu, V. Koltun, Multi-scale context aggregation by dilated convolutions, arXiv preprint arXiv:1511.07122 (2015).
- [59] R. Bousseljot, D. Kreiseler, A. Schnabel, Nutzung der ekg-signaldatenbank cardiodat der ptb über das internet, Biomedizinische Technik/Biomedical Engineering 40 (s1) (1995) 317–318 (1995).
- [60] B. Liu, J. Liu, G. Wang, K. Huang, F. Li, Y. Zheng, Y. Luo, F. Zhou, A novel electrocardiogram parameterization algorithm and its application in myocardial infarction detection, Computers in Biology and Medicine 61 (2015) 178–184 (2015).
- [61] D. Ballabio, F. Grisoni, R. Todeschini, Multivariate comparison of classification performance measures, Chemometrics and Intelligent Laboratory Systems 174 (2018) 33–44 (2018).
- [62] F. Chollet, et al., Keras, <https://keras.io> (2015).
- [63] D. P. Kingma, J. Ba, Adam: A method for stochastic optimization, arXiv preprint arXiv:1412.6980 (2014).
- [64] J. Brownlee, Better Deep Learning: Train Faster, Reduce Overfitting, and Make Better Predictions, Machine Learning Mastery, 2018 (2018).
- [65] J. Cho, K. Lee, E. Shin, G. Choy, S. Do, How much data is needed to train a medical image deep learning system to achieve necessary high accuracy? (2015). [arXiv:arXiv:1511.06348](https://arxiv.org/abs/1511.06348).
- [66] A. Noorian, N. J. Dabanloo, S. Parvaneh, Wavelet based method for localization of myocardial infarction using the electrocardiogram, in: Computing in Cardiology 2014, IEEE, 2014, pp. 645–648 (2014).
- [67] T. N. Sainath, R. J. Weiss, A. Senior, K. W. Wilson, O. Vinyals, Learning the speech front-end with raw waveform CLDNNs, in: Sixteenth Annual Conference of the International Speech Communication Association, 2015 (2015).
- [68] P. J. Lynch, Coronary circulation, with coronary arteries labeled in red text and other landmarks in blue text (2010). URL https://commons.wikimedia.org/wiki/File:Coronary_arteries.svg
- [69] L. Biel, O. Pettersson, L. Philipson, P. Wide, ECG analysis: A new approach in human identification, IEEE Transactions on Instrumentation and Measurement 50 (2001) 808–812 (2001).
- [70] N. Patchett, Spatial orientation of EKG leads, Wikimedia Commons (2015). URL https://commons.wikimedia.org/wiki/File:EKG_leads.png