EMSENSE	**Complaint handling Form**	**DOCUMENT NO.**	ETPL/F/15
		REVISION NO.	01
		EFFECTIVE DATE.	03/10/2026

Complaint No:

Date:

Customer Information	
Customer name	
Customer contact	
Customer address	
Contact name	
Designation	
Customer PO No.	
Invoice No.	
Date and time of event	

Device Identification	
Product	
Device ID	
Serial Number	
IFU Revision available at site	
Installation/handover date	
Last service date, if applicable	
Device physically available for investigation?	

Event / Complaint Classification:

Tick all that apply:

☐ No Flow
☐ Low / Inadequate Flow
☐ Overflow / Excessive Flow
☐ Value Error
☐ Tube Error
☐ Air Bubble Alarm
☐ Empty Fluid / No Drop Alarm
☐ Battery / Charging Issue
☐ Display / Control Issue
☐ Connectivity / Mobile / Web Issue
☐ Device Mounting Issue
☐ Motor / Tube-Control Issue
☐ Physical Damage
☐ Unexpected Device Shutdown
☐ Other: _____

Description of complaint	
Describe exactly what was observed. Do not record only “device not working.	
Displayed alarm/error message:	
What was the device expected to do?	
What actually occurred?	

Infusion and setup Information	
IV fluid/container type	
Initial fluid volume	
Approx. remaining volume when event occurred	
IV administration set brand/type, if available:	
New IV set used for this infusion?	☐Yes ☐No ☐Unknown
Drip chamber within specified compatibility requirements?	☐Yes ☐No ☐Unknown
Drip chamber fluid level:	
IV stand / fluid-container height	
Tube calibration selected	
Programmed flow rate	
Programmed volume / time:	
Manual regulator/roller clamp opened gradually at startup?	☐Yes ☐No ☐Unknown

Immediate Patient / Clinical Information	
Was a patient connected when the event occurred?	☐Yes ☐No
Did the event result in or potentially contribute to patient harm?	☐ No known harm ☐Potential harm / near miss ☐ Patient harm reported ☐ Unable to determine
Brief description, where applicable	
Immediate clinical action was taken according to hospital procedure?	☐Yes ☐No ☐Not applicable

Note: *Clinical assessment and patient management remain under the responsibility of the treating healthcare team and applicable hospital procedures.*

Complete Infusion Pathway Check

Before attributing a flow-related event to the device, record the observations:

Check	Normal	Abnormal	Not Checked	Observation
IV fluid available in container				
Fluid container / bottle condition				
IV stand / gravity head condition				
Drip chamber compatibility				
Drip chamber fluid level				
Drip chamber correctly positioned in DripSet				
Manual regulator position				
IV tubing free from kink/twist/compression				
IV line correctly primed				
No unintended air in IV line				
Correct tube calibration selected				
Tube correctly centred/inserted/locked				
DripSet securely mounted				
Sensor area unobstructed				
Direct/high-intensity light exposure				
Device power/battery condition				
Downstream/patient-side pathway checked				

Immediate Action Taken

☐ Infusion paused/stopped as appropriate
☐ IV pathway checked
☐ Tubing repositioned
☐ IV set replaced
☐ Fluid container replaced
☐ Device repositioned correctly
☐ Device restarted after correction
☐ Alternate device used
☐ DripSet removed from service
☐ Device quarantined for investigation
☐ Emsensing informed
☐ Other: _____

Result after corrective action:

Emsensing Technical Investigation	
Device received for investigation	
Date received:	
Physical inspection	☐ Normal ☐ Abnormal
Device powered ON/self-check	☐ Pass ☐ Fail ☐ Not performed
Alarm / event log reviewed	☐ Yes ☐ No ☐ Not available
Functional test performed	☐ Pass ☐ Fail ☐ Not performed
Flow-control function checked	☐ Pass ☐ Fail ☐ Not performed
Sensor function checked	☐ Pass ☐ Fail ☐ Not performed
Motorized tube-control mechanism checked	☐ Pass ☐ Fail ☐ Not performed
Battery / power system checked	☐ Pass ☐ Fail ☐ Not performed
Complaint reproduced during investigation	☐ Yes ☐ No ☐ Intermittently ☐ Unable to test
Investigation observations	

Investigation Conclusion

Select only after completing the investigation:

☐ Confirmed device malfunction
☐ Consumable / IV administration set related
☐ Installation / mounting related
☐ Tube positioning / calibration related
☐ IV pathway / flow-condition related
☐ Power / battery related
☐ Environmental condition related
☐ User-operation / IFU-deviation related
☐ Patient-side / downstream condition potentially contributed
☐ No fault found / device met tested requirements
☐ Unable to determine from available evidence
☐ Other: _____

Evidence supporting conclusion:

Root Cause / CAPA Assessment

Root cause identified? ☐ Yes ☐ No ☐ Investigation ongoing

Root cause:

Correction / containment action:

Corrective action required? ☐ Yes ☐ No

Preventive action required? ☐ Yes ☐ No

CAPA Reference No., if opened: _____

Retraining required? ☐ Yes ☐ No

IFU / labelling review required? ☐ Yes ☐ No

Design / software / manufacturing review required? ☐ Yes ☐ No

Regulatory / Safety Assessment

Potential reportable adverse event?
☐ No ☐ Yes ☐ Requires Regulatory/QA assessment

Escalated to QA/RA: ☐ Yes ☐ No

QA/RA Reference: _____

Regulatory reporting decision / reference, if applicable:

Complaint Closure

Final action:
☐ Device returned to service
☐ Device repaired
☐ Device replaced
☐ Device retained for further investigation
☐ Device withdrawn from use
☐ User retraining completed
☐ Other: _____

Hospital informed of investigation outcome: ☐ Yes ☐ No

Date communicated: _____

Closure comments:

Responsibility	Name	Signature	Date
Investigation performed by			
Technical reviewer			
QA/RA reviewer			
Complaint closure approved by			