Malindi District Hospital
WorklistManager Summary Report

Director: Juma Mangi   **Date:** Fri, Nov 8, 2013
Operator: WARIO   **Time:** 11:08:41 AM
Software: WorklistManager 5.0.1

Sample Information		Location	Data Files	Status

1
Sample ID:	66			
Assay:	MultiSET			
Panel:	MALINDI/3/4/45			
Instr. Settings:	Calib File.LNW			
Collected:	Friday, November 8, 2013 10:49:07 AM			
CD3/CD4/CD45 TruC		1/1	3372801.01	OK

2
Sample ID:	67			
Assay:	MultiSET			
Panel:	MALINDI/3/4/45			
Instr. Settings:	Calib File.LNW			
Collected:	Friday, November 8, 2013 10:49:34 AM			
CD3/CD4/CD45 TruC		1/2		Error

Tube will not pressurize.

3
Sample ID:	68			
Assay:	MultiSET			
Panel:	MALINDI/3/4/45			
Instr. Settings:	Calib File.LNW			
Collected:	Friday, November 8, 2013 10:55:09 AM			
CD3/CD4/CD45 TruC		1/3	09103.01	OK

Did not get specified number of events.

4
Sample ID:	69			
Assay:	MultiSET			
Panel:	MALINDI/3/4/45			
Instr. Settings:	Calib File.LNW			
Collected:	Not Yet Run.			
CD3/CD4/CD45 TruC		1/4		Skipped

5
Sample ID:	70			
Assay:	MultiSET			
Panel:	MALINDI/3/4/45			
Instr. Settings:	Calib File.LNW			
Collected:	Not Yet Run.			
CD3/CD4/CD45 TruC		1/5		Skipped

6   Sample ID: 3   Worklist aborted
    Assay: MultiSET
    Panel: MALINDI/3/4/45
    Instr. Settings: Calib File.LNW
    Collected: Friday, November 8, 2013 11:08:41 AM
    CD3/CD4/CD45 TruC   1/6   Aborted