

BPM Alarm Settings

Current

Channel	Value	LOLO	LO	HI	HIHI	LOLO	LO	HI	HIHI
2C21A Current	0.000	-2	-1	8	9	MAJOR	MINOR	MINOR	MAJOR
2C24A Current	0.000	-2	-1	8	9	MAJOR	MINOR	MINOR	MAJOR
2H01 Current	0.000	-2	-1	8	9	MAJOR	MINOR	MINOR	MAJOR
Faraday Cup Current	-0.002	0	0	0	0	NO_ALARM	NO_ALARM	NO_ALARM	NO_ALARM

Position

Channel	Value	Min Current	Mean Pos	Minor Delta	Major Delta	Minor Severity	Major Severity
2C21A XPOS	0.000	4	-0.35	0.12	0.17	MINOR	MAJOR
2C21A YPOS	0.000	4	0.06	0.12	0.17	MINOR	MAJOR
2C24A XPOS	0.000	4	0.67	0.12	0.17	MINOR	MAJOR
2C24A YPOS	0.000	4	0.25	0.12	0.17	MINOR	MAJOR
2H00 XPOS	0.000	4	1.45	0.12	0.17	NO_ALARM	NO_ALARM
2H00 YPOS	0.000	4	0.5	0.15	0.2	NO_ALARM	NO_ALARM
2H01 XPOS	0.000	4	0.61	0.15	0.2	MINOR	MAJOR
2H01 YPOS	0.000	4	0.9	0.15	0.2	MINOR	MAJOR