


NOTES

- 1 ELEMENTS CROSS THE ELECTRON BEAM AS HEAD (FROM LEFT-TO-RIGHT)
- 2 MACHINE CENTER IS 400.0948 METERS UPSTREAM OF HALL CENTER\*
- 3 CONTENTS:  
PAGE 1: COMPONENT POSITIONING FOR REFERENCE BY COUNTING-ROOM & MCC OPERATORS  
PAGE 2: COMPONENT IDENTIFICATION FOR INSTALLATION BY TECHNICAL STAFF  
PAGE 3: IDENTIFICATION OF MATERIALS IN THE BEAM, & BEAMLINE SHIELDING
- 4 BEAM-TUBE IDENTIFICATION:  
A - TUBE ID  
B - TUBE OD  
C - DOWNSTREAM FLANGE SIZE  
D - LENGTH  
CONFLAT FLANGES UNLESS SPECIFIED

3-WIRE, X,Y,45°



FOR CONTINUATION  
SEE DRAWING NO.  
28405-0034

- 24.2 M : SHIELD WALL, UPSTREAM ALCOVE
- 23.755 M : 2003H RASTER MAGNET, 66844-02226
- 23.247 M : 2003V RASTER MAGNET, 66844-02226
- 22.769 M : 2024 Tagger HARP: 3-WIRE, 45°, Y,X
- 22.195 M : 2X COUNTERS, UPSTREAM
- 22.005 M : VALVE #VBV2C24A
- 21.925 M : 2024A PUMP-SET : VTP2C24A

- 16.213 M : COLLIMATOR BOX : EMPTY, 20mm, NO HOLE, 12mm
- 15.827 M : VALVE #VBV2H00

- 14.003 M : 2H00 BPM PM-M15 ASSY, 58432-0156
- 13.647 M : 2H00 QA MAGNET ASSY, 22232-0001
- 13.202 M : 2H00 QA MAGNET ASSY, 22232-0001
- 12.859 M : 2H00 BD MAGNET ASSY, MAG-002-0003-0001
- 12.636 M : 2H00 BD MAGNET ASSY, MAG-002-0003-0001
- 12.388 M : 2H00 PUMP-SET : VTP2H00

- 7.443 M : 2H01 BPM : 1PM2H01
- 7.166 M : 2H01 BPM : 1PM2H01V
- 6.867 M : 2H01 BPM : 1PM2H01H
- 6.507 M : 2H01 BPM : 3-WIRE, X,Y,45°
- 6.241 M : VALVE #VBF2H01A
- 5.957 M : VALVE #VBV2H01A
- 5.868 M : 4X COUNTERS, MIDSTREAM
- 4.750 M : SACLAY PUMP-SET : VTP2H01A

- 1.303 M : SOLENOID CENTER
- .000 M : HALL CENTER

- 7.923 M : 4X COUNTERS, DOWNSTREAM

- 17.658 M : HPS DIPOLE MAGNET (OFF BEAMLINE)

- 20.8 M : 2H02 PUMP-SET : VTP2H02
- 21.0 M : SHIELD WALL, DOWNSTREAM ALCOVE

- 23.463 M : MAGNET (DECOMMISSIONED)

- 25.157 M : HPS PHOTON DUMP
- 26.165 M : DOWNSTREAM VIEWER

- 28.1 M : SHIELD WALL, FARADAY CUP
- 29.757 M : RETRACTABLE BEAM STOPPER
- 31.170 M : FARADAY CUP

THIS DRAWING IS THE PROPERTY OF THE U.S. GOVERNMENT AND IS TO BE REPRODUCED IN WHOLE OR IN PART BY ANY PERSON OR ORGANIZATION WITHOUT PERMISSION OF THE U.S. GOVERNMENT.

DO NOT SCALE DRAWING

SEE PARTS LIST

02 HALL B LINE  
00 INSTALLATION TEMPERATURE  
BM2101-02-00-0000

DATE: 06/17/2024

BY: [Signature]

CHK: [Signature]

APP: [Signature]

1 OF 4