


**Standardized Protocol for Method Validation/Verification**  
Standard Operating Procedure  
Quality Assurance Unit  
Laboratory Services Section - Austin

**Table of Contents**

I. Purpose .....	2
II. Scope .....	2
III. Definitions .....	3
IV. Responsibilities and Authority .....	4
V. Validation and Verification and Guidelines .....	4
VI. Validation Sample Size .....	8
VII. Validation Records .....	9
VIII. References and Supporting Guidance Documents (not otherwise listed) .....	9

## I. Purpose

This document establishes guidelines for the minimum requirements to perform a validation or verification study. Validations must be performed for all non-standard and laboratory-developed methods. Verifications must be performed for all unmodified standard methods such as EPA and FDA official methods. The DSHS Laboratory is regulated and/or accredited by/to ISO 17025, TNI, CAP-CLIA, FDA, EPA, and the USDA. There are many commonalities in the validation and verification requirements of the applicable regulatory and/or accrediting bodies that the DSHS is subject to; this procedure consolidates the common requirements as applicable to all validations/verifications and isolates any additional requirements that are specific to a single standard that may be considered too cumbersome and/or irrelevant to areas not subject to those specific requirements. The conceptual view of the validation and verification process is shown in Figure 1 below.



## II. Scope

This procedure applies to all testing areas of the laboratory regardless of regulatory body. All laboratory tests must be validated or verified before being placed into routine use for testing and reporting of patient, animal, environmental, or surveillance samples.

<sup>1</sup> Mattocks, Christopher J et al. "A standardized framework for the validation and verification of clinical molecular genetic tests." *European journal of human genetics: EJHG* vol. 18,12 (2010): 1276-88. doi:10.1038/ejhg.2010.101

### III. Definitions

- A. Accuracy – Closeness of agreement between a measured value and a true value.
- B. Analytic Measurement Range (AMR) - The range of analyte values that a method can directly measure on the sample without any dilution, concentration, or other pretreatment not part of the usual assay process.
- C. Analytical Sensitivity – The smallest quantity of an analyte that can be reproducibly distinguished from background levels. Positive agreement as compared to reference method.
- D. Analytical Specificity – The ability of a method to detect only the analyte it is designed to detect. Negative agreement as compared to reference method.
- E. AOAC – Association of Official Analytical Chemists
- F. ASTM – American Society for Testing and Materials
- G. CAP – College of American Pathologists. Deemed to be an accreditation body by CLIA and currently directs the Laboratory Accreditation Program (LAP), established in 1961.
- H. CLIA- Clinical Laboratory Improvement Amendments of 1988. Responsible under the Centers for Medicare & Medicaid Services (CMS), an agency within the US Department of Health and Human Services for the regulation of clinical laboratories in the United States.
- I. Detection limit – A detection limit is the lowest amount of analyte in a sample which can be detected but not necessarily quantitated as an exact value. It is often called the limit of detection (LOD) which is the lowest concentration level that can be determined statistically different from a blank at a specified level of confidence. It is determined from the analysis of sample blanks. Method detection limit (MDL) is the minimum concentration of a substance than can be measured and reported with 99% confidence that the analyte concentration is greater than zero. It is determined from analysis of a sample in a given matrix containing the analyte.
- J. Diagnostic Sensitivity – The percentage of subjects with the target condition whose test values are positive.
- K. Diagnostic Specificity – The percentage of subjects without the target condition whose test values are negative
- L. FDA – U.S. Food and Drug Administration
- M. Limit of Quantitation (LOQ) – This is the level above which quantitative results may be determined with acceptable accuracy and precision.
- N. Linearity – Linearity is the ability of the method to elicit results that are directly proportional to analyte concentration within a given range.

- O. Non-standard Method – Refers to a method that is not taken from authoritative and validated sources. This includes methods from scientific journals and unpublished laboratory-developed methods.
- P. Precision – The ability of the laboratory to duplicate results time after time on different days and with different operators. Measures random error and the precision or imprecision can be expressed in CV% from the calculated standard deviation SD and mean. Repeat measurements of samples at varying concentrations, within-run and between run over a period of time should be performed.
- Q. Qualitative Method – A method that identifies analyte(s) based on chemical, biological, or physical properties; method of analysis whose response is either the presence or absence of the analyte detected either directly or indirectly in a certain amount of sample.
- R. Quantitative Method – A method that provides an estimate of the amount of analyte present in the test sample, expressed as a numerical value in appropriate units, with trueness and precision which are fit for the purpose.
- S. Range – A range is the interval between the upper and lower concentration of analyte in sample for which it has been demonstrated that the analytical procedure has an acceptable level of accuracy, precision, and linearity.
- T. Reference Range – The set of values that is deemed normal for a sample population.
- U. Reportable Range – The range of analyte concentrations that can be measured with acceptable accuracy and precision to include the Analytic Measurement Range (AMR). This can be determined by a linearity study for quantitative methods.
- V. Ruggedness or robustness – Ruggedness is a measure of an analytical procedure's capacity to remain unaffected by small, but deliberate variations in method parameters and provides an indication of its reliability during normal usage.
- W. Validation – A validation is the process of establishing the performance characteristics and limitations of a method and the identification of the influences which may change these characteristics and to what extent.
- X. Verification – The one-time process performed to determine or to confirm a test's expected performance compared to actual results produced by the laboratory.

#### **IV. Responsibilities and Authority**

Verification and validation studies must be planned and performed by competent personnel. The Laboratory Director is responsible for reviewing and approving all validation and verification plans and final reports. Any modifications to validation or verification plans, such as: changes to the procedure, acceptance criteria, or the decision to exclude data, must be approved in writing by a Branch Manager or Lead Technical Manager and the appropriate Quality Assurance Officer prior to making the modification.

#### **V. Validation and Verification and Guidelines**

##### References:

- TNI Standard EL-V1M2-2016-Rev2.1 Section 5.4.5
- TNI Standard EL-V1M4-2017-Rev2.2 Section 1.5 (Chemical Testing)
- TNI Standard EL-V1M5-2016-Rev2.0 Section 1.5 (Microbiological Testing)

- TNI Standard EL-V1M6-2016-Rev2.0 Section 1.5 (Radiochemical Testing)
- ISO 17025:2017 Section 7.2.2
- CAP Common Checklist items: COM.40250, COM.40350, COM.40475
- 42 CFR § 493.1253 Standard: Establishment and verification of performance specifications.

Standard methods, such as approved procedures from ASTM, FDA, EPA, etc., require only a verification study prior to use in the laboratory. ISO 17025:2017, TNI 2016, and CAP-CLIA require that procedures used without modification be subject to independent verification by the laboratory where testing will occur prior to being placed into routine use. A verification is a provision of objective evidence that a given item fulfills specified requirements. A verification study demonstrates that the laboratory can meet or exceed the method specifications of the manufacturer or published standard method.

Non-standard methods, standard methods used outside their intended scope or otherwise modified, and laboratory-developed methods/tests require validation. Laboratory validations shall be a planned activity assigned to qualified and authorized personnel and shall be as extensive as is necessary to meet defined method performance specifications.

Validation and verification plans shall include predefined acceptance criteria. Any change to the plan, acceptance criteria, or the decision to exclude any data from the final report must be documented, technically justified, and approved by a QA Officer and Branch Manager or Lead Technical Manager. When applicable, validation and verification study results should be compared to the reference method or 'gold standard' in which the new method is intended to replace or improve. These comparisons may include statistical analyses such as confidence intervals, linear regression, F-test, and *t*-test.

Validations shall comply with the requirements included in the references above and should generally evaluate the following method performance specifications as applicable:

- Accuracy
- Precision
- Reportable Range
- Reference Range
- Sensitivity
- Specificity
- False Positive/Negative Rate
- Interferences
- Carryover
- Detection Limit
- Limit of Quantitation
- Linearity
- Ruggedness
- Sample and reagent stability

Ruggedness is a quality characteristic that contains a potentially unlimited number of sources of variability that can be tested as part of a validation or verification. Some of the sources of variability that may be considered for investigation during a study are:

- Instrument to instrument
- Operator to operator
- Changes in sample quantity or concentration
- Reagent lot variability

- Time of day
- Temperature variations

The size and scope of validation and verification studies vary based on available time, cost, amount of testing material/template available, future use of method, and whether the method is qualitative or quantitative. When planning a validation or verification study the regulatory and/or accreditation body requirements will form the basis of the plan. Often there is a lack of clear guidance in these regulations and more information is needed to adequately plan a validation or verification. In these cases, it is necessary to refer to consensus standards such as those available from AOAC, ASTM, and CLSI (among others) for further guidance. If the regulatory/accreditation body requirements and the consensus standards fail to adequately address any 'grey area' in the validation/verification plan, the final step is to apply sound technical and statistical practices in consultation with internal and/or external subject matter experts to address any remaining issues. Records of these consultations shall be retained as part of the final validation/verification report.

Table 1 summarizes the general guidelines for performance characteristic evaluation of microbiological test methods from AOAC. Table 2 is adapted from the National Association of Testing Authorities (NATA) and summarizes the general guidelines for performance characteristic evaluation of chemical test methods. These tables should be used as guides when planning ISO 17025 and/or TNI validation and verification studies. Additionally, all TNI validations shall comply with Section 1.5 of the applicable module listed in the references section above. Table 3 contains validation and verification performance characteristics that must be determined for clinical testing under CAP/CLIA regulations as applicable.

**Table 1: Microbiological Test Method Performance Characteristic Guidelines<sup>2</sup>**

<b>Performance Characteristic</b>	<b>Identification</b>	<b>Quantitative</b>	<b>Qualitative (P/A)</b>	<b>Verification</b>
<b>Accuracy</b>	Yes	Yes	No	No
<b>Matrix Effects/Interferences</b>	No	Yes	Yes	Yes
<b>Precision</b>	No	Yes	No	Yes
<b>Selectivity</b>	No	Yes	Yes	No
<b>Specificity</b>	Yes	Yes	Yes	No
<b>Inclusivity</b>	Yes	Yes	Yes	No
<b>Exclusivity</b>	Yes	Yes	Yes	No
<b>False-Positive Rate</b>	No	Yes	Yes	No
<b>False-Negative Rate</b>	No	Yes	Yes	No
<b>LOD (sensitivity)</b>	No	Yes	Yes	No
<b>LOQ</b>	No	Yes	No	No
<b>Ruggedness</b>	Yes	Yes	Yes	No
<b>Linearity/Range</b>	No	Yes	No	No

<sup>2</sup> AOAC International. "How to Meet ISO 17025 Requirements for Method Verification." (2007).

**Table 2: Chemical Test Method Performance Characteristic Guidelines<sup>3</sup>**

<b>Characteristics to be evaluated</b>	<b>Validation</b>		<b>Verification</b>	
	Quantitative Method	Qualitative Method	Quantitative Method	Qualitative Method
<b>Accuracy</b>	Yes	Yes	Yes	Yes
<b>Limit of Detection</b>	Yes	No	Yes	No
<b>Limit of Quantitation</b>	Yes	No	Yes	No
<b>Sensitivity</b>	Yes	Yes	Yes	Yes
<b>Selectivity</b>	Yes	Yes	Yes	Yes
<b>Range/Linearity</b>	Yes	No	Yes	No
<b>Matrix Effects</b>	Yes	Yes	Yes	Yes
<b>Precision (repeatability/reproducibility)</b>	Yes	Yes	Yes	Yes
<b>Ruggedness</b>	Yes	Yes	No	No
<b>Measurement Uncertainty</b>	Yes	No	Yes*	No

\*In cases where a well-recognized test method specifies limits to the values of the major sources of measurement uncertainty and specifies the form of presentation of the calculated results, the laboratory is considered to have satisfied the measurement uncertainty requirements by following the test method and reporting instructions.

Note: Many regulatory/accreditation agencies require ongoing verification of some of the performance specifications listed above. These ongoing quality control measures should be incorporated into method SOPs and routine instrument maintenance plans at the appropriate intervals as applicable.

**Table 3: Requirements for Clinical Tests (CAP-CLIA)**

<b>Performance Characteristic</b>	<b>Validation (Laboratory Developed Tests and Modified FDA Tests)</b>	<b>Verification (FDA Approved Tests)</b>
<b>Accuracy</b>	Yes	Yes
<b>Precision</b>	Yes	Yes
<b>Sensitivity</b>	Yes	No
<b>Specificity</b>	Yes	No
<b>Reportable Range or AMR*</b>	Yes	Yes
<b>Reference Range</b>	Yes	No
<b>Interferences (COM.40500)</b>	Yes	No

\*CLSI EP28-A3c describes the process for determining this performance characteristic

A confusion matrix also should be included for all qualitative clinical tests. An example of a confusion matrix is show in Figure 2 below.

<sup>3</sup> NATA Technical Note 17 “Guidelines for the validation and verification of quantitative and qualitative test methods” (2012)

**Figure 2: Example Confusion Matrix**

Actual Values	Predicted Values		
	Positive	Negative	
Positive	True Positive (TP)	False Negative (FN)	<b>Sensitivity</b> $\frac{TP}{(TP + FN)}$
Negative	False Positive (FP)	True Negative (TN)	<b>Specificity</b> $\frac{TN}{(TN + FP)}$
	<b>Precision</b> $\frac{TP}{TP + FP}$	<b>Negative Predictive Value</b> $\frac{TN}{TN + FN}$	<b>Accuracy</b> $\frac{TP + TN}{(TP + TN + FP + FN)}$

There are situations where a validation or verification may need to be repeated. Some examples are: relocating an instrument, placing a new instrument into service, and modifications to methods. Some general guidelines for determining if a modification to a method is substantial enough to require a new validation or verification are provided in the references below:

- Appendix 5 and Appendix 6 of the FDA Guidelines for the Validation of Chemical Methods in Food, Feeds, Cosmetics, and Veterinary Products, (3<sup>rd</sup> ed. 2019)
- Attachment A of ORA Laboratory Procedure 5.4.5: Methods, Method Verification, and Validation

## VI. Validation Sample Size

### References:

- CLSI EP5-A2
- CLSI EP10-A3
- CLSI EP12-A2
- CLSI EP15-A2
- CAP Common Checklist COM.40350
- FDA Guidelines for the Validation of Analytical Methods for the Detection of Microbial Pathogens in Foods and Feeds, (3<sup>rd</sup> ed. 2019)
- FDA Guidelines for the Validation of Chemical Methods in Food, Feeds, Cosmetics, and Veterinary Products, (3<sup>rd</sup> ed. 2019)
- B. Magnusson and U. Örnemark (eds.) Eurachem Guide: The Fitness for Purpose of Analytical Methods – A Laboratory Guide to Method Validation and Related Topics, (2<sup>nd</sup> ed. 2014)

The number of samples included in a validation or verification study determines the statistical validity of the results and thus determines how much confidence can be placed in the results of the validation study. Definitive guidelines for validation/verification study sample numbers are limited. The references listed above give general guidelines for the number of samples that should be included in a study with the exception of the CAP, which sets the minimum at 20 samples for validation studies unless fewer samples are explicitly authorized by the Laboratory Director. Validations and verifications are a balance between costs, risks, and technical limitations. Generally, a higher quantity of test samples is desired and produces increased levels of statistical significance, however, it may not always be possible to have a large sample



size due to lack of availability or prohibitive costs. Table 4 below demonstrates the increased statistical validity of a qualitative test as the number of samples increases.

**Table 4: Samples required to determine false positive/negative rates (qualitative test)<sup>4</sup>**

FN or FP Rate	Confidence Level			
	80%	90%	95%	99%
<1%	161	230	299	459
<2%	80	114	149	228
<5%	32	45	59	90
<10%	16	22	29	44

## VII. Validation Records

The validation/verification plan template (DSHS Quality Assurance XXX Lab-Wide Validation/Verification plan template) and the validation/verification report template (DSHS Quality Assurance XXX Lab-Wide Validation/Verification report template) may be used for validations and verifications as applicable.

The following records shall be retained for all validations and verifications:

- A. Validation/verification procedure/plan;
- B. Specification of the requirements, including acceptance criteria;
- C. Determination of the performance characteristics of the method;
- D. Results obtained (including an explanation and investigation of discrepant/discordant results and any excluded data); and
- E. A statement on the validity of the method, detailing its fitness for the intended use.

Note: All validations and verifications (regardless of accrediting/regulatory body) shall comply with the following technical records requirement from ISO 17025: *'Records shall contain sufficient information to facilitate, if possible, identification of factors affecting the uncertainty and to enable the test to be repeated under conditions as close as possible to the original.'* This requires traceability to all aspects of the work performed including, but not limited to the following:

- Identity of personnel performing the work;
- Identity of all instruments, including pipettes;
- Reagent lot numbers and expiration dates; and
- Environmental conditions

Validation and verification records shall be kept for at least the life of the test plus two years.

## VIII. References and Supporting Guidance Documents (not otherwise listed)

- A. Jones, S., Carley, S., Harrison, M. An introduction to power and sample size estimation. *Emerg Med J* **20**, 453-458 (2003)
- B. Molecular Diagnostic Assay Validation (*Update to the 2009 AMP Molecular Diagnostic Assay Validation White Paper*), Association for Molecular Pathology, September 2014
- C. CLIA-Compliant Analytical Method Validation Plan and Template for LRN-C Laboratories, APhL (2013)

<sup>4</sup> FDA Guidelines for the Validation of Chemical Methods in Food, Feeds, Cosmetics, and Veterinary Products, (3rd ed. 2019)

- D. FSIS Guidance for Test Kit Manufacturers, Laboratories: Evaluating the Performance of Pathogen Test Kit Methods, USDA (2010)
- E. B. Magnusson and U. Örnemark (eds.) Eurachem Guide: The Fitness for Purpose of Analytical Methods – A Laboratory Guide to Method Validation and Related Topics, (2nd ed. 2014). ISBN 978-91-87461-59-0. Available from [www.eurachem.org](http://www.eurachem.org)
- F. V. Barwick (ed.), Planning and Reporting Method Validation Studies – Supplement to Eurachem Guide on the Fitness for Purpose of Analytical Methods (2019). Available from <http://www.eurachem.org>"
- G. Westgard J. O.: Basic Method Validation, Westgard Quality Corporation
- H. Sarewitz S.J.: CAP Accreditation Requirements for Validating Laboratory Tests, 7/9/13
- I. Lawrence Jennings, Vivianna M. Van Deerlin, and Margaret L. Gulley (2009) Recommended Principles and Practices for Validating Clinical Molecular Pathology Tests. Archives of Pathology & Laboratory Medicine: May 2009, Vol. 133, No. 5, pp. 743-755.