12.4 Biomedical Techniques

Blood Cell Counting using Haemocytometer

The haemocytometer is a thick glass slide with a counting chamber in the middle. The counting chamber contains two grids with improved Neubaur rulings of 3 by 3 mm primary square. The primary square is further subdivide into 9 secondary squares, each 1 by 1 mm. The four corner squares are used for the white blood cell count which are further subdivided into 16 tertiary squares. The central secondary square is divided into 25 tertiary squares, each of which measure 0.2 by 0.2 mm, each single tertiary square is further divided into 16 smaller squares. The five black squares along with the shaded squares in the centre are used for platelet count, while the five black squares alone are used for red blood cells counting (Figure12.10).

Diluting fluid

The blood cells are diluted in specific diluting fluid to keep the cells intact. RBC diluting fluid (Hayem's) is isotonic with blood, hence haemolysis does not occur. The blood is diluted 1:200 times with RBC diluting fluid and the cells are counted under 45X objective of the microscope.

The diluting fluid used for WBC count is Turk's solution which contains glacial acetic acid and Gentian violet. The glacial acetic acid lyses the red blood cells and the Gentian violet stains the nuclei of the leucocytes. The blood is diluted 1:20 times and the cell are counted under 10X objective of the microscope. The total number of cells counted is expressed in mm^3.

Blood cell counting using hemocytometer

1. The blood is collected till the 0.5 graduation in the pipette.
2. The diluting fluid is taken till the graduations 11 and 101 of the WBC and RBC pipette respectively.
3. The blood is diluted and mixed well with the respective diluting fluid by rotating the pipette horizontally several times.




Figure 12.10 Hemocytometer, RBC pipette and WBC pipette

TN_GOVT_IX_Zoology_Ch12.indd 231 09-01-2020 5.20.45 PM