

Electrophoresis apparatus: The electrophoretic apparatus consists of a reservoir tank to fill running buffer, transparent insulating cover, gel plates, spacers and gel comb to form wells. Platinum electrodes provide even current with the help of a regulated power pack. The gel is packed in-between two glass plates with the help of spacers. Clear wells are obtained using comb. Samples are layered in the little slots cut in the top of the gel slab using gel comb. Buffer is cautiously layered over the samples, and a voltage is applied to the gel using power pack for a period of usually 1-3 h. The proteins migrate in the gel depending upon their electrophoretic mobility, which is dependent on the size.





Protein samples and marker loaded in vertical SDS - PAGE system

Direction of migration of samples in vertical SDS - PAGE System



SDS - PAGE gel after Coomassie blue staining

Fig.10.10 SDS-PAGE

Unit 10.indd 267 21-05-2018 15:13:14