
Before a burette is filled with the solution, it is thoroughly washed, so that no greasy matter is sticking inside or outside the burette. Then rinse the burette with a small volume of titrant solution to be taken in it. Discard this solution through the stop-cock. After that,the burette is filled with titrant solution with the help of a funnel inserted in the top. The funnel must then be taken out after filling the burette.Care must be taken that no air bubbles remain in the narrow bottom tip of the burette. If air bubbles found in it, should be removed by running out the solution rapidly into the reagent bottle. Burette reading forms the most important aspect of the experiment, therefore, burette reading should be read very carefully.

In case of colourless solutions lower meniscus is read, while in case of coloured solutions, level is read from the upper meniscus. This is due to the reason that in case of coloured solutions lower meniscus is not visible clearly.Take reading of the burette placing your eye exactly in front of meniscus (Fig.) of the solution.



Pipette

The pipette is used for accurate measurements of definite volume of solution. It consists of a long narrow tube with cylindrical bulb in the middle and a jet at its lower end. On the upper part of the stem, there is an etched circular mark. Before a pipette is filled with the solution, it is washed and thoroughly rinsed with the solution to be measured with it. The upper part of pipette is then held by the thumb and middle finger of the right hand, the lower end is dipped into the liquid and the solution is sucked into the pipette until the liquid level is about 2 cm above the mark. The open end of pipette is then closed with index finger. The liquid is allowed to run slowly until the lower edge of meniscus just touches the mark. The solution is then allowed to run freely out of the pipette in the titration flask.

Practical_Bio-Chem-EM.indd 266 4/8/2020 3:28:52 PM