
Rapid slide test:

1. Clean the glass slide or test card supplied in the kit well and make it dry.
2. Label the circles (1, 2, 3, 4, 5 and 6) in the test card as O, H, AH, BH, Negative control and Positive control
3. Place a drop of undiluted test serum in each of the four labelled circle (1, 2, 3 and 4) ie O, H, AH and BH and place a drop of Negative control serum in circle 5 and Positive control in circle 6.
4. Place a drop of antigen O, H, AH and BH in circle 1, 2, 3, and 4 respectively and no antigen in circle 5 and O/H antigen in circle 6.
5. Mix the content of each circle with a separate wooden applicator stick and spread to fill the whole area of the individual circle.
6. Rock the test card for a minute and observe for agglutination.

Diagram



Observation

Agglutination was observed in O and H side within a minute which indicates the presence of antibodies in the serum sample against Salmonella typhi antigens.

Proceed for quantitative slide test or tube test for the quantitative estimation of the titre of the antibody.

Result

Qualtative widal test was carried out using rapid slide agglutination method. Antibodies against O and H antigens of Salmonella typhi were detected in the serum.

XII_Micro_Biology_Practical_Manual.indd 303 15-04-2019 17:13:16