
Family Planning Methods



Chemical Method

a. Foams
b. Creams, jellies and pastes
c. Soluble films

I. Mechanical Family Planning Method

a. Male condoms
b. Female condoms
c. Diaphragms
d. Cervical cap
e. Intrauterine devices (IUD)
f. Sponge

Hormonal Contraceptive Method

a. Contraceptive skin patch
b. Vaginal ring
c. Pills (Combined and Minipill)
d. Injection
e. Implant
f. Emergency Pill
g. Intrauterine device

COPPER T:

Copper T is a small T-shaped, barium-sulphate incorporated, polythene device that is placed inside uterus to obtain birth control. The placement is done with a plastic syringe called the IUD inserter.

A fine copper wire weighing 120 mg, with a surface area of 208 mm^2, is wound round the upright limb of T. Two fine filaments are attached to the lower tip of the vertical limb. Copper T along with its inserter is supplied in a pre-sterilized packet. Copper T is inserted on the 6th day following the menstrual period. The ideal time for postpartum insertion of Copper T is immediately after delivery .Copper T is introduced 12 weeks after abortion, a doctor or a trained paramedical person carries out the insertion.



Use of copper-T may lead to irregular bleeding, more cramps and pain during menstrual cycle. It is a widely preferred method of contraception. Complication may lead to miscarriage, preterm or infection rarely.

Types of IUDS

There are two types of IUDs:

- Nonmedicated Intrauterine Devices
- Medicated intrauterine devices

XII_Nursing_Vocational_Unit 5.indd 171 1/11/2020 2:46:18 PM