



LCR 999



Preliminary Data

Features

- Large LCD Display
- 0.3% Basic Accuracy, Resolution : 0.01%
- Primary and Secondary Parameter Dual Display
- 100Hz, 120Hz and 1KHz, 10KHz Test Frequencies
- Complete Four-Terminal Measurement
- Auto Power Off Function
- Battery and External Power Supply
- Sweep Frequency Correction Function
- 4 Bins Comparator

Specifications

Test Parameter	L-Q, C-D, R-Q, I-ZI-Q	
Test Frequency	100Hz, 120Hz, 1KHz	
Test Level	0.3Vrms (1 ± 10%)	
Basic Accuracy	0.3%	
Measuring Speed	3meas/sec (approx.)	
Equivalent Circuit	Series and Parallel	
Measuring Range	L	100Hz/120Hz 1μH - 9999H
		1KHz 0.1μH - 999.9H
		10KHz 0.01μH - 99.99H
	C	100Hz/120Hz 1pF - 9999μF
		1KHz 0.1pF - 999.9μF
		10KHz 0.01pF - 99.99μF
	R, IZI	0.0001Ω - 9.999MΩ
	Q	0.0001 - 9999
Ranging Mode	D	0.0001 - 9.999
	Δ%	-999.99% - 999.99%
	Auto and Hold	
	Display Mode	Direct, Δ%, ΔABS
Measurement Terminals	5 Terminals	
Calibration Function	Open and Short Sweep Zeroing	
Comparator Function	4 Bins : NC, P1, P2, P3	

MECO INSTRUMENTS PRIVATE LTD.

Plot No. EL-1, MIDC Electronic Zone, TTC Indl. Area, Mahape, Navi Mumbai 400 710. (INDIA)
 Tel : 0091-22-27673311-16 (Sales), 27673300 (Board) • Fax : 0091-22-27673310, 27673330
 Email : sales@mecoinst.com • Web : www.mecoinst.com