




Analog and Digital Insulation Tester

50th Anniversary



Model	Range	Test Voltage DC
MC901	0 - 20 M Ohms	100 V
MC902	0 - 50 M Ohms	250 V
MC903	0 - 100 M Ohms	500 V
MC904	0 - 500 M Ohms	500 V
MC941	0 - 1000 M Ohms	500 V
MC906	0 - 200 M Ohms	1000 V
MC907	0 - 500 M Ohms	1000 V
MC981	0 - 2000 M Ohms	1000 V

Features

- Single Person Push Button Operation
- High Accuracy $\pm 5\%$ of Indicated Value in Effective Range
- Terminal Voltage : $> 85\%$ of Rated Voltage for Insulation Resistance
- Measurement from 10% to 100% of the Insulation Resistance Range
- Scale Length : 80mm (approx.)
- ABS Resin Yellow Case with Acrylic Meter Cover
- Meets Requirement of IEC1010, Installation Category II, 1000V Phase Earth
- Designed to Generally Conform to IS 10656-1983
- 8 Models with various Voltage & MOhm Ranges
- Carrying Case

Features

- $3\frac{1}{2}$ Digit 17mm LCD Display
- High Accuracy $\pm (3\% \text{ rdg} + 2 \text{ dgt})$
- Single Person Push Button Operation
- Designed to Generally Conform to IS 10656-1983
- Meets Requirement of IEC1010, Installation Category II, 1000V Phase - Earth
- Terminal Voltage more than 85% of rated Voltage for Insulation Resistance Measurement from 10% to 100% of the Insulation Resistance Range
- 5 Models with various Voltage & MOhm Ranges
- Carrying Case

DIT 99A 100 V
20 MΩ

DIT 99B 250 V
200 MΩ

DIT 99C 500 V
200 MΩ

DIT 99D 1000 V
200 MΩ

DIT 99E 1000 V
2000 MΩ



An ISO 9001-2008 Company

MECO INSTRUMENTS PRIVATE LTD.

Plot No. EL-1, MIDC Electronic Zone, TTC Indl. Area, Mahape, Navi Mumbai 400 710. (INDIA)
Tel : 0091-22-27673311-16 (Sales), 27673300 (Board) ● Fax : 0091-22-27673310, 27673330
Email : sales@mecoinst.com ● Web : www.mecoinst.com

Visit us at



ELECRAMA-2012

18 - 22 January 2012
NSE, Mumbai

Hall 1
Stall P35



MIDDLE EAST
ELECTRICITY

07 - 09 February 2012
Dubai