example
COMMUNITY AUDIOLOGY

2 Demonstration Road, Sydney NSW 2000
02 5550 0142
hearing@communityaudiology.example.invalid

3 November 2026

Dr James Smith
Sydney Aged Care

**Re: Jane Doe | DOB 1 November 1935 | JD-19351101**

Sydney Aged Care

Dear Dr Smith,

## Hearing-aid service check

I saw Jane at Sydney Aged Care today for a brief hearing-aid service check after she reported that speech sounded muffled through the left aid. Her daughter Emily attended with Jane's agreement. This was a troubleshooting visit.

Jane has used hearing aids since September 2025 and would like to follow conversation more easily at meals. No new audiogram was obtained today.

| Check | Finding / action |
|---|---|
| External device inspection | Left wax guard obstructed; replaced. Right wax guard clear. |
| Listening check | Both aids sounded clear after servicing. |
| Patient response | Jane reported clearer speech with the left aid. |
| Programming / audiometry | No programming change or threshold testing undertaken. |

The immediate concern was consistent with a device maintenance problem. A satisfactory listening check is not a measurement of hearing thresholds and does not exclude a change in hearing. We did not undertake vestibular testing, and the reported dizziness should continue to be assessed by the medical team.

## Recommendations

1. Facility staff can help Jane check that the aids are seated and switched on before meals. Please report any return of muffled sound or discomfort.

2. Continue the booked hearing-aid review on 18 November, bringing both aids and any observations about communication in daily activities.

3. Use a quiet setting and written medication information for important discussions, allowing Jane time to respond.

Yours sincerely,
**Olivia Reed**
Audiologist