RegScoutX result report

Requirements and gap analysis under the MDR and the EU AI Act for AI-based medical device software

Assessment identifier	0faf35c7-ec36-434c-83ee-f84307cba12b
Revision	1 of 1
State	draft
Catalog version check	catalog versions match
Report language	English
Generated at	2026-09-28T22:14:43.686699+00:00

Contents

1. Assessment identification
2. Catalog and source versions used
3. Product profile and scope
4. Regulatory classification
5. Management summary
6. Gap overview
7. Results by theme cluster
8. Results by life-cycle phase
9. Results per requirement
10. Measures and evidence
11. Open items and review needs
12. Traceability and method
13. Limits of this report
14. Version and revision information

1. Assessment identification

Assessment identifier	0faf35c7-ec36-434c-83ee-f84307cba12b
Revision	1 of 1
State	draft
Origin of revision	created
Base revision	none
Revision created	2026-09-28T22:14:00.232632+00:00
Revision last changed	2026-09-28T22:14:35.832159+00:00
Domain fingerprint of the assessment	52ebd2bfc814506fc1722a9d1cea8caf66978eb2852cd53b5b1aa688968b9de1
Report schema version	1.0
Content hash of the report	fdfb9862fa527cb756e1c53ba52af89001eeab6dfabf86236e08424d34558ba2
Report language	English
Generated at	2026-09-28T22:14:43.686699+00:00

The assessment is uniquely identified by identifier, revision and domain fingerprint. Generation time and language are metadata; the content hash changes only with the domain content, the language or the schema version.

2. Catalog and source versions used

Catalog versions of the reported revision

Catalog	Version
Source register	1.6
Catalog of regulatory requirements (RR)	0.26
Catalog of decision rules (RUL)	0.2
Data model	0.8
Applicability mapping of requirements	0.4
Measure catalog (RR_MEASURE)	0.1

Check against the loaded catalogs: catalog versions match.

3. Product profile and scope

The product profile comprises 41 controlled inputs. Free text is not a decision input.

Recorded product inputs

Question	Input key	Answer
Does the system meet the definition of an AI system?	ai_system_status	Yes
How is the AI system related to the MDR product listed in Annex I Section A point 11?	annex_i_product_relation	The AI system is itself the MDR product
Is the AI system intended for any Annex III use case relevant to the RegScoutX MDSW scope?	annex_iii_use_case	No Annex III use case
To which date does the assessment refer (reference date 2 August 2028 for Article 6(1) systems)?	assessment_date	2028-09-01
Is the device a custom-made device?	custom_made_device	no
Does the software do more with data than storage, archival, communication, simple search or lossless compression? A search function that contributes to achieving a medical purpose does not count as simple search.	data_action_type	Data processing beyond that
Can decisions based on this information have an impact that may cause death or an irreversible deterioration of a person's state of health?	decision_impact_death_or_irreversible_deterioration	no
Can decisions based on this information have an impact that may cause a serious deterioration of a person's state of health or a surgical intervention?	decision_impact_serious_deterioration_or_surgery	no
Does the software drive a (hardware) medical device or influence its use – and if so, does it additionally have its own medical purpose?	drives_or_influences_device	Does not drive or influence a medical device
Does the software drive, or influence the use of, an active device intended to administer and/or remove medicinal products, body liquids or other substances to or from the body?	drives_substance_administration_or_removal_device	no
How does the manufacturer provide the instructions for use electronically?	eifu_provision	No electronic instructions for use
Could further MDR classification rules not asked here apply to the product – for example Rules 9 or 10 for hardware or combination products?	further_unmodelled_rule_possibly_applicable	no
Which IEC 62304 route applies to the software?	iec62304_route	regular life-cycle processes (Clauses 5 to 9)
Which IEC 62366-1 route applies to the user interface?	iec62366_ui_route	regular route under Clauses 5.1 to 5.9
Is the device an implantable device?	implantable_device	no
Is this data processing performed for the benefit of individual patients – and not only to aggregate population data, for generic diagnostic or treatment pathways, literature, atlases, models, templates, epidemiological studies or registers?	individual_patient_benefit	yes
Is the device an investigational device?	investigational_device	no
Is the product software, i.e. a set of instructions that processes input data and creates output data?	is_software	yes
Does the organization claim justified exclusions of requirements in Clauses 6 to 8 under ISO 13485 Clause 1?	iso13485_clause1_exclusion_claimed	no
Does the software provide information within the definition of an in vitro diagnostic medical device, i.e. concerning physiological or pathological processes or states, congenital physical or mental impairments, the predisposition to a medical condition or disease, safety and compatibility with potential recipients, predicted treatment response or reactions, or the definition or monitoring of therapeutic measures?	ivdr_information_purpose	no
Is the device intended for use by lay persons?	lay_use	yes
When was the high-risk system placed on the market or put into service (reference date 2 August 2028)?	market_or_service_date	2028-09-01

Question	Input key	Answer
Does the organization outsource processes?	outsourced_processes	no
Is the software itself intended to fulfil one of the medical purposes set out in the MDR – e.g. diagnosis, prediction, monitoring or treatment of a disease?	own_medical_purpose	yes
Does the software provide information which is used to take decisions with diagnostic or therapeutic purposes?	provides_information_for_diagnostic_or_therapeutic_decisions	yes
Does the organization purchase products or services (purchasing under ISO 13485 Clause 7.4)?	purchasing_performed	yes
Does the organization use computer software in the quality management system?	qms_software_used	yes
Is safety information used as a risk-control measure?	safety_information_as_risk_control	yes
Does a relevant Article 2 exclusion apply?	scope_exclusion	No exclusion
Is there a territorial or operator nexus under Article 2(1)?	scope_nexus	Product manufacturer placing the AI system on the market or putting it into service together with its product under its own name or trademark
Are changes made to the software?	software_change	no
Are there problem reports or has a software problem been identified?	software_problem_reported	no
Which IEC 62304 software safety class is assigned to the software system?	software_safety_class	Class B
Does the software contain SOUP items?	soup_used	yes
Does one of the following apply: the software is an MDR Annex XVI product, it is an accessory for a medical device, or it drives a (hardware) medical device or influences its use?	special_route	None of these
Is the software part of an active therapeutic device with an integrated or incorporated diagnostic function which significantly determines the patient management by the device, such as a closed loop system?	therapeutic_with_integrated_diagnostic_function	no
Is that third-party assessment required solely because of risks unrelated to health and safety?	third_party_ca_only_non_health_safety	No
Is third-party conformity assessment required for the MDR product for placing on the market or putting into service?	third_party_ca_required	Yes
Is training required for safe use?	training_required_for_safe_use	no
Is a usability test planned for the user-interface evaluation?	usability_test_planned	yes
Is the software intended to be used for contraception or prevention of the transmission of sexually transmitted diseases?	used_for_contraception_or_sti_prevention	no

4. Regulatory classification

The classification follows the catalogued decision rules only. Open review points and missing inputs are never presented as a decision.

Decision areas of the regulatory classification

Decision area	Status	Outcome	Details
Qualification as medical device software	decided	Medical device software (MDSW); MDR/IVDR boundary follows	none
MDR/IVDR boundary	decided	MDSW within the scope of the MDR	none
MDR classification	decided	Final MDR class: strictest reliably determined class candidate (only without a review outcome on the path)	MDR class IIa
AI Act applicability	decided	High-risk AI system; high-risk requirements in effect	applicable from 2028-08-02

Qualification as medical device software

Applied decision rules with source references

Rule	Primary law	Guidance
RUL-001	SRC-001#Art.2/Nr.1	SRC-004#Abschnitt 2; SRC-004#Abschnitt 3.3
RUL-002	SRC-001#Art.1/Abs.2; SRC-001#Anh.XVI; SRC-001#Art.2/Nr.2; SRC-001#Anh.VIII/Kap.II/Abschn.3.3/Satz.1	SRC-004#Abschnitt 3.3
RUL-003	SRC-001#Art.2/Nr.1	SRC-004#Abschnitt 3.1; SRC-004#Abschnitt 3.3
RUL-004	SRC-001#Art.2/Nr.1	SRC-004#Abschnitt 3.3
RUL-005	SRC-001#Art.2/Nr.1; SRC-001#Art.2/Nr.12	SRC-004#Abschnitt 3.1; SRC-004#Abschnitt 3.3

MDR/IVDR boundary

Applied decision rules with source references

Rule	Primary law	Guidance
RUL-006	SRC-010#Art.2/Nr.2	SRC-004#Abschnitt 3.4

MDR classification

Applied decision rules with source references

Rule	Primary law	Guidance
RUL-009	SRC-001#Anh.VIII/Kap.II/Abschn.3.3/Satz.1; SRC-001#Anh.VIII/Kap.II/Abschn.3.3/Satz.2	SRC-004#Abschnitt 4.1
RUL-010	SRC-001#Anh.VIII/Kap.III/Abschn.6.3/Regel.11/UAbs.1	SRC-004#Abschnitt 4.2.1
RUL-011	SRC-001#Anh.VIII/Kap.III/Abschn.6.3/Regel.11/UAbs.1	SRC-004#Abschnitt 4.2.1
RUL-012	SRC-001#Anh.VIII/Kap.III/Abschn.6.3/Regel.11/UAbs.1	SRC-004#Abschnitt 4.2.1
RUL-016	SRC-001#Anh.VIII/Kap.III/Abschn.7.2/Regel.15	SRC-004#Abschnitt 4.2.4
RUL-017	SRC-001#Anh.VIII/Kap.III/Abschn.7.9/Regel.22	SRC-004#Abschnitt 4.2.5
RUL-018	SRC-001#Anh.VIII/Kap.III/Abschn.6.4/Regel.12; SRC-001#Anh.VIII/Kap.II/Abschn.3.3/Satz.1	SRC-004#Abschnitt 4.2.2
RUL-019	SRC-001#Anh.VIII/Kap.III/Abschn.6.5/Regel.13	SRC-004#Abschnitt 4.2.3
RUL-020	SRC-001#Anh.VIII/Kap.II/Abschn.3.5	SRC-004#Abschnitt 4.1; SRC-004#Abschnitt 4.2

AI Act applicability

Applied decision rules with source references

Rule	Primary law	Guidance
RUL-027	SRC-002#Art.3/Nr.1	none
RUL-028	SRC-002#Art.2/Abs.1	none
RUL-029	SRC-002#Art.2/Abs.3; SRC-002#Art.2/Abs.6; SRC-002#Art.2/Abs.8	none
RUL-030	SRC-002#Art.6/Abs.1/Buchst.a; SRC-002#Anh.I/Abschn.A/Nr.11	none
RUL-032	SRC-002#Art.6/Abs.1/Buchst.b	none
RUL-033	SRC-002#Art.6/Abs.1c	none
RUL-034	SRC-002#Art.113/Buchst.c/Ziff.ii	none
RUL-035	SRC-002#Art.111/Abs.2	none
RUL-038	SRC-002#Art.6/Abs.2; SRC-002#Anh.III/Nr.1/Buchst.a; SRC-002#Anh.III/Nr.1/Buchst.b; SRC-002#Anh.III/Nr.1/Buchst.c; SRC-002#Anh.III/Nr.5/Buchst.a; SRC-002#Anh.III/Nr.5/Buchst.c; SRC-002#Anh.III/Nr.5/Buchst.d	none
RUL-043	SRC-002#Art.6	none

5. Management summary

Of 959 catalogued requirements, 542 are applicable, 128 are not applicable and for 289 applicability is unresolved. There are 15 gaps; 523 applicable requirements are not yet assessed.

Assessment flow

Product profile	Applicability	Requirements	Gaps	Evidence	Report (revision)
41	**542 of 959**	**831 of 959**	**15 of 542**	**8 of 542**	**1 of 1**

Key figures of the assessment

Figure	Count
Requirements in the catalog	959
Applicable	542
Not applicable	128
Applicability unresolved	289
Fulfilled (user statement with evidence reference)	4
Partially fulfilled	4
Not fulfilled	4
Not assessed	523
Evidence review required	7
Gaps	15
Measure assignment open (applicable requirements)	284
Confirmation of earlier assessments required	0

The key figures describe the recorded state of the assessment. They are not a statement of conformity.

6. Gap overview

The gap status combines applicability with the user's structured fulfilment assessment. A missing assessment or a missing evidence reference is never counted as “not fulfilled”.

Distribution of applicable requirements by gap status



fulfilled (user statement with evidence reference): 4 · partially fulfilled (user statement with evidence reference): 4 · not fulfilled (user statement): 4 · evidence review required: 7 · not assessed: 523

Gap status of all requirements

Gap status	Count	Share of applicable
fulfilled (user statement with evidence reference)	4	0.7%
partially fulfilled (user statement with evidence reference)	4	0.7%
not fulfilled (user statement)	4	0.7%
not assessed	523	96.5%
evidence review required	7	1.3%
not applicable	128	not included
applicability unresolved	289	not included

A gap exists for partial or missing fulfilment and for claimed fulfilment without an evidence reference. Requirements that are not applicable or unresolved are not included in the share.

7. Results by theme cluster

Every requirement belongs to exactly one theme cluster of the controlled vocabulary.

Applicability per Theme cluster

Theme cluster	Total	Applicable	Not applicable	Unresolved
Quality and Risk Management	363	199	37	127
Data Governance	21	1	0	20
Transparency	98	19	32	47
Human Oversight	10	9	0	1
Robustness	16	8	6	2
Post-Market Monitoring	138	92	13	33
Software Life Cycle	207	136	30	41
Usability	106	78	10	18

Gap status per Theme cluster (applicable requirements only)

Theme cluster	Fulfilled	Partially fulfilled	Not fulfilled	Evidence review	Not assessed	Gaps
Quality and Risk Management	1	1	2	2	193	5
Data Governance	0	0	0	0	1	0
Transparency	1	0	0	1	17	1
Human Oversight	0	1	0	0	8	1
Robustness	0	0	1	1	6	2
Post-Market Monitoring	1	1	0	1	89	2
Software Life Cycle	0	0	1	1	134	2
Usability	1	1	0	1	75	2

8. Results by life-cycle phase

A requirement can belong to several life-cycle phases and is then counted in each of them. The phases are taken from the controlled vocabulary.

Applicability per Life-cycle phase

Life-cycle phase	Total	Applicable	Not applicable	Unresolved
Planning and Conception	339	260	22	57
Development and Realisation	444	250	58	136
Verification and Validation	276	162	24	90
Conformity Assessment	101	84	4	13
Placing on the Market, Putting into Service and Operation	272	145	48	79
Post-Market Monitoring and Surveillance	257	160	28	69
Modification and Maintenance	489	233	97	159

Gap status per Life-cycle phase (applicable requirements only)

Life-cycle phase	Fulfilled	Partially fulfilled	Not fulfilled	Evidence review	Not assessed	Gaps
Planning and Conception	2	1	2	4	251	7
Development and Realisation	1	2	2	4	241	8
Verification and Validation	0	0	0	2	160	2
Conformity Assessment	0	0	0	1	83	1
Placing on the Market, Putting into Service and Operation	1	0	1	3	140	4
Post-Market Monitoring and Surveillance	1	1	0	1	157	2
Modification and Maintenance	0	1	2	4	226	7

9. Results per requirement

Applicable and unresolved requirements are shown in full with requirement text, sources, derivation, measure and evidence; requirements that are not applicable are listed with their reason in an overview.

Quality and Risk Management (326)

RR-007 · Establishment of the Risk Management System

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that a risk management system is established for each high-risk AI system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.9/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-008 · Implementation of the Risk Management System

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the risk management system of the high-risk AI system is implemented.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-002#Art.9/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-009 · Documentation of the Risk Management System

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the risk management system of the high-risk AI system is documented.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-002#Art.9/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-010 · Maintenance of the Risk Management System

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the risk management system of the high-risk AI system is maintained.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-011 · Lifecycle Planning of the Risk Management Process

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the risk management process is planned as a continuous iterative process for the entire lifecycle of the high-risk AI system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.9/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-012 · Continuous Iterative Operation of the Risk Management Process

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the risk management process is run continuously and iteratively throughout the entire lifecycle of the high-risk AI system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-013 · Regular Systematic Review of the Risk Management System

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the risk management system is subject to regular systematic review.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-014 · Updating of the Risk Management System

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the risk management system is updated as part of the continuous iterative lifecycle process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-015 · Identification of Known and Foreseeable Risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall identify known and reasonably foreseeable risks that the high-risk AI system may pose to health, safety or fundamental rights when used in accordance with its intended purpose, insofar as they fall within the risk scope defined by Article 9(3).

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Buchst.a`
Supplementary sources	`SRC-002#Art.9/Abs.2/Satz.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-016 · Analysis of Known and Foreseeable Risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall analyse known and reasonably foreseeable risks that the high-risk AI system may pose to health, safety or fundamental rights when used in accordance with its intended purpose, insofar as they fall within the risk scope defined by Article 9(3).

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Buchst.a`
Supplementary sources	`SRC-002#Art.9/Abs.2/Satz.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-017 · Estimation of Risks During Intended Use and Foreseeable Misuse

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall estimate the risks that may emerge when the high-risk AI system is used in accordance with its intended purpose and under conditions of reasonably foreseeable misuse, insofar as they fall within the risk scope defined by Article 9(3).

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Buchst.b`
Supplementary sources	`SRC-002#Art.9/Abs.2/Satz.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-018 · Evaluation of Risks During Intended Use and Foreseeable Misuse

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall evaluate the risks that may emerge when the high-risk AI system is used in accordance with its intended purpose and under conditions of reasonably foreseeable misuse, insofar as they fall within the risk scope defined by Article 9(3).

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Buchst.b`
Supplementary sources	`SRC-002#Art.9/Abs.2/Satz.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-019 · Evaluation of Additional Risks from Post-Market Data

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The provider shall evaluate other risks that may arise on the basis of the analysis of data from the post-market monitoring system, insofar as those risks fall within the risk scope defined by Article 9(3).

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.9/Abs.2/Buchst.c`
Supplementary sources	`SRC-002#Art.9/Abs.2/Satz.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-020 · Adoption of Appropriate Targeted Risk Management Measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall adopt appropriate and targeted risk management measures to address the risks identified under Article 9(2)(a) within the risk scope limited by Article 9(3).

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.2/Buchst.d`
Supplementary sources	`SRC-002#Art.9/Abs.2/Satz.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-021 · Consideration of Interactions Between Regulatory Requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall give due consideration, when applying risk management measures, to the effects and possible interactions arising from the combined application of the requirements in Chapter III, Section 2, so that risks are effectively minimised while an appropriate balance is maintained in implementing those measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.4`
Supplementary sources	`SRC-002#Art.9/Abs.2/Buchst.d; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-022 · Acceptability of Hazard-Specific Residual Risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that each relevant residual risk associated with a specific hazard is judged acceptable after application of the risk management measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.5/UAbs.1`
Supplementary sources	`SRC-002#Art.9/Abs.2/Buchst.d; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-023 · Acceptability of the Overall Residual Risk

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the overall residual risk of the high-risk AI system is judged acceptable after application of the risk management measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.5/UAbs.1`
Supplementary sources	`SRC-002#Art.9/Abs.2/Buchst.d; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-024 · Elimination or Reduction of Risks Through Design and Development

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

When determining the most appropriate risk management measures, the provider shall ensure that identified and evaluated risks are eliminated or reduced, insofar as technically feasible, through adequate design and development of the high-risk AI system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.5/UAbs.2/Buchst.a`
Supplementary sources	`SRC-002#Art.9/Abs.5/UAbs.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-025 · Mitigation and Control Measures for Non-Eliminable Risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

Where appropriate, the provider shall implement adequate mitigation and control measures for risks that cannot be eliminated.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.5/UAbs.2/Buchst.b`
Supplementary sources	`SRC-002#Art.9/Abs.5/UAbs.2; SRC-002#Art.9/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-026 · Training of Deployers as a Risk Management Measure

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> When determining the most appropriate risk management measures, the provider shall, where appropriate, ensure appropriate training of deployers.

Theme cluster	Quality and Risk Management
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.9/Abs.5/UAbs.2/Buchst.c`
Supplementary sources	`SRC-002#Art.9/Abs.5/UAbs.2; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Ensure and demonstrate competence or training (MEAS_COMPETENCE)
Planned evidence type	Competence or training record
Evidence references	none
Open items	Resolve applicability

RR-027 · Testing to Identify Appropriate Risk Management Measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the high-risk AI system is tested for the purpose of identifying the most appropriate and targeted risk management measures, at appropriate times throughout the development process and in any event before it is placed on the market or put into service, against predefined metrics and probabilistic thresholds appropriate to its intended purpose.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.9/Abs.6/Satz.1; SRC-002#Art.9/Abs.8/Satz.1; SRC-002#Art.9/Abs.8/Satz.2`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-028 · Testing for Intended-Purpose Performance and Section 2 Compliance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that testing confirms that the high-risk AI system performs consistently in accordance with its intended purpose and complies with the requirements of Chapter III, Section 2.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-002#Art.9/Abs.6/Satz.2`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-029 · Consideration of Minors and Other Vulnerable Groups

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> When implementing the risk management system, the provider shall consider whether, in view of its intended purpose, the high-risk AI system is likely to adversely affect persons under the age of 18 or, where appropriate, other vulnerable groups.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.9/Abs.9`
Supplementary sources	none
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-063 · Consideration of Deployer Characteristics and Context of Use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall, with a view to eliminating or reducing risks related to the use of the high-risk AI system, give due consideration to the technical knowledge, experience, education and training to be expected from the deployer, as well as the presumable context in which the system is intended to be used.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-002#Art.9/Abs.5/UAbs.3`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-081 · Establishment of a Compliance-Ensuring Quality Management System

Applicability **applicable** · Gap status **not fulfilled (user statement)** · Assessment state **active** · Measure **measure assignment open (review required)**

> The provider shall put in place a quality management system that ensures compliance with the Regulation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The user assessed the requirement as not fulfilled.
User fulfilment assessment	not fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Close gap; Review measure assignment

RR-082 · Systematic and Orderly QMS Documentation

Applicability **applicable** · Gap status **fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assigned by rule**

> The provider shall document the quality management system in a systematic and orderly manner in the form of written policies, procedures and instructions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	Fallstudie: Nachweis zu RR-082 (CS-DOC-RR-082; 1.0; identifier cdf2f10e-bcf9-4f15-ad26-d543b0553426)
Open items	none

RR-083 · Strategy for Regulatory Compliance

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes a strategy for regulatory compliance, including compliance with the applicable conformity assessment procedures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1/Buchst.a`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-089 · Quality Control Techniques, Procedures and Systematic Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes techniques, procedures and systematic actions to be used for quality control of the high-risk AI system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.c`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-090 · Quality Assurance Techniques, Procedures and Systematic Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes techniques, procedures and systematic actions to be used for quality assurance of the high-risk AI system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.c`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-092 · Applicable Technical Specifications and Standards

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes the technical specifications, including standards, to be applied.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1/Buchst.e`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-093 · Alternative Means to Ensure Compliance with Section 2

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

Where the relevant harmonised standards are not applied in full or do not cover all relevant requirements set out in Section 2, the provider shall ensure that the quality management system includes the means to be used to ensure that the high-risk AI system complies with those requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1/Buchst.e`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-095 · Inclusion of the Risk Management System in the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes the risk management system referred to in Article 9.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1/Buchst.g`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-099 · Management of Communications

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes the handling of communication with national competent authorities, other relevant authorities, including those providing or supporting access to data, notified bodies, other operators, customers and other interested parties.

Theme cluster	Quality and Risk Management
Life-cycle phases	Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.17/Abs.1/Buchst.j`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-100 · Record-Keeping Systems and Procedures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes systems and procedures for record-keeping of all relevant documentation and information.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.17/Abs.1/Buchst.k`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-101 · Resource Management

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes resource management, including security-of-supply related measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.l`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-102 · Accountability Framework and Responsibilities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes an accountability framework setting out the responsibilities of management and other staff with regard to all aspects listed in Article 17(1).

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.17/Abs.1/Buchst.m`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-103 · Proportionate QMS Implementation According to Organisational Size

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall implement the aspects referred to in Article 17(1) proportionately to the size of its organisation, in particular where the provider is an SME, including a start-up, or an SMC.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.17/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-104 · Required Degree of Rigour and Level of Protection

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall in all cases respect the degree of rigour and the level of protection required to ensure the compliance of its high-risk AI system with the Regulation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-002#Art.17/Abs.2/Satz.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-116 · Establishment of the Risk Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall establish a risk management system as described in Section 3 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.10/Abs.2`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-117 · Documentation of the Risk Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the risk management system described in Section 3 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.10/Abs.2`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-118 · Implementation of the Risk Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall implement the risk management system described in Section 3 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.2`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-119 · Maintenance of the Risk Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall maintain the risk management system described in Section 3 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.2`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-120 · Procedures to Maintain Conformity in Series Production

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that procedures are in place to keep series production in conformity with the requirements of the MDR.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-121 · Timely Consideration of Relevant Changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall adequately take into account, in a timely manner, changes in device design or characteristics and changes in harmonised standards or common specifications by reference to which conformity of the device is declared.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-122 · Establishment of the Quality Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> For devices other than investigational devices, the manufacturer shall establish a quality management system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-123 · Documentation of the Quality Management System

Applicability **applicable** · Gap status **partially fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assigned by rule**

> For devices other than investigational devices, the manufacturer shall document the quality management system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	Fallstudie: Nachweis zu RR-123 (CS-DOC-RR-123; 1.0; identifier b14f91dc-177d-471c-953f-2351cd18d49e)
Open items	Close gap

RR-124 · Implementation of the Quality Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> For devices other than investigational devices, the manufacturer shall implement the quality management system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-125 · Maintenance of the Quality Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> For devices other than investigational devices, the manufacturer shall maintain the quality management system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-126 · Keeping the Quality Management System Up to Date

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> For devices other than investigational devices, the manufacturer shall keep the quality management system up to date.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-127 · Continual Improvement of the Quality Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall continually improve the quality management system.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-128 · Effective and Proportionate Assurance of MDR Compliance by the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the quality management system shall ensure compliance with the MDR in the most effective manner and in a manner that is proportionate to the risk class and the type of device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-129 · Organisational Scope of the Quality Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the quality management system shall cover all parts and elements of the manufacturer's organisation dealing with the quality of processes, procedures and devices.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.2/Satz.1`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-130 · Governance of QMS Structures and Resources

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the quality management system shall govern the structure, responsibilities, procedures, processes and management resources required to implement the principles and actions necessary to achieve compliance with the MDR.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/UAbs.2/Satz.2`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-131 · Management Responsibility in the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses management responsibility.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.c`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-132 · Resource Management Including Supplier and Subcontractor Control

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses resource management, including the selection and control of suppliers and subcontractors.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.d`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-133 · Inclusion of Risk Management in the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses risk management as set out in Section 3 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.e`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-134 · Product Realisation within the Quality Management System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses product realisation, including planning, design, development, production and service provision.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Art.10/Abs.9/Buchst.g`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-138 · Management of Corrective and Preventive Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses the management of corrective and preventive actions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.1`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-139 · Verification of the Effectiveness of Corrective and Preventive Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses verification of the effectiveness of corrective and preventive actions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.1`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-140 · Monitoring and Measurement of Output

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system includes processes for monitoring and measurement of output.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.m`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-141 · Data Analysis within the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system includes processes for data analysis.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.m`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-142 · Product Improvement within the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system includes processes for product improvement.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.m`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-143 · Risk Management as a Continuous Iterative Process across the Device Lifecycle

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall conduct risk management as a continuous iterative process throughout the entire lifecycle of the device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.3/Satz.2`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-144 · Regular Systematic Updating of Risk Management

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall subject risk management to regular systematic updating.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.3/Satz.2`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-145 · Establishment of a Risk Management Plan for Each Device

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish a risk management plan for each device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.a`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-146 · Documentation of the Risk Management Plan for Each Device

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the risk management plan for each device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.a`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-147 · Identification of Known and Foreseeable Hazards

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify the known and foreseeable hazards associated with each device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.b`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-148 · Analysis of Known and Foreseeable Hazards

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall analyse the known and foreseeable hazards associated with each device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.b`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-149 · Estimation of Risks in Intended Use and Reasonably Foreseeable Misuse

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall estimate the risks associated with, and occurring during, the intended use and during reasonably foreseeable misuse.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.c`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-150 · Evaluation of Risks in Intended Use and Reasonably Foreseeable Misuse

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall evaluate the risks associated with, and occurring during, the intended use and during reasonably foreseeable misuse.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.c`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-151 · Elimination or Control of the Assessed Risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall eliminate or control the risks referred to in point (c) in accordance with the requirements of Section 4 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.d`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Buchst.c; SRC-001#Anh.I/Nr.3/Satz.3; SRC-001#Anh.I/Nr.4`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-152 · Evaluation of the Impact of Production and Post-Market Information on the Risk Profile

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall evaluate the impact of information from the production phase and, in particular, from the post-market surveillance system, on hazards and the frequency of occurrence thereof, on estimates of their associated risks, as well as on the overall risk, benefit-risk ratio and risk acceptability.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.e`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-153 · Amendment of Control Measures Where Necessary

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where the evaluation referred to in point (e) shows it to be necessary, the manufacturer shall amend control measures in line with the requirements of Section 4 of Annex I.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.3/Buchst.f`
Supplementary sources	`SRC-001#Anh.I/Nr.3/Buchst.e; SRC-001#Anh.I/Nr.3/Satz.3; SRC-001#Anh.I/Nr.4`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-154 · Safety Principles and State of the Art for Risk Control Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that risk control measures adopted for the design and manufacture of the device conform to safety principles, taking account of the generally acknowledged state of the art.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.4/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-155 · Acceptability of Hazard-Specific Residual Risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall manage risks so that the residual risk associated with each individual hazard is judged acceptable.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.4/Satz.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-156 · Acceptability of the Overall Residual Risk

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall manage risks so that the overall residual risk is judged acceptable.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.4/Satz.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-157 · Compliance with the Priority Order for Risk Control Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

When selecting the most appropriate solutions, the manufacturer shall follow the prescribed order of priority: first eliminating or reducing risks as far as possible through safe design and manufacture, then, where appropriate, taking adequate protection measures for risks that cannot be eliminated, and thereafter providing information for safety and, where appropriate, training to users.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.4/Satz.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-158 · Elimination or Reduction of Risks as Far as Possible Through Safe Design and Manufacture

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall eliminate or reduce risks as far as possible through safe design and manufacture, with risk reduction being pursued as far as possible without adversely affecting the benefit-risk ratio.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.4/Buchst.a`
Supplementary sources	`SRC-001#Anh.I/Nr.4/Satz.3; SRC-001#Anh.I/Nr.2`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-159 · Protective Measures for Risks That Cannot Be Eliminated

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where appropriate, the manufacturer shall take adequate protection measures in relation to risks that cannot be eliminated, including alarms where necessary.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.4/Buchst.b`
Supplementary sources	`SRC-001#Anh.I/Nr.4/Satz.3`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-161 · User Training as a Risk Control Measure

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the manufacturer shall, within the prescribed risk-control hierarchy, provide training to users.

Theme cluster	Quality and Risk Management
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Anh.I/Nr.4/Buchst.c`
Supplementary sources	`SRC-001#Anh.I/Nr.4/Satz.3`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Ensure and demonstrate competence or training (MEAS_COMPETENCE)
Planned evidence type	Competence or training record
Evidence references	none
Open items	Resolve applicability

RR-171 · Consideration of Risk-Management Principles Including Information Security in Software Development and Manufacture

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that, in the development and manufacture of the software, the principles of risk management, including information security, are taken into account.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-378 · Determine risk-control measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall determine suitable risk-control measures to reduce a risk requiring control to an acceptable level.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.1`
Supplementary sources	`SRC-007#Abschnitt 7.2.1`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-379 · Document risk-control measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The manufacturer shall document the selected risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.1`
Supplementary sources	`SRC-007#Abschnitt 7.2.1`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-380 · Verify implementation of risk-control measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall verify implementation of the applicable risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.2`
Supplementary sources	`SRC-007#Abschnitt 7.3.1; SRC-008#Anhang C/C.2.4`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-381 · Document verification of implementation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall document verification of the implementation of risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.2`
Supplementary sources	`SRC-007#Abschnitt 7.3.1; SRC-008#Anhang C/C.2.4`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-382 · Evaluate or re-evaluate residual risk

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

After implementation of risk-control measures, the manufacturer shall evaluate residual risk against the applicable acceptability criteria and repeat that evaluation when an applicable usability/UOUP reassessment trigger occurs.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.3`
Supplementary sources	`SRC-008#Abschnitt 5.9; SRC-008#Anhang C/C.2.5`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-383 · Document the result of the residual-risk evaluation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall document the result of the residual-risk evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.3`
Supplementary sources	`SRC-008#Anhang C/C.2.5`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-384 · Document a QMS in accordance with ISO 13485 and applicable regulatory requirements

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assigned by rule**

The organization shall document a QMS in accordance with ISO 13485 and applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Add evidence reference

RR-385 · Maintain the effectiveness of the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall maintain the effectiveness of the QMS.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-386 · Establish, implement and maintain every requirement, procedure, activity and arrangement required to be documented

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall establish, implement and maintain every requirement, procedure, activity and arrangement required to be documented.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Resolve applicability

RR-387 · Document the roles assumed under applicable regulatory requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall document the roles assumed under applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Conformity Assessment; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-388 · Determine the processes needed for the QMS and their implementation within the organization

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall determine the processes needed for the QMS and their implementation within the organization.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.2(a)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-389 · Apply a risk-based approach to the control of appropriate QMS processes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall apply a risk-based approach to the control of appropriate QMS processes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.2(b)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-390 · Determine the sequence and interaction of QMS processes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall determine the sequence and interaction of QMS processes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.2(c)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-391 · Determine criteria and methods for effective operation and control of each QMS process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall determine criteria and methods for effective operation and control of each QMS process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.3(a)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-392 · Ensure availability of resources and information needed to operate and monitor QMS processes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall ensure availability of resources and information needed to operate and monitor QMS processes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.3(b)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-393 · Implement actions necessary to achieve planned results and maintain process effectiveness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall implement actions necessary to achieve planned results and maintain process effectiveness.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.3(c)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-394 · Monitor and, where appropriate, measure and analyse QMS processes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall monitor and, where appropriate, measure and analyse QMS processes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.3(d)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-395 · Establish and maintain records needed to demonstrate conformity with the standard and applicable regulatory requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall establish and maintain records needed to demonstrate conformity with the standard and applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.3(e)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-396 · Manage and control QMS processes in accordance with ISO 13485 and applicable regulatory requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall manage and control QMS processes in accordance with ISO 13485 and applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-397 · Evaluate the impact of QMS process changes on the QMS and on medical devices

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall evaluate the impact of QMS process changes on the QMS and on medical devices.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.4(a)–(b)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-398 · Control changes to QMS processes in conformity with ISO 13485 and applicable regulatory requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall control changes to QMS processes in conformity with ISO 13485 and applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.4(c)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-403 · Document procedures for validation of computer software used in the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for validation of computer software used in the QMS.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`qms_software_used = yes`
Evaluated conditions	`MDR_MDSW: true; QMS_SOFTWARE_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Assess fulfilment

RR-404 · Validate QMS computer software before first use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall validate QMS computer software before first use.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 4.1.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`qms_software_used = yes`
Evaluated conditions	`MDR_MDSW: true; QMS_SOFTWARE_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Assess fulfilment

RR-405 · Revalidate QMS computer software after changes where appropriate

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall revalidate QMS computer software after changes where appropriate.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-406 · Make software validation and revalidation activities proportionate to the associated risk

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall make software validation and revalidation activities proportionate to the associated risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`qms_software_used = yes`
Evaluated conditions	`MDR_MDSW: true; QMS_SOFTWARE_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-407 · Maintain records of QMS software validation activities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of QMS software validation activities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.1.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`qms_software_used = yes`
Evaluated conditions	`MDR_MDSW: true; QMS_SOFTWARE_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-408 · Maintain the required QMS documentation set, including policy, objectives, manual, procedures, records and regulatory documentation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall maintain the required QMS documentation set, including policy, objectives, manual, procedures, records and regulatory documentation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-409 · Document a QMS manual covering scope, exclusions/non-applicability, procedures or references, process interactions and documentation structure

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall document a QMS manual covering scope, exclusions/non-applicability, procedures or references, process interactions and documentation structure.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-410 · Establish and maintain a medical-device file for each device type or group containing or referencing required conformity and product information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall establish and maintain a medical-device file for each device type or group containing or referencing required conformity and product information.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Include in the technical documentation (MEAS_TECHNICAL_DOCUMENTATION)
Planned evidence type	Technical documentation
Evidence references	none
Open items	Resolve applicability

RR-411 · Maintain a document-control procedure covering approval, update, revision status, availability, identification, external documents, loss prevention and obsolete documents

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall maintain a document-control procedure covering approval, update, revision status, availability, identification, external documents, loss prevention and obsolete documents.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-412 · Have document changes reviewed and approved by an authorized person with appropriate background information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall have document changes reviewed and approved by an authorized person with appropriate background information.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-413 · Define retention of at least one obsolete document copy in line with device lifetime, record retention and regulatory requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall define retention of at least one obsolete document copy in line with device lifetime, record retention and regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-414 · Maintain records as evidence of conformity and effective QMS operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records as evidence of conformity and effective QMS operation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-415 · Document controls for identification, storage, protection, integrity, retrieval, retention and disposition of records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document controls for identification, storage, protection, integrity, retrieval, retention and disposition of records.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-416 · Keep records legible, identifiable and retrievable and preserve identification of record changes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall keep records legible, identifiable and retrievable and preserve identification of record changes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-417 · Retain records for at least the defined or regulatory device lifetime and not less than two years after release

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall retain records for at least the defined or regulatory device lifetime and not less than two years after release.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 4.2.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-418 · Top management shall demonstrate commitment to QMS development, implementation and effectiveness through the specified management actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall demonstrate commitment to QMS development, implementation and effectiveness through the specified management actions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-419 · Ensure customer and applicable regulatory requirements are determined and met

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall ensure customer and applicable regulatory requirements are determined and met.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-420 · Ensure an appropriate quality policy covering compliance, QMS effectiveness, objectives, communication and continuing suitability

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall ensure an appropriate quality policy covering compliance, QMS effectiveness, objectives, communication and continuing suitability.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-421 · Establish measurable quality objectives consistent with the quality policy, including objectives needed for regulatory and product requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall establish measurable quality objectives consistent with the quality policy, including objectives needed for regulatory and product requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-422 · Plan the QMS to meet Clause 4.1 requirements and quality objectives

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Top management shall plan the QMS to meet Clause 4.1 requirements and quality objectives.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 5.4.2(a)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-423 · Maintain QMS integrity when QMS changes are planned and implemented

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

Top management shall maintain QMS integrity when QMS changes are planned and implemented.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.4.2(b)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-424 · Define, document and communicate responsibilities and authorities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

Top management shall define, document and communicate responsibilities and authorities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-425 · Document relationships of quality-relevant personnel and ensure necessary independence and authority

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

Top management shall document relationships of quality-relevant personnel and ensure necessary independence and authority.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-426 · Appoint a management representative with the specified QMS responsibilities and authority

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Top management shall appoint a management representative with the specified QMS responsibilities and authority.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.5.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-427 · Establish appropriate internal communication processes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Top management shall establish appropriate internal communication processes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.5.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-428 · Ensure communication regarding QMS effectiveness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Top management shall ensure communication regarding QMS effectiveness.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.5.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-429 · Document procedures for management review

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for management review.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.6.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Assess fulfilment

RR-430 · Review the QMS at planned intervals for continuing suitability, adequacy and effectiveness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall review the QMS at planned intervals for continuing suitability, adequacy and effectiveness.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.6.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-431 · Maintain management-review records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> Top management shall maintain management-review records.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.6.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-432 · Include the specified feedback, complaint, regulatory, audit, CAPA and regulatory-change inputs in management review

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization or top management shall include the specified feedback, complaint, regulatory, audit, CAPA and regulatory-change inputs in management review.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.6.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-433 · Record management-review outputs, including required QMS/product improvements, regulatory changes and resource decisions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization or top management shall record management-review outputs, including required QMS/product improvements, regulatory changes and resource decisions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 5.6.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-434 · Determine and provide resources needed for QMS implementation/effectiveness and regulatory/customer compliance

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall determine and provide resources needed for QMS implementation/effectiveness and regulatory/customer compliance.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-435 · Ensure personnel affecting product quality are competent through appropriate education, training, skills and experience

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall ensure personnel affecting product quality are competent through appropriate education, training, skills and experience.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Ensure and demonstrate competence or training (MEAS_COMPETENCE)
Planned evidence type	Competence or training record
Evidence references	none
Open items	Assess fulfilment

RR-436 · Document procedures for competence, required training and quality awareness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for competence, required training and quality awareness.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Ensure and demonstrate competence or training (MEAS_COMPETENCE)
Planned evidence type	Competence or training record
Evidence references	none
Open items	Assess fulfilment

RR-437 · Determine competence needs, provide competence actions, evaluate effectiveness, ensure quality awareness and maintain competence records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall determine competence needs, provide competence actions, evaluate effectiveness, ensure quality awareness and maintain competence records.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.2(a)–(e)`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-438 · Document infrastructure requirements, including necessary hardware/software and information systems, needed for product conformity

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document infrastructure requirements, including necessary hardware/software and information systems, needed for product conformity.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-439 · Document quality-relevant infrastructure maintenance requirements and intervals

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document quality-relevant infrastructure maintenance requirements and intervals.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-440 · Maintain infrastructure-maintenance records

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain infrastructure-maintenance records.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 6.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-441 · Document work-environment requirements needed for product conformity

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document work-environment requirements needed for product conformity.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 6.4.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-442 · Where the environment can affect product quality, document procedures for its monitoring and control

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall, where the environment can affect product quality, document procedures for its monitoring and control.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 6.4.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Resolve applicability

RR-443 · Plan and develop processes needed for product realization consistently with other QMS processes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall plan and develop processes needed for product realization consistently with other QMS processes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-444 · Document risk-management process(es) in product realization

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document risk-management process(es) in product realization.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-445 · Maintain records of risk-management activities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of risk-management activities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-446 · Define as appropriate product requirements, resources, V&V/monitoring activities, acceptance criteria and required evidence in product-realization planning and document the result

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall define as appropriate product requirements, resources, V&V/monitoring activities, acceptance criteria and required evidence in product-realization planning and document the result.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.1(a)–(d)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-453 · Plan and document customer communication on product information, orders, feedback/complaints and advisory notices

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall plan and document customer communication on product information, orders, feedback/complaints and advisory notices.

Theme cluster	Quality and Risk Management
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 7.2.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-454 · Communicate with regulatory authorities in accordance with applicable regulatory requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall communicate with regulatory authorities in accordance with applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Conformity Assessment; Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 7.2.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-497 · Document procedures ensuring purchased product conforms to specified purchasing requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall document procedures ensuring purchased product conforms to specified purchasing requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Resolve applicability

RR-498 · Evaluate and select suppliers based on capability, performance, impact on device quality and device risk

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall evaluate and select suppliers based on capability, performance, impact on device quality and device risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-499 · Plan supplier monitoring and re-evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall plan supplier monitoring and re-evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-500 · Monitor supplier performance against purchasing requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall monitor supplier performance against purchasing requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-501 · Use monitoring results as input to supplier re-evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall use monitoring results as input to supplier re-evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-502 · Address supplier nonconformity proportionately to device risk and regulatory compliance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall address supplier nonconformity proportionately to device risk and regulatory compliance.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-503 · Maintain records of supplier evaluation, selection, monitoring, re-evaluation and resulting actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of supplier evaluation, selection, monitoring, re-evaluation and resulting actions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-504 · Define purchasing information including, as appropriate, product, acceptance, process, personnel-qualification and QMS requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall define purchasing information including, as appropriate, product, acceptance, process, personnel-qualification and QMS requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-505 · Ensure adequacy of purchasing requirements before communicating them to the supplier

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall ensure adequacy of purchasing requirements before communicating them to the supplier.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.4.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-506 · Where applicable, establish a written requirement for suppliers to notify relevant purchased-product changes before implementation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall, where applicable, establish a written requirement for suppliers to notify relevant purchased-product changes before implementation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-507 · Maintain relevant purchasing documents and records to the extent required for traceability

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall maintain relevant purchasing documents and records to the extent required for traceability.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.4.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-508 · Determine and implement purchased-product verification proportionate to supplier evaluation and product risk

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall determine and implement purchased-product verification proportionate to supplier evaluation and product risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 7.4.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-509 · Determine the impact of known purchased-product changes on product realization or the medical device

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall determine the impact of known purchased-product changes on product realization or the medical device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.4.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-510 · Specify supplier-site verification activities and product-release method when such verification is intended

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall specify supplier-site verification activities and product-release method when such verification is intended.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.4.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-511 · Maintain records of purchased-product verification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of purchased-product verification.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.4.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no; purchasing_performed = yes`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PURCHASING_PERFORMED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-512 · Plan, perform, monitor and control production and service provision so the product conforms to specification

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall plan, perform, monitor and control production and service provision so the product conforms to specification.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-518 · Validate production or service processes whose output cannot be subsequently verified and whose deficiencies become apparent only later

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall validate production or service processes whose output cannot be subsequently verified and whose deficiencies become apparent only later.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.5.6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-519 · Document a process-validation procedure covering evaluation/approval, qualification, methods, acceptance criteria, records, revalidation and change approval

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document a process-validation procedure covering evaluation/approval, qualification, methods, acceptance criteria, records, revalidation and change approval.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-520 · Validate computer software used in production or service provision before first use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall validate computer software used in production or service provision before first use.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PRODUCTION_SERVICE_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-521 · Where appropriate, revalidate computer software used in production or service provision after changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall, where appropriate, revalidate computer software used in production or service provision after changes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PRODUCTION_SERVICE_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-522 · Make validation/revalidation of such software proportionate to risk, including its effect on product-specification conformity

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall make validation/revalidation of such software proportionate to risk, including its effect on product-specification conformity.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; PRODUCTION_SERVICE_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-523 · Maintain records of validation results, conclusions and necessary actions

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of validation results, conclusions and necessary actions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-524 · Document product-identification procedures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document product-identification procedures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-525 · Identify the product by suitable means throughout product realization

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall identify the product by suitable means throughout product realization.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-526 · Identify and maintain product status with respect to monitoring and measurement requirements throughout product realization

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall identify and maintain product status with respect to monitoring and measurement requirements throughout product realization.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-527 · Where regulatorily required, maintain a documented system for assignment of unique device identification

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall, where regulatorily required, maintain a documented system for assignment of unique device identification.

Theme cluster	Quality and Risk Management
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.8`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-528 · Document traceability procedures defining regulatory traceability extent and required records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall document traceability procedures defining regulatory traceability extent and required records.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 7.5.9.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-529 · Identify, verify, protect and safeguard customer property under organizational control

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall identify, verify, protect and safeguard customer property under organizational control.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.10`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; CUSTOMER_PROPERTY_PROVIDED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-530 · Report and record loss, damage or unsuitability of customer property

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall report and record loss, damage or unsuitability of customer property.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.10`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; CUSTOMER_PROPERTY_PROVIDED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-531 · Document procedures to preserve product conformity during processing, storage, handling and distribution

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall document procedures to preserve product conformity during processing, storage, handling and distribution.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.11`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Resolve applicability

RR-532 · Determine monitoring and measurements and the means needed to demonstrate product conformity

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall determine monitoring and measurements and the means needed to demonstrate product conformity.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-533 · Document procedures ensuring monitoring and measurement are performed consistently with their requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures ensuring monitoring and measurement are performed consistently with their requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-534 · Validate computer software used for monitoring or measurement before first use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall validate computer software used for monitoring or measurement before first use.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; MONITORING_MEASUREMENT_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-535 · Where appropriate, revalidate computer software used for monitoring or measurement after changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall, where appropriate, revalidate computer software used for monitoring or measurement after changes.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; MONITORING_MEASUREMENT_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-536 · Make validation/revalidation of such software proportionate to the associated risk

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall make validation/revalidation of such software proportionate to the associated risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; MONITORING_MEASUREMENT_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-537 · Maintain records of results, conclusions and necessary actions from such software validation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of results, conclusions and necessary actions from such software validation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true; MONITORING_MEASUREMENT_SOFTWARE_USED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-538 · Plan and implement necessary monitoring, measurement, analysis and improvement processes, including appropriate methods

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall plan and implement necessary monitoring, measurement, analysis and improvement processes, including appropriate methods.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-551 · Conduct internal audits at planned intervals to assess QMS conformity and effective implementation/maintenance

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall conduct internal audits at planned intervals to assess QMS conformity and effective implementation/maintenance.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-552 · Document a procedure for audit responsibilities, planning, conduct, records and reporting

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document a procedure for audit responsibilities, planning, conduct, records and reporting.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-553 · Plan the audit programme considering process status, significance and prior audit results

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall plan the audit programme considering process status, significance and prior audit results.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-554 · Define and record audit criteria, scope, frequency and methods

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall define and record audit criteria, scope, frequency and methods.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-555 · Ensure auditor objectivity and impartiality and prohibit auditing one's own work

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall ensure auditor objectivity and impartiality and prohibit auditing one's own work.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-556 · Maintain records of audits and results, including audited processes/areas and conclusions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of audits and results, including audited processes/areas and conclusions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Assess fulfilment

RR-557 · Implement necessary corrections and corrective actions without undue delay

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The responsible management shall implement necessary corrections and corrective actions without undue delay.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-558 · Perform follow-up including verification of actions and reporting of verification results

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The responsible management shall perform follow-up including verification of actions and reporting of verification results.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-559 · Monitor and, where appropriate, measure QMS processes to demonstrate their ability to achieve planned results

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall monitor and, where appropriate, measure QMS processes to demonstrate their ability to achieve planned results.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-560 · Take corrections and corrective action, where appropriate, when planned process results are not achieved

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall take corrections and corrective action, where appropriate, when planned process results are not achieved.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-566 · Identify and control nonconforming product to prevent unintended use or delivery

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall identify and control nonconforming product to prevent unintended use or delivery.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-567 · Document procedures for identification, documentation, segregation, evaluation and disposition of nonconforming product, including responsibilities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for identification, documentation, segregation, evaluation and disposition of nonconforming product, including responsibilities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-568 · Determine whether investigation of a nonconformity is necessary

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall determine whether investigation of a nonconformity is necessary.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.3.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-569 · Notify external parties responsible for the nonconformity, where applicable

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall notify external parties responsible for the nonconformity, where applicable.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.3.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-570 · Maintain records of the nonconformity, evaluation, investigation, decisions and subsequent actions

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall maintain records of the nonconformity, evaluation, investigation, decisions and subsequent actions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-571 · Address pre-delivery nonconformity by correction, preventing intended use, or permitted concession

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall address pre-delivery nonconformity by correction, preventing intended use, or permitted concession.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 8.3.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-572 · Permit concession only with justification, approval and regulatory compliance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall permit concession only with justification, approval and regulatory compliance.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 8.3.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-573 · Record the concession and identity of the approving person

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall record the concession and identity of the approving person.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 8.3.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-574 · Take action proportionate to actual or potential effects of nonconformity identified after delivery or start of use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall take action proportionate to actual or potential effects of nonconformity identified after delivery or start of use.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-577 · Perform rework under documented procedures considering possible adverse effects on the product

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall perform rework under documented procedures considering possible adverse effects on the product.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Resolve applicability

RR-578 · Subject rework procedures to the same review and approval as the original procedure

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall subject rework procedures to the same review and approval as the original procedure.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-579 · Verify reworked product against applicable acceptance criteria and regulatory requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall verify reworked product against applicable acceptance criteria and regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-580 · Maintain records of rework

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of rework.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-581 · Document processes and methods for identifying, collecting and analysing data to demonstrate QMS suitability, adequacy and effectiveness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document processes and methods for identifying, collecting and analysing data to demonstrate QMS suitability, adequacy and effectiveness.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-582 · Include at least feedback, product conformity, process/product trends, supplier, audit and where appropriate service data in analysis

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall include at least feedback, product conformity, process/product trends, supplier, audit and where appropriate service data in analysis.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Plan and conduct audit (MEAS_AUDIT)
Planned evidence type	Audit record
Evidence references	none
Open items	Resolve applicability

RR-583 · Use findings of QMS unsuitability, inadequacy or ineffectiveness as input to improvement

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall use findings of QMS unsuitability, inadequacy or ineffectiveness as input to improvement.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-584 · Maintain records of data-analysis results

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of data-analysis results.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-585 · Identify changes necessary to maintain QMS suitability/effectiveness and device safety/performance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall identify changes necessary to maintain QMS suitability/effectiveness and device safety/performance.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-586 · Implement the changes identified as necessary

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall implement the changes identified as necessary.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-587 · Eliminate causes of existing nonconformities through timely, proportionate corrective action to prevent recurrence

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall eliminate causes of existing nonconformities through timely, proportionate corrective action to prevent recurrence.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-588 · Document a corrective-action process covering review, root cause, need, planning/implementation, adverse-impact verification and effectiveness evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document a corrective-action process covering review, root cause, need, planning/implementation, adverse-impact verification and effectiveness evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.2(a)–(f)`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-589 · Maintain records of investigations and corrective actions taken

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of investigations and corrective actions taken.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-590 · Identify causes of potential nonconformities and determine proportionate preventive action to prevent occurrence

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall identify causes of potential nonconformities and determine proportionate preventive action to prevent occurrence.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-591 · Document a preventive-action process covering potential nonconformities/causes, need, planning/implementation, adverse-impact verification and effectiveness evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall document a preventive-action process covering potential nonconformities/causes, need, planning/implementation, adverse-impact verification and effectiveness evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.3(a)–(e)`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-592 · Maintain records of investigations and preventive actions taken

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of investigations and preventive actions taken.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.5.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-593 · Establish an ongoing risk-management process

Applicability **applicable** · Gap status **not fulfilled (user statement)** · Assessment state **active** · Measure **measure assignment open (review required)**

The manufacturer shall establish an ongoing risk-management process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user assessed the requirement as not fulfilled.
User fulfilment assessment	not fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Close gap; Review measure assignment

RR-594 · Implement the risk-management process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall implement the risk-management process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-595 · Document the risk-management process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the risk-management process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-596 · Maintain the risk-management process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall maintain the risk-management process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-597 · Ensure the process identifies device-related hazards and hazardous situations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure the process identifies device-related hazards and hazardous situations.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1(a)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-598 · Ensure the process estimates associated risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure the process estimates associated risks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1(b)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-599 · Ensure the process evaluates associated risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure the process evaluates associated risks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1(b)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-600 · Ensure the process controls risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure the process controls risks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1(c)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-601 · Monitor the effectiveness of risk-control measures within the process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall monitor the effectiveness of risk-control measures within the process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1(d)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-602 · Apply the risk-management process throughout the device lifecycle

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall apply the risk-management process throughout the device lifecycle.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-603 · Incorporate applicable parts of the risk-management process into an existing documented product-realization process

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall incorporate applicable parts of the risk-management process into an existing documented product-realization process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-604 · Ensure appropriate resources for the risk-management process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Top management shall ensure appropriate resources for the risk-management process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-605 · Assign competent personnel to risk-management activities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Top management shall assign competent personnel to risk-management activities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-606 · Establish and document a policy for determining risk-acceptability criteria

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> Top management shall establish and document a policy for determining risk-acceptability criteria.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-607 · Ensure that the policy reflects applicable regulatory and standards bases and available information, including state of the art and known stakeholder concerns

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall ensure that the policy reflects applicable regulatory and standards bases and available information, including state of the art and known stakeholder concerns.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-608 · Review the suitability of the risk-management process at planned intervals to ensure continuing effectiveness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> Top management shall review the suitability of the risk-management process at planned intervals to ensure continuing effectiveness.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-609 · Document decisions and actions resulting from the review

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Top management shall document decisions and actions resulting from the review.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-610 · Ensure risk-management personnel are competent through education, training, skills and experience appropriate to their tasks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> Personnel entrusted with risk-management tasks shall be competent through education, training, skills and experience appropriate to their tasks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Ensure and demonstrate competence or training (MEAS_COMPETENCE)
Planned evidence type	Competence or training record
Evidence references	none
Open items	Assess fulfilment

RR-611 · Where applicable, ensure relevant knowledge and experience of the device, its use, technologies or risk-management methods

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where applicable, relevant knowledge and experience of the device, its use, technologies or risk-management methods shall be ensured for personnel entrusted with risk-management tasks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-612 · Maintain appropriate competence records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall maintain appropriate competence records.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-613 · Include the risk-management plan in the risk-management file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall include the risk-management plan in the risk-management file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-614 · Define the scope, device and applicable lifecycle phases of planned risk-management activities in the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall define the scope, device and applicable lifecycle phases of planned risk-management activities in the plan.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(a)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-615 · Assign responsibilities and authorities in the risk-management plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall assign responsibilities and authorities in the risk-management plan.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(b)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-616 · Define requirements for review of risk-management activities in the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall define requirements for review of risk-management activities in the plan.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(c)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-617 · Define risk-acceptability criteria, including criteria for cases where probability of harm cannot be estimated

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall define risk-acceptability criteria, including criteria for cases where probability of harm cannot be estimated.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(d)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-618 · Define the method for evaluating overall residual risk in the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall define the method for evaluating overall residual risk in the plan.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(e)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-619 · Define criteria for acceptability of overall residual risk in the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall define criteria for acceptability of overall residual risk in the plan.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(e)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-620 · Plan activities to verify implementation of risk-control measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall plan activities to verify implementation of risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(f)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-621 · Plan activities to verify effectiveness of risk-control measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall plan activities to verify effectiveness of risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(f)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-622 · Plan activities for collecting relevant production and post-production information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall plan activities for collecting relevant production and post-production information.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(g)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-623 · Plan activities for reviewing relevant production and post-production information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall plan activities for reviewing relevant production and post-production information.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.4(g)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-624 · Record lifecycle changes to the risk-management plan in the risk-management file

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall record lifecycle changes to the risk-management plan in the risk-management file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Resolve applicability

RR-625 · Establish a risk-management file for the device

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall establish a risk-management file for the device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 4.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Resolve applicability

RR-626 · Maintain the risk-management file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain the risk-management file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-627 · Ensure traceability for each hazard to risk analysis, risk evaluation, implementation/verification of controls and residual-risk evaluation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall ensure traceability for each hazard to risk analysis, risk evaluation, implementation/verification of controls and residual-risk evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Conformity Assessment; Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 4.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-628 · Perform a risk analysis for each device in accordance with 5.2–5.5

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall perform a risk analysis for each device in accordance with 5.2–5.5.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-629 · Record implementation of planned risk-analysis activities in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record implementation of planned risk-analysis activities in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-630 · Record risk-analysis results in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record risk-analysis results in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-631 · Identify and describe the analysed device in the risk-analysis documentation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify and describe the analysed device in the risk-analysis documentation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.1(a)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-632 · Document the persons and organization performing the risk analysis

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the persons and organization performing the risk analysis.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.1(b)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-633 · Document the scope and date of the risk analysis

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the scope and date of the risk analysis.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.1(c)`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-634 · Document the intended use of the device

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the intended use of the device.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-635 · Document reasonably foreseeable misuse

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document reasonably foreseeable misuse.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-636 · Maintain intended-use and foreseeable-misuse documentation in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain intended-use and foreseeable-misuse documentation in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-637 · Identify qualitative and quantitative characteristics that could affect device safety

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify qualitative and quantitative characteristics that could affect device safety.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-638 · Document the identified qualitative and quantitative safety-related characteristics

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the identified qualitative and quantitative safety-related characteristics.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-639 · Where applicable, establish limits for safety-related characteristics

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, establish limits for safety-related characteristics.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-640 · Maintain safety-related-characteristic documentation in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain safety-related-characteristic documentation in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-641 · Identify known and foreseeable hazards considering intended use, foreseeable misuse, safety characteristics and normal/fault conditions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify known and foreseeable hazards considering intended use, foreseeable misuse, safety characteristics and normal/fault conditions.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-642 · Document the identified known and foreseeable hazards

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the identified known and foreseeable hazards.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-643 · Consider reasonably foreseeable sequences or combinations of events that could lead to hazardous situations for each hazard

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall consider reasonably foreseeable sequences or combinations of events that could lead to hazardous situations for each hazard.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-644 · Identify hazardous situations resulting from reasonably foreseeable sequences or combinations of events

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall identify hazardous situations resulting from reasonably foreseeable sequences or combinations of events.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-645 · Document hazardous situations resulting from reasonably foreseeable sequences or combinations of events

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall document hazardous situations resulting from reasonably foreseeable sequences or combinations of events.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-646 · Maintain Clause 5.4 documentation in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain Clause 5.4 documentation in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-647 · Estimate associated risks for each identified hazardous situation using available information or data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall estimate associated risks for each identified hazardous situation using available information or data.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-648 · Where probability of harm cannot be estimated, prepare a list of possible consequences for risk evaluation and control

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where probability of harm cannot be estimated, prepare a list of possible consequences for risk evaluation and control.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-649 · Record risk-estimation results in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall record risk-estimation results in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-650 · Record the system used to categorize probability and severity qualitatively or quantitatively in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall record the system used to categorize probability and severity qualitatively or quantitatively in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-006#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-651 · Evaluate each estimated risk using the plan criteria and determine whether it is acceptable

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall evaluate each estimated risk using the plan criteria and determine whether it is acceptable.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 6`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-652 · Treat an acceptable estimated risk as residual risk

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall treat an acceptable estimated risk as residual risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-653 · Record risk-evaluation results in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record risk-evaluation results in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 6`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-654 · Perform a benefit-risk analysis of residual risk when further risk reduction is not feasible

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall perform a benefit-risk analysis of residual risk when further risk reduction is not feasible.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-655 · Implement selected risk-control measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall implement selected risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-656 · Verify effectiveness of risk-control measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify effectiveness of risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-657 · Record results of effectiveness verification in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record results of effectiveness verification in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-658 · Consider further risk-control measures when residual risk remains unacceptable

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall consider further risk-control measures when residual risk remains unacceptable.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-659 · Record benefit-risk analysis results in the RM file

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall record benefit-risk analysis results in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Resolve applicability

RR-660 · Review risk-control effects for newly introduced hazards/hazardous situations and changes to previously estimated risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall review risk-control effects for newly introduced hazards/hazardous situations and changes to previously estimated risks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-661 · Process all new or increased risks in accordance with Clauses 5.5–7.4

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall process all new or increased risks in accordance with Clauses 5.5–7.4.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-662 · Record results of the review of risks arising from controls

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record results of the review of risks arising from controls.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.5`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-663 · Review risk-control activities for completeness across all hazardous situations and completion of all activities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall review risk-control activities for completeness across all hazardous situations and completion of all activities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-664 · Record the results of the completeness review in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record the results of the completeness review in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 7.6`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-665 · After implementation and verification of all controls, evaluate overall residual risk against the benefit of the intended use considering all residual risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, after implementation and verification of all controls, evaluate overall residual risk against the benefit of the intended use considering all residual risks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Conformity Assessment; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 8`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-668 · Record results of the overall residual-risk evaluation in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record results of the overall residual-risk evaluation in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Conformity Assessment; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 8`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-669 · Before commercial release, review execution of the RM plan and confirm at least proper implementation, acceptable overall residual risk and suitable production/post-production information methods

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall, before commercial release, review execution of the RM plan and confirm at least proper implementation, acceptable overall residual risk and suitable production/post-production information methods.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Conformity Assessment
Primary source	`SRC-006#Abschnitt 9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Resolve applicability

RR-670 · Record risk-management review results as a risk-management report

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall record risk-management review results as a risk-management report.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Conformity Assessment
Primary source	`SRC-006#Abschnitt 9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Resolve applicability

RR-671 · Maintain the risk-management report

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain the risk-management report.

Theme cluster	Quality and Risk Management
Life-cycle phases	Conformity Assessment; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 9`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-672 · Include the risk-management report in the risk-management file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall include the risk-management report in the risk-management file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Conformity Assessment
Primary source	`SRC-006#Abschnitt 9`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-673 · Assign responsibility for the review in the RM plan to persons with appropriate authority

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall assign responsibility for the review in the RM plan to persons with appropriate authority.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 9`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-687 · Where information is safety-relevant, review the RM file and determine whether existing risks require reassessment and/or new risks require assessment

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The manufacturer shall, where information is safety-relevant, review the RM file and determine whether existing risks require reassessment and/or new risks require assessment.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.4(1)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Resolve applicability

RR-688 · If residual risk is no longer acceptable, evaluate the impact on previously implemented risk-control measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, if residual risk is no longer acceptable, evaluate the impact on previously implemented risk-control measures.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.4(1)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-689 · Record all decisions and actions resulting from the safety-relevant reassessment in the RM file

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The manufacturer shall record all decisions and actions resulting from the safety-relevant reassessment in the RM file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.4(1)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Resolve applicability

RR-690 · Evaluate the impact of safety-relevant information on previously implemented risk-management activities

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall evaluate the impact of safety-relevant information on previously implemented risk-management activities.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.4(2)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-691 · Use the results of that evaluation as input to top-management review of the suitability of the risk-management process

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The manufacturer or top management shall use the results of that evaluation as input to top-management review of the suitability of the risk-management process.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.4(2)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Resolve applicability

RR-692 · Demonstrate the ability to create medical device software that consistently meets customer and applicable regulatory requirements

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assignment open (review required)**

> The manufacturer shall demonstrate the ability to create medical device software that consistently meets customer and applicable regulatory requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 4.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Add evidence reference; Review measure assignment

RR-730 · Plan or reference performance of software risk management, including SOUP-related risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall plan or reference performance of software risk management, including SOUP-related risks.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-735 · Provide a procedure to identify relevant technology-induced error categories and evidence that they do not contribute to unacceptable risk

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall provide a procedure to identify relevant technology-induced error categories and evidence that they do not contribute to unacceptable risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Verification and Validation
Primary source	`SRC-007#Abschnitt 5.1.12`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-750 · Where appropriate, include risk-control measures in software requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include risk-control measures in software requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.2.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-751 · Re-evaluate the device risk analysis when software requirements are established

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall re-evaluate the device risk analysis when software requirements are established.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-752 · Update the risk analysis where appropriate following re-evaluation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall update the risk analysis where appropriate following re-evaluation.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.2.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-803 · Evaluate known remaining anomalies for contribution to unacceptable risk

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall evaluate known remaining anomalies for contribution to unacceptable risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-813 · Address use of the software risk-management process in the maintenance plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall address use of the software risk-management process in the maintenance plan.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.1/c`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-822 · Evaluate each problem report for safety impact

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall evaluate each problem report for safety impact.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.1.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-834 · Identify software components that could contribute to hazardous situations identified by ISO 14971 risk analysis

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify software components that could contribute to hazardous situations identified by ISO 14971 risk analysis.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-835 · Identify possible software-related causes contributing to hazardous situations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify possible software-related causes contributing to hazardous situations.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.1.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-836 · Where SOUP is a possible cause, evaluate published anomaly lists for the version in use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where SOUP is a possible cause, evaluate published anomaly lists for the version in use.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.1.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B; soup_used = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true; SOUP_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-837 · Document possible software-related causes in the risk-management file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document possible software-related causes in the risk-management file.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.1.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-838 · Include software-implemented risk controls in software requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall include software-implemented risk controls in software requirements.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 7.2.2/a`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-841 · Review each risk-control measure for possible creation of new hazardous situations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall review each risk-control measure for possible creation of new hazardous situations.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.3.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-842 · Where appropriate, document traceability from hazardous situation through software cause to control and verification

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall, where appropriate, document traceability from hazardous situation through software cause to control and verification.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.3.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-846 · Perform risk-management activities under 7.1–7.3 required by the change analyses

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall perform risk-management activities under 7.1–7.3 required by the change analyses.

Theme cluster	Quality and Risk Management
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 7.4.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-868 · Where appropriate, update the risk-management file following problem resolution

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall, where appropriate, update the risk-management file following problem resolution.

Theme cluster	Quality and Risk Management
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Resolve applicability

RR-893 · Identify known or foreseeable use-related hazards and hazardous situations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify known or foreseeable use-related hazards and hazardous situations.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-894 · When identifying use-related hazards and hazardous situations, consider the Use Specification/user profiles, available information on similar interfaces and identified use errors

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, when identifying use-related hazards and hazardous situations, consider the Use Specification/user profiles, available information on similar interfaces and identified use errors.

Theme cluster	Quality and Risk Management
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-895 · Record the results of this hazard analysis in the UE File

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall record the results of this hazard analysis in the UE File.

Theme cluster	Quality and Risk Management
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	none
Open items	Assess fulfilment

RR-952 · Identify usability-engineering data needed to determine use-related residual risk

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall identify usability-engineering data needed to determine use-related residual risk.

Theme cluster	Quality and Risk Management
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

Data Governance (21)

RR-003 · Examination of Potential Biases in Data Sets

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate for the intended purpose for training, validation and testing data sets include an examination of potential biases that could affect the health and safety of persons, negatively affect fundamental rights, or lead to discrimination prohibited under Union law, in particular where data outputs influence inputs for future operations. For the development of high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.f`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-004 · Measures to Detect Potential Biases

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the data governance and management practices appropriate for the intended purpose for training, validation and testing data sets include appropriate measures to detect potential biases identified pursuant to Article 10(2), point (f). For the development of high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.g`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.2/Buchst.f; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-005 · Measures to Prevent Potential Biases

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the data governance and management practices appropriate for the intended purpose for training, validation and testing data sets include appropriate measures to prevent potential biases identified pursuant to Article 10(2), point (f). For the development of high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.g`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.2/Buchst.f; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-006 · Measures to Mitigate Potential Biases

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that the data governance and management practices appropriate for the intended purpose for training, validation and testing data sets include appropriate measures to mitigate potential biases identified pursuant to Article 10(2), point (f). For the development of high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.g`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.2/Buchst.f; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-030 · Data Governance and Management Practices Appropriate to the Intended Purpose

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that training, validation and testing data sets are subject to data governance and management practices appropriate to the intended purpose of the high-risk AI system. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-031 · Consideration of Relevant Design Choices

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular the relevant design choices. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.a`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-032 · Consideration of Data Collection Processes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular the data collection processes. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.b`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-033 · Consideration of Data Origin

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular the origin of the data. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.b`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-034 · Consideration of the Original Purpose of Personal Data Collection

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Assessment state **inactive (requirement currently not applicable)** · Measure **measure assignment open (review required)**

> Where personal data are used, the provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular the original purpose of the data collection. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.b`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	not fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-035 · Consideration of Relevant Data-Preparation Operations

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular relevant data-preparation processing operations. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.c`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-036 · Formulation of Data-Related Assumptions

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular the formulation of assumptions, in particular with respect to the information that the data are intended to measure and represent. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.d`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-037 · Assessment of Required Data Sets

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular an assessment of the availability, quantity and suitability of the data sets that are needed. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.e`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-038 · Identification of Compliance-Relevant Data Gaps and Shortcomings

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular the identification of relevant data gaps or shortcomings that prevent compliance with this Regulation. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.h`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-039 · Determination of How Data Gaps and Shortcomings Can Be Addressed

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that the data governance and management practices appropriate to the intended purpose concern in particular how relevant data gaps or shortcomings that prevent compliance with this Regulation can be addressed. For high-risk AI systems not using techniques involving the training of AI models, this requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.2/Buchst.h`
Supplementary sources	`SRC-002#Art.10/Abs.2/Satz.1; SRC-002#Art.10/Abs.2/Satz.2; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-040 · Relevance of Data Sets

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that training, validation and testing data sets are relevant in view of the intended purpose. This characteristic may be satisfied at the level of individual data sets or a combination of data sets. For high-risk AI systems not using techniques involving the training of AI models, the requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.3/Satz.1`
Supplementary sources	`SRC-002#Art.10/Abs.3/Satz.3; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-041 · Sufficient Representativeness of Data Sets

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that training, validation and testing data sets are sufficiently representative in view of the intended purpose. This characteristic may be satisfied at the level of individual data sets or a combination of data sets. For high-risk AI systems not using techniques involving the training of AI models, the requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.3/Satz.1`
Supplementary sources	`SRC-002#Art.10/Abs.3/Satz.3; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-042 · Freedom from Errors of Data Sets to the Best Extent Possible

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that training, validation and testing data sets are, to the best extent possible, free of errors in view of the intended purpose. This characteristic may be satisfied at the level of individual data sets or a combination of data sets. For high-risk AI systems not using techniques involving the training of AI models, the requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.3/Satz.1`
Supplementary sources	`SRC-002#Art.10/Abs.3/Satz.3; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-043 · Completeness of Data Sets to the Best Extent Possible

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that training, validation and testing data sets are, to the best extent possible, complete in view of the intended purpose. This characteristic may be satisfied at the level of individual data sets or a combination of data sets. For high-risk AI systems not using techniques involving the training of AI models, the requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.3/Satz.1`
Supplementary sources	`SRC-002#Art.10/Abs.3/Satz.3; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-044 · Appropriate Statistical Properties of Data Sets

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that training, validation and testing data sets have appropriate statistical properties, including, where applicable, as regards the persons or groups of persons in relation to whom the high-risk AI system is intended to be used. Those properties may be satisfied at the level of individual data sets or a combination of data sets. For high-risk AI systems not using techniques involving the training of AI models, the requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.3/Satz.2`
Supplementary sources	`SRC-002#Art.10/Abs.3/Satz.3; SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-045 · Consideration of Specific Settings in Data Sets

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The provider shall ensure that training, validation and testing data sets take into account, to the extent required by the intended purpose, the relevant characteristics or elements particular to the specific geographical, contextual, behavioural or functional setting within which the high-risk AI system is intended to be used. For high-risk AI systems not using techniques involving the training of AI models, the requirement applies only to testing data sets.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.10/Abs.4`
Supplementary sources	`SRC-002#Art.10/Abs.6; SRC-002#Art.16/Buchst.a`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-094 · Data Management Systems and Procedures

Applicability **applicable** · Gap status **not assessed** · Assessment state **active** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes systems and procedures for data management, including the relevant data operations performed before and for the purpose of placing the high-risk AI system on the market or putting it into service.

Theme cluster	Data Governance
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.f`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The fulfilment assessment is explicitly recorded as not assessed.
User fulfilment assessment	not assessed
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

Transparency (66)

RR-001 · Transparency for Interpretation and Appropriate Use of System Outputs

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Assessment state **inactive (requirement currently not applicable)** · Measure **measure assignment open (review required)**

> The provider shall ensure that a high-risk AI system is designed and developed so that its operation is sufficiently transparent to enable deployers to appropriately interpret and use the system's outputs.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.13/Abs.1/Satz.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-002 · Appropriate Type and Degree of Transparency

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The provider shall ensure that an appropriate type and degree of transparency are provided for a high-risk AI system so that the relevant obligations of the provider and deployer under Section 3 can be fulfilled.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.13/Abs.1/Satz.2`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-046 · Instructions for Use and Information Quality

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the high-risk AI system is accompanied by instructions for use in an appropriate digital format or otherwise that include concise, complete, correct and clear information that is relevant, accessible and comprehensible to deployers.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.2`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-047 · Provider and Authorised Representative Details

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use contain its identity and contact details and, where applicable, the identity and contact details of its authorised representative.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.a`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-048 · Information on Additional Characteristics, Capabilities and Performance Limitations

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use contain information on additional characteristics, capabilities and limitations of performance of the high-risk AI system insofar as they are not already covered by the information expressly specified in Article 13(3), point (b), subpoints (i) to (vii).

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-049 · Information on the Intended Purpose

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state the intended purpose of the high-risk AI system.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.i`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-050 · Information on Tested, Validated and Expected Performance Levels

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state the levels of accuracy, including the relevant metrics, robustness and cybersecurity referred to in Article 15 against which the high-risk AI system has been tested and validated and which can be expected.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.ii`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-051 · Information on Circumstances Affecting Expected Performance Levels

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state any known and foreseeable circumstances that may have an impact on the expected level of accuracy, robustness or cybersecurity of the high-risk AI system.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.ii`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-052 · Information on Risk-Relevant Circumstances of Use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state any known or foreseeable circumstances related to use in accordance with the intended purpose or under conditions of reasonably foreseeable misuse that may lead to risks to health, safety or fundamental rights referred to in Article 9(2).

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.iii`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-053 · Information on Technical Capabilities for Explaining Outputs

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where applicable, the provider shall ensure that the instructions for use state the technical capabilities and characteristics of the high-risk AI system to provide information relevant to explaining its output.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.iv`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-054 · Information on Performance for Specific Persons or Groups

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the provider shall ensure that the instructions for use contain information on the performance of the high-risk AI system regarding specific persons or groups of persons on which the system is intended to be used.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.v`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-055 · Information on Input Data and Data Sets Used

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the provider shall ensure that the instructions for use contain specifications for the input data or other relevant information concerning the training, validation and testing data sets used, taking into account the intended purpose.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.vi`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-056 · Information Enabling Interpretation and Appropriate Use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where applicable, the provider shall ensure that the instructions for use contain information enabling deployers to interpret the output of the high-risk AI system and use the system appropriately.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.b/Ziff.vii`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.13/Abs.3/Buchst.b; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-057 · Information on Pre-Determined Changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state any changes to the high-risk AI system and its performance that were pre-determined by the provider at the time of the initial conformity assessment.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.c`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-058 · Information on Human Oversight Measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use contain the human oversight measures referred to in Article 14, including the technical measures put in place to facilitate interpretation of the outputs of the high-risk AI system by deployers.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.d`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-059 · Information on Required Computational and Hardware Resources

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state the computational and hardware resources required for the high-risk AI system.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.e`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-060 · Information on Expected Lifetime

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state the expected lifetime of the high-risk AI system.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.e`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-061 · Information on Maintenance and Care Measures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The provider shall ensure that the instructions for use state all necessary maintenance and care measures, including their frequency, required to ensure the proper functioning of the high-risk AI system, including as regards software updates.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.e`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-062 · Description of Mechanisms for Handling Logs

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where relevant, the provider shall ensure that the instructions for use describe the mechanisms included in the high-risk AI system that allow deployers to properly collect, store and interpret the logs in accordance with Article 12.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.13/Abs.3/Buchst.f`
Supplementary sources	`SRC-002#Art.13/Abs.3; SRC-002#Art.16/Buchst.a`
Reason for applicability	The evidence documents no classification of this requirement as regular or conditional. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-076 · Declaration of Accuracy Levels and Accuracy Metrics

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assigned by rule**

> The provider shall ensure that the levels of accuracy and the relevant accuracy metrics of the high-risk AI system are declared in the accompanying instructions for use.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.15/Abs.3`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Add evidence reference

RR-160 · Provision of Information for Safety

Applicability **applicable** · Gap status **fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assignment open (review required)**

> The manufacturer shall, within the prescribed risk-control hierarchy, provide information for safety (warnings, precautions and contra-indications).

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Anh.I/Nr.4/Buchst.c`
Supplementary sources	`SRC-001#Anh.I/Nr.4/Satz.3`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	Fallstudie: Nachweis zu RR-160 (CS-DOC-RR-160; 1.0; identifier 1f968f7e-0505-4ef8-9742-2e9e6b2cc06e)
Open items	Review measure assignment

RR-162 · Information of Users About Residual Risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall inform users of any residual risks.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Anh.I/Nr.4/Satz.4`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-183 · Identification of the Device

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Manufacturer shall ensure that each device is accompanied by the information needed to identify it.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-184 · Identification of the Manufacturer

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

Each device shall be accompanied by information needed to identify its manufacturer.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-185 · Relevant Safety Information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that each device is accompanied by any safety information relevant to the user or, as appropriate, any other person.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-186 · Relevant Performance Information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that each device is accompanied by any performance information relevant to the user or, as appropriate, any other person.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-187 · Availability of Manufacturer Information on the Website

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where the manufacturer has a website, the manufacturer shall ensure that the information needed to identify the device and its manufacturer and the relevant safety and performance information are made available on that website.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Anh.I/Nr.23.1/Satz.2`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MANUFACTURER_WEBSITE: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-188 · Keeping Website Information Up to Date

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where the manufacturer has a website, the manufacturer shall ensure that the information made available there to identify the device and its manufacturer and the relevant safety and performance information are kept up to date.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Satz.2`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MANUFACTURER_WEBSITE: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-191 · Suitability of the Content

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the content of the label and instructions for use is appropriate to the particular device, its intended purpose and the technical knowledge, experience, education or training of the intended user(s).

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-196 · Placement of Required Label Information on the Device

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the information required on the label is provided on the device itself; if this is not practicable or appropriate, some or all of the information may appear on the packaging for each unit and/or on the packaging of multiple devices.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-198 · Provision of Instructions for Use with the Device

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall ensure that instructions for use are provided together with the device. This does not apply where, cumulatively, the device is a class I or class IIa device, it can be used safely without instructions for use, and nothing to the contrary is provided for elsewhere in Section 23.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.d`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-199 · Free Additional Copies of the Instructions for Use upon Request

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where multiple devices are supplied to a single user and/or location and the purchaser requests further copies of the instructions for use, the manufacturer shall ensure that they are provided free of charge.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.e`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-201 · Inclusion of Residual Risks in Manufacturer Information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where there are residual risks which are required to be communicated to the user and/or other person, the manufacturer shall ensure that they are included as limitations, contra-indications, precautions or warnings in the information supplied by the manufacturer.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.g`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-204 · Explanation of Symbols and Identification Colours Where No Harmonised Standard or CS Exists

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where the manufacturer uses a symbol or identification colour and no relevant harmonised standard or common specification exists for the area concerned, the manufacturer shall ensure that the symbol or identification colour is described in the documentation supplied with the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.h`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; SYMBOL_WITHOUT_HS_OR_CS: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-266 · Specification of the Intended Purpose

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the instructions for use clearly specify the intended purpose of the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-267 · Specification of Indications

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where appropriate for the device, the manufacturer shall ensure that the instructions for use clearly specify the indications.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.b`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-268 · Specification of Contra-Indications within the Intended-Purpose Context

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where appropriate for the device, the manufacturer shall ensure that the instructions for use clearly specify the contra-indications in the context of the intended purpose.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.b`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-269 · Specification of Patient Target Groups

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where appropriate for the device, the manufacturer shall ensure that the instructions for use clearly specify the patient target group or groups.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.b`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-270 · Specification of Intended Users

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where appropriate for the device, the manufacturer shall ensure that the instructions for use clearly specify the intended users.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.b`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-271 · Provision of Performance Characteristics

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the instructions for use contain the performance characteristics of the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.e`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-272 · Information for Verification of Device Suitability

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where applicable, the manufacturer shall provide in the instructions for use the information enabling a healthcare professional to verify whether the device is suitable.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.f`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-273 · Information for Selection of Corresponding Software

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where applicable, the manufacturer shall provide in the instructions for use the information enabling a healthcare professional to select the software corresponding to the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.f`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-274 · Information for Selection of Corresponding Accessories

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where applicable, the manufacturer shall provide in the instructions for use the information enabling a healthcare professional to select the accessories corresponding to the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.f`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-275 · Information on Residual Risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where residual risks exist, the manufacturer shall ensure that they are stated in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.g`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; RESIDUAL_RISK_PRESENT: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-276 · Information on Contra-Indications

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where contra-indications exist, the manufacturer shall ensure that they are stated in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.g`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-277 · Information on Undesirable Side-Effects

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where undesirable side-effects exist, the manufacturer shall ensure that they are stated in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.g`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-278 · Patient Information on Risks, Contra-Indications and Side-Effects

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where information concerning residual risks, contra-indications or undesirable side-effects is to be conveyed to the patient, the manufacturer shall include that patient information in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.g`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-279 · Specifications Required for Appropriate Use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall ensure that the instructions for use contain the specifications required by the user to use the device appropriately, including the relevant accuracy information where applicable to a device with a measuring function.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.h`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-280 · Information on Required Special Facilities

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where special facilities are required for use of the device, the manufacturer shall specify those requirements in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.j`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-281 · Information on Required Special Training

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where special training is required for use of the device, the manufacturer shall specify those requirements in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.j`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-282 · Information on Required Particular Qualifications

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where particular qualifications are required of the device user or other persons involved, the manufacturer shall specify those requirements in the instructions for use.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.j`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-283 · Information for Verification of Proper Installation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall provide in the instructions for use the information needed to verify whether the device has been properly installed.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.k`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-284 · Information for Verification of Safe and Intended Readiness for Operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall provide in the instructions for use the information needed to verify whether the device is ready to perform safely and as intended by the manufacturer.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.k`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-285 · Information on Preventive and Regular Maintenance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where relevant, the manufacturer shall specify in the instructions for use the nature and frequency of required preventive and regular maintenance.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.k`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-286 · Information on Necessary Calibration

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where calibration is necessary, the manufacturer shall provide in the instructions for use the calibration information required to ensure proper and safe operation of the device throughout its intended lifetime.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.k`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; CALIBRATION_REQUIRED: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-287 · Methods for Eliminating Risks during Installation, Calibration or Servicing

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where persons involved in installing, calibrating or servicing the device may encounter risks, the manufacturer shall specify in the instructions for use methods for eliminating those risks.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.k`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-288 · General Safety, Warning and Use-Limitation Information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall provide in the instructions for use information enabling the user and/or patient to be informed of relevant warnings, precautions, contra-indications, measures to be taken and limitations of use concerning the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.s`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-289 · Information Enabling the User to Brief the Patient

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where relevant, the manufacturer shall provide sufficient information in the instructions for use to enable the user to brief the patient on warnings, precautions, contra-indications, measures to be taken and limitations of use concerning the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.s`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-290 · Information for Malfunction or Safety-Relevant Performance Changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the instructions for use shall contain warnings, precautions and/or measures to be taken in the event of a device malfunction or a change in performance that may affect safety.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.s`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-291 · Information on Foreseeable External Influences and Environmental Conditions

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the instructions for use shall contain warnings, precautions and/or measures concerning reasonably foreseeable external influences or environmental conditions that may affect safe use of the device.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.s`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-292 · Information on Interference Risks during Diagnostic or Therapeutic Procedures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the instructions for use shall contain warnings, precautions and/or measures concerning interference risks arising from the reasonably foreseeable presence of the device during diagnostic investigations, evaluations, therapeutic treatment or other procedures.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.s`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-293 · Minimum Hardware Requirements for Intended Software Operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> For medical device software, the manufacturer shall specify in the instructions for use the minimum hardware requirements necessary to run the software as intended.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.ab`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-294 · Minimum Requirements for IT Network Characteristics

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For medical device software, the manufacturer shall specify in the instructions for use the minimum requirements concerning IT network characteristics necessary to run the software as intended.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.ab`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-295 · Minimum Requirements for IT Security Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For medical device software, the manufacturer shall specify in the instructions for use the minimum requirements concerning IT security measures necessary to run the software as intended, including the required protection against unauthorised access.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.4/Buchst.ab`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-666 · Where overall residual risk is acceptable, inform users of significant residual risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where overall residual risk is acceptable, inform users of significant residual risks.

Theme cluster	Transparency
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 8`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-667 · Include necessary information in accompanying documentation to disclose significant residual risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall include necessary information in accompanying documentation to disclose significant residual risks.

Theme cluster	Transparency
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 8`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

Human Oversight (10)

RR-064 · Effective Human Oversight Through System Design

Applicability **applicable** · Gap status **partially fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assignment open (review required)**

> The provider shall ensure that a high-risk AI system is designed and developed so that it can be effectively overseen by natural persons during the period in which it is in use, including through appropriate human-machine interface tools.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.14/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	Fallstudie: Nachweis zu RR-064 (CS-DOC-RR-064; 1.0; identifier b07cf2ed-3422-48af-8935-149abb0d85c0)
Open items	Close gap; Review measure assignment

RR-065 · Proportionality of Human Oversight Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the human oversight measures are commensurate with the relevant risks, the level of autonomy and the context of use of the high-risk AI system.

Theme cluster	Human Oversight
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-002#Art.14/Abs.3`
Supplementary sources	`SRC-002#Art.14/Abs.2; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-066 · Ensuring Human Oversight Through Appropriate Types of Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that human oversight is provided through either one or both of the following types of measures: measures identified by the provider before placing the high-risk AI system on the market or putting it into service and, where technically feasible, built into the system, or measures identified by the provider before that time that are appropriate to be implemented by the deployer.

Theme cluster	Human Oversight
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-002#Art.14/Abs.3; SRC-002#Art.14/Abs.3/Buchst.a; SRC-002#Art.14/Abs.3/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-067 · Understanding the System's Capabilities and Limitations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the high-risk AI system is provided to the deployer in such a way that natural persons assigned human oversight are enabled, as appropriate and proportionate, to properly understand the relevant capabilities and limitations of the system.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.4/Buchst.a`
Supplementary sources	`SRC-002#Art.14/Abs.4; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-068 · Monitoring of System Operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the high-risk AI system is provided to the deployer in such a way that natural persons assigned human oversight are enabled, as appropriate and proportionate, to duly monitor its operation, including with a view to detecting and addressing anomalies, dysfunctions and unexpected performance.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.4/Buchst.a`
Supplementary sources	`SRC-002#Art.14/Abs.4; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-069 · Awareness of Automation Bias

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the high-risk AI system is provided to the deployer in such a way that natural persons assigned human oversight are enabled, as appropriate and proportionate, to remain aware of the possible tendency to automatically rely or over-rely on the system's output, in particular where the system provides information or recommendations for decisions to be taken by natural persons.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.4/Buchst.b`
Supplementary sources	`SRC-002#Art.14/Abs.4; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-070 · Correct Interpretation of System Output

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the high-risk AI system is provided to the deployer in such a way that natural persons assigned human oversight are enabled, as appropriate and proportionate, to correctly interpret its output, taking into account, for example, the interpretation tools and methods available.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.4/Buchst.c`
Supplementary sources	`SRC-002#Art.14/Abs.4; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-071 · Ability Not to Use the System or to Override Its Output

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the high-risk AI system is provided to the deployer in such a way that natural persons assigned human oversight are enabled, as appropriate and proportionate, in any particular situation, to decide not to use the system or otherwise to disregard, override or reverse its output.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.4/Buchst.d`
Supplementary sources	`SRC-002#Art.14/Abs.4; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-072 · Intervention and Safe Interruption of System Operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the high-risk AI system is provided to the deployer in such a way that natural persons assigned human oversight are enabled, as appropriate and proportionate, to intervene in its operation or interrupt it through a stop button or a similar procedure that allows the system to come to a halt in a safe state.

Theme cluster	Human Oversight
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.4/Buchst.e`
Supplementary sources	`SRC-002#Art.14/Abs.4; SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-073 · Two-Person Verification for Remote Biometric Identification

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

For a high-risk AI system referred to in Annex III, point 1(a), the provider shall ensure that the human oversight measures are designed so that no action or decision is taken by the deployer on the basis of the identification resulting from the system unless that identification has been separately verified and confirmed by at least two natural persons with the necessary competence, training and authority; the exception in Article 14(5), second subparagraph, remains unaffected.

Theme cluster	Human Oversight
Life-cycle phases	Planning and Conception; Development and Realisation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-002#Art.14/Abs.5/UAbs.1`
Supplementary sources	`SRC-002#Anh.III/Nr.1/Buchst.a; SRC-002#Art.14/Abs.5/UAbs.2; SRC-002#Art.16/Buchst.a`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Ensure and demonstrate competence or training (MEAS_COMPETENCE)
Planned evidence type	Competence or training record
Evidence references	none
Open items	Resolve applicability

Robustness (10)

RR-074 · Appropriate and Lifecycle-Consistent Level of Accuracy

Applicability **applicable** · Gap status **not fulfilled (user statement)** · Assessment state **active** · Measure **measure assignment open (review required)**

> The provider shall ensure that a high-risk AI system is designed and developed so that it achieves an appropriate level of accuracy and performs consistently in that respect throughout its lifecycle.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-002#Art.15/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The user assessed the requirement as not fulfilled.
User fulfilment assessment	not fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Close gap; Review measure assignment

RR-075 · Appropriate and Lifecycle-Consistent Level of Robustness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that a high-risk AI system is designed and developed so that it achieves an appropriate level of robustness and performs consistently in that respect throughout its lifecycle.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-002#Art.15/Abs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-077 · Resilience to Errors, Faults and Inconsistencies

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the high-risk AI system is as resilient as possible regarding errors, faults or inconsistencies that may occur within the system or the environment in which it operates, in particular due to interaction with natural persons or other systems.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-002#Art.15/Abs.4/UAbs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-078 · Technical and Organisational Robustness Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that technical and organisational measures are taken in this regard.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-002#Art.15/Abs.4/UAbs.1`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-079 · Elimination or Reduction of Feedback-Loop Risks

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> For a high-risk AI system that continues to learn after being placed on the market or put into service, the provider shall ensure that the system is developed so as to eliminate or reduce as far as possible the risk of possibly biased outputs influencing input for future operations through feedback loops.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.15/Abs.4/UAbs.3`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-080 · Appropriate Mitigation of Feedback Loops

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> For a high-risk AI system that continues to learn after being placed on the market or put into service, the provider shall ensure that the system is developed in such a way as to ensure that any such feedback loops are duly addressed with appropriate mitigation measures.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-002#Art.15/Abs.4/UAbs.3`
Supplementary sources	`SRC-002#Art.16/Buchst.a`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-165 · Ensuring Repeatability of Programmable Medical Devices

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that devices incorporating electronic programmable systems, including software, and software that are devices in themselves are designed to ensure repeatability.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Add evidence reference; Review measure assignment

RR-166 · Ensuring Reliability of Programmable Medical Devices

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that devices incorporating electronic programmable systems, including software, and software that are devices in themselves are designed to ensure reliability.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-167 · Ensuring Performance in Line with Intended Use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that devices incorporating electronic programmable systems, including software, and software that are devices in themselves are designed to ensure performance in line with their intended use.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.1/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-168 · Control of Consequences of a Single-Fault Condition

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

For devices within the scope of Annex I, Section 17.1, the manufacturer shall, in the event of a single fault condition, adopt appropriate means to eliminate or reduce as far as possible consequent risks or impairment of performance.

Theme cluster	Robustness
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.1/Satz.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

Post-Market Monitoring (125)

RR-096 · Establishment of the Post-Market Monitoring System

Applicability **applicable** · Gap status **not assessed** · Assessment state **active** · Measure **measure assigned by rule**

The provider shall ensure that the quality management system includes the setting-up of a post-market monitoring system in accordance with Article 72.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.17/Abs.1/Buchst.h`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The fulfilment assessment is explicitly recorded as not assessed.
User fulfilment assessment	not assessed
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-097 · Implementation of the Post-Market Monitoring System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the quality management system includes the implementation of the post-market monitoring system in accordance with Article 72.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.17/Abs.1/Buchst.h`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-098 · Maintenance of the Post-Market Monitoring System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the quality management system includes the maintenance of the post-market monitoring system in accordance with Article 72.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.17/Abs.1/Buchst.h`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-105 · Establishment of the Post-Market Monitoring System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall establish a post-market monitoring system for the high-risk AI system.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.72/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-106 · Documentation of the Post-Market Monitoring System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall document the post-market monitoring system.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.72/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-107 · Proportionate Design of the Post-Market Monitoring System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the manner in which the post-market monitoring system is established and documented is proportionate to the nature of the AI technologies and the risks of the high-risk AI system.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.72/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-108 · Active and Systematic Collection of Suitable Post-Market Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the post-market monitoring system actively and systematically collects, throughout the lifetime of the high-risk AI system, relevant data on its performance which allow the provider to evaluate the continuous compliance of the high-risk AI system with the requirements set out in Chapter III, Section 2.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.72/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.72/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-109 · Active and Systematic Documentation of Suitable Post-Market Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the post-market monitoring system actively and systematically documents, throughout the lifetime of the high-risk AI system, relevant data on its performance which allow the provider to evaluate the continuous compliance of the high-risk AI system with the requirements set out in Chapter III, Section 2.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.72/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.72/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-110 · Active and Systematic Analysis of Suitable Post-Market Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the post-market monitoring system actively and systematically analyses, throughout the lifetime of the high-risk AI system, relevant data on its performance which allow the provider to evaluate the continuous compliance of the high-risk AI system with the requirements set out in Chapter III, Section 2.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.72/Abs.2/Satz.1`
Supplementary sources	`SRC-002#Art.72/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-111 · Analysis of Interaction with Other AI Systems

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where relevant, the provider shall ensure that post-market monitoring includes an analysis of the interaction of the high-risk AI system with other AI systems.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.72/Abs.2/Satz.2`
Supplementary sources	`SRC-002#Art.72/Abs.1`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-112 · Post-Market Monitoring System Based on a Post-Market Monitoring Plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the post-market monitoring system is based on a post-market monitoring plan.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.72/Abs.3/Satz.1`
Supplementary sources	`SRC-002#Art.72/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-113 · Post-Market Monitoring Plan as Part of the Technical Documentation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The provider shall ensure that the post-market monitoring plan forms part of the technical documentation referred to in Annex IV.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Conformity Assessment
Primary source	`SRC-002#Art.72/Abs.3/Satz.2`
Supplementary sources	`SRC-002#Art.72/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Include in the technical documentation (MEAS_TECHNICAL_DOCUMENTATION)
Planned evidence type	Technical documentation
Evidence references	none
Open items	Assess fulfilment

RR-114 · Use of the Post-Market Monitoring Plan Template for an Integrated Arrangement

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where, under the conditions set out in Article 72(4), first subparagraph, the provider of a high-risk AI system chooses, as appropriate, to integrate the necessary elements of Article 72(1) to (3) into an existing post-market monitoring system and plan established under the applicable Union harmonisation legislation, the provider shall use for that integration the template referred to in Article 72(3).

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-002#Art.72/Abs.4/UAbs.1`
Supplementary sources	`SRC-002#Art.72/Abs.1; SRC-002#Art.72/Abs.2; SRC-002#Art.72/Abs.3/Satz.1; SRC-002#Art.72/Abs.3/Satz.2; SRC-002#Art.72/Abs.3/Satz.3`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Resolve applicability

RR-115 · Equivalent Level of Protection for an Integrated Post-Market Monitoring Arrangement

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where, under the conditions set out in Article 72(4), first subparagraph, the provider of a high-risk AI system chooses to integrate the necessary elements of Article 72(1) to (3) into an existing post-market monitoring system and plan established under the applicable Union harmonisation legislation, the integrated arrangement shall achieve an equivalent level of protection.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance
Primary source	`SRC-002#Art.72/Abs.4/UAbs.1`
Supplementary sources	`SRC-002#Art.72/Abs.1; SRC-002#Art.72/Abs.2; SRC-002#Art.72/Abs.3/Satz.1; SRC-002#Art.72/Abs.3/Satz.2`
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Resolve applicability

RR-135 · Setting-Up of the Post-Market Surveillance System as a QMS Aspect

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assigned by rule**

> For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses the setting-up of a post-market surveillance system in accordance with Article 83.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.10/Abs.9/Buchst.i`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Add evidence reference

RR-136 · Implementation of the Post-Market Surveillance System as a QMS Aspect

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses the implementation of the post-market surveillance system in accordance with Article 83.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.i`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-137 · Maintenance of the Post-Market Surveillance System as a QMS Aspect

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For devices other than investigational devices, the manufacturer shall ensure that the quality management system addresses the maintenance of the post-market surveillance system in accordance with Article 83.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.10/Abs.9/Buchst.i`
Supplementary sources	`SRC-001#Art.10/Abs.9/UAbs.1/Satz.3; SRC-001#Art.10/Abs.9/UAbs.3`
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`investigational_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_INVESTIGATIONAL_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-296 · Planning of the PMS System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall plan a post-market surveillance system for each device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-297 · Establishment of the PMS System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish a post-market surveillance system for each device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-298 · Documentation of the PMS System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the post-market surveillance system established for each device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-299 · Implementation of the PMS System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall implement the post-market surveillance system for each device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-300 · Maintenance of the PMS System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain the post-market surveillance system for each device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-301 · Updating of the PMS System

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall keep the post-market surveillance system for each device up to date.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-302 · Proportionate and Device-Appropriate PMS Design

Applicability **applicable** · Gap status **fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assignment open (review required)**

> The manufacturer shall design the post-market surveillance system in a manner proportionate to the risk class and appropriate for the type of device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	Fallstudie: Nachweis zu RR-302 (CS-DOC-RR-302; 1.0; identifier e3333ce7-3c17-44d0-8aa0-3471480f39f3)
Open items	Review measure assignment

RR-303 · Integration of the PMS System into the QMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the post-market surveillance system is an integral part of its quality management system referred to in Article 10(9).

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-304 · Active and Systematic Collection of PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the PMS system actively and systematically gathers relevant data on the quality, performance and safety of the device throughout its lifetime.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-305 · Active and Systematic Recording of PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the PMS system actively and systematically records relevant quality, performance and safety data throughout the device lifetime.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-306 · Active and Systematic Analysis of PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the PMS system actively and systematically analyses relevant quality, performance and safety data throughout the device lifetime.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-307 · Drawing Necessary PMS Conclusions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the PMS system draws the necessary conclusions from the relevant post-market surveillance data.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-308 · Determination of Preventive and Corrective Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the PMS system determines any necessary preventive and corrective actions on the basis of PMS results.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-309 · Implementation of Preventive and Corrective Actions within PMS

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the PMS system implements the preventive and corrective actions identified as necessary.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-310 · Monitoring of Preventive and Corrective Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that the PMS system monitors the preventive and corrective actions identified through post-market surveillance.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-311 · Updating the Benefit-Risk Determination Using PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use data gathered through its PMS system to update the benefit-risk determination.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-312 · Improvement of Risk Management Using PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use data gathered through its PMS system to improve risk management referred to in Chapter I of Annex I.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-313 · Updating Design and Manufacturing Information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use PMS data to update the design and manufacturing information of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-314 · Updating the Instructions for Use Using PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall use PMS data to update the instructions for use.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-315 · Updating Labelling Using PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall use PMS data to update the labelling of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-316 · Updating the Clinical Evaluation Using PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use data gathered through the PMS system to update the clinical evaluation.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.c`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-317 · Updating the Summary of Safety and Clinical Performance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

Where a summary of safety and clinical performance under Article 32 is required for the device, the manufacturer shall use PMS data to update that summary.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.d`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-318 · Identification of Action Needs from PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall use PMS data to identify needs for preventive, corrective or field safety corrective action.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.e`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-319 · Identification of Usability Improvement Options

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use PMS data to identify options for improving the usability of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.f`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-320 · Identification of Performance Improvement Options

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use PMS data to identify options for improving the performance of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.f`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-321 · Identification of Safety Improvement Options

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall use PMS data to identify options for improving the safety of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.f`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-322 · Contribution to PMS of Other Devices

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where relevant, the manufacturer shall use data gathered through its PMS system to contribute to the post-market surveillance of other devices.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.g`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-323 · Detection of Trends from PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall use PMS data to detect trends relevant to post-market surveillance.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.h`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-324 · Reporting of Trends under Article 88

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where a trend meets the reporting conditions of Article 88, the manufacturer shall report that trend on the basis of PMS data in accordance with Article 88.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Buchst.h`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-325 · Corresponding Update of the Technical Documentation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall update the technical documentation accordingly in light of the PMS results and resulting updates.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.3/Satz.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Include in the technical documentation (MEAS_TECHNICAL_DOCUMENTATION)
Planned evidence type	Technical documentation
Evidence references	none
Open items	Assess fulfilment

RR-326 · Implementation of Appropriate Measures when Action Is Needed

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where PMS identifies a need for preventive or corrective action or both, the manufacturer shall implement the appropriate measures.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.4/Satz.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-327 · Information of Competent Authorities about Required PMS Action

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where PMS identifies a need for preventive or corrective action, the manufacturer shall inform the competent authorities concerned.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.4/Satz.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-328 · Information of the Notified Body about Required PMS Action

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where PMS identifies a need for preventive or corrective action and a notified body is applicable, the manufacturer shall inform that notified body.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.4/Satz.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-329 · Reporting of Serious Incidents and FSCAs

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where a serious incident is identified or a field safety corrective action is implemented, the manufacturer shall report it in accordance with Article 87.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.83/Abs.4/Satz.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-330 · PMS System Based on a PMS Plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the post-market surveillance system referred to in Article 83 is based on a post-market surveillance plan.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Art.84/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-331 · PMS Plan as Part of the Technical Documentation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For devices other than custom-made devices, the manufacturer shall ensure that the post-market surveillance plan forms part of the technical documentation specified in Annex II.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Conformity Assessment
Primary source	`SRC-001#Art.84/Satz.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	`custom_made_device = no`
Evaluated conditions	`MDR_MDSW: true; NOT_CUSTOM_MADE_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Include in the technical documentation (MEAS_TECHNICAL_DOCUMENTATION)
Planned evidence type	Technical documentation
Evidence references	none
Open items	Assess fulfilment

RR-332 · Demonstration of Article 83 Compliance in the PMS Plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall demonstrate in the post-market surveillance plan that it complies with the obligations referred to in Article 83.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-333 · PMS Plan for Serious Incidents and FSCAs

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall address the collection and utilisation of available information concerning serious incidents, including information from PSURs, and field safety corrective actions.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-334 · PMS Plan for Non-Serious Incidents and Undesirable Side-Effects

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall address the collection and utilisation of records concerning non-serious incidents and data on any undesirable side-effects.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-335 · PMS Plan for Trend-Reporting Information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall address the collection and utilisation of available information from trend reporting.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-336 · PMS Plan for Literature, Databases and Registers

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall address the collection and utilisation of relevant specialist or technical literature, databases and/or registers.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-337 · PMS Plan for User, Distributor and Importer Information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall address the collection and utilisation of information, including feedback and complaints, provided by users, distributors and importers.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-338 · PMS Plan for Public Information on Similar Devices

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The PMS plan shall address the collection and utilisation of publicly available information about similar medical devices.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-339 · Proactive and Systematic PMS Collection Process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The PMS plan shall include a proactive and systematic process for collecting the information referred to in point (a).

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-340 · Correct Characterisation of Device Performance

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The collection process specified in the PMS plan shall enable a correct characterisation of the performance of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-341 · Comparison with Similar Devices

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The collection process specified in the PMS plan shall enable comparison between the device and similar devices available on the market.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-342 · Methods for Assessment of Collected PMS Data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify effective and appropriate methods and processes for assessing the collected PMS data.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-343 · Indicators and Threshold Values for Continuous Reassessment

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify suitable indicators and threshold values for the continuous reassessment of the benefit-risk analysis and risk management referred to in Section 3 of Annex I.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-344 · Methods for Investigation of Complaints

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify effective and appropriate methods and tools for investigating complaints.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-345 · Methods for Analysis of Market-Related Experience

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify effective and appropriate methods and tools for analysing market-related experience collected in the field.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-346 · Procedures for Incidents Subject to Trend Reporting

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify methods and protocols for managing incidents subject to trend reporting under Article 88, including methods for determining statistically significant increases and the applicable observation period.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-347 · Procedures for Effective PMS Communication

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify methods and protocols for effective communication with competent authorities, notified bodies, economic operators and users.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-348 · Reference to Procedures for Fulfilling PMS Obligations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall refer to the procedures for fulfilling the manufacturer's obligations under Articles 83 and 84 and, where applicable to the device, Article 86.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-349 · Systematic Identification of Appropriate PMS Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify systematic procedures for identifying appropriate measures, including corrective actions where appropriate.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-350 · Systematic Initiation of Appropriate PMS Measures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall specify systematic procedures for initiating appropriate measures identified through post-market surveillance, including corrective actions where required.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-351 · Tools for Identifying Devices Requiring Corrective Action

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall provide effective tools for tracing and identifying devices that may require corrective action.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-352 · PMCF Plan or Justification of Non-Applicability

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The PMS plan shall contain either a post-market clinical follow-up plan in accordance with Part B of Annex XIV or a justification as to why post-market clinical follow-up is not applicable.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-001#Anh.III/Abschn.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-360 · Preparation of the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For a class IIa, IIb or III device, the manufacturer shall prepare a periodic safety update report for each device and, where relevant, for each category or group of devices.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-361 · Summary of PMS Analysis Results in the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall summarise in the PSUR the results of the analyses of PMS data gathered pursuant to the post-market surveillance plan referred to in Article 84.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-362 · Summary of PMS Analysis Conclusions in the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall summarise in the PSUR the conclusions of the analyses of PMS data gathered pursuant to the post-market surveillance plan referred to in Article 84.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-363 · Rationale for Preventive and Corrective Actions in the PSUR

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall include in the PSUR the rationale for any preventive and corrective actions taken.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Resolve applicability

RR-364 · Description of Preventive and Corrective Actions in the PSUR

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall describe in the PSUR any preventive and corrective actions taken.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Resolve applicability

RR-365 · Benefit-Risk Conclusions in the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

Throughout the lifetime of the device, the manufacturer shall set out in the PSUR the conclusions of the benefit-risk determination.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1/Buchst.a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-366 · Main PMCF Findings in the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

Throughout the lifetime of the device, the manufacturer shall set out in the PSUR the main findings of the PMCF.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1/Buchst.b`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-367 · Device Sales Volume in the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> Throughout the lifetime of the device, the manufacturer shall set out in the PSUR the volume of sales of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1/Buchst.c`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-368 · Estimate of the Device-Using Population in the PSUR

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> Throughout the lifetime of the device, the manufacturer shall set out in the PSUR an estimate of the size and other characteristics of the population using the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1/Buchst.c`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-369 · Device Usage Frequency in the PSUR

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

Where practicable, throughout the lifetime of the device the manufacturer shall set out in the PSUR the frequency of use of the device.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1/Buchst.c`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Resolve applicability

RR-371 · Needs-Based PSUR Update for Class IIa

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

For a class IIa device, the manufacturer shall update the PSUR when necessary.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Article 86 for class IIa
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Resolve applicability

RR-372 · At Least Biennial PSUR Update for Class IIa

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For a class IIa device, the manufacturer shall update the PSUR at least every two years.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Article 86 for class IIa
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-373 · PSUR as Part of the Technical Documentation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For a class IIa, IIb or III device other than a custom-made device, the manufacturer shall ensure that the PSUR forms part of the technical documentation specified in Annexes II and III.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: PSUR under Article 86 (classes IIa, IIb, III)
Decision-relevant inputs	`custom_made_device = no`
Evaluated conditions	`MDR_MDSW: true; MDR_CLASS_IIA_IIB_III: true; NOT_CUSTOM_MADE_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Include in the technical documentation (MEAS_TECHNICAL_DOCUMENTATION)
Planned evidence type	Technical documentation
Evidence references	none
Open items	Assess fulfilment

RR-376 · Availability of the PSUR to the Notified Body

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

For a device falling under Article 86 that is not referred to in Article 86(2), the manufacturer shall make the PSUR available to the notified body involved in the conformity assessment.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Article 86 devices outside paragraph 2
Decision-relevant inputs	`implantable_device = no`
Evaluated conditions	`MDR_MDSW: true; ART86_OUTSIDE_ABS2_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-377 · Availability of the PSUR to Competent Authorities

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

For a device falling under Article 86 that is not referred to in Article 86(2), the manufacturer shall make the PSUR available to competent authorities upon request.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-001#Art.86/Abs.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: Article 86 devices outside paragraph 2
Decision-relevant inputs	`implantable_device = no`
Evaluated conditions	`MDR_MDSW: true; ART86_OUTSIDE_ABS2_DEVICE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006; RUL-009; RUL-010; RUL-011; RUL-012; RUL-016; RUL-017; RUL-018; RUL-019; RUL-020`
Decisions used	MDR/IVDR boundary; MDR classification; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Resolve applicability

RR-516 · Evaluate servicing records for complaint status and, where appropriate, improvement input

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall evaluate servicing records for complaint status and, where appropriate, improvement input.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-539 · Collect and monitor information on fulfilment of customer requirements

Applicability **applicable** · Gap status **partially fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assignment open (review required)**

The organization shall collect and monitor information on fulfilment of customer requirements.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	Fallstudie: Nachweis zu RR-539 (CS-DOC-RR-539; 1.0; identifier b85c2f69-243d-40ae-91fd-f772c701ce67)
Open items	Close gap; Review measure assignment

RR-540 · Document methods for obtaining and using that information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document methods for obtaining and using that information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-541 · Maintain a documented feedback process covering collection of production and post-production data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall maintain a documented feedback process covering collection of production and post-production data.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-542 · Use feedback information as potential input to risk management, product requirements, product realization and improvement

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall use feedback information as potential input to risk management, product requirements, product realization and improvement.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-543 · Include evaluation of regulatorily required post-production experience in the feedback process

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall include evaluation of regulatorily required post-production experience in the feedback process.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-544 · Document procedures for timely complaint handling in accordance with regulatory requirements and cover the specified activities and responsibilities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for timely complaint handling in accordance with regulatory requirements and cover the specified activities and responsibilities.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-545 · Document the rationale when a complaint is not investigated further

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document the rationale when a complaint is not investigated further.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-546 · Document corrections or corrective actions resulting from complaint handling

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document corrections or corrective actions resulting from complaint handling.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-547 · Exchange relevant information with the external party when external activities contributed to a complaint

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall exchange relevant information with the external party when external activities contributed to a complaint.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-548 · Maintain complaint-handling records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain complaint-handling records.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Assess fulfilment

RR-549 · Where regulatory reporting or advisory notices are required, document procedures for reporting to relevant regulatory authorities

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall, where regulatory reporting or advisory notices are required, document procedures for reporting to relevant regulatory authorities.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record PMS or vigilance activity (MEAS_PMS_VIGILANCE)
Planned evidence type	PMS or vigilance record
Evidence references	none
Open items	Resolve applicability

RR-550 · Maintain records of reporting to regulatory authorities

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of reporting to regulatory authorities.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.2.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-575 · Document procedures for advisory notices in accordance with regulatory requirements and keep them capable of execution at any time

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall document procedures for advisory notices in accordance with regulatory requirements and keep them capable of execution at any time.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-005#Abschnitt 8.3.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Resolve applicability

RR-576 · Maintain records of actions for post-delivery nonconformity and advisory notices

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall maintain records of actions for post-delivery nonconformity and advisory notices.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 8.3.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-674 · Establish a system for active collection and review of relevant production and post-production information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall establish a system for active collection and review of relevant production and post-production information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-675 · Document the system for active collection and review of relevant production and post-production information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall document the system for active collection and review of relevant production and post-production information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-676 · Maintain the system for active collection and review of relevant production and post-production information

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall maintain the system for active collection and review of relevant production and post-production information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-677 · Consider appropriate methods for collection and processing of information when establishing the system

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall consider appropriate methods for collection and processing of information when establishing the system.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception
Primary source	`SRC-006#Abschnitt 10.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-678 · Where applicable, collect information generated during production and production-process monitoring

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where applicable, collect information generated during production and production-process monitoring.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Development and Realisation; Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.2(a)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-679 · Where applicable, collect user-generated information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, collect user-generated information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.2(b)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-680 · Where applicable, collect information generated through installation, use and servicing

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, collect information generated through installation, use and servicing.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.2(c)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-681 · Where applicable, collect supply-chain information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, collect supply-chain information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.2(d)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-682 · Where applicable, collect publicly available information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, collect publicly available information.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.2(e)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-683 · Where applicable, collect information on generally acknowledged state of the art

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, collect information on generally acknowledged state of the art.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-006#Abschnitt 10.2(f)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-684 · Consider the need to actively collect and review publicly available information on similar devices and products

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall consider the need to actively collect and review publicly available information on similar devices and products.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-685 · Review collected information for safety relevance, including new hazards, unacceptable individual risks, unacceptable overall residual risk or changes in state of the art

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall review collected information for safety relevance, including new hazards, unacceptable individual risks, unacceptable overall residual risk or changes in state of the art.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-686 · Record results of the information review in the RM file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall record results of the information review in the RM file.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-006#Abschnitt 10.3`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: ISO 14971:2019 (SRC-006) requirements including cross-norm merged RR within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-811 · Address procedures for receiving, documenting, evaluating, resolving and tracing post-release feedback

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall address procedures for receiving, documenting, evaluating, resolving and tracing post-release feedback.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.1/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-812 · Define criteria for determining whether feedback constitutes a problem

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall define criteria for determining whether feedback constitutes a problem.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance
Primary source	`SRC-007#Abschnitt 6.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-817 · Monitor feedback on released medical device software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall monitor feedback on released medical device software.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-007#Abschnitt 6.2.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-818 · Document feedback on released software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document feedback on released software.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-007#Abschnitt 6.2.1.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-819 · Evaluate feedback to determine whether a software problem exists

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall evaluate feedback to determine whether a software problem exists.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance
Primary source	`SRC-007#Abschnitt 6.2.1.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-829 · Where locally required, inform users about problems, consequences and available changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where locally required, inform users about problems, consequences and available changes.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-830 · Where locally required, inform competent authorities about problems, consequences and available changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where locally required, inform competent authorities about problems, consequences and available changes.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-869 · Analyse problem reports for trends

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall analyse problem reports for trends.

Theme cluster	Post-Market Monitoring
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

Software Life Cycle (177)

RR-084 · Procedures for Management of Modifications

Applicability **applicable** · Gap status **not assessed** · Assessment state **active** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes procedures for the management of modifications to the high-risk AI system.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-002#Art.17/Abs.1/Buchst.a`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	The fulfilment assessment is explicitly recorded as not assessed.
User fulfilment assessment	not assessed
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-085 · Design Techniques, Procedures and Systematic Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes techniques, procedures and systematic actions to be used for the design of the high-risk AI system.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.b`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-086 · Design Control Techniques, Procedures and Systematic Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes techniques, procedures and systematic actions to be used for design control of the high-risk AI system.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.b`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-087 · Design Verification Techniques, Procedures and Systematic Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The provider shall ensure that the quality management system includes techniques, procedures and systematic actions to be used for design verification of the high-risk AI system.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.b`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-088 · Development Techniques, Procedures and Systematic Actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes techniques, procedures and systematic actions to be used for the development of the high-risk AI system.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.c`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-091 · Examination, Testing and Validation Procedures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The provider shall ensure that the quality management system includes examination, test and validation procedures to be carried out before, during and after the development of the high-risk AI system, including the frequency with which they are to be carried out.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation
Primary source	`SRC-002#Art.17/Abs.1/Buchst.d`
Supplementary sources	`SRC-002#Art.17/Abs.1`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: Provider obligations for high-risk AI systems under SRC-002
Decision-relevant inputs	none
Evaluated conditions	`AI_ACT_HIGH_RISK_SYSTEM: true`
Decision rules	`RUL-027; RUL-028; RUL-029; RUL-030; RUL-032; RUL-033; RUL-034; RUL-035; RUL-038; RUL-043`
Decisions used	AI Act applicability
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-169 · Software Development and Manufacture in Accordance with the State of the Art

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that, for devices that incorporate software or for software that are devices in themselves, the software is developed and manufactured in accordance with the state of the art.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-170 · Consideration of Development Life-Cycle Principles

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that, in the development and manufacture of the software, the principles of the development life cycle are taken into account.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-172 · Consideration of Verification Principles

Applicability **applicable** · Gap status **not fulfilled (user statement)** · Assessment state **active** · Measure **measure assigned by rule**

> The manufacturer shall ensure that, in the development and manufacture of the software, the principles of verification are taken into account.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user assessed the requirement as not fulfilled.
User fulfilment assessment	not fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Close gap

RR-173 · Consideration of Validation Principles

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall ensure that, in the development and manufacture of the software, the principles of validation are taken into account.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.17.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Assess fulfilment

RR-447 · Determine customer, intended-use, regulatory, training and other necessary product requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall determine customer, intended-use, regulatory, training and other necessary product requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.2.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-448 · Review product requirements before supply commitment and ensure the required review outcomes

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assigned by rule**

> The organization shall review product requirements before supply commitment and ensure the required review outcomes.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Add evidence reference

RR-449 · Maintain records of requirement-review results and follow-up actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of requirement-review results and follow-up actions.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-450 · Clarify non-written customer requirements before acceptance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall clarify non-written customer requirements before acceptance.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.2.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-451 · Update applicable documents when product requirements change

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall update applicable documents when product requirements change.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.2.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-452 · Ensure relevant personnel are made aware of changed requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall ensure relevant personnel are made aware of changed requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.2.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-455 · Document procedures for design and development

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for design and development.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-456 · Plan and control product design and development

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall plan and control product design and development.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-457 · Document and, where appropriate, update development planning covering stages, reviews, V&V/transfer, responsibilities, traceability and resources

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall document and, where appropriate, update development planning covering stages, reviews, V&V/transfer, responsibilities, traceability and resources.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.2(a)–(f)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Resolve applicability

RR-458 · Determine design and development inputs relating to product requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall determine design and development inputs relating to product requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-459 · Maintain records of design and development inputs

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of design and development inputs.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-460 · Include functional, performance, usability, safety, regulatory, standards, risk-management and other essential requirements in design inputs

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall include functional, performance, usability, safety, regulatory, standards, risk-management and other essential requirements in design inputs.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.3(a)–(e)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-461 · Review design inputs for adequacy

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall review design inputs for adequacy.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-462 · Approve design inputs

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall approve design inputs.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-463 · Ensure requirements are complete, unambiguous, verifiable or validatable and non-conflicting

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall ensure requirements are complete, unambiguous, verifiable or validatable and non-conflicting.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-005#Abschnitt 7.3.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-464 · Ensure design outputs fulfil inputs, provide downstream information and acceptance criteria, and define characteristics essential for safe and intended use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall ensure design outputs fulfil inputs, provide downstream information and acceptance criteria, and define characteristics essential for safe and intended use.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.4(a)–(d)`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-465 · Make design outputs verifiable against design inputs

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall make design outputs verifiable against design inputs.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-466 · Approve design outputs before release

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall approve design outputs before release.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-467 · Maintain records of design outputs

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall maintain records of design outputs.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-468 · Conduct systematic design reviews at appropriate stages according to planned documented arrangements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall conduct systematic design reviews at appropriate stages according to planned documented arrangements.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Resolve applicability

RR-469 · Include relevant functions and necessary specialists in design reviews

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall include relevant functions and necessary specialists in design reviews.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Assess fulfilment

RR-470 · Maintain the required design-review records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain the required design-review records.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Conduct review or management review (MEAS_REVIEW)
Planned evidence type	Review record
Evidence references	none
Open items	Assess fulfilment

RR-471 · Perform design verification under planned documented arrangements to confirm outputs meet input requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall perform design verification under planned documented arrangements to confirm outputs meet input requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-472 · Document verification plans with methods, acceptance criteria and, where appropriate, statistical methods and sample-size rationale

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall document verification plans with methods, acceptance criteria and, where appropriate, statistical methods and sample-size rationale.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Resolve applicability

RR-473 · Verify output-to-input conformity with the intended connection or interface to other medical devices

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall verify output-to-input conformity with the intended connection or interface to other medical devices.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.6`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-474 · Maintain records of verification results, conclusions and necessary actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of verification results, conclusions and necessary actions.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-475 · Perform design validation under planned documented arrangements to demonstrate fitness for specified application or intended use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall perform design validation under planned documented arrangements to demonstrate fitness for specified application or intended use.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-476 · Document validation plans with methods, acceptance criteria and, where appropriate, statistics and sample-size rationale

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall document validation plans with methods, acceptance criteria and, where appropriate, statistics and sample-size rationale.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Resolve applicability

RR-477 · Perform design validation on representative product

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall perform design validation on representative product.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-478 · Record the rationale for selecting the representative validation product

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall record the rationale for selecting the representative validation product.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-479 · Perform the clinically or performance-related evaluation required by applicable regulatory requirements as part of design validation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The organization shall perform the clinically or performance-related evaluation required by applicable regulatory requirements as part of design validation.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Conformity Assessment
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-480 · Validate intended-use requirements in the connected or interfaced configuration

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall validate intended-use requirements in the connected or interfaced configuration.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-481 · Complete design validation before release for customer use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall complete design validation before release for customer use.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-482 · Maintain records of validation results, conclusions and necessary actions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall maintain records of validation results, conclusions and necessary actions.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.3.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-483 · Document design-transfer procedures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document design-transfer procedures.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-484 · Verify design-output suitability for manufacture before final production specifications

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall verify design-output suitability for manufacture before final production specifications.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-485 · Ensure production capability can meet product requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall ensure production capability can meet product requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-486 · Record design-transfer results and conclusions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall record design-transfer results and conclusions.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-005#Abschnitt 7.3.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-487 · Document procedures for control of design and development changes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall document procedures for control of design and development changes.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-488 · Determine the significance of a change for function, performance, usability, safety, regulatory requirements and intended use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall determine the significance of a change for function, performance, usability, safety, regulatory requirements and intended use.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-489 · Identify design changes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall identify design changes.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-490 · Review changes before implementation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall review changes before implementation.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9(a)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-491 · Verify changes before implementation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall verify changes before implementation.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9(b)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-492 · Where appropriate, validate changes before implementation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall, where appropriate, validate changes before implementation.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9(c)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record validation (MEAS_VALIDATION)
Planned evidence type	Validation record
Evidence references	none
Open items	Resolve applicability

RR-493 · Approve changes before implementation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall approve changes before implementation.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9(d)`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-494 · Assess impacts of design changes on affected products/components, risk-management inputs/outputs and product-realization processes

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The organization shall assess impacts of design changes on affected products/components, risk-management inputs/outputs and product-realization processes.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-495 · Maintain records of design changes, their review and necessary actions

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain records of design changes, their review and necessary actions.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-496 · Maintain a design and development file for each device type or group containing or referencing design-control and change records

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The organization shall maintain a design and development file for each device type or group containing or referencing design-control and change records.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.3.10`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Include in the technical documentation (MEAS_TECHNICAL_DOCUMENTATION)
Planned evidence type	Technical documentation
Evidence references	none
Open items	Resolve applicability

RR-513 · Where appropriate, document installation requirements and acceptance criteria for installation verification

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall, where appropriate, document installation requirements and acceptance criteria for installation verification.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-514 · Maintain records of installation and installation verification performed

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall maintain records of installation and installation verification performed.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 7.5.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-515 · Where servicing is a specified requirement, document necessary servicing procedures and means to verify product requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall, where servicing is a specified requirement, document necessary servicing procedures and means to verify product requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-517 · Maintain records of servicing performed by the organization or its supplier

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> The organization shall maintain records of servicing performed by the organization or its supplier.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-005#Abschnitt 7.5.4`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

RR-561 · Monitor and measure product characteristics at applicable realization stages to verify fulfilment of product requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The organization shall monitor and measure product characteristics at applicable realization stages to verify fulfilment of product requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-005#Abschnitt 8.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-562 · Maintain evidence of conformity with acceptance criteria

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The organization shall maintain evidence of conformity with acceptance criteria.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 8.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-563 · Record the identity of the person responsible for product release

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall record the identity of the person responsible for product release.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 8.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-564 · Where appropriate, identify measuring equipment used in the records

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The organization shall, where appropriate, identify measuring equipment used in the records.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 8.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-565 · Permit product release or service provision only after satisfactory completion of planned documented arrangements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The organization shall permit product release or service provision only after satisfactory completion of planned documented arrangements.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-005#Abschnitt 8.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: ISO 13485:2016 (SRC-005) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iso13485_clause1_exclusion_claimed = no`
Evaluated conditions	`MDR_MDSW: true; ISO13485_NO_CLAUSE1_EXCLUSION: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-693 · Assign each software system safety class A, B or C based on the worst-case risk of harm

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall assign each software system safety class A, B or C based on the worst-case risk of harm.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 4.3/a`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-694 · Document the assigned safety class of each software system in the risk-management file

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the assigned safety class of each software system in the risk-management file.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 4.3/c`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-695 · When decomposing software, inherit the parent safety class or document a rationale for a different class including effective segregation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall, when decomposing software, inherit the parent safety class or document a rationale for a different class including effective segregation.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 4.3/d`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-696 · Document a component safety class when it differs from that of its parent component

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document a component safety class when it differs from that of its parent component.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 4.3/e`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-697 · For a group of software components, apply the processes and tasks required by the highest class present or document a rationale for using a lower class

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall, for a group of software components, apply the processes and tasks required by the highest class present or document a rationale for using a lower class.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 4.3/f`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the risk management file (MEAS_RISK_MANAGEMENT_FILE)
Planned evidence type	Risk management file
Evidence references	none
Open items	Assess fulfilment

RR-716 · Establish a software development plan for the software development process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish a software development plan for the software development process.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-717 · Make the development process appropriate to its scope, extent and applicable safety classes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall make the development process appropriate to its scope, extent and applicable safety classes.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation
Primary source	`SRC-007#Abschnitt 5.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-718 · Fully define the software development life-cycle model or reference it in the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall fully define the software development life-cycle model or reference it in the plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-719 · Specify or reference the software processes used for development

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall specify or reference the software processes used for development.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-720 · Specify planned deliverables, including documentation, in the development plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall specify planned deliverables, including documentation, in the development plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-721 · Plan traceability among system requirements, software requirements, system tests and software-implemented risk controls

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall plan traceability among system requirements, software requirements, system tests and software-implemented risk controls.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1/c`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-722 · Address configuration and change management including SOUP and development-support software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall address configuration and change management including SOUP and development-support software.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1/d`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-723 · Address software problem resolution and lifecycle-wide problem handling in the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall address software problem resolution and lifecycle-wide problem handling in the plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.1/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-724 · Update the development plan as appropriate to development progress

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall update the development plan as appropriate to development progress.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.1.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Resolve applicability

RR-725 · Reference system requirements as inputs to software development

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall reference system requirements as inputs to software development.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.3/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-726 · Include or reference procedures coordinating software and system development

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include or reference procedures coordinating software and system development.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.3/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-728 · Plan or reference software integration, including SOUP, and integration testing

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall plan or reference software integration, including SOUP, and integration testing.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-729 · Specify or reference verification deliverables, tasks, milestones and acceptance criteria

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall specify or reference verification deliverables, tasks, milestones and acceptance criteria.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-731 · Specify or reference planned software documentation, including identification, purpose and document-control responsibilities

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall specify or reference planned software documentation, including identification, purpose and document-control responsibilities.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-732 · Plan software configuration management including controlled items, activities, responsibilities, relationships and control timing

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall plan software configuration management including controlled items, activities, responsibilities, relationships and control timing.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.9`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-733 · Place development tools, items or settings that can affect the software under configuration control

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall place development tools, items or settings that can affect the software under configuration control.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.1.10`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-734 · Plan configuration-management control of configuration items before verification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall plan configuration-management control of configuration items before verification.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.1.11`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-736 · Derive software-system requirements from system requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall derive software-system requirements from system requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-737 · Document the derived software requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the derived software requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-738 · Where appropriate, include functionality and performance requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include functionality and performance requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/a`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-739 · Where appropriate, include system input and output requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include system input and output requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/b`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-740 · Where appropriate, include requirements for interfaces with other systems

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include requirements for interfaces with other systems.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/c`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-741 · Where appropriate, include requirements for software-controlled alarms, warnings and user messages

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where appropriate, include requirements for software-controlled alarms, warnings and user messages.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/d`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-742 · Where appropriate, include data-security requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where appropriate, include data-security requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/e`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-744 · Where appropriate, include data-definition and database requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include data-definition and database requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/g`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-745 · Where appropriate, include requirements for installation, site acceptance and maintenance

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include requirements for installation, site acceptance and maintenance.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/h`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-746 · Where appropriate, include requirements for operating and maintenance methods

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include requirements for operating and maintenance methods.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/i`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-747 · Where appropriate, include requirements for IT-network aspects

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include requirements for IT-network aspects.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/j`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-748 · Where appropriate, include operator-maintenance requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include operator-maintenance requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/k`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-749 · Where appropriate, include applicable regulatory requirements

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include applicable regulatory requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/l`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-753 · Re-evaluate existing requirements, including system requirements, following software requirements analysis

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall re-evaluate existing requirements, including system requirements, following software requirements analysis.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.2.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-754 · Update existing requirements where appropriate following re-evaluation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall update existing requirements where appropriate following re-evaluation.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.2.5`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-755 · Verify software requirements for implementation of system/risk-control requirements, consistency, unambiguity, testability, identification and traceability

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify software requirements for implementation of system/risk-control requirements, consistency, unambiguity, testability, identification and traceability.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-756 · Document verification of software requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document verification of software requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.2.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-757 · Implement software requirements in a documented architecture describing structure and components

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall implement software requirements in a documented architecture describing structure and components.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 5.3.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-758 · Develop an architecture for internal and external software interfaces

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall develop an architecture for internal and external software interfaces.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 5.3.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-759 · Document the architecture for internal and external software interfaces

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the architecture for internal and external software interfaces.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 5.3.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-760 · Specify functional and performance requirements necessary for intended use of each SOUP component

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify functional and performance requirements necessary for intended use of each SOUP component.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.3.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B; soup_used = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true; SOUP_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-761 · Specify system hardware and software required for proper SOUP operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify system hardware and software required for proper SOUP operation.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-007#Abschnitt 5.3.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B; soup_used = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true; SOUP_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-764 · Verify the architecture for implementation of requirements, interface support and proper SOUP operation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify the architecture for implementation of requirements, interface support and proper SOUP operation.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.3.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-765 · Document verification of the software architecture

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document verification of the software architecture.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.3.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-766 · Decompose software until represented by software units

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall decompose software until represented by software units.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 5.4.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-772 · Implement each software unit

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall implement each software unit.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 5.5.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-773 · Establish strategies, methods and procedures for software-unit verification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish strategies, methods and procedures for software-unit verification.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Verification and Validation
Primary source	`SRC-007#Abschnitt 5.5.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-774 · Where verification uses testing, evaluate adequacy of the test procedures

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall, where verification uses testing, evaluate adequacy of the test procedures.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.5.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Resolve applicability

RR-775 · Where appropriate, establish acceptance criteria before integration

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, establish acceptance criteria before integration.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Verification and Validation
Primary source	`SRC-007#Abschnitt 5.5.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-776 · Ensure software units meet established acceptance criteria

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure software units meet established acceptance criteria.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.5.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-778 · Perform verification of software units

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall perform verification of software units.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.5.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-779 · Document software-unit verification results

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document software-unit verification results.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.5.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-780 · Integrate software units according to the integration plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall integrate software units according to the integration plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 5.6.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-781 · Verify software integration against the integration plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify software integration against the integration plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-782 · Retain records evidencing software-integration verification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall retain records evidencing software-integration verification.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-783 · Test integrated software components according to the integration plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall test integrated software components according to the integration plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-784 · Document integration-test results

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document integration-test results.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-785 · Assess whether integrated software functions as intended

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall assess whether integrated software functions as intended.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-786 · Evaluate the adequacy of the integration-test procedure

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall evaluate the adequacy of the integration-test procedure.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-787 · Perform suitable regression testing during software integration

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall perform suitable regression testing during software integration.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-788 · Document integration tests with results/anomalies, repeatability evidence and tester identity

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document integration tests with results/anomalies, repeatability evidence and tester identity.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.6.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-789 · Enter integration anomalies into the software problem-resolution process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall enter integration anomalies into the software problem-resolution process.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.6.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-790 · Define a test set with inputs, expected outputs, pass/fail criteria and procedures covering all software requirements

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall define a test set with inputs, expected outputs, pass/fail criteria and procedures covering all software requirements.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.1/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-791 · Perform the defined software-system tests

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall perform the defined software-system tests.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.1/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-792 · Evaluate adequacy of verification strategies and test procedures

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall evaluate adequacy of verification strategies and test procedures.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-793 · Enter anomalies found in software-system testing into the problem-resolution process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall enter anomalies found in software-system testing into the problem-resolution process.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.7.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-797 · Verify that all software requirements have been tested or otherwise verified

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify that all software requirements have been tested or otherwise verified.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.4/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-798 · Record traceability between software requirements and tests or other verification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record traceability between software requirements and tests or other verification.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.4/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-799 · Verify that test results meet required pass/fail criteria

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify that test results meet required pass/fail criteria.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.4/c`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-800 · Document software-system testing with the specified information needed for repeatability

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document software-system testing with the specified information needed for repeatability.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation
Primary source	`SRC-007#Abschnitt 5.7.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-801 · Before release, ensure software verification is complete and its results evaluated

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall, before release, ensure software verification is complete and its results evaluated.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation
Primary source	`SRC-007#Abschnitt 5.8.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-802 · Document all known remaining anomalies

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document all known remaining anomalies.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-804 · Document released medical-device-software versions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document released medical-device-software versions.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-805 · Document the procedure and environment used to produce released software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the procedure and environment used to produce released software.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-806 · Ensure all development- or maintenance-plan activities and tasks are complete

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure all development- or maintenance-plan activities and tasks are complete.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-807 · Ensure the associated documentation is complete

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure the associated documentation is complete.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-808 · Archive software, configuration items and documentation for at least the longer applicable retention period

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall archive software, configuration items and documentation for at least the longer applicable retention period.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-809 · Establish procedures for reliable delivery without damage or unauthorized modification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish procedures for reliable delivery without damage or unauthorized modification.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 5.8.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and document procedure (MEAS_DOCUMENTED_PROCEDURE)
Planned evidence type	Documented procedure
Evidence references	none
Open items	Assess fulfilment

RR-810 · Establish a software maintenance plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish a software maintenance plan.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-814 · Address use of the software problem-resolution process for post-release problems

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall address use of the software problem-resolution process for post-release problems.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.1/d`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-815 · Address use of configuration management for changes to existing software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall address use of configuration management for changes to existing software.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.1/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-816 · Address procedures for evaluating and implementing SOUP upgrades, fixes, patches and obsolescence

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall address procedures for evaluating and implementing SOUP upgrades, fixes, patches and obsolescence.

Theme cluster	Software Life Cycle
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.1/f`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B; soup_used = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true; SOUP_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-823 · Determine whether a software change is required to address the problem

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall determine whether a software change is required to address the problem.

Theme cluster	Software Life Cycle
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.1.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-824 · Process problem reports using the software problem-resolution process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall process problem reports using the software problem-resolution process.

Theme cluster	Software Life Cycle
Life-cycle phases	Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-825 · Analyse change requests for impact on the organization, released software and interfacing systems

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall analyse change requests for impact on the organization, released software and interfacing systems.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Control change through change request (MEAS_CHANGE_CONTROL)
Planned evidence type	Change record
Evidence references	none
Open items	Assess fulfilment

RR-826 · Evaluate change requests modifying released software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall evaluate change requests modifying released software.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Control change through change request (MEAS_CHANGE_CONTROL)
Planned evidence type	Change record
Evidence references	none
Open items	Assess fulfilment

RR-827 · Approve change requests modifying released software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall approve change requests modifying released software.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Control change through change request (MEAS_CHANGE_CONTROL)
Planned evidence type	Change record
Evidence references	none
Open items	Assess fulfilment

RR-828 · Identify approved change requests affecting released software

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall identify approved change requests affecting released software.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.2.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Control change through change request (MEAS_CHANGE_CONTROL)
Planned evidence type	Change record
Evidence references	none
Open items	Assess fulfilment

RR-831 · Identify Clause 5 activities requiring repetition because of a change

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify Clause 5 activities requiring repetition because of a change.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.3.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-832 · Repeat the Clause 5 activities identified as requiring repetition because of a change

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall repeat the Clause 5 activities identified as requiring repetition because of a change.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.3.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-833 · Release software changes in accordance with 5.8

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall release software changes in accordance with 5.8.

Theme cluster	Software Life Cycle
Life-cycle phases	Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 6.3.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-839 · Assign contributing components a safety class based on the controlled risk

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall assign contributing components a safety class based on the controlled risk.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation
Primary source	`SRC-007#Abschnitt 7.2.2/b`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-840 · Develop components implementing software risk controls in accordance with Clause 5

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall develop components implementing software risk controls in accordance with Clause 5.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-007#Abschnitt 7.2.2/c`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_BC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-847 · Establish a scheme for unique identification of configuration items and versions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall establish a scheme for unique identification of configuration items and versions.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-848 · Control the identification/versioning scheme according to development and configuration planning

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall control the identification/versioning scheme according to development and configuration planning.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.1.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-849 · Document title, manufacturer and unique identifier of each SOUP configuration item

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document title, manufacturer and unique identifier of each SOUP configuration item.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.1.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B; soup_used = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true; SOUP_USED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-850 · Document the software-system configuration including configuration items and versions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the software-system configuration including configuration items and versions.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.1.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-852 · Implement an approved change as specified

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall implement an approved change as specified.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-853 · Identify activities that must be repeated because of a change

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify activities that must be repeated because of a change.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-854 · Perform required repeated activities, including necessary software-safety reclassification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall perform required repeated activities, including necessary software-safety reclassification.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.2.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-855 · Verify changes and repeat verifications invalidated by the change

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall verify changes and repeat verifications invalidated by the change.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.2.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

RR-856 · Retain records linking change requests, relevant problem reports and change approval

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall retain records linking change requests, relevant problem reports and change approval.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.2.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-857 · Retain a retrievable history of controlled configuration items including system configuration

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall retain a retrievable history of controlled configuration items including system configuration.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 8.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-864 · Where appropriate, inform involved parties of the software problem

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, inform involved parties of the software problem.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-865 · Approve change requests according to the change-control process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall approve change requests according to the change-control process.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Control change through change request (MEAS_CHANGE_CONTROL)
Planned evidence type	Change record
Evidence references	none
Open items	Assess fulfilment

RR-866 · Implement approved change requests according to the change-control process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall implement approved change requests according to the change-control process.

Theme cluster	Software Life Cycle
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Control change through change request (MEAS_CHANGE_CONTROL)
Planned evidence type	Change record
Evidence references	none
Open items	Assess fulfilment

RR-867 · Retain records of problem reports, resolution and verification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall retain records of problem reports, resolution and verification.

Theme cluster	Software Life Cycle
Life-cycle phases	Development and Realisation; Verification and Validation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Prepare and update report (MEAS_REPORT)
Planned evidence type	Report
Evidence references	none
Open items	Assess fulfilment

RR-870 · Verify problem resolution for closure, reversal of adverse trends, implementation of required changes and absence of newly caused problems

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall verify problem resolution for closure, reversal of adverse trends, implementation of required changes and absence of newly caused problems.

Theme cluster	Software Life Cycle
Life-cycle phases	Verification and Validation; Post-Market Monitoring and Surveillance; Modification and Maintenance
Primary source	`SRC-007#Abschnitt 9.7`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Perform and record verification or testing (MEAS_VERIFICATION)
Planned evidence type	Verification or test record
Evidence references	none
Open items	Assess fulfilment

Usability (96)

RR-163 · Reduction of Use-Error Risks Related to Ergonomic Features and Intended Use Environment

Applicability **applicable** · Gap status **partially fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assignment open (review required)**

> The manufacturer shall, in eliminating or reducing risks related to use error, reduce as far as possible the risks related to the ergonomic features of the device and the environment in which the device is intended to be used; such risk reduction shall be pursued as far as possible without adversely affecting the benefit-risk ratio.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.5/Buchst.a`
Supplementary sources	`SRC-001#Anh.I/Nr.5; SRC-001#Anh.I/Nr.2`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	Fallstudie: Nachweis zu RR-163 (CS-DOC-RR-163; 1.0; identifier 2bb7c959-bae6-41d4-b464-a5f3f600ab26)
Open items	Close gap; Review measure assignment

RR-164 · Consideration of Intended User Characteristics and Use Environment for Use-Error Risks

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall, in eliminating or reducing risks related to use error, give consideration to the technical knowledge, experience, education, training and use environment, where applicable, and the medical and physical conditions of intended users.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.5/Buchst.b`
Supplementary sources	`SRC-001#Anh.I/Nr.5`
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-174 · Design and Manufacture for Intended Performance in Lay Use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that devices intended for lay use are designed and manufactured to perform appropriately for their intended purpose, taking account of lay persons' skills and available means and reasonably anticipated variation in technique and environment.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-175 · Understandability of Manufacturer Information and Instructions for Lay Persons

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall ensure that the information and instructions it provides are easy for lay persons to understand.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-176 · Applicability of Manufacturer Information and Instructions for Lay Persons

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure that the information and instructions it provides are easy for lay persons to apply.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-177 · Safe Use by the Intended User at All Stages of the Procedure

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure through design and manufacture that the device can be used safely by the intended user at all stages of the procedure, if necessary after appropriate training and/or information.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-178 · Accurate Use by the Intended User at All Stages of the Procedure

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure through design and manufacture that the device can be used accurately by the intended user at all stages of the procedure, if necessary after appropriate training and/or information.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-179 · Minimisation of Risk of Handling Error by the Intended User

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall ensure through design and manufacture that the risk of error by the intended user in handling the device is reduced as far as possible.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-180 · Minimisation of Risk of Erroneous Interpretation of Results

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

If interpretation of results is applicable, the manufacturer shall ensure through design and manufacture that the risk of erroneous interpretation by the intended user is reduced as far as possible.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.2`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true; RESULT_INTERPRETATION_BY_USER: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-181 · Procedure for Verification of Intended Performance at the Time of Use

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

Where appropriate, the manufacturer shall ensure that the device includes a procedure enabling the lay person, at the time of use, to verify that it will perform as intended.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-182 · Warning When No Valid Result Is Provided

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where appropriate and if applicable, the manufacturer shall ensure that the device includes a procedure by which the lay person is warned when the device has failed to provide a valid result.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.22.3`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR Annex I section 22 for intended lay use
Decision-relevant inputs	`lay_use = yes`
Evaluated conditions	`MDR_MDSW: true; LAY_USE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-189 · Suitability of the Medium

Applicability **applicable** · Gap status **evidence review required** · Assessment state **active** · Measure **measure assigned by rule**

> The manufacturer shall ensure that the medium of the label and instructions for use is appropriate to the particular device, its intended purpose and the technical knowledge, experience, education or training of the intended user(s).

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment but assigned no evidence reference (evidence gap).
User fulfilment assessment	partially fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Add evidence reference

RR-190 · Suitability of the Format

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the format of the label and instructions for use is appropriate to the particular device, its intended purpose and the technical knowledge, experience, education or training of the intended user(s).

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-192 · Suitability of Legibility

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the legibility of the label and instructions for use is appropriate to the particular device, its intended purpose and the technical knowledge, experience, education or training of the intended user(s).

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-193 · Suitability of Location

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the location of the label and instructions for use is appropriate to the particular device, its intended purpose and the technical knowledge, experience, education or training of the intended user(s).

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-194 · Understandability of the Instructions for Use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall ensure that the instructions for use are written in terms readily understood by the intended user.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-195 · Supplementation with Drawings and Diagrams

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

> Where appropriate, the manufacturer shall ensure that the instructions for use are supplemented with drawings and diagrams.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.a/Satz.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Resolve applicability

RR-197 · Human-Readable Labelling

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall ensure that labels are provided in a human-readable format.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.c`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Design labelling or instructions for use accordingly (MEAS_LABELLING_AND_INSTRUCTIONS)
Planned evidence type	Labelling and instructions for use
Evidence references	none
Open items	Assess fulfilment

RR-202 · Use of Internationally Recognised Symbols

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where appropriate, the manufacturer shall ensure that the information supplied by the manufacturer takes the form of internationally recognised symbols.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.h`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-203 · Conformity of Used Symbols and Identification Colours

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> Where the manufacturer uses a symbol or identification colour and a relevant harmonised standard or common specification exists for the area concerned, the manufacturer shall ensure that the symbol or identification colour conforms to it.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-001#Anh.I/Nr.23.1/Buchst.h`
Supplementary sources	none
Reason for applicability	The input required for a local condition is missing. Scope: MDR gap corpus for MDSW
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true; SYMBOL_WITH_HS_OR_CS: unknown`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-743 · Where appropriate, include requirements for the software-implemented user interface

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where appropriate, include requirements for the software-implemented user interface.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-007#Abschnitt 5.2.2/f`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62304:2006+AMD1:2015 (SRC-007) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62304_route = REGULAR_LIFECYCLE; software_safety_class = B`
Evaluated conditions	`MDR_MDSW: true; IEC62304_REGULAR_ROUTE: true; IEC62304_CLASS_ABC: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-872 · Establish a usability engineering process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall establish a usability engineering process.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-873 · Document the usability engineering process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the usability engineering process.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-874 · Implement the usability engineering process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall implement the usability engineering process.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-875 · Maintain the usability engineering process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall maintain the usability engineering process.

Theme cluster	Usability
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-876 · Address safety-relevant user interactions across applicable use phases

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall address safety-relevant user interactions across applicable use phases.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation; Placing on the Market, Putting into Service and Operation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-877 · Have usability engineering activities planned, performed and documented by competent persons

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

> The manufacturer shall have usability engineering activities planned, performed and documented by competent persons.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-878 · Incorporate or reference applicable usability engineering activities in an existing documented product-development process

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall incorporate or reference applicable usability engineering activities in an existing documented product-development process.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 4.1.1`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-879 · Address safety information used as a risk-control measure within the usability engineering process

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall address safety information used as a risk-control measure within the usability engineering process.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation; Verification and Validation
Primary source	`SRC-008#Abschnitt 4.1.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`safety_information_as_risk_control = yes`
Evaluated conditions	`MDR_MDSW: true; SAFETY_INFORMATION_AS_RISK_CONTROL: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-880 · Determine whether such safety information is perceptible, understandable and supports correct use

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall determine whether such safety information is perceptible, understandable and supports correct use.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 4.1.3`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`safety_information_as_risk_control = yes`
Evaluated conditions	`MDR_MDSW: true; SAFETY_INFORMATION_AS_RISK_CONTROL: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-881 · Record usability engineering process results in the Usability Engineering File

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record usability engineering process results in the Usability Engineering File.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 4.2`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	none
Open items	Assess fulfilment

RR-882 · Establish a Use Specification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall establish a Use Specification.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-883 · Include the intended medical indication

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include the intended medical indication.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-884 · Include intended patient groups

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include intended patient groups.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-885 · Include intended body part or tissue type

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include intended body part or tissue type.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-886 · Include the intended user profile

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include the intended user profile.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-887 · Include the intended use environment

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include the intended use environment.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-888 · Include the operating principle

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall include the operating principle.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.1`
Supplementary sources	none
Reason for applicability	The parent regulatory scope is active; no local condition documented. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	none
Evaluated conditions	`MDR_MDSW: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-889 · Identify safety-related user-interface characteristics as part of the ISO 14971 risk analysis

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify safety-related user-interface characteristics as part of the ISO 14971 risk analysis.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-890 · Include prescribed primary operating functions in the identification

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall include prescribed primary operating functions in the identification.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-891 · Identify possible user-interface-related use errors

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify possible user-interface-related use errors.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-892 · Record the identified safety characteristics and possible use errors in the UE File

Applicability **applicable** · Gap status **fulfilled (user statement with evidence reference)** · Assessment state **active** · Measure **measure assigned by rule**

The manufacturer shall record the identified safety characteristics and possible use errors in the UE File.

Theme cluster	Usability
Life-cycle phases	Development and Realisation
Primary source	`SRC-008#Abschnitt 5.2`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	The user stated (partial) fulfilment and assigned at least one evidence reference.
User fulfilment assessment	fulfilled
User justification	Synthetische Fallstudienangabe (Block 6, Szenario laut REFERENCE_ASSESSMENT.yaml)
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	Fallstudie: Nachweis zu RR-892 (CS-DOC-RR-892; 1.0; identifier 21960125-3485-4c3b-b732-a9ff90e96037)
Open items	none

RR-896 · Identify reasonably foreseeable hazard-related use scenarios

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify reasonably foreseeable hazard-related use scenarios.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-897 · Describe all tasks and their sequence for each hazard-related use scenario

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall describe all tasks and their sequence for each hazard-related use scenario.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-898 · Specify the severity of associated harm for each hazard-related use scenario

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify the severity of associated harm for each hazard-related use scenario.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.4`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-899 · Select hazard-related use scenarios for summative evaluation using one of the permitted selection schemes

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall select hazard-related use scenarios for summative evaluation using one of the permitted selection schemes.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-900 · Document a summary of the selection scheme

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document a summary of the selection scheme.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-901 · Document the rationale for the selection scheme

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall document the rationale for the selection scheme.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-902 · Document the results of applying the selection scheme

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the results of applying the selection scheme.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.5`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-903 · Establish a User Interface Specification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall establish a User Interface Specification.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-904 · Maintain the User Interface Specification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall maintain the User Interface Specification.

Theme cluster	Usability
Life-cycle phases	Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-905 · Consider the Use Specification when establishing the UI Specification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall consider the Use Specification when establishing the UI Specification.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-906 · Consider known or foreseeable use errors

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall consider known or foreseeable use errors.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-907 · Consider hazard-related use scenarios

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall consider hazard-related use scenarios.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-908 · Specify testable technical UI requirements, including UI elements implementing selected risk controls

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify testable technical UI requirements, including UI elements implementing selected risk controls.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Development and Realisation
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-909 · Specify whether accompanying documentation is required

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify whether accompanying documentation is required.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-910 · Specify whether device-specific training is required

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify whether device-specific training is required.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-911 · Record the User Interface Specification in the UE File

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record the User Interface Specification in the UE File.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.6`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	none
Open items	Assess fulfilment

RR-912 · Establish a User Interface Evaluation Plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall establish a User Interface Evaluation Plan.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-913 · Maintain the evaluation plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall maintain the evaluation plan.

Theme cluster	Usability
Life-cycle phases	Planning and Conception; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.7.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Establish and maintain plan (MEAS_PLAN)
Planned evidence type	Plan
Evidence references	none
Open items	Assess fulfilment

RR-914 · Document the objective of planned evaluations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document the objective of planned evaluations.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-915 · Determine methods for planned formative and summative evaluations

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall determine methods for planned formative and summative evaluations.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-916 · Document intended user groups for usability testing

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall document intended user groups for usability testing.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Assess fulfilment

RR-917 · Describe test environment and other use conditions based on the Use Specification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall describe test environment and other use conditions based on the Use Specification.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-918 · Specify whether accompanying documentation will be available during testing

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify whether accompanying documentation will be available during testing.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-919 · Specify whether device-specific training precedes the test and define the minimum interval

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify whether device-specific training precedes the test and define the minimum interval.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-920 · Record the UI Evaluation Plan in the UE File

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record the UI Evaluation Plan in the UE File.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.1`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	none
Open items	Assess fulfilment

RR-921 · Specify formative-evaluation methods

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify formative-evaluation methods.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.2/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-922 · Specify the part of the UI to be evaluated formatively

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify the part of the UI to be evaluated formatively.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.2/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-923 · Specify when each formative evaluation will occur

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify when each formative evaluation will occur.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.2/c`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-924 · Specify summative evaluation method(s) for each selected scenario

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify summative evaluation method(s) for each selected scenario.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-925 · Justify that the selected method(s) provide objective evidence

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall justify that the selected method(s) provide objective evidence.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/a`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-926 · Specify the UI part subject to summative evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify the UI part subject to summative evaluation.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/b`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-927 · Where applicable, define criteria for perceptibility, comprehension and support of correct use by safety information

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where applicable, define criteria for perceptibility, comprehension and support of correct use by safety information.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/c`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-928 · Specify availability of accompanying documentation during summative evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify availability of accompanying documentation during summative evaluation.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/d`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-929 · Specify training to be provided during summative evaluation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify training to be provided during summative evaluation.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/d`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-930 · Specify how test participants represent intended user profiles

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify how test participants represent intended user profiles.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-931 · Justify grouping of participants used to determine sample size

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall justify grouping of participants used to determine sample size.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-932 · Specify test environment/use conditions and justify their representativeness

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify test environment/use conditions and justify their representativeness.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-933 · Define correct use for each selected hazard-related use scenario

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall define correct use for each selected hazard-related use scenario.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-934 · Specify data-collection methods for subsequent analysis of use errors and use difficulties

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall specify data-collection methods for subsequent analysis of use errors and use difficulties.

Theme cluster	Usability
Life-cycle phases	Planning and Conception
Primary source	`SRC-008#Abschnitt 5.7.3/e`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE; usability_test_planned = yes`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true; USABILITY_TEST_PLANNED: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-935 · Design the UI in accordance with the UI Specification

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall design the UI in accordance with the UI Specification.

Theme cluster	Usability
Life-cycle phases	Development and Realisation
Primary source	`SRC-008#Abschnitt 5.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-936 · Implement the UI according to the specification, including required accompanying documentation and training provisions

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall implement the UI according to the specification, including required accompanying documentation and training provisions.

Theme cluster	Usability
Life-cycle phases	Development and Realisation
Primary source	`SRC-008#Abschnitt 5.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-937 · Apply appropriate usability engineering methods, including formative evaluation, during design and implementation

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall apply appropriate usability engineering methods, including formative evaluation, during design and implementation.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Verification and Validation
Primary source	`SRC-008#Abschnitt 5.8`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-938 · Record results of formative evaluations in the UE File

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall record results of formative evaluations in the UE File.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.8`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	none
Open items	Resolve applicability

RR-939 · Repeat applicable Clause 5 activities when new usability-related issues are discovered

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall repeat applicable Clause 5 activities when new usability-related issues are discovered.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.8`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-942 · Perform summative evaluation on the final or production-equivalent UI for every selected scenario according to the plan

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

> The manufacturer shall perform summative evaluation on the final or production-equivalent UI for every selected scenario according to the plan.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-943 · When using summative data from comparable interfaces, justify its applicability

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, when using summative data from comparable interfaces, justify its applicability.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-944 · Record summative-evaluation results in the UE File

Applicability **applicable** · Gap status **not assessed** · Measure **measure assigned by rule**

The manufacturer shall record summative-evaluation results in the UE File.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	Document in the usability engineering file (MEAS_USABILITY_ENGINEERING_FILE)
Planned evidence type	Usability engineering file
Evidence references	none
Open items	Assess fulfilment

RR-945 · Analyse summative-evaluation data

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall analyse summative-evaluation data.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-946 · Identify all use errors and use difficulties observed

Applicability **applicable** · Gap status **not assessed** · Measure **measure assignment open (review required)**

The manufacturer shall identify all use errors and use difficulties observed.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	The parent scope is active and all local conditions are met. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	No fulfilment assessment exists for the applicable requirement.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Assess fulfilment; Review measure assignment

RR-947 · Determine the cause when a use error or difficulty can lead to a hazardous situation

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall determine the cause when a use error or difficulty can lead to a hazardous situation.

Theme cluster	Usability
Life-cycle phases	Verification and Validation
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-948 · Repeat applicable Clause 5 activities when new usability-related issues are identified

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall repeat applicable Clause 5 activities when new usability-related issues are identified.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-949 · Where no new use errors, hazards, hazardous situations or scenarios have been identified, determine whether further UI safety improvement is necessary and practicable

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

> The manufacturer shall, where no new use errors, hazards, hazardous situations or scenarios have been identified, determine whether further UI safety improvement is necessary and practicable.

Theme cluster	Usability
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-950 · Where further improvement is necessary and practicable, re-enter the process at Clause 5.6

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assignment open (review required)**

The manufacturer shall, where further improvement is necessary and practicable, re-enter the process at Clause 5.6.

Theme cluster	Usability
Life-cycle phases	Development and Realisation; Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	No measure planned; assignment open
Planned evidence type	none
Evidence references	none
Open items	Resolve applicability

RR-951 · If further improvement is unnecessary or impracticable, document the rationale

Applicability **applicability unresolved** · Gap status **applicability unresolved** · Measure **measure assigned by rule**

The manufacturer shall, if further improvement is unnecessary or impracticable, document the rationale.

Theme cluster	Usability
Life-cycle phases	Verification and Validation; Modification and Maintenance
Primary source	`SRC-008#Abschnitt 5.9`
Supplementary sources	none
Reason for applicability	A documented local condition is not modelled in machine-readable form. Scope: IEC 62366-1:2015+AMD1:2020 (SRC-008) requirements within the RegScoutX scope of MDR medical device software
Decision-relevant inputs	`iec62366_ui_route = REGULAR_ROUTE`
Evaluated conditions	`MDR_MDSW: true; IEC62366_REGULAR_ROUTE: true`
Decision rules	`RUL-001; RUL-002; RUL-003; RUL-004; RUL-005; RUL-006`
Decisions used	MDR/IVDR boundary; Qualification as medical device software
Reason for gap status	Applicability is unresolved; no fulfilment assessment is used.
User fulfilment assessment	none
Planned measure	Record result (MEAS_RECORD)
Planned evidence type	Record
Evidence references	none
Open items	Resolve applicability

Requirements not applicable (128)

Requirements not applicable, with reason

Requirement	Title	Theme cluster	Reason	Decisions used
RR-200	Non-Paper Instructions for Use Only within the Applicable Implementing Framework	Transparency	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-205	Paper IFU for Foreseeable Lay Use	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-206	Documented eIFU Risk Assessment	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-207	User Knowledge, Experience and Needs	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-208	Use Environment	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-209	User Knowledge of Display Hardware and Software	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-210	Access to Electronic Resources	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-211	Performance of Anti-Tampering Safeguards	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-212	Safety and Back-Up Mechanisms for Faults	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-213	Foreseeable Medical Emergencies	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-214	Impact of Temporary Unavailability	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-215	Safety Measures for Unavailability	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-216	Evaluation of Paper-IFU Provision Period	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-217	Website Compatibility with Display Devices	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-218	Management of Different IFU Versions	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-219	Updating Based on Post-Market Experience	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-220	Demonstration of At Least Equivalent Safety	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-221	Provision in All Relevant Member States	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-222	System for Providing Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-223	Paper IFU Free of Additional Cost	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-224	Timely Provision of Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-225	Information on Foreseeable Medical Emergencies	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-226	Start-Up Information for Built-In Display Systems	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-227	Proper Design of eIFU	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-228	Proper Functioning of eIFU	Robustness	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-229	Verification Evidence	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-230	Validation Evidence	Quality and Risk Management	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-231	No Impairment of Safe Use by eIFU Display	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-232	Information on Software Requirements	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-233	Information on Hardware Requirements	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-234	System for Indicating IFU Revisions	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-235	User Notification of Safety-Related Revisions	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-236	Ten-Year eIFU Availability	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-237	Availability for Two Years Beyond Final Expiry	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-238	Fifteen-Year eIFU Availability	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-239	eIFU Availability on Manufacturer Website	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-240	Member-State-Determined Official Union Language	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-241	Availability of All Issued eIFU Versions	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-242	Availability of Publication Dates	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-243	Clear Notice of Electronic Instead of Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-244	Notice at the Software Access Location	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-245	Information on Accessing eIFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-246	Placement of eIFU Access Information	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-247	Information Needed to View eIFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-248	Basic UDI-DI and/or UDI-DI	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-249	Additional Device Identification Information	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-250	Manufacturer Contact Details	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-251	Information on Requesting Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-252	Information on Timing and Free Provision of Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-253	Complete eIFU in Text Form	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-254	At Least the Same Information as Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-255	Additional Website Access	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-256	Commonly Used Format	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-257	Readability with Freely Available Software	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-258	Protection Against Unauthorised Access	Robustness	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-259	Protection Against Content Tampering	Robustness	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-260	Minimisation of Server Downtime	Robustness	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-261	Minimisation of Display Errors	Robustness	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-262	Stability of the Internet Address	Robustness	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-263	Direct Accessibility of the Internet Address	Usability	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-264	Provision of the eIFU Internet Address to the UDI Database	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-265	Content Consistency of Additional eIFU with Paper IFU	Transparency	The parent regulatory scope is not active.	MDR/IVDR boundary; Qualification as medical device software
RR-353	Preparation of the Post-Market Surveillance Report	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-354	Summary of PMS Analysis Results	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-355	Summary of PMS Analysis Conclusions	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-356	Rationale for Preventive and Corrective Actions Taken	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-357	Description of Preventive and Corrective Actions Taken	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-358	Updating the PMS Report When Necessary	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-359	Availability of the PMS Report to the Competent Authority	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-370	At Least Annual PSUR Update for Class IIb and III	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-374	PSUR Documentation for Custom-Made Devices	Post-Market Monitoring	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-375	Electronic Submission of the PSUR to the Notified Body	Post-Market Monitoring	The parent regulatory scope is not active.	MDR/IVDR boundary; MDR classification; Qualification as medical device software
RR-399	Monitor and control outsourced processes affecting product conformity	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-400	Retain responsibility for conformity when a relevant process is outsourced	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-401	Make outsourced-process controls proportionate to risk and the external party's capability	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-402	Use written quality agreements for such outsourced processes	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-698	Investigate feedback on legacy software, including post-production information on accidents and near-accidents	Post-Market Monitoring	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-699	Consider integration of legacy software into the overall device architecture in risk management for continued use	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-700	Evaluate the continuing validity of risk-control measures implemented in legacy software	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-701	Identify hazardous situations associated with continued use of legacy software	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-702	Identify possible software-related causes of those hazardous situations	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-703	Define risk-control measures for possible causes related to continued use of legacy software	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-704	Perform a safety-class-based gap analysis of available deliverables against 5.2, 5.3, 5.7 and Clause 7	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-705	Evaluate the continuing validity of available legacy-software deliverables	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-706	Where gaps exist, evaluate the possible risk reduction from generating missing deliverables and performing related activities	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-707	Determine which missing deliverables must be generated	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-708	Determine which related activities must be performed	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-709	Provide at least software-system test records as a deliverable	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-710	Establish a plan for generating identified missing deliverables	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-711	Execute the plan for generating the identified missing deliverables	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-712	Address use of the Clause 9 problem-resolution process in the gap-closing plan	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-713	Perform changes to legacy software in accordance with Clause 6	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-714	Document the version of legacy software in use	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-715	Document a rationale for continued use of legacy software based on the results of Clause 4.4	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-727	Include or reference development standards, methods and tools used for class C components	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-762	Define software-component separation essential to risk control	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-763	Demonstrate that the separation essential to risk control is effective	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-767	Develop sufficiently detailed design for each software unit	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-768	Document the detailed design of each software unit	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-769	Document sufficiently detailed design of all software-unit interfaces	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-770	Verify that detailed software design implements and is consistent with the architecture	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-771	Document verification of detailed software design	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-777	Where applicable, establish additional acceptance criteria for behaviour, resources, error handling and related characteristics	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-794	After changes, repeat, modify or add tests to verify effectiveness of problem correction	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-795	After changes, test for unintended side effects	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-796	After changes, perform relevant software risk-management activities under 7.4	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-820	Record each identified problem as a problem report	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-821	Ensure problem reports include actual or potential harm events and specification deviations	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-843	Analyse software changes, including SOUP, for newly introduced causes of hazardous situations	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-844	Analyse software changes for need for additional software risk controls	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-845	Analyse software/SOUP changes for interference with existing risk controls	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-851	Modify controlled configuration items only in response to approved change requests	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-858	Create a problem report for each detected software problem	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-859	Include criticality and other information needed for resolution in the problem report	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-860	Investigate the problem and, where possible, identify its cause	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-861	Evaluate the safety relevance of the problem using the software risk-management process	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software

Requirement	Title	Theme cluster	Reason	Decisions used
RR-862	Document results of problem investigation and safety evaluation	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-863	Create a change request for required corrective action or document the rationale for no action	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-871	For change-related testing, retesting or regression testing, document the specified test information completely	Software Life Cycle	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-940	Where training is required for safe use, establish training capability for the expected operating life	Usability	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-941	Implement the training capability for safe use using at least one of the permitted provision routes	Usability	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-953	Retain the Use Specification established under 5.1 in the UE File	Usability	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-954	Review available post-production data including complaints and incident/near-incident reports	Post-Market Monitoring	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-955	Record relevant use-error and field information indicating possible usability-related safety issues in the UE File	Post-Market Monitoring	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-956	Review the device risk analysis	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-957	Ensure usability-related hazards and hazardous situations have been identified	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-958	Ensure those hazards and hazardous situations are documented	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software
RR-959	Verify that all relevant risks have been reduced to an acceptable level as specified by the risk assessment	Quality and Risk Management	The parent scope is active, a local condition is not met.	MDR/IVDR boundary; Qualification as medical device software

10. Measures and evidence

Measures and planned evidence types are taken from the versioned measure catalog. They describe which kind of evidence is expected, not whether evidence exists.

Planned measures for applicable requirements

Planned measure	Planned evidence type	Applicable requirements	of which with evidence reference
Plan and conduct audit	Audit record	7	0
Control change through change request	Change record	6	0
Ensure and demonstrate competence or training	Competence or training record	3	0
Establish and document procedure	Documented procedure	17	1
Design labelling or instructions for use accordingly	Labelling and instructions for use	21	0
Establish and maintain plan	Plan	48	0
Perform and record PMS or vigilance activity	PMS or vigilance record	23	0
Record result	Record	56	1
Prepare and update report	Report	15	0
Conduct review or management review	Review record	4	0
Document in the risk management file	Risk management file	19	0
Include in the technical documentation	Technical documentation	4	0
Document in the usability engineering file	Usability engineering file	6	1
Perform and record validation	Validation record	3	0
Perform and record verification or testing	Verification or test record	26	0

For 284 applicable requirements no measure is assigned (assignment open); this is listed as an open item and not assumed.

Evidence register

Recorded evidence references

Title	Document identifier	Version	Location	Measure	Linked requirements
Fallstudie: Nachweis zu RR-001	CS-DOC-RR-001	1.0	none	none	RR-001
Fallstudie: Nachweis zu RR-064	CS-DOC-RR-064	1.0	none	none	RR-064
Fallstudie: Nachweis zu RR-082	CS-DOC-RR-082	1.0	none	MEAS_DOCUMENTED_PROCEDURE	RR-082
Fallstudie: Nachweis zu RR-123	CS-DOC-RR-123	1.0	none	MEAS_RECORD	RR-123
Fallstudie: Nachweis zu RR-160	CS-DOC-RR-160	1.0	none	none	RR-160
Fallstudie: Nachweis zu RR-163	CS-DOC-RR-163	1.0	none	none	RR-163
Fallstudie: Nachweis zu RR-205	CS-DOC-RR-205	1.0	none	MEAS_LABELLING_AND_INSTRUCTIONS	RR-205
Fallstudie: Nachweis zu RR-302	CS-DOC-RR-302	1.0	none	none	RR-302
Fallstudie: Nachweis zu RR-539	CS-DOC-RR-539	1.0	none	none	RR-539
Fallstudie: Nachweis zu RR-892	CS-DOC-RR-892	1.0	none	MEAS_USABILITY_ENGINEERING_FILE	RR-892

The planned evidence type from the measure catalog, the evidence references recorded by the user and the project evidence of the demonstrator are kept separate; only the first two appear in this report.

11. Open items and review needs

Open items comprise review points from decision rules, incomplete inputs, unresolved applicability, requirements not yet assessed, earlier assessments to be confirmed, missing evidence references and open measure assignments.

Open review points from decision rules

Decision area	Rule	Outcome	Cause
No open review points from decision rules.			

Incomplete inputs

Decision area	Rule	Input key	Question
No incomplete inputs.			

Unresolved applicability by reason

Reason	Count
A documented local condition is not modelled in machine-readable form.	212
The evidence documents no classification of this requirement as regular or conditional.	61
The input required for a local condition is missing.	16

Not assessed	523
Confirmation of earlier assessments required	none
Evidence review required	RR-076; RR-135; RR-165; RR-189; RR-384; RR-448; RR-692
Measure assignment open (applicable requirements)	284

12. Traceability and method

RegScoutX is rule-based and deterministic; no generative AI is used. Identical inputs, assessments and catalog versions yield the same result and the same domain fingerprint.

Derivation chain per requirement: inputs, evaluated conditions, decision rules and decision areas, applicability, requirement, gap status, planned measure and evidence reference. Section 9 shows the chain for every applicable and unresolved requirement.

Applicability follows only the versioned applicability mapping and the decision results; requirement texts are not evaluated. Cases that cannot be decided remain unresolved and are never presented as applicable or not applicable.

The report is created from the stored revision without recomputing decisions, applicability or gap status. If the catalog versions match, the stored state is checked against a re-evaluation before the report is generated.

Source references are locators of the versioned source register. Primary law and guidance are shown separately.

13. Limits of this report

This report is the result of a rule-based demonstrator. It does not replace legal advice, certification or a conformity assessment by an authority or a notified body.

The report covers only the versioned catalogs and theme clusters listed in section 2. Requirements outside this scope are not assessed; no statement is made about the completeness of all requirements applicable to the product.

Fulfilment statuses, justifications and evidence references are statements of the user. RegScoutX does not check document contents; an existing evidence reference does not prove that a requirement is fulfilled.

14. Version and revision information

This report is based on revision 1. Stored revisions are never modified; reopening and re-assessment against current catalogs create new revisions.

Revisions of the assessment

Revision	State	Origin	Based on	Created	Last changed	Catalog versions
1	draft	created	none	2026-09-28T22:14:00.232632+00:00	2026-09-28T22:14:35.832159+00:00	as reported revision