



Calibration qualified

Equipment Name	Sourdproof Box
Control no	97W-0526C
Số SN	FN-CNSB-CPEII-0526
Date	March 31, 2023
Due	September 30, 2023
Spec	Max \leq -45dB
Result	PASS
Signed by	HAI
Department	CPEII-TE

	open	close				Max	Result
		lan 1	lan 2	lan 3	spec		
Value (Db)	100	41	40	41	\leq -45dB	41	Pass