

Performance Work Statement (PWS)
Waters Technologies (TA Instruments) Calibration and Maintenance Services

1.0 Description/Background:

The Non-Metallic and Low Observable Materials Branch, Code 6170, of the Naval Surface Warfare Center, Carderock Division (NSWCCD) requires calibration, performance maintenance, and technical support services for government-owned Waters Technologies Corporation (TA Instruments) thermal analysis equipment currently utilized by the Branch. This specialized equipment is essential for testing qualified materials and supporting program requirements for multiple NAVSEA projects. The procurement ensures that all systems are properly verified, serviced, and calibrated to original manufacturer specifications to prevent project delays and ensure continuous fleet support.

2.0 Objective:

The Contractor shall provide professional preventative maintenance and calibration services for the specified thermal analysis equipment. All calibrations performed under this contract shall be certified and valid for a period of one (1) year from the date of service.

The contract scope covers the following equipment modules, aligned with the manufacturer's quote (Quote # 24579363) options and matching active government contract inventories (excluding Section 2.4)

2.1 Discovery DSC 2500 System:

2.1.1 Part Number: 999800.705 (PM Visit Charge DDSC DSC2500 DSC-X3)

2.1.2 Serial Number: DSC24-03162

2.1.3 Service Required: Performance Maintenance Visit (PMV) including labor and certified calibration with traceable standards.

2.2 Discovery TGA 5500 System:

2.2.1 Part Number: 999800.559 (PM Visit Charge Q5000IR DTGA TGA5500)

2.2.2 Serial Number: 5500-1116

2.2.3 Service Required: Performance Maintenance Visit (PMV) including labor and certified calibration with traceable standards.

2.3 Discovery Q1000/Q2000 DSC250 System:

2.3.1 Part Number: 999800.558 (PM Visit Charge Q1000/Q2000 DSC250)

2.3.2 Serial Number: 00034

2.3.3 Service Required: Performance Maintenance Visit (PMV) including labor and certified calibration with traceable standards.

2.4 Discovery Q500 TGA550 System:

2.4.1 Part Number: 999800.560 (PM Visit Charge Q500 TGA550)

2.4.2 Serial Number: 0500-0016

2.4.3 Service Required: Performance Maintenance Visit (PMV) including labor and certified calibration with traceable standards.

2.5 Thermal Analysis Modules (On-Site Labor):

2.5.1 Part Number: 999800.500 (Labor Thermal Analysis)

2.5.2 Quantity: 4 Hours

2.5.3 Service Required: On-site technical support for connection setup, diagnostic checks, and system validation of the thermal analysis modules.

3.0 Performance Maintenance & Annual Calibration/Certification:

The Contractor shall perform on-site performance maintenance and annual calibration services on the equipment listed in Section 2.0. The services performed by the Contractor's service engineer must guarantee that all instruments meet original factory performance specifications.

Key maintenance actions include:

3.1 Complete system inspection, cleaning, and replacement of standard worn items.

3.2 Verification of gas cooling accessories (GCA), furnaces, and auto-sampler alignments.

3.3 Calibration using NIST-traceable reference standards.

3.4 Validation testing to verify baseline flatness, sensitivity, and measurement precision.

3.5 Provision of formal calibration certificates and comprehensive maintenance check sheets.

4.0 Special Qualifications:

The Contractor and any field service engineers assigned to execute this PWS must be factory-trained, authorized, and certified directly by the original equipment manufacturer (Waters Technologies Corporation / TA Instruments). The engineers must possess demonstrated competence in servicing high-precision thermal analysis instrumentation, including Discovery series DSC and TGA modules, Horizontal Gas Purge systems.

5.0 Deliverables:

The Contractor shall deliver the physical services and corresponding documentation for each piece of serviced equipment. All electronic reports shall be submitted directly to the Technical Point of Contact (TPOC) upon completion of the on-site visits.

Number	Name
A001	Inspection Report
A002	Calibration Report

6.0 Period, Place of Performance, and Delivery Information:

6.1 Period of Performance: One (1) year from the date of contract award.

6.2 Place of Performance: All services, including the 4-hour technical support labor and Performance Maintenance Visits (PMV), shall be performed on-site at the government facility:

All services, including the 4-hour technical support labor and Performance Maintenance Visits (PMV), shall be performed on-site at the government facility:

6.3 Delivery & Shipping Information: Any replacement parts, accessories, or materials shipped in support of this contract must be routed through the central receiving building during standard receiving hours (0600-1400 Monday through Thursday; 0600-1330 Friday):

6.4: Shipping Address: NSWCCD, Central Receiving (Attn: Code 6170 / Bldg. 143), 9500 MacArthur Blvd, West Bethesda, MD 20817-5700.

7.0 Security Requirements: Unclassified

8.0 Government Technical Points of Contracts:
