

TEST REPORT IEC 60068-2-78:2012: Damp heat , steady state	
Date of issue.....:	2024-07-12
Standard.....:	IEC 60068-2-78:2012: Damp heat , steady state
Procedure deviation.....:	N/A
Non-standard test method.....:	N/A
Type of test object.....:	Infrared No-Touch Forehead Thermometer
Trademark.....:	N/A
Model/type reference.....:	PT9L
Report No	ETX202303-07-078-05
Testing period.....:	2024.07.10-2024.07.11
Manufacturer.....:	ANDON HEALTH CO., LTD.
Address.....:	No.3 JinPing Street, YaAn Road, Nankai District, Tianjin, China
Test laboratory..... Address.....	Tianxu(Tianjin)Testing Technology Service Co.,LTD 101, First floor, Building L, No.3 JinPing Street, Ya An Road, Nankai District, Tianjin, China

Complied by: _____ (Kang Yanli) Date: _____ 3

Reviewed by: _____ (Guan Qingxiang) Date: _____ 4

Clause	Requirement + Test	Result - Remark	Verdict
Damp heat test	The sample is in normal working condition and put into the programmable constant temperature and humidity test chamber. The temperature and humidity of the test box are adjusted to 40 ° C and 95% RH. After the temperature and humidity are stable, it is maintained for 14.8 hours.	/	

1



2



3



4

ATTACHMENT FILE 1 photos of the EUT



--- End of Test Report ---