

## pH

Name of the test	pH
Range of testing (ppm)	4 - 10
Correlation	0.85
Method	Colorimetric
Calibration point	4 ppm
Time taken	2 minutes
Storage	Keep in a dry and cool place and away from sunlight.
Safety	Avoid contact with skin and use gloves to perform a test.
Shelf Life	1 year from the date of manufacturing.
Limitations	This test does not work with turbid or coloured water samples.

\*Tests to be performed in a well lit ambience with natural light

\*These tests do not work with turbid or coloured water samples