

Fluoride

Name of the test	Fluoride
Range of testing (ppm)	0 - 1.5
Unit	mg/L
Correlation	0.85
Method	Colorimetric
Calibration point	0 ppm
Time taken	2 minutes
Storage	Keep in a dry and cool place and away from sunlight.
Safety	Avoid contact with skin and use gloves to perform a test.
Shelf Life	1 year from the date of manufacturing.
Limitations	Free chlorine in the water gives a false positive reading and inaccurate results when temperature is above 41 C.

*Tests to be performed in a well lit ambience with natural light

*These tests do not work with turbid or coloured water samples