

## Residual Chlorine

|                         |  |
|-------------------------|--|
| Name of the test        | Residual Chlorine  |
| Range of testing (ppm)  | 0 - 3  |
| Unit                    | mg/L   |
| Correlation to standard | 0.80   |
| Method                  | Colorimetric   |
| Calibration point       | 0.2 ppm  |
| Time taken              | Immediate  |
| Storage                 | Keep in a cool, dry place away from sunlight.                                    |
| Safety                  | Avoid contact with skin. Use gloves to perform a test.                           |
| Shelf life              | 1 year from the date of manufacturing.   |
| Limitations             | This test should be carried out immediately, as soon as the sample is collected. |

\*Tests to be performed in a well lit ambience with natural light

\*These tests do not work with turbid or coloured water samples