Dosimetry, half-lifetime and statistical analysis of radioactive decays of Cs-137 and Ba-137

Parussel, Helke and Scholl, Maria

The Radiation of a Cs-137 sample was measured at different distances to confirm the law of squares(?). The threshold voltage associated with ~~the~~ Geiger-Mueller counter was determined. ~~With respect to the Geiger counter,~~ The responsitivity of the Geiger counter to γ- radiation to 0.2% was also determined. The half-life $T_{\frac{1}{2}} = (150.0 \pm 1.4)$s of ~~the~~ Ba-137m was measured experimentally practically ~~and compared with the literature value~~. A χ^2 test was conducted for the ~~a~~ decay of ~~in a~~ trioxide uranium tile ~~if it follows a Poisson distribution on a 5% level~~. The decay was measured with a gate time of t=10s and t=1s. It was determined that the radioactive decay follows a Poisson distribution. The same was executed ~~if it follows~~ to a Gaussian distribution ~~on a 5% level~~ which was rejected based on the 5% χ^2 test.
Date of execution: April 1st, 2022
Date of submission: April 8th 2022

On a 5% significance level

test

1 Introduction

Radioactivity, the property of unstable atomic nuclei to spontaneously emit ionizing radiation to the environment, is a process that ~~still~~ plays an important role today. The effects of ionizing radiation can be very dangerous to the human body. The emitted radiation, usually in alpha, beta or gamma rays, when penetrating matter, excites ~~the hit~~ atoms or molecules and transfers energy. This transfer can spatially displace atoms, knock electrons out of their atomic shells, and break chemical bonds. ~~In addition,~~ It is important to protect oneself adequately and take precautions when working with radioactive materials because, unlike other radiation such as heat radiation, the effect is not felt immediately. However, ionizing radiation and atomic decays also have very beneficial properties; by using the radiation it is possible to destroy cancer cells ~~specifically in medicine~~ and radioactivity is the basis of or also ~~in~~ nuclear reactors. In this paper, the properties of radioactive decay, the half-life and some other studies on the Geiger counter are examined.

2 Fundamental theory

One can predict radioactive decays only statistically. In the case of an individual atom, it is not possible to ~~determine~~ foresee the exact time of decay. However, it is possible to calculate the probability of a decay during a period of time. Hence it is possible to determine the number of atoms N(t) remaining after a period of time if the number of atoms N_0 at $t = 0$s is known as well as the decay constant λ:

$$N(t) = N_0\, e^{\lambda t} \tag{1}$$

The half life $T_{1/2}$ is the ~~moment when~~ time where half of the atoms has decayed and can be ascertained by the means of the decay constant λ:

$$\lambda = \frac{ln(2)}{T_{1/2}} \tag{2}$$

The activity of the radioactive preparation is described via $A(t) = \frac{dN}{dt}$ and was measured in the following experiment.

3 Dosimetry and radiation exposure

While conducting experiments with radiation, it is very important to monitor the radiation exposure as there is a limit to which one should be subjected. For people working with radioactive radiation there are ~~other~~ higher radiation limits ~~as~~ than for the average citizen but they are closely supervised. Those limits cannot be exceeded and if so, they cannot continue working until their limit per period of time is low enough. Those limits vary with the physical condition of the regarding person.
This limit is described by the equivalent dose H and ~~is described~~ calculated via

$$H = q \cdot D. \tag{3}$$

Whereas q scores the intensity of the radioactive radiation, it extends from q=1 for β- and γ-radiation to q=20 for α-radiation.
~~So~~ In this part of the experiment we examined the radiation exposure during the whole conduction. The equivalent dose H is listed with the according distance d to the probe in tab. 1.

d/[m]	$\frac{dH}{dt}$ / [μSv/h]	H_{4h}/[μSv]
$\approx$ 0.05	6.54±0.16	26.16±0.64
1	0.250±0.054	1.00±0.22

Table 1: Equivalent dose of the radioactive preparation depending on the distance between the preparation and the measuring tool, per hour and 4 hours.

The period of time until reaching 5mSv being 1m apart from the radioactive preparation amounts to t=(2.00±0.43)·10^4 h. The relative ratio between the period of time reaching the 5mSv and the 4h executing the experiment numbers to p=(0.0200±0.0043)%. Then the radiation exposure was determined and is ~~listed~~ given in tab. 2.

$H_{3.28h}$/[μSv]	H_{1a}/[mSv]
0.375± 0.092	1.00 ± 0.25

Table 2: Radiation exposure during the experiment plus the one for a whole year in the laboratory.

It is visible, that the radiation exposure during the experiment and even extrapolated for a year does not overstep the radiation limit of H=1mSv for an average person.

4 Activity of Cs-137

The activity A(t) is described ~~via~~ by

$$A(t) = \frac{dN}{dt} = A_0 e^{-\lambda t} \tag{4}$$

It describes the decays per ~~a~~ period of time ~~2~~. Knowing the half-time and the date of manufacture, we can calculate the activity of the Cs-137 ~~preparation~~ probe of today:
With the given data from tab. 3 and eq. [4] the activity is calculated to

$$A = (1.3607 \pm 0.0010)\, MBq$$

Date of manufacture	21.08.1978
Date of execution	01.04.2022
Activity A_0	3.7 $\cdot 10^6 Bq$
Half-lifetime	30.2a

Table 3: Given data to calculate the activity of Cs-137.

5 Measurements with a Geiger-Müller counter

5.1 Function of a Geiger-Müller counter

The Geiger-Müller counter measures power surges which are triggered by γ-radiation ionising atoms. The Geiger-Müller counter is a capacitor with a positive charged wire in the middle surrounded ~~??????~~ by ~~a encased by~~ a negative charged cylinder~~??????~~. The electric field is stronger in the middle because of the cylindric-structure:

$$E = \frac{1}{r}\frac{U}{ln(r_1/r_2)} \tag{5}$$

The counter is filled with argon ~~and alcohol vapour~~. ???
The γ-radiation ionises the gas. Thereupon free electrons create an ionisation avalanche near the positive charged wire and the following power surge creates a voltage surge at a resistor connected in between. This voltage surge is transferred to another capacitor which is connected to an amplifier. Since the ionised atoms move slower to the negative charged cylinder ~~as~~ than the free electrons to the positive charged wire because they weigh more, the electric field between those ionised atoms and the positive charged wire is decreased as the ionised atoms shield the negative charged cylinder. The ionisation avalanche stops. When the ionised atoms reach the negative~~ly~~ charged cylinder, a new ionisation avalanche can be caused.

5.2 Characteristics of a Geiger-Müller counter

The activity of a radioactive ~~preparation~~ probe is measured ~~by~~ through the voltage surge created at a resistor according to the Ohm-law. This power surge ~~is proportional to~~ depends on the applied voltage. If this voltage is too small, the electric field in the Geiger-Müller counter may be too small to create an ionisation avalanche and no ionisation and subsequently ~~the~~ no activity of the ~~preparation~~ probe can~~not~~ be measured. The ~~applied~~ minimum voltage ~~from~~ at which a current surge can be detected is called threshold voltage.
To ~~get~~ determine the threshold voltage, we increased the applied voltage until we could detect voltage ~~im~~pulses.

The threshold voltage amounts U=(516±4)V according to fig. 1. The error results from the uncertainty of adjusting the applied voltage.



Figure 1: The measured impulses are portrayed in reference to the input voltage. Voltage surges are detected ~~at~~ above a threshold voltage of U=(516±4)V.

To measure enough voltage impulses in the following parts of the experiment, we decided to use an input voltage of U=560V.

5.3 Null effect

In order to determine the actual radiation activity of materials, it is important to know beforehand that a so-called zero effect exists. This effect means that although there are no radiating samples in the vicinity, the Geiger-Mueller counter still somehow detects radiation. The reason for these measurements is cosmic radiation, as well as partially occurring radioactive isotopes, which have such a short half-life that they decay immediately, but are still detected. ~~However, if one wants to know exactly how the material under investigation radiates, it is necessary to know the zero effect and how much background radiation is present. For this purpose,~~ To measure the zero-effect, the measuring device is shielded with ~~the help of~~ lead walls and the detected radiation is recorded. Subsequently, this measurement is repeated at the workplace without shielding. Both procedures are carried out over a period of $t = 5$min with a voltage of $U_{tube} = 560$V. The resulting values are shown in Table 4.
As expected, the detected radiation decreases strongly inside the lead shielding. Proportionally, the value of radiation outside to the inside decreased by 49.74%.

5.4 Measuring the half- life of Ba-137m

The Cs-137 used in the experiments can enter two states by β- radiation. One is the stable Ba-137, but

counter location	null effect [min^{-1}]
working space	38.2
inside shielding	19.0

Table 4: Measuring the null effect for a Geiger counter at different locations at the workplace

the other is the excited state of Ba-137m. By subsequent radiation of γ-rays, the Ba-137m also reaches its ground state. In order to obtain this excited barium, it is now possible to separate the Ba-137m from the Cs-137 due to the different solubilities. Subsequently, the half-life of the dissolved sample was measured. For this purpose the Ba-137m was placed in front of the Leybold Geiger counter and the number of measured decays was noted during a gate time of $t = 10$s.



Figure 2: Decay rate of a Ba-137m sample fitted with an exponential function. The exponent of the fit allows us to calculate the half life of Ba-137m as $T_{\frac{1}{2}} = (150,01,4)s$.

The resulting data is shown in fig. 2.

In order to calculate the half-life with the measured decays, an exponential curve was fitted to the data in fig. 2 which provides us $\lambda = 0.004620 \pm 0.000044$. With λ and formula 2 the half-life of Ba- 137m can be determined to $T_{\frac{1}{2}} = (150.0 \pm 1.4)$s. This ~~practically~~ experimentally determined value is not consistent within its error with the theoretical value of $T_{1/2}$=153.12s. Warum? Fehler zu klein?

5.5 Responsiveness of the Geiger-counter

In this part of the experiment, the response of the Geiger counter is calculated. In reality, not every incoming γ- particle interacts with the gas inside the tube. The ratio of the incoming and detected radiation is called the responsiveness of the counter with respect to gamma radiation. To obtain this value for the Leybold counter used here, the Cs-137 sample is taken from the lead cube and placed at a distance of $d = (0.80 \pm 0.01)$m in front of the counter. Then the number of detected impulses is measured in a time interval of $t = 10$s. The responsiveness $\gamma = (0.2030 \pm 0.0010)\%$ was determined by the formula

$$\gamma = \frac{N}{A\,t\,\frac{F}{4\,\Pi\,d^2}} \tag{6}$$

Here F corresponds to the sensitive area of the counting tube, N to the measured pulse rate in $t = 10$s and A is the calculated activity in section 4.

6 Analysis of radioactive decays

6.1 Poisson and Gaussian distribution

Formula 1 describes the expected value of the number of non-decayed atoms after a period of time t. However, the radioactive decay is to be described stochastically and the number of decays only scatters around the expected value. To get a better understanding of this distribution, it is useful to take a closer look at its behavior:

The probability of an atom to decay is given by the Poisson distribution with

$$P(k) = \frac{\mu^k}{k!} e^{-\mu} \tag{7}$$

Whereas k is the number if decays detected in the gate time and μ is the mean decay number detected. For big mean $\mu = p_{decay} \cdot N_{atoms}$ the Poisson distribution transitions into a Gaussian distribution and if the mean μ is constant, whereby $N_{atoms} \to \infty$ and $p_{decay} \to 0$. The probability of the number of decays during a gate time is given by

$$P(k) = \frac{1}{\sqrt{2\pi\sigma}} e^{-\frac{(k-\mu)^2}{2\sigma}} \tag{8}$$

The standard derviation is given by $\sigma = \sqrt{\mu}$.

6.2 Poisson and Gaussian distribution with gate time t=10s

For this part of the experiment an radioactive orange tile was put in a plumb box with the counting tube Leybold. It was put in a plumb box to suppress the background radiation. The number of decays during a gate time of t=10s was measured for 1000 gate times. To ascertain if the detected number of impulses follows a Gaussian or Poisson distribution, we conduct a χ^2-test. At first the number of detected impulses for a certain decay number per gate time was listed and the theoretical expected values were determined for Gaussian and Poisson distribution. Then the number of classes was compiled. The number of detected decays for each class is portrayed in fig. 3. The mean μ accounts to $\mu = 7.12 \pm 0.37$. The mean is pretty high so we examine a χ^2-test for a Poisson and Gaussian distribution. χ^2 indicates the deviation of a distribution. The smaller the value of χ^2 the more likely that the data follows the examined distribution. The calculated values are listed in tab. 5.

Null hypotesis	χ^2	$\chi^2_{5\%,max}$
Poisson distribution	12.70	23.03
Gaussian distribution	23.91	21.03

Table 5: χ^2-values are listed to ascertain if the number decays of a radioactive tile during a gate time of t=10s follows a Poisson and Gaussian distribution.

The χ^2 value with a significance level indicates that if the calculated χ^2 is higher, the deviation is too big and the distribution does not follow the examined one. The $\chi^2_{5\%,max}$ indicates the threshold of χ^2. If we calculate a χ^2 higher than the one of the significance level of 5% and conclude that it does not follow a Poisson or Gaussian distribution, there is a 5% chance left that it actually does.

As we can see the calculated χ^2 for Poisson distribution is smaller than the value of a significance level of 5%. For the Gaussian distribution, the calculated χ^2 is bigger than the one with a significance level of 5% and this hypothesis needs to be rejected. In conclusion, this distribution follows a Poisson distribution and not a Gaussian.

To stress this assumption, our measured data is plotted with its cumulated relative frequency along the number of measured decays during the gate time in fig. 4. Since the probabilities do not follow the linear fit and the residuum is not by chance, the data does not likely follow a Gaussian distribution. The ascertained mean based on fig. 4 is listed in tab. 6. The error of the mean ascertained by the help of fig. 4 is determined by the help of the Gaussian bell and its probability that 68.2% of the data lies within this mean with its error.



Figure 3: The measured frequency of a number of decays during a gate time of t=10s, as well as the Poisson and Gaussian distribution, is depicted.



Figure 4: Cumulated relative frequency mapped onto measured decays during a gate time if t=10s.

Calculated mean	7.12±0.37
Ascertained mean	6.9±2.8

Table 6: Listed are the out of the data calculated mean and the one ascertained by the help on fig. 4.

graphically

6.3 Poisson and Gaussian distribution with a gate time of t=1s

In this part we also used the radioactive orange tile with the counting tube Leybold in a plumb box but measured now the number of decays during a gate time of 1s. The mean numbers to $\mu = 0.7 \pm 1.1$.

Null hypotesis	χ^2	$\chi^2_{5\%,max}$
Poisson distribution	5.14	7.81
Gaussian distribution	175.03	5.99

Table 7: χ^2-values are listed to ascertain if the number decays of a radioactive tile during a gate time of t=1s follows a Poisson and Gaussian distribution.

This mean is pretty small, smaller than the one in section 6.2. Hence, the distribution will probably not follow a Gaussian distribution as it transforms into one for big mean μ. The χ^2 values are listed in tab. 7.

As assumed, our data does not follow a Gaussian but a Poisson distribution. The impulses during a gate time of t=1s and the measured frequency is visible in fig. 5.



Figure 5: The measured frequency of a number of decays during a gate time of t=1s, as well as the Poisson distribution, is depicted.

7 Results

For radiation exposure we measured the dose rate $\frac{dH}{dt}$ in a distance of d=0.05 as $\frac{dH}{dt} = (6.54 \pm 0.16)\frac{\mu Sv}{h}$ and the dose for 4 hours during the conducted experiment as H_{4h}=(26.16 ± 0.64)μSv. In one meter distance,

the dosis rate numbers to $\frac{dH}{dt} = (0.250 \pm 0.054)\frac{\mu Sv}{h}$ and the dose for 4 hours accounts to $H_{4h} = 1.00 \pm 0.22)\mu$Sv. The relative ratio between reaching a limit of 5mSv and the dose rate in 4 hours aggregate~~s~~ to p=(0.0200±0.0043)%.
Then the activity of a Cs-137 probe was calculated to A=(1.3607±0.0010)MBq. With this activity we determined the responsiveness of a Geiger-Müller counter to $\gamma = (0.2030 \pm 0.0010)\%$.
The voltage threshold of the used Geiger-Müller counter was ascertained to $U_{thres} = (516 \pm 4)V$. For the null effect the background radiation was measured by the help of the Geiger-Müller counter: at first at the working space to 38.2 counts per minute and inside a plumb shielding to 19.0 counts per minute. The deviation is 49.74%.
Then the half-life-time of Ba- 137m was determined to $T_{1/2} = (150.0 \pm 1.4)$s.
Lastly, the number of decays of a radioactive orange tile was ascertained for t=10s and t=1s. The mean of decays to the regarding gate time was calculated to $\mu_{1s} = (0.7 \pm 1.1)$, $\mu_{10s,calc.} = (7.12 \pm 0.37)$ and $\mu_{10s,\cancel{ascert.}\,\text{grph}} = (6.9 \pm 2.8)$. For both gate times the χ^2 values acknowledged that ... follow Poisson, don't follow Gauss,

References

[1] Physikalisches Grundpraktikum, Universitat Würzburg, Modul C2, Versuch 46, "Dosimetrie, Halbwertszeit, 2 - Test", 2022

[2] Meschede, Dieter, Gerthsen Physik, 25. Auflage, Springer-Verlag, Berlin, 2015

[3] Half life of Ba-137m, https://www.internetchemie.info/isotop.php?Kern=Ba-137m, last visited April 8th, 2022, 11:48 am

[4] Physikalisches Grundpraktikum, Universität Würzburg, "Begriffsbestimmungen Strahlenschutz", 2022