

Bone Density (DXA) Report

BONETWIN SYNTHETIC DXA

SYNTHETIC DEMO - NOT A MEDICAL RECORD

Study details

Subject pseudonym: SYNTH-BONE-001
Scan date: 2026-08-16
Facility: Synthetic Imaging Center C

Scanner manufacturer: GE Healthcare
Scanner model: Lunar iDXA (synthetic)
Report type: DXA BMD

Site	BMD g/cm2	T-score	Z-score	Use
Left Total Hip	0.735	-1.7	-0.5	YES
Left Femoral Neck	0.661	-1.8	-0.6	YES
Lumbar Spine L1-L4	0.807	-2.1	-0.8	REVIEW

Source evidence text

Left Total Hip BMD 0.735 g/cm2; T-score -1.7; Z-score -0.5; Confidence 0.99; Longitudinal YES
Left Femoral Neck BMD 0.661 g/cm2; T-score -1.8; Z-score -0.6; Confidence 0.96; Longitudinal YES
Lumbar Spine L1-L4 BMD 0.807 g/cm2; T-score -2.1; Z-score -0.8; Confidence 0.91; Longitudinal NO

Review note

New scanner metadata and lumbar exclusion are flagged for source-backed review.