

Bone Density (DXA) Report

BONETWIN SYNTHETIC DXA

SYNTHETIC DEMO - NOT A MEDICAL RECORD

Study details

Subject pseudonym: SYNTH-BONE-001
Scan date: 2022-06-08
Facility: Synthetic Imaging Center A

Scanner manufacturer: Hologic
Scanner model: Discovery A (synthetic)
Report type: DXA BMD

Site	BMD g/cm2	T-score	Z-score	Use
Left Total Hip	0.756	-1.5	-0.4	YES
Left Femoral Neck	0.683	-1.6	-0.5	YES
Lumbar Spine L1-L4	0.822	-2.0	-0.8	REVIEW

Source evidence text

Left Total Hip BMD 0.756 g/cm2; T-score -1.5; Z-score -0.4; Confidence 0.99; Longitudinal YES
Left Femoral Neck BMD 0.683 g/cm2; T-score -1.6; Z-score -0.5; Confidence 0.96; Longitudinal YES
Lumbar Spine L1-L4 BMD 0.822 g/cm2; T-score -2.0; Z-score -0.8; Confidence 0.78; Longitudinal NO

Review note

Synthetic positioning artifact flagged at the lumbar site for human review.