

Bone Density (DXA) Report

BONETWIN SYNTHETIC DXA

SYNTHETIC DEMO - NOT A MEDICAL RECORD

Study details

Subject pseudonym: SYNTH-BONE-001
Scan date: 2026-04-12
Facility: Synthetic Imaging Center B

Scanner manufacturer: Hologic
Scanner model: Horizon A (synthetic)
Report type: DXA BMD

Site	BMD g/cm2	T-score	Z-score	Use
Left Total Hip	0.742	-1.6	-0.4	YES
Left Femoral Neck	0.668	-1.7	-0.5	YES
Lumbar Spine L1-L4	0.811	-2.1	-0.8	REVIEW

Source evidence text

Left Total Hip BMD 0.742 g/cm2; T-score -1.6; Z-score -0.4; Confidence 0.98; Longitudinal YES
Left Femoral Neck BMD 0.668 g/cm2; T-score -1.7; Z-score -0.5; Confidence 0.88; Longitudinal YES
Lumbar Spine L1-L4 BMD 0.811 g/cm2; T-score -2.1; Z-score -0.8; Confidence 0.97; Longitudinal NO

Review note

Scanner metadata differs from the prior synthetic study; clinician review requested.