	Topic	Page	Include?
A	Introduction	10	
B	Integrated Clinical Environment (ICE) Clinical Scenarios	11	
UC.1	ICE.1 - Safety Interlocks	11	
UC.2	ICE.2 - Synchronization with safety interlock	12	
UC.3	ICE.3 - Process control (workflow)	14	
UC.4	ICE.4 - Smart alarm system	14	
UC.5	ICE.5 - Decision support	15	
UC.6	ICE.6 - Physiological Closed Loop Control (PCLC)	17	
UC.7	ICE.7 - Medical Device Plug-and-Play Interoperability (MD PnP)	18	
C	SDC/QH – Quiet Hospital (QH) Scenario and Use Cases	20	
UC.8	SDC/QH.1 – Device alert signal delegation to single-pt. alert aggregator	20	
UC.9	SDC/QH.2 – Single pt. alert aggregator alert signal delegation to multi-pt. aggregator	21	
UC.10	SDC/QH.3 – Device alert signal delegation to Alert Manager (ACM AM)	21	
UC.11	SDC/QH.4 – Multi-pt. aggregator to Alert Manager (ACM AM)	21	
UC.12	SDC/QH.5 – SDC to FHIR® Gateway.	21	
UC.13	SDC/QH.6 – Alert Manager (ACM AM) to care-giver Alert Communicator (ACM AC)	21	
UC.14	SDC/QH.7 – Alert Communicator (ACM AC) failure	21	
UC.15	SDC/QH.8 – Alert Manager (ACM AM) failure	21	
UC.16	SDC/QH.9 – Multi-pt. aggregator failure	22	
UC.17	SDC/QH.10 – Single pt. aggregator failure	22	
D	SDC/PDP - Preeclampsia (PDP) Scenario and Use Cases	23	
UC.18	SDC/PDP.1 - In-hospital, at home and mobile / clinic care contexts	23	
UC.19	SDC/PDP.2 - Acute point-of-care medical and personal health devices	23	
UC.20	SDC/PDP.3 - Lab results including point of care devices and transactions	23	
UC.21	SDC/PDP.4 - Location tracking and Device Identification / Association Management	23	
UC.22	SDC/PDP.5 - A Cloud-based CDS system and “locally” networked systems & applications	24	
E	SDC/FESS – Endoscopic Surgery Scenario and Use Cases	25	
UC.23	SDC/FESS.1 – Surgeon view of patient vitals	25	
UC.24	SDC/FESS.2– Surgeon control of OR table and lights	25	
UC.25	SDC/FESS.3– Surgeon control of surgical tools	25	
UC.26	SDC/FESS.4 – Device reports technical issue to responsible BMET	26	
UC.27	SDC/FESS.5 – Seamless exchange of Medical Devices	26	
F	NITRD – Medical Device Scenario and Use Cases	27	
UC.28	NITRD.1 – Seamless changes of medical devices	28	
UC.29	NITRD.2 – Capture of data and settings	28	
UC.30	NITRD.3 – Supervisory control established	28	
UC.31	NITRD.4 – Autonomous patient therapy	28	
UC.32	NITRD.5 – Data flows through the Continuum of Care	28	
UC.33	NITRD.6 – Capture of equipment configurations	28	
UC.34	NITRD.7 – Black Box Recorder	28	
G	OR.NET Use Cases	29	
UC.35	ORNET.001 - Integration of anesthesia video data into (anesthesiologic) patient monitoring via radio transmission	29	
UC.36	ORNET.004 - Indicates a warning from one OSCB-compliant device to another	29	
UC.37	ORNET.005 - Transfer of the control of device A by a device B, which is actually prohibited	29	
UC.38	ORNET.006 - User-specific workflow-dependent adjustment of the height of the operating table / foot step	29	
UC.39	ORNET.009 - Display of electrophysiological signals on a monitor by a surgeon	29	
UC.40	ORNET.010 - Use of a foot switch as an input device	29	

UC.41	ORNET.012 - Display of force and moment sensor (KMS) for knee axis determination on central monitor	29	
UC.42	ORNET.016 - Collision avoidance of devices	29	
UC.43	ORNET.017 - Control of a desired manipulator by a common control console	29	
UC.44	ORNET.018 - Power control of the milling machine at the middle	29	
UC.45	ORNET.019 - Linking the surgical light to the endoscope	29	
UC.46	ORNET.024 - Storage of intraoperative image data in the EHR	29	
UC.47	ORNET.029 - Intraoperative input of monitoring data to EHR	29	
UC.48	ORNET.030 - Display of monitoring data in microscope image	29	
UC.49	ORNET.032 - Insertion of relevant information into the surgical microscope according to the clinical situation	29	
UC.50	ORNET.035 - Central presentation of relevant device data / documents / patient data close to the site	29	
UC.51	ORNET.040 - Single check-desk	30	
UC.52	ORNET.042 - prevention of explosions	30	
UC.53	ORNET.043 - Controlling the US Dissector with the Human Machine Interface (MMI) of the microscope	30	
UC.54	ORNET.050 - Service Total Log	30	
UC.55	ORNET.051 - Visualization of instrument position	30	
UC.56	ORNET.051 - Visualization of instrument position	30	
UC.57	ORNET.067 - Fusion imaging intraoperatively US - CT / MRT	30	
UC.58	ORNET.068 - Image-based communication.	30	
UC.59	ORNET.069 - Monitoring of patient and device parameters, especially alarms	30	
UC.60	ORNET.070 - Control of the anesthesiologically relevant device parameters.	30	
UC.61	ORNET.072 - Import of clinical data from the EHR	30	
UC.62	ORNET.073 - Storage of digital patient records in the EHR.	30	
UC.63	ORNET.078 - Operating room preparation	30	
UC.64	ORNET.079 - remote diagnostics remote maintenance firmware update	30	
UC.65	ORNET.082 - Video display of selected parameters from external OSCB-compliant devices	30	
UC.66	ORNET.084 - Transfer of a 3D volume image to a navigation system	30	
UC.67	ORNET.086 - Storing an Endoscope Video	30	
UC.68	ORNET.090 - Position adjustment C-arm and operating table	30	
UC.69	ORNET.091 - Insertion of tumor borders into the microscope.	30	
UC.70	ORNET.095 - Control of the endoscope light source via the navigation MMI	30	
UC.71	ORNET.096 - Integration of B-Mode ultrasound into navigation	30	
UC.72	ORNET.097 - Integration of 3D ultrasound into navigation.	31	
UC.73	ORNET.100 - uploading patient strain data to the medical IT subnet	31	
UC.74	ORNET.101 - Coupling of the footing height to the seat height	31	
UC.75	ORNET.102 - Detection of OSCB compliant devices	31	
UC.76	ORNET.103 - Failure of OSCB compliant devices	31	
UC.77	ORNET.104 - Logout of OSCB-compliant devices	31	
UC.78	ORNET.105 - Automatic feeding of patient strain data to OSCB compliant device	31	
UC.79	ORNET.106 - Enter the operator ID in the IT subnet	31	
UC.80	ORNET.107 - Automatic input of operator code in OSCB-compliant device.	31	
UC.81	ORNET.108 - Transfer of image data from the PACS to a monitor	31	
UC.82	ORNET.109 - Create intraoperative screenshots.	31	
UC.83	ORNET.110 - Transfer registration information from the navigation device to the robotic system	31	
UC.84	ORNET.111 - Display additional information on the monitor of the navigation device	31	
UC.85	ORNET.112 - Load planning image data from the PACS into the robotic system	31	
UC.86	ORNET.114 - Recording and archiving of video data for documentation via a Frame grabber application	31	

UC.87	ORNET.115 - Recording and archiving of video data for documentation via an IP data stream	31	
UC.88	ORNET.116 - View of DICOM® objects on any terminal with display function	31	
UC.89	ORNET.117 - Patient and surgical centered DICOM object display	31	
UC.90	ORNET.118 - Storage of DICOM objects in DICOM object-recording systems	32	
UC.91	ORNET.119 - Control from a DICOM object viewer to change the layout of a display screen	32	
UC.92	ORNET.120 - Linking the OP light to the C-arm	32	
UC.93	ORNET.121 - Reconstruction of the OP Situs by laser measuring point.	32	
UC.94	ORNET.122 - Power control of the vacuum cleaner on the central pipe.	32	
UC.95	ORNET.123 - Derivation of the alarms of the surgical alarm system to the operating theater	32	
UC.96	ORNET.124 - Alarm-triggered manipulation of the RGB room light	32	
UC.97	ORNET.125 - Show warning signals in videos of minimally / non-invasive imaging procedures	32	
UC.98	ORNET.126 - Re-adjustment of preoperative CT data using intraoperative DVT	32	
UC.99	ORNET.127 - Automatic orientation of the stereo camera	32	
UC.100	ORNET.128 - Logging the Instrument Position During Optical Navigation	32	
UC.101	ORNET.129 - Transfer of a log to the logistics system	32	
UC.102	ORNET.130 - Integration of neuromonitoring of the ENT area into the optical navigation	32	
UC.103	ORNET.131 - Adjust the position of the operating table using a navigated pointer	32	
UC.104	ORNET.132 - Common control interface for all devices.	32	
UC.105	ORNET.133 - Automatic documentation of device parameters: blood volume	32	
UC.106	ORNET.134 - Automatic documentation of device parameters: surface dose.	32	
UC.107	ORNET.135 - Combination of rotational angiography with electrophysiological measuring site	32	
UC.108	ORNET.136 - CT / MRI linkage with TEE ultrasound in endoscopic mitral valve surgery	32	
UC.109	ORNET.137 - Common operation of all room air-conditioning units from sterile workplaces	33	
UC.110	ORNET.138 - Availability of all patients Information from sterile workplace	33	
UC.111	ORNET.139 - Power control of an ultrasonic breaker	33	
UC.112	ORNET.140 - Integration of neuromonitoring of the brain into the risk areas of optical navigation	33	
UC.113	ORNET.141 - Integration of intraoperative ultrasound data of neurosurgery into the risk areas of optical navigation	33	
UC.114	ORNET.142 - Adaptation of device parameters to vital parameters of patient	33	
UC.115	ORNET.143 - Adaptation of device parameters to patient 's vital parameters	33	
UC.116	ORNET.144 - Integration of the Laparoscopy and Endoscopy Tower	33	
UC.117	ORNET.145 - Linking the surgical column and the operating table	33	
UC.118	ORNET.146 - Warning if pressure is too strong on the ultrasound device	33	
UC.119	ORNET.147 - Change the US dissector parameters	33	
UC.120	ORNET.148 - Remote triggering of US dissector device functionalities	33	
UC.121	ORNET.149 - US dissector informs the microscope of its current trigger state	33	
UC.122	ORNET.150 - Remote tripping of US dissector device functionalities	33	
UC.123	ORNET.151 - Acoustic and visual feedback of intraoperative neuromonitoring during surgical procedures	33	
UC.124	ORNET.152 - Transfer of video and image data	33	
UC.125	ORNET.153 - Read image data from the PACS	33	
UC.126	ORNET.154 - Uploading image data to the PACS	33	
UC.127	ORNET.155 - File import from the EHR	33	
UC.128	ORNET.156 - File export to the EHR	33	
H	IHE PCD - Device to Enterprise Communication (DEC) Use Cases	34	
UC.129	DEC.1 - Communicate patient identified DEC data to EMR/EHR	34	
UC.130	DEC.2 - Communicate validated periodic DEC data to EMR/EHR	34	
UC.131	DEC.3 - Use Cases for Automatic Patient Demographics Acquisition	34	
UC.132	DEC.3.1 - Patient ID known in ADT, locally available	35	
UC.133	DEC.3.2 - Patient ID known in ADT, not locally available	35	
UC.134	DEC.3.3 - Patient ID not known in ADT, locally available	35	

UC.135	DEC.3.4 - Patient ID not known in ADT, not locally available.	35	
UC.136	DEC.3.5 Other Clinical Examples	36	
UC.137	DEC.3.5a – Association of Patient ID and Medical Device – via ID List	36	
UC.138	DEC.3.5b - Association of Patient ID and Medical Device – via patient wristband	36	
UC.139	DEC.3.5c - Association of Patient ID and Medical Device – via RFID tags	36	
I	IHE PCD - DEC Subscribe to Patient Data (SPD) Use Cases	37	
UC.140	SPD.1: Communicate patient identified data to EMR/EHR	37	
UC.141	SPD.2 - Communicate validated periodic data to EMR/EHR	37	
UC.142	SPD.3 - Subscribe to patient data at specific periodic interval.	37	
UC.143	SPD.4 - Subscribe to patient data for specific patients.	37	
UC.144	SPD.5 - Subscribe to patient data for patients from a specific location.	37	
UC.145	SPD.6 - Subscribe to patient data for a specific device or class of devices	37	
UC.146	SPD.7 - Subscribe to patient data for specific parameters or class of parameters.	38	
UC.147	SPD.8 - Request a snapshot of current or most recent patient data.	38	
J	IHE PCD - Alert Communication Management (ACM) Use Cases	39	
UC.148	ACM.1 - Location Sourced	39	
UC.149	ACM.2 - Identified Patient Source	39	
UC.150	ACM.3 - Same as ACM.1/ACM.2 with Escalation with Cancel at Alert Source	40	
UC.151	ACM.4 - Same as ACM.1/ACM.2 with Escalation with Cancel at Communication Endpoint	40	
UC.152	ACM.5 - Same as ACM.1/ACM.2 with Escalation with Cancel at AM	40	
UC.153	ACM.6 - Information with no destination other than logging by the AM	40	
UC.154	ACM.7 - Equipment Sourced Alert	40	
K	IHE PCD - Medical Equipment Management Device Management (MEMDMC) Use Cases	41	
UC.155	MEMDMC.1 - Equipment Observations to CMMS	41	
UC.156	MEMDMC.2 - Equipment Observations to CMMS	41	
L	IHE PCD - Medical Equipment Management Location Services (MEMLS) Use Cases	43	
UC.157	MEMLS.1 - Communication of location observations in conjunction with other non-location related transactions	43	
UC.158	MEMLS.2 - Communication of location observations in conjunction with LS specific events	43	
M	IHE PCD - Waveform Content Module (WCM) Use Cases	45	
UC.159	WCM.1 - Alarm Waveform Snapshot	45	
UC.160	WCM.2 - Real-Time Waveform Viewing	45	
UC.161	WCM.3 - Archived Waveform Viewing	45	
UC.162	WCM.4 - Mixed Snapshot and Continuous Waveform Viewing	45	
UC.163	WCM.5 - Waveform Snapshot to EHR	45	
UC.164	WCM.6 - 12 Lead ECG	45	
N	IHE PCD - Retrospective Data Query (RDQ) Use Cases	46	
UC.165	RDQ.1 - Query for all retrospective data on a single patient.	46	
UC.166	RDQ.2 - Query for all retrospective data on multiple patients.	46	
UC.167	RDQ.3 - Query for retrospective data on a single patient within a specified time interval.	47	
UC.168	RDQ.4 - Query for retrospective data on multiple patients within a specified time interval.	47	
UC.169	RDQ.5 - Query for retrospective data on 1 or more parameter elements on a single patient.	47	
UC.170	RDQ.6 - Query for retrospective data on 1 or more parameter elements on multiple patients.	48	
UC.171	RDQ.7 - Query for retrospective data on 1 or more parameter elements on a single patient within a specified time interval.	48	
UC.172	RDQ.8 - Query for retrospective data on 1 or more parameter elements on multiple patients within a specified time interval.	48	
O	IHE PCD - Implantable Device Cardiac Observation (IDCO) Use Cases	50	
UC.173	IDCO.1 - Implantable Cardiac Device In-Clinic Follow-up	50	
UC.174	IDCO.2 - Implantable Cardiac Device In-Clinic Follow-up with Networked Programmer that Translates Information	51	

UC.175	IDCO.3 - Implantable Cardiac Device Remote Followup	51	
UC.176	IDCO.4 - Remote Monitoring of Implanted Cardiac Devices	51	
P	IHE PCD - Point-of-Care Infusion Verification (PIV) Use Cases	53	
UC.177	PIV.1 – Transfer of infusion parameters from BPOC to infusion device	53	
Q	IHE PCD - Infusion Pump Event Communication (IPEC)	55	
UC.178	IPEC.1: Communicate event data to EMR/EHR	55	
R	IHE PCD - Point-of-Care Identity Management (PCIM) Use Cases	56	
UC.179	PCIM.1 - Associating Device with Patient	56	
UC.180	PCIM.2 - Disassociating Device from Patient	57	
UC.181	PCIM.3 - Query for the Devices for a Patient	58	
UC.182	PCIM.4 - Query the Associated Patient for a Device	58	
UC.183	PCIM.5 - Device Registrant Registers a Device	58	
UC.184	PCIM.6 - Query the Device Registrant for a list of candidate devices for an association	59	
S	CEN/TC 251/PT5-021 (VITAL) Use Cases	60	
UC.185	VITAL.1 - Data Logger - Single Device	62	
UC.186	VITAL.2 - Data Logger - Multiple Devices	64	
UC.187	VITAL.3 - Real-Time Data Display	67	
UC.188	VITAL.4 - Patient Alarm Monitoring	68	
UC.189	VITAL.5 - Remote Control	70	
UC.190	VITAL.6 - Patient Viewing Interoperability	71	
UC.191	VITAL.7 - Patient Monitoring Interoperability	73	
UC.192	VITAL.8 - Patient Data Exchange (Off-line)	74	
UC.193	VITAL.9 - Patient Data Exchange (On-line/Interactive)	76	
T	ONC/AHIC Common Device Connectivity Use Cases	79	
UC.194	AHIC.1 - Configure and register a device to communicate with an EHR.	80	
UC.195	AHIC.2 - Associate patient ID and device information within an EHR.	80	
UC.196	AHIC.3 - Communicate measurement information to the EHR	80	
UC.197	AHIC.4 - Communicate device meta-data with each measurement to the EHR.	81	
UC.198	AHIC.5 - Communicate measurement intervals, etc. within the EHR.	81	
UC.199	AHIC.6 - Query the device or device intermediary for additional information.	81	
UC.200	AHIC.7 – Gracefully recover from a lapse in EHR connectivity.	82	
UC.201	AHIC.8 - Communicate standardized alarm types to the EHR.	82	
UC.202	AHIC.9 - Set limits and safeguards for device settings from the EHR to a device.	82	
UC.203	AHIC.10 - Wirelessly communicate PoC device information from the device to a device intermediary or EHR.	82	
U	“Special” Patient Monitoring Use Cases	90	
UC.204	PM.1 - Synchronized Cardioversion	90	
UC.205	PM.2 - Intra-Aortic Balloon Pump Synchronization	90	
UC.206	PM.3 - Catheter Insertion Procedure	91	
UC.207	PM.4 - The User “Cockpit”	91	