

# IHE Change Proposal

## Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-DOM-xxx (assigned by Domain Technical Committee)
Change Proposal Status:	Submitted
Date of last update:	8-12-2011
Person assigned:	(assigned by Domain Technical Committee)

## Change Proposal Summary information:

Example Text Change - OBR	
Submitter's Name(s) and e-mail address(es):	Stephen Swanson ( <a href="mailto:sswanson@sjm.com">sswanson@sjm.com</a> ) Gene Hageman ( <a href="mailto:gene.hageman@medtronic.com">gene.hageman@medtronic.com</a> ) Alexander Kraus ( <a href="mailto:alexander.kraus@biotronik.com">alexander.kraus@biotronik.com</a> ) Ken Hoyme ( <a href="mailto:ken.hoyme@bsci.com">ken.hoyme@bsci.com</a> )
Submission Date:	April 05, 2012
Integration Profile(s) affected:	Implantable Device – Cardiac – Observation
Actor(s) affected:	None
IHE Technical Framework or Supplement modified:	IHE Patient Care Device (PCD) Technical Framework; Volume 2; Revision 1.0; August 12, 2011
Volume(s) and Section(s) affected:	Volume 2 (PCD TF-2); Section 3.9.4.1.2.4
Rationale for Change:	
1. The example value provided in Table 3.2.4.1.4-2: OBR Segment does not match the value defined in the IEEE 11073 Nomenclature. The Change Proposal is to bring the example text in-line with the IEEE value.	

## Proposed Changes:

*Replace Table 3.2.4.1.4-2 Row 7 by the following:*

**Table 3.2.4.1.4-2: OBR Segment**

Name	Seq	DT	Len	Opt	Rep	Min	Max	Tbl	Fixed Val	Ex Val
<i>identifier</i>	1	ST	20	R		1	1			Remote <del>follow-up</del>