

IHE Change Proposal

Tracking Information

IHE Domain	Patient Care Devices
Change Proposal ID	CP-PCD-047-00
Change Proposal Status	Submitted
Date of last update	2010-03-20
Person assigned	J. Rhoads

Change Proposal Summary Information

Title:	Make MSH-21 Required and List Message Profile Identifiers
Submitter's Name(s) and e-mail address(es):	John Rhoads (john.rhoads@philips.com)
Submission Date:	2010-12-21
Integration Profile(s) affected:	all PCD profiles
Actor(s) affected:	all PCD actors
IHE Technical Framework or Supplement modified:	IHE PCD Technical Framework, Final Text and Trial Use versions
Volume(s) and Section(s) affected:	Vol. 2
Rationale for Change: The rationale for this change is as follows: To allow specific identification of the IHE PCD HL7 profile that the message is claiming performance to, field MSH-21 Message Profile Identifier will be required in PCD messages. Message Profile Identifiers in the form of ISO OIDs from the IHE PCD identifier node are listed for all profiles in final text or in trial use are listed	